

BOOK91R

Notes:

"T" on front

"7-49/8-54" on front

"NB-WI-1" on front

"T" on spine

--Please note that this is a very old book, (8.5x14) and had to be reduced and darkened in order to pick up all data. Sorry for the darkness on some pages.

--Please note that pages 1-4 and 99-102 have been torn out from this logbook.

Blank pages: page opposite page 5, 6, 12-14, 18, 21, 28, 31, 32, 36-39, 48, 114-123, 128-139, 141-149, 151

- page 22 has long graph taped to it
- page 25 has big graph taped to it
- pages 52, 80, 83, and 95 each have drawing attached to it
- pages 66/67 have long sheet of paper between pages
- page 69 has (8.5x11) graph taped to it
- page 71 has (8.5x11) graph taped to it
- page 79 has graph taped
- pages 86/87 has (8.5x11) sheet between pages
- page 88 has big graph taped
- page 105 has 2 drawings taped
- page 109 has graph taped
- pages 114/115 has drawing between pages
- page 124 has paper glued to it
- 2 (8.5x11) sheets (1 graph) taped to inside back cover sheet
- 7 (8.5x11) sheets of paper taped to inside back of book

Scanned by:

Sheila Finch

RSICC /Oak Ridge National Lab.

September 14, 1999

SUBJECT	PAGE
DEVELOPMENT	
Test Generator for Recorders	40
Rectifier Supply - relays 202	49
Liquid Level Indicator	81, 81, 90-91,
Amp. circuit for Brush	83-86
Brush use 202	78
P.M. RT counter	61-75, 88-89
Alarm	57-59
Test Input Recorders	61
G.E. relay characteristics	76
Speed Controller	78
Log N. battery supply	60
Rod Drop	15-19
Sealing Paints	94
SERVICING	
F-05	5-11
Scalers - 202	20-23, 30
	50
→ Scalers - 215	76, 125, 126
Linear Amplifier	39, 33
Rate Meter	35, 60-61
Reed	51
D.C. Amps (202)	77, 92, 93
Log N	47
— SCOPE CALIBRATION —	158
Two-fold counter	27, 61
Poppy	41-46
Neutron Survey Meter	29, 26, 33
	52-56
East End Monitor	140 -

General Maint.

95



Standard Blank Book

No. 67½

Journals Double \$ and Cts. no Units

Cash Books " " " "

Day Books " " " "

S. E. Ledgers " " " "

D. E. Ledgers Full Page Form "

Long Day Books "

Records with Margin Line

In 150, 200, 300 and 500 Pages

Made in U. S. A.

TO REORDER THIS BOOK, SPECIFY
NUMBER, RULING AND THICKNESS
AS INDICATED ON BACKBONE OF BOOK

A BOORUM & PEASE PRODUCT

5

ELECTRONIC MAINTENANCE

- 7-19-49 COUNTER #1 - ERRATIC SENSITIVITY to DEAD
 Rx - Replaced 6SN7 and 6AR7 - R₁ 35K 100Ω
 RK 10K-2W & 7.5K 2W Black Resistor - HAK
- 7-26-49 D.C. Amp #4 - EXCESSIVE sensitivity to TRANSIENTS
 Rx - Panned down B⁺ to ground
 1MFD cathode 125T7 regulator tube to ground
 1MFD across B⁺ to ground at rectifier section
 0.1MFD B⁺ to B⁻ HAK-105
- 7-28-49 D.C. Amp #5 - Alarm circuit difficult to reset
 Rx - 0.1MFD - 600WV across G₁ to K of 2050 vibrator
 R.F. oscillation of B⁺ supply - R₁ changed
 125A7 regulator tube to 125T7
 8mh
- 9-30-49 Vibrating Reed Electronometer - Dead
 C-11 shorted - Replaced
 This shorted capacitor had
 destroyed the 6Y6 and 6Y3 in
 regulated power supply. Replaced Tubes
 Replaced 5 other tubes in amp circuit
 4mh
- 9/30/49 Photo Multiplier placed in service after
 minor circuit changes (jots instead of grids)
 in battery circuit)
 3mh
 1 - fixed resistor on 951-A voltage divider
 opened - replaced
- 10/4/49 Vibrating Reed bypass capacitor on plate coupling
 capacitor shorted & replaced HAK
 4mh
- 10/7/49 Counter Time interval control switch checked
 all screws tightened and contacts repaired left
 2mh
- 10/11/49 Vibrating Reed tubes checked & replaced
 #3 BF's Counter
 3mh
 7mh

10-13-49 Counter # 2 - Dead
4mh (a) Replaced a leaky 8-450 μ fd electrolytic capacitor in the pre amp.
(b) Replaced a 500- Ω $\frac{1}{2}$ W Carbon resistor (grid limiting resistor) in 6SN7 stage of main amp.

11/8/49 Photomultiplier radiation alarm
3mh Symptom - gradual drop off of sensitivity reported.
DFC check batteries as \sim 220V 1.5 meg resistors
1.8 - 3 mega - Replaced batt & all 1.5 meg resistors (1.6-1.8 mega range of available resistors from stores which were used) values checked after soldering in all but one 1.8 mega, that one 1.6 mega #1 & #2 dynodes had 1.0 meg resistors whose values were checked as ok. With max auxiliary sensitivity trip point (10mg) was \sim 8 inches Read by AC box controls to 3" KE

11/10/49 #4 D.C. AMPLIFIER - Replaced 6-45 volt batteries
3mh 1-12SH7, 2-12J5, 1-5Z3; cleaned ion chamber of dust - remedied our sensitivity and our control. (4/5)

11-11-49 #5 DC Amp. Replaced 1-45V battery 1-1 $\frac{1}{2}$ V, and 1-6V battery (all locs).
1mh Replaced 1 $\frac{1}{2}$ V in Brown and 1-7F7 tubes.

11/15/49^{2mh} Replaced 5Z3 rectifier, checked batteries wpt.

11-16-49^{1h} Replaced 300V battery in Photo-Multiplier

11-17-49^{1h} Photomultiplier returned to shop.

11-21-49^{1h} Firing circuit for scalars - replaced:
5V4, 2, 6SJ7's and 3, 2050's tubes rectifiers to have shorted out.

12-13-49 Vibrating Reed - 6SJ7; blank circuit 6SN7; repaired oscillator - wiring, needles, Fox 6 buttons

12/21 Reed Electrometer - tube insulators replaced and batteries in head suspended in air by thin bars - appears to improve stability.

JAN 10-1950 Reed Electrometer - checked tubes, replaced input 6AK5's - Replaced 2ea 6J5 in main amplifier. Removed reed head from BF₃ ion chamber, and reinstalled on Fron 8 chamber. End of test started late Nov. 1949.
6 hrs. #4 D.C. Amp. - placed in panel rack. Replaced 5 low 450-ohm batteries. (M.F.B.)

JAN 20, 1950 #4 D.C. Amplifier - checked tubes and batteries, loose cable connector could have caused trouble.
12 hrs #1 - Brown lined Counter - checked all tubes - replaced 6-46 6567 & V.R.105 - soldered loose 110V AC wire to switch.
#2 - Brown lined Counter - checked all tubes - replaced 26S67 26S67 - 4ea 6SN7 - 1ea V.R.105 - Reduced slightly the high voltage, scalar etc. H.A.K. & M.F.B.
#5 - Electrometer - shift does not seem excessive.

MARCH 1, 1950 #1 - Counter - Checked all tubes, replaced 1-6AC7 and 1-6SN7
2 hrs #2 - Counter - Checked all tubes, replaced 2-6AC7, 26SN7
1-6567
#3 - B.C. Amplifier - Checked all tubes and batteries. Replaced 2-807, 2-12J5, 1-6J5; 5-45V batteries. Replaced Brown recorder pen with a clean one and replaced 1 1/2 batteries in Brown. Tightened ground connection to water pipe reduced transient response.
#4 - D.C. Amplifier - Checked all tubes and batteries. Replaced 1-807, 1-6J5 and 1-954 (VII) preamp tubes. 3-45V Batteries. Changed A.C. connectors. Replaced 2-1 1/2 batteries in Spectromat recorder. (M.F.B.)
#5 - Electrometer - Remained on all day - instability noted in AM but left in 3 hours. Not fixed.
H.A.K. & M.F.B.

3-9-50 #6 Red Electrometer - checked tubes replaced
 657's and 655's reading low. Attempted adjust-
 ments on instrument. ^{ERR 2/8} Osc. adj. and gain settings
 set instruments in operating condition. Packed with
 full capacitance in. Changed R23 & R25
 from 82k each to 100k each, giving
 better peaking point, in capacitor range.
 2 hrs. HAK (ERR)

3-13-50 #1 scales - tripped register on 40 & 32 at times instead
 2 numbers of 64. Behaviour erratic. Replaced 657's on
 scale units reached 16 & 32. MJB - ERR.

3-15-50 #4 DC Amplifier - feature to gain adjust.
 6 months Checked tubes - all ok.
 Checked batteries - replaced 3 45-volt units.
 Checked pre-amplifier - tubes checked ok. Cleaned
 tubes & input resistor. Packed 1/4 month.
 Made adjustments for balance -
 MJB. ERR

3-20-50 - fets & survey meter - reading low.
 checked batteries replaced several 22.5 v. batteries
 and filament batteries. Instruments still seem
 poor in response.
 - #4 DC Amplifier - safety trip circuit bad. Trip point
 can not be set. ^{checked 655 & 2050} replaced 2050. Still not able
 to set trip point. Noted that could not coarse adj
 in such a position that trip light go on when chassis
 was struck or floor hit.
 - linear amplifier #2. Noted vibration, mechanical,
 thought possibly to have been "frying" capacitor. Checked
 instrument & noted noise ceased when one capacitor
 can was touched. Tightened capacitor on chassis. -
 S.P.L.

3-27-50
(PM)#5 - Electrometer - sensitivity low -

Checked batteries. Replaced 1.50 filament battery (reading 1.4 v) by new 1.50 battery. Then helped. ✓

#4 D.C. Amplifier - Meter affected by output scale change. Meter deflection peculiar. As scales lowered from 1000 deflection normal (increasing) down to 100 or 50 settings; then deflection of scale negative!! Setting diffused with zero swing, adjustment. Checked batteries, replacing 1.45, #738. Checked tubes - o.k.

Checked components in general circuit check. Noted few questionable sensitive components.

3-28-50

- Rechecked #4 - still as good -

Put another unit in place and tested, after putting in good batteries & tubes. Found to be good. Checked 3rd unit with little success. Trouble with Spardomax.

Put in good #4 unit. -

4-5-58 Instrument check:

#1 2 min cts: 12+8; 12+52; 12+35 ✓ 12+32

#2 2 min cts: 16+12; 16+9; 16+20 ✓ 16+44

Both instruments seem to be OK.

#4 D.C. Amplifier - subject to transients -

Set trip points: Meter x50 scale. Output x100.

Wants to try at 60 m sec (62 on meter)

@ 55 (58) not tri.

57 (60)

Tripped at 59 (62) reset fine trip

Tripped @ 63 (66)

@ 60 way trip when spurious pulse trip relay.

Seems safe at 55 - Trips above 55.

#3 OK


#5 OK.

#6 OK.

#7 OK.

4-12-58 #7 PM alarm - inoperative - replaced lead 3000's battery.

12/8/50



Time sec's (between arrival of pulse at counter & release of skin rod).

- 1.60
- 1.57
- 1.18
- 1.18
- 1.21
- 1.45

80 on trip scale. output meter
 Background = 10
 avg = 1.225

Background = 80 by adding Y under source.

- 1.20
- 1.09
- 1.05
- 1.22

trips at 80 on output meter
 1.14 avg
 .25

Background = 63 from Y's

- 0.98
- 1.02
- 0.96
- 1.01

0.99 sec avg

Background = 62 from Y's

Time without magnet - blocks across K-10 relay points

0.70 sec. (Time between arrival of source at
 chamber and opening of relay in DC
 Amplifier, i.e. deenergizing magnet).
 0.69
~~0.49~~
 0.78
 0.79
 0.79

0.70 sec avg.

Background = 40 from Y's

1.05 sec.

1.03

1.02

0.98

0.82

0.98

0.98 sec avg

Background = 10 (8 some removed)

1.18 sec

1.21

1.00

1.15

1.00

1.28

1.13 sec avg

~~45~~
~~1.13~~

Total time including 12" free fall of loaded
chain rod in air

1.87

1.81

1.64

1.63

 1.80 sec.

} Rod jumped causing approx .2 sec delay
possible avg = 1.60 sec.

1.60

 1.25

 .35

Total time inc. 12" free fall = 1.80 sec.

time no fall (1/32") = 1.25 sec

free fall .55 sec

Instrument alone no mag 1.13

Release time of mag .12 sec 10 background

" " " .16 sec 40 background

.29 sec 62 background

 3) .57

 .19 sec avg

With Instrument time ≤ 0.1 sec

total time = 0.70 sec estimated

Shim Rod Drop Time Tests:

Dry Run: 1020

0.4	24 in drop	- rejected
0.71	"	reasonable
0.81	"	

Wet Run	17.5" drop	} should repeat
0.62 sec		
1.50 sec		

Dry	17.5" drop
1.71 sec	
1.13 sec	
1.06 sec	

	15.5"	(note limit on set up 2")
0.52		
0.44		

WV

20 General Servicing -

Poppy Check - 4-28-57

Time	Source	Meters	Time	Source	Meters
0907	12kg	13	0910	98%in	1
5		0	5		11
10		0	10		4
15		0	15		0
20		0	20		5
25		5	25		14
30		2	30		13
35		3	35		5
40		3	40		8
45		2	45		10
50		8	50		19
55		3	55		9
0908		5	0911		12
05		0	5		7
10		4	10		8
15		5	15		13
20		0	20		9
25		0	25		6
30		5	30		3
35		0	35		7
40		8	40		11
45		5	45		11
50		11	50		6
55		5	55		5
0909		2	0912		12
05		0	5		8
10		5	10		0
15		3	15		1
20		5	20		8
25		0	25		8
30		0	30		6

Total 99

240

21 samples

36

avg: 3.19

avg: 7.74

①

②

③

12-20-50 (Scale Counting Rates) - source in "doghouse"
 West side 7 AF₂ South side 2 AF₂ East side 3 AF₂

reject	$21^{64} + 3$	$36^{64} + 38$	$10^{64} + 57$
	$5^{64} + 43$	$3^{64} + 40$	$12^{64} + 29$

Source down - between slugs - no water -

$15^{64} + 4$	$(784^{64} + 8)$	$27^{64} + 12$
$12^{64} + 2$	$(765^{64} + 46)$	$27^{64} + 29$

reject - loose coupling

Source down 4" from bottom (water) 93.7 cm above tank

3^{64}
 $3 \frac{3}{4}$
 $4^{64} 5$

2^{64}
 $2 \frac{1}{4}$

$2^{64} 0$
 $2^{64} \frac{1}{4}$

Source in "doghouse" water in tank

0.5

2.5

0.5

0.5


3.5

0.5



KUFFEL & ESSER
New York, N.Y.


Model Engineering Co. Inc.



KUFFEL & ESSER CO., N. Y. NO. 358-61
Small Engineering Co. Inc. 27th Street, New York, N. Y.

12
+29
12
+29

*Pulse Height
Input*



Age	Sex	Age	Sex	Region	Sub	Year	Day	P. Day
0.5	V	69	-1	20/23	27	57	8-1	51/53
0.5	V	69	-1	25	-	11	8-1	53/55
0.5	V	64	-1	25	55	61	8-1	100
"	"	"	"	26	-	-	4-1	25/27
0.5	V	69	-1	48	27	25	4-1	27/29
"	"	"	"	49	-	-	4-1	27/29
"	"	64	-1	100	ot.	-	4-1	28/30
0.5	-	32	-1	25	ot.	-	4-1	21/33
"	"	32	-1	25	N.C.	-	4-1	33/35
0.5	-	32	-1	36	N.C.	-	4-1	35/37
"	"	32	-1	36	-	-	4-1	37/39
0.5	-	82	-1	100	-	-	4-1	39/41
"	"	16	-1	26/28	-	-	4-1	41/43
0.5	-	16	-1	32/36	-	-	4-1	43/44
0.5	-	16	-1	36/38	-	-	4-1	44/45
0.5	-	16	-1	41/43	-	-	4-1	45/46
0.5	-	16	-1	45/47	-	-	16-1	21/23
0.5	-	16	-1	51/53	-	-	8-1	24/26
0.5	-	16	-1	100	-	-	4-1	25/27
0.5	-	8	-1	26/28	-	-	-	-
"	"	8	-1	29/31	-	-	-	-
"	"	12	-1	32/34	-	-	-	-
"	"	13	-1	34/36	-	-	-	-
"	"	14	-1	36/38	-	-	-	-
"	"	15	-1	39/41	-	-	-	-
"	"	16	-1	42/44	-	-	-	-
"	"	17	-1	44/46	-	-	-	-
"	"	18	-1	46/48	-	-	-	-
"	"	19	-1	48/50	-	-	-	-
"	"	20	-1	50/52	-	-	-	-

-15-57 Alvin Suck.

7-19-57 *Wave Number* calibration -

S: Pb 221

$N_1 = 3.54 \times 10^7$

$\frac{50^\circ \times 100}{100} \text{ eff} = 89.7\%$

$N = N_1 e^{-\mu x} = (3.54 \times 10^7) e^{-\frac{0.49 \times 6}{100}}$

$= 3.54 \times 10^7 e^{-0.297} = 3.54 \times 10^7 \times 0.74 = 2.62 \times 10^7$

$A = \sqrt{N} = 12571^2$


Flux = $\frac{N}{A} = \frac{N_1}{A} = \frac{3.48 \times 10^7}{1.257 \times 10^4 \text{ cm}^2} = \frac{277}{1.257} \times 10^3$

Counting $x \pm 3\sigma$	r (cm)	r ²	Flux (cp)	Flux (value)	Flux (value)
8	6.5	42×10^1	$\frac{2.47 \times 10^6}{3.6 \times 10^1}$	$.967 \times 10^5$	9.67×10^4
19.5	17	289×10^1	$\frac{2.45 \times 10^6}{4 \times 10^2}$	$.612 \times 10^4$	6.12×10^3
16.4	18	324×10^1	$\frac{2.45 \times 10^6}{9 \times 10^2}$	$.272 \times 10^4$	2.72×10^3
14.8	20	400×10^1	$\frac{2.45 \times 10^6}{3.6 \times 10^3}$	$.681 \times 10^3$	6.81×10^2
12.6	30	900×10^1	$\frac{2.45 \times 10^6}{1.6 \times 10^3}$	1.53×10^3	1.53×10^3
10.3	40	1600×10^1	$\frac{2.45 \times 10^6}{2.5 \times 10^3}$	$.98 \times 10^3$	9.8×10^2
8.1	50	2500×10^1	$\frac{2.45 \times 10^6}{3.6 \times 10^3}$	$.681 \times 10^3$	6.81×10^2
7.3	60	3600×10^1	$\frac{2.45 \times 10^6}{4.9 \times 10^3}$	$.498 \times 10^3$	4.98×10^2
5.8	70	4900×10^1	$\frac{2.45 \times 10^6}{6.4 \times 10^3}$	$.383 \times 10^3$	3.83×10^2
4.6	80	6400×10^1	$\frac{2.45 \times 10^6}{8.1 \times 10^3}$	$.302 \times 10^3$	3.02×10^2
4.1	90	8100×10^1	$\frac{2.45 \times 10^6}{1.00 \times 10^4}$	$.245 \times 10^3$	2.45×10^2
3.7	100	10000×10^1	$\frac{2.45 \times 10^6}{1.44 \times 10^4}$	$.170 \times 10^3$	1.70×10^2
3.0	120	14400×10^1	$\frac{2.45 \times 10^6}{1.96 \times 10^4}$	$.125 \times 10^3$	1.25×10^2
2.5	140	19600×10^1	$\frac{2.45 \times 10^6}{2.56 \times 10^4}$	$.96 \times 10^2$	9.6×10^1
2.0	160	25600×10^1	$\frac{2.45 \times 10^6}{3.24 \times 10^4}$	$.756 \times 10^2$	7.56×10^1
2.0	180	32400×10^1	$\frac{2.45 \times 10^6}{3.96 \times 10^4}$	$.618 \times 10^2$	6.18×10^1
2.0	200	40000×10^1	$\frac{2.45 \times 10^6}{4.84 \times 10^4}$	$.506 \times 10^2$	5.06×10^1
2.0	230	52900×10^1	$\frac{2.45 \times 10^6}{5.29 \times 10^4}$	$.463 \times 10^2$	4.63×10^1
2.0	260	67600×10^1	$\frac{2.45 \times 10^6}{6.76 \times 10^4}$	$.362 \times 10^2$	3.62×10^1
2.0	290	84100×10^1	$\frac{2.45 \times 10^6}{8.41 \times 10^4}$	$.291 \times 10^2$	2.91×10^1
2.0	320	102400×10^1			2.42×10^1
2.0	350	122500×10^1			2.34×10^1
2.0	380	144400×10^1			2.42×10^1
2.0	410	168100×10^1			2.07×10^1

Phys 111

calibration outside west end.

Neutron Flux (n/s/cm²)



n/s/cm²

Meter Reading (m)

X3

X20

Neutron Survey Meter

Calibration

719/sr

Source: Cobalt-60

Source: P8221

7-16-57 Neutron Meter Calibration

$$\frac{N}{N_0} = e^{-\lambda t}$$

Sources: $395^{101250} \quad 887 \times 10^6$
 7.88×10^6
 $16.75 \times 10^6 \text{ n/s}$

$t_h = 140 \text{ d}$ $2 = e^{-\lambda \cdot 140}$

$10/12/50 - 7/18/51 = 261 \text{ d}$ $279 \cdot .693 = \lambda \cdot 140$ $\lambda = \frac{.693}{279}$

$N = (16.75 \times 10^6) (e^{-\frac{.693}{279} \cdot 261}) = \frac{4.67 \times 10^6 \text{ n/s}}{4.19 \times 10^6}$

$A = 4\pi r^2 = 12.57 r^2$

Flux $n/cm^2/sec = \frac{N}{A}$

r (cm)	r ²	A	N/A n/cm ²
10	100	1257	3.67 x 10 ³
15	225	1886	1.63 x 10 ³
20	400	2513	9.17 x 10 ²
25	625	3142	5.97 x 10 ²
30	900	3771	4.08 x 10 ²
40	1600	5027	2.29 x 10 ²
50	2500	6283	1.46 x 10 ²
60	3600	7539	1.02 x 10 ²
70	4900	8796	7.49 x 10 ¹
80	6400	10052	5.74 x 10 ¹
90	8100	11309	4.16 x 10 ¹
100	10000	12566	3.67 x 10 ¹

$N = 4.67 \times 10^6$
 $\frac{4.67 \times 10^6}{12.57 r^2} = \frac{3.72 \times 10^5}{r^2}$

results

r	r ²	Flux	Meters	Eq 1,1
10	100	3.34 x 10 ³	—	2.9
15	225	1.48 x 10 ³	—	2.1
20	400	8.34 x 10 ²	19.7	2.0
25	625	5.34 x 10 ²	17.0	
30	900	3.71 x 10 ²	16.0	
40	1600	2.08 x 10 ²	12.5	
50	2500	1.33 x 10 ²	10.5	
60	3600	92.8	8.5	
70	4900	68.2	6.9	
80	6400	52.2	6.4	
90	8100	41.2	5.7	
100	10000	33.4	4.8	

calibration in 210

10w. 1/2 emitted.

7-19-57 G. Shute calc - $2 \times 10^{11} \text{ g/s} = 3 \times 10^{11} \text{ g/s}$

$$\text{Flux} = \frac{3 \times 10^{11}}{4\pi r^2} = \frac{3 \times 10^{11}}{1.26 \times 10^6 \times (6.55)^2 \times 10^6}$$

$\theta = 85'$

$$= \frac{3 \times 10^{11}}{5.26 \times 10^{12}} = 5.7 \times 10^{-2} = 5.7 \times 10^{-2} \text{ g/cm}^2$$

$\frac{85 \times 30 \text{ cm}}{2.5 \times 10^2}$

Gamma lens check no refl. to 5m Ra @ 6 cm.
 Cadman shield - no change in readings.

7-31-57 2 fold lead counter - HV X tube burnout

2X2A checks OK. -

2C53 checks bad -

Noted large collection of bugs in unit,
many at HV points.

Removed X tube, Hallenon P1932A

8-9-57 ^{summary} Replaced X tube - coated H.V. receptacles with
plexiglas dope to insulate against future
bug shorts.




In checking found HV did not adjust. Checked
divider circuit. Noted 10 Mega resistance from
filament of 2X2A to ground. 2X2A replaced!

(There was 10. MΩ ~~to~~ between 18V and 3 (W3)
which was used as terminal for 2C53 Htro in unit)
OK.

C₁ - C₂ Linear Amplifier Voltages & Waveforms. 29

Linear Amplifier - characteristics

Waveforms Input: 1 μ sec, neg, 0.5V, repeats 2ke-



Voltages: Simpson B+ to amp strip 300v
Reduced B+ 225v

6S57	E _{s1}	77	E _{p1}	96	E _{c1}	10
6SA7	E _{s2}	112	E _{p2}	40	E _{c2}	10.0V
6SA7	E _{s3}	179	E _{p3}		E _{c3}	42
6SA7	E _{s4}	103	E _{p4}	185	E _{c4}	10
6AG7		116		36		
6AC7			300v			42v
6SN7	g ₁	42	P ₂₁₀	k ₁	46	
	g ₂	130	P ₃₀₀	k ₂	140	

Resistor on 300v point $\approx 1/2$ cm amp gain $\times 100$ at 1

30 C₁ - C₂ - C₃ Reproducibility of Count Check: 11-1-57

A.M. C₁: Met lat head bare

C₂: GE head in paraffin

Two minute counts

C₃: Met lat head in paraffin

Scale	C ₁ count	%m	Scale	C ₂ count	%m	Scale	C ₃ count	%m
17 ⁶¹	1149	574.5						
19 ³⁴	1250	625						
18 ²⁶	1178	589						
18 ¹⁹	1171	585.5						
18 ⁵²	1204	602						
19 ⁴⁰	1256	628	72 ¹¹	2812.5		33 ³⁷		1074.5
19 ⁴²	1258	629	68 ⁵¹	2201.5		32 ⁵		1026.5
19 ⁰⁵	1221	610.5	67 ⁴⁹	2153.5		32 ¹¹		1029.5
19 ³⁴	1250	625	70 ¹⁰	2245.0		32 ¹⁹		1031.0
18 ¹³	1165	582.5	68 ¹⁰	2181.0		32 ³²		1046.0
Avg:		605.1		2218.7				1040.3
max dev:		-30.6		+93.8				+34.2
% dev:		5.1%		4.3%				3.0%

P.M. C₁: GE head in paraffin

C₂: Met lat head bare


C₃: Met lat head in paraffin (uncharged)

Scale	C ₁ %m	Scale	C ₂ %m	Scale	C ₃ %m
43 ³	1377.5	28 ⁴³	917.5	33 ³⁴	1073
44 ³⁷	1426.5	28 ⁴⁴	918	33 ¹²	1062
44 ⁶³	1439.5	28 ⁴³	917.5	33 ³⁸	1075
44 ⁴⁹	1400.5	26 ⁴⁹	836.5	33 ¹⁷	1064.5
42 ⁵³	1370.5	27 ⁵¹	889.5	31 ⁶¹	1022.5
Avg	1402.9		895.8		1053.4
max dev	+36.6		-59.3		-36.9
% dev	2.61%		6.62%		3.48%

12-13-57 Check 204-B LA. Characteristics

input: pulse from NICE pulse gen
 input pulse negative - 1 μ s @ 10⁵ c/s

	am-amp	main	diff
input lv.:	scope gain x 10 - 1	att x 1	2 cm
output:	scope gain x 1 - 1	att x 1	2 cm



LA. .5, 2 μ s - Δ
 .8 μ s Δ
 1.2 μ s Δ

Pa Source

50.0 mph 1.0 ft

12.5 mph @ 2.0 ft

5.55 mph 3.0 ft

3.125 mph 4.0 ft

~~50~~ 10.0 ft

1.3875 6.0 ft

200 .5 ft


200 .25

12-20-37 RCL Rate Meter

operation checks with G/M Tube

source dist		Sl.	R.M.	Sc	sum	HV
2.5'	8 m/ft	.5	100	(230)	5	840
3.0	5.55	.4	80	(230)	5	840
2.0'	12.15	.65	135	(230)	5	840
1.5'	22.2	.93	180	(230)	5	(840)

2/5/2 Signal Generator, pulse type, for checking response of recorder -




2/2 Checked H/D Speedmax -
using following set values for R_1 -

$$R_1 = 220k$$

$$R_2 = 7.5M$$

$$R_K = 10K$$

C C-D decade capacitor -



3-3-52 Poppy (K-25) - bkg check - no HV. in cable.

counts/min (1m) :	Channel 1	Channel 2
	5	4
	3	5
	2	7
	3	3
	5	1
	8	2
	8	5
	4	8
	2	2
	2	7
	<u>130</u> Avg 3.0	<u>144</u> 4.4

Seems to be a difference between channels.
 Run voltage check using Tronax 870m.
 No pre-amp. tubes still.

Input receptacles: pin A and at selector switch 396v.

- A 4.1 & 2 270v
- C 5.25 v ac.
- D at sel. switch common 0 v.

VT 1	pin 3,5 (K)	5.6v	VT 2:	3,5	1.9v
6557	4 (G)	probe connected avg = 0.2v	6557	4	0
	6 (Sc)	2.6v		6	16.8v
	8 (P)	3.8v		8	2.8v
	E _A	270v		E _B	295v

VT 3	3,5	0 v.	VT 4	5(G) = 3.5 *	20.5v
6557	4	0 v	2050		
	6	13.4v			
	8	27.8v			
	E _B	29.6			

E_B values high: check to see if C₁₅ & R₃₄ are in ckt. Missing
 put on pre amp checked pre-amp voltages at input recpt.

- pin A 29.5
 - B 2.65
 - D 14.5
 - K 10.4
- voltages too high for pre-amp

Calculations - current values:

Pre-amp: $I_R = \frac{19.5 \times 10^{-5}}{1.01} = 1.93 \text{ ma}$
 $I_S = \frac{5 \times 10^{-3}}{27} = .185 \text{ ma}$

VT 1: $I_K = \frac{5.6 \times 10^{-1}}{51 \times 10^2} = 1.1 \text{ ma}$
 $I_G = \frac{270 - 26}{10^6} = 2.44 \times 10^{-4} \text{ a. } .24 \text{ ma}$

$I_P = \frac{270 - 38}{2.7 \times 10^5} = \frac{232}{2.7} \times 10^{-7} = 86 \text{ ma}$ ✓ check ✓

Poppy (6-25) - 100V, 200V, 300V, 400V, 500V, 600V, 700V, 800V, 900V, 1000V

Channel	Channel
1	1
2	2
3	3
4	4
5	5
6	6
7	7
8	8
9	9
10	10
11	11
12	12
13	13
14	14
15	15
16	16
17	17
18	18
19	19
20	20
21	21
22	22
23	23
24	24
25	25
26	26
27	27
28	28
29	29
30	30
31	31
32	32
33	33
34	34
35	35
36	36
37	37
38	38
39	39
40	40
41	41
42	42
43	43
44	44
45	45
46	46
47	47
48	48
49	49
50	50

Seems to be a difference between channels
 Run voltage check using nominal 570V
 no pre-amp - cables still.
 Input receptacles: pin A ^{unmarked} and at collector marked 296V
 B 210V
 C 220V ac

Direct control common 10V

VT 1	pin 2 (G)	570V	VT 2 (P)	10V
6557	4 (G)	26V	6 (G)	188V
	6 (G)	26V	8 (P)	28V
	8 (P)	28V	En	270V
	En	270V		
VT 3	5 (G)	0V	VT 4	5 (G)
6557	4	0V	7050	
	6	18.4V		
	8	27.5V		
	En	296V		

En values high - check to see if C1 & P3 are in ckt. ^{pinning}
 put on pre-amp - checked pre-amp voltages at input repts.
 pin A 295
 B 265
 D 19.5
 K 10.4

Calculations - current values

Pre-amp: $I_p = \frac{195 \times 10^{-3}}{10} = 19.5 \times 10^{-3} \text{ A} = 19.5 \text{ mA}$
 $I_s = \frac{103 \times 10^{-3}}{10} = 10.3 \times 10^{-3} \text{ A} = 10.3 \text{ mA}$
 VT 1: $I_p = \frac{5.6 \times 10^{-3}}{10} = 0.56 \times 10^{-3} \text{ A} = 0.56 \text{ mA}$
 $I_s = \frac{270 - 26}{10} = 24.4 \times 10^{-3} \text{ A} = 24.4 \text{ mA}$
 $I_p = \frac{270 - 38}{2.7 \times 10^3} = \frac{232}{2.7} \times 10^{-3} = 86 \text{ mA}$ check

Current calculations -

$$\text{current thru } R_1 = \frac{25}{20} \times 10^{-3} = 1.25 \text{ ma check}$$

$$VT2: I_{R11} = \frac{1.9}{115} \times 10^{-3} = 1.3 \text{ ma}$$

$$I_{R8} = \frac{2.93 \times 10^2}{10^5} = 2.9 \text{ ma check ?}$$

$$I_{R9} = \frac{9}{2.7} \times 10^{-5} = 33 \mu\text{a}$$

$$I_{R10} = \frac{16.8}{68} \times 10^{-3} = .25 \text{ ma}$$

$$I_{R11} = \frac{295-17}{100} = \frac{278 \times 10^2}{10^4} = 27 \text{ ma } \left. \begin{array}{l} \\ \\ \end{array} \right\} \text{ check } \checkmark$$

$$VT2: I_{R11} = \frac{295-18}{100} = 2.77 \times 10^{-4} = .277 \text{ ma}$$

$$I_p = \frac{295-27}{2.7 \times 10^5} = \frac{2.68 \times 10^{-3}}{2.7} \mu\text{a}$$

$$1.1 + 1.2 + 3.0 + .3 + .3 + 1 = 5.9 \text{ ma}$$

$$295 - 200 = 95 \text{ V}$$

Inverted Resistor Chain:

Voltage readings:

Input receptacle A (at aw) 252v

B 226 V

C 5.3AC

D at aw 210 V

E 13.0 V

VT 1	3.5	4.7	VT 2	3.5	1.54	VT 3	3.5	0
	4	-1.2		4	0		4	0
	6	23.5		6	13.4		6	16.4
	8	32		8	23.3		8	23.4
	B	226		B	244		B	244

$$VT4: 5 = -35$$

$$3 = 295$$

$$B = 297$$

$.2 \times 10^{-3} \times 500 \times 10^3 = 100 \times 10^0$

Exp pump
(1700)

Partic. ^{dropping} ~~deposition~~ for pump (270k .1 pfd)
Blg still rather high

24-52 Shunted R₉₀ .68 Meg with 39 Meg for radium sensitivity.
Seems better operating @ 2800V (meter ^{low} vert). Plugged
scales in for check of blg & sample counts.

Time	Sample	Count	Count	Rate
3m	Blg	0 ¹² 12	12	4 c/m
"	99	3 10	58	19.3 c/m
"	Blg	0 14	14	4.7 c/m
"	99	4 4	68	22.7 c/m
1m	516	5 1	81	81 c/m
"	"	5 12	92	92 c/m
"	"	5 9	89	89 c/m

4.25
21.0
4.82/1
87.3
20.07/1

Summ of 200 counts:

Blg	99	516
4	7	30
1	8	30
1	18	25
1	7	32
0	4	25
1	5	22
3	3	35
2	8	24
2	7	25
3	9	26
18	71	282
Avg	18	7.1
	384/1	1567/1

Blg	10 m	3011	43	4.2 c/m
99	"	1412	236	23.6 c/m

57.09/1

next page

2/4 Check on rate meter - scale examination removed.
 ayell. 99% sample for 15 sec. (max. reading recorded) 15 sec nit - note
 blk for 15 sec (note max) 15 sec lag.

s	15	s	34	s	11	s	18	
b	5	b	10.8	b	9	b	5	
s	9	s	20	s	10	s	20	
b	2	b	5	b	5	b	8	
$\sum s = 368$	$s_c = 18.15$	s	24	s	24	s	30	
$\sum b = 117$	$b_c = 5.85$	b	5	b	10.8	b	5	
3.10%	s	24	s	17	s	14	s	17
	b	2	b	10.8	b	2	b	7
	s	13	s	12	s	24	s	25
	b	7	b	5	b	10	b	7

ayell sample 20, Hank 20, blk 20 (max).

s	34	s	27	s	21	s	23	s	26
b	9	b	7	b	12	b	7	b	10
s	24	s	29	s	16	s	23	s	17
b_{10}^{10}	10	b	5	b	10	b	5	b	8
$\sum s = 240$	$s_c = 24.0$								
$\sum b = 82$	$b_c = 8.2$					} 2.92/1			

Above readings taken with chassis^(low) on work bench. —
 Placed in dolly, panel up chassis vert.
 Excess blk, again noted — change from above.
 Hard to detect 100 cpm. — Still needs work. if
 to be used in cart. of or horizontal surface chassis would
 ok. —

3-13-52 Poppy Servicing -

Early morning checks showed HP poppy in fair shape -
good for 100 & more for same.

Poppy (K-25) responds to operation peculiar - shielded
whole under side of chassis.

4-4-52 log N amplifier check

cal. at ORNL setting -

ORNL

C-C 16 1/2" s.w. d.h.



pp. machine 0022

s.w. d.h.

0027

log house reduces radiation by factor of 2, with lead brick as shield found change in power from 001 to 002

Chamber 2 1/2' from d.h. opening



(center)

Source on p.p.

width 0.0042

27" 0.0012

18" 0.002

13 1/2" 0.003

7 1/2" 0.007

6 1/2" 0.009

5 1/2" 0.011

app 0.018

Chamber
600V on

source "i" 4" from

4" from 0.02

6" 0.015

8" 0.010

10" 0.007

12" 0.0055

14" 0.004

16" 0.003

20" 0.0023

25" 0.0018

30" 0.0012

35" 0.00085

40" 0.00070

45" 0.00062

d.h. 0.00057


3000V - 0.017 max

4-17-52 Rectifier for relays.

Acct. 1 Fantriel BE304M -80v 4a-

1 each relay 2128 $R_c = 540 \Omega$ R_c of other type is same.

bridge cbt. 1-2:1 step down 250W Xform



$$V = 46v$$

$$A = 82ma$$

} no filter capacitor

18 relays in parallel -

$$18 \times 82 \times 10^{-3} = 14.77 \times 10^{-1} = 1.477 \text{ amps}$$

30 Ω 48v.

1.53 amps -

$$2.25 \times 30 = 75 W$$

70ma / Ω @ 3105 Ω 4-18-52 1010 - setup as above diagram with R_c equiv. of

30 ohms. Fan cooling resistors

$$V = 40.4v$$

$$A = 1.25 A$$

1210 Ω -

$$\left\{ \begin{array}{l} V = 40.8 \\ A = 1.25 \end{array} \right\}$$

5-9-52 Neutron Detector Rechecked - For RCR - LRM

MV adj: counter clockwise full bias
 Bias: counter clockwise full bias

MV bias: 22.40.

bias reduced to free running point 2.39 volts

then turned back & set at 2.50.

meter was on bias setting no source in room.

400.

3.7 begin to get a pulse or pick up pulses

3.4 pick up 11-source


messy - to be taken up later.

5-9-52 Robert Reed Check:

- 1. output norm - OK
- 2. ~~Output Norm~~ / 100 sec - 5V .46 - .40 mV .06 mV
- Drift Rate } drift rate .0006 mV/sec
- 100 sec .46 - .42 : .0002 mV/sec

- 3. Contact Potential - /
- 4. Unbalance - -.2 mV. OK.
- 5. Over-all Gain Check -
- Test - .31 mV Sec: 11 mV

All Tubes
CK512AX



	MV MAX i	MV CONF i		
5-28-52				
T_{41}	-1.23	—	T_{51}	-58
$\frac{e_1}{C}$	$\frac{-22.8}{-28.2}$	$\frac{-1.26}{-7.0}$	T_{52}	$\frac{-1.05}{-1.06}$
T_{42}	11	13	T_{54}	11
T_{43}	44	27.8	T_{55}	13

T_1 amp in - input V 1.2 1.5 μ
 T_2 disc 2.2 μ 2.5 V
 T_3 in 2.2 V 2.5 μ
 $T_{4,5}$ MV

* Just in case

recheck of gains in subband light

T_{12} - 2.7 T_{22} 2.9 T_{32} -1.26 T_{42} -22.7 T_{52} -1.2
 T_{11} -2.28 ϕ 2.5 ϕ -1.58 D -23.4

T_2 : 1-3 μ T_1 loop 2 - .62 μ
 2 - .18 μ 3 - .63 μ
 8 (1.1.)

HV - discuss.

-4.

sig - 6.2 μ 60 μ

Observed oscillation on T_2 -3 when disc set defocus on incoming 10 pulses. T_3 removed (pair 2, 4, 5).
 Replaced T_2 - new tube behaved similarly.

tube not amplifying grid neg plate neg by observing scope trace. Changed plate resistance to 560k put in new 100 capacitor - kill this out out internal oscillation. Operation seems to be improved.

Put patch leads on top side of etal Disc control & read discriminator volts. Meter in charge half glass - Adjusted -

A. MV adj. (disc at max bias)

i. turn clockwise until meter indicates free-running mv.
 ii. back off so as to stop meter deflection (may get 50-5 pa deflection with proper adjustment.)

B. Discriminator -

5-20-52 Winter survey check-up -
adjustments -

B-discriminator:

i - turn clockwise til "background" causes meter reading.

ii back off to level respecting "bkg" (reading $\leq 1 \mu a$)

Disc bias Meter x3

max 4.23 volts

@ 3.75 volts 1.0

with source in position $m = 15 \mu a$

used Disc.

to 0.5 more linking

mv adj. 2.6v bias

disc 4.2v re-pulse. read 15 μa at 4.3' (R3)
4.5v.

2 x 300

3 x 30

6-3-52 Continued check. Source not in room (in paraffin rm 201 W201)

Re-check of mv point - bias ≈ 2.52 v bias set @ 2.6v.

with disc field in (max. bias) meter reads $\approx 2.3 \mu a$.

Attached voltmeter to discriminator pot.

max 24.2v.

drift up @ 4.05v. drift increases as disc lower -

S 221. bias 4.05-4.10 disc.

disc.	source	meter scale	disc.	meter scale	disc.
set at 4.5v:	bkg.	.25 μa -3	2440	15	3.2
19.35	100 cm.	2 μa 3	5950	10	5.1
26.4	150 cm.	4 μa 3			1
38.1	125 cm.	3 μa 3	with source in paraffin		.8
54.5	100 cm.	5.5 3.		180	3.7
93	80 cm.	7.5 3		150	4.5
140.7	65	10 3		125	5.5
238	50	12 3		100	8-12
372	40	12 3		80	8-12
484	35	16 3		65	10-16
661	20	16 3		50	12-20
952	15	2.1 30		40	0.5, 3
1478	10	2.1 30		35	2.0 30

disc. 4.25

Source	meter scale	disc.
30	2.5	30
25	2.8	20
20	4.2	20
15	11.3	20
	5.9	20
10	9.0	20
	1.9	20
6	2.5	20
	11.5	20

* 180	1.25	3	4.25
* 150	1.50		
* 125	2.0		
100	2.5		
* 80	3.5		
* 65	4.0		
50	5.0		
40	6.0		
20	10.0		
20	16.0	3	
	11.5	30	
10	2.2	30	
	0.5	3	4.5

Source	meter scale	disc.	4.05
180	4-6	3	
	5.0		
150	2-9		
	6.0		
125	5-12		
	8		
100	6-15		
	10.5		
	13.5		
80	1.0	20	
	1.2	20	
65	1.5	3	
50	2.5	3	
	2.1	30	
40	2.4		
30	2.4		
	2.5	20	
20	6.0	20	
	7.0	20	
10	12.2	20	
	10.3	20	
6	9.0	30	
40	1.5	3	4.5

bly - 0.3 ya. 45
 drift p. 3.9.
 disc 5.0
 20. 10.0
 20 5.0
 disc 5.5
 20 0.3
 20 0.3 - cut off
 out of point.
 @ 10. disc. very close to
 response @ 5.4.
 for source.

Source P13 2.1 p 25

$$N_0 = 3.59 \times 10^7 \times 14.93 \text{ c.}$$

$$N = N_0 e^{-\lambda t} = N_0 e^{-\frac{0.693}{140} \times t(326)}$$

$$N = 3.59 \times 10^7 e^{-1.57}$$

$$= 3.59 \times 10^7 \times 0.208$$

$$= 7.47 \times 10^6$$

Source	meter scale	disc.	4.05
180	3.24 x 10 ⁴	1.35 x 10 ¹	
150	2.25 x 10 ⁴	2.64 x 10 ¹	
125	1.56 x 10 ⁴	3.91 x 10 ¹	
100	10 ⁴	5.65 x 10 ¹	
80	6.4 x 10 ³	8.0 x 10 ¹	
65	4.22 x 10 ³	14.97 x 10 ¹	
50	2.5 x 10 ³	23.9 x 10 ¹	
40	1.6 x 10 ³	37.2 x 10 ¹	
35	1.23 x 10 ³	48.1 x 10 ¹	
20	9.0 x 10 ²	66.1 x 10 ¹	
25	6.25 x 10 ²	95.2 x 10 ¹	
20	4.0 x 10 ²	149.7 x 10 ¹	
15	2.25 x 10 ²	244 x 10 ¹	
10	10 ²	565 x 10 ¹	
6	36 x 10 ¹	165 x 10 ⁴	

	41			42		43		44	45
	D	m	3	D	M	B	M	DM	M
123	150	3.5	3	6		2.5		3	2.5
130.3	140	7		8		5		4	5
135	140	8		12.5		8		6.5	5
152	80	12		12.5		11		9.5	7.5
238	50	1.5		0.5		19		16	13
272	40					0.5		0.5	15.5
161	30								0.5

same as hill Run 22

In checking PM work C.P. → (Had noted in B, on page, signal drop off at v² 1100 v.) 57

6-19-52 Photomultiplier Testing:

- (A) 5810 in closed case - (5.0 mag units in divider)
- (B) 5819 in sample changer (10.01 90)

High Voltage Supply - ship built

(A) Canada - 10⁶ a PB 221 source near PM.

V	Source	Pattern
600	max sens.	1/2 - 2/3 an.
700	"	2 - 2 1/2 an
750	"	3
800	4 att K2	2-3
850	4 att	3-3 1/2
900	X4	3
1000	"	2 1/4

> 1100 signal over loads - no drop off

Pattern appears to be satisfactory

(B) R₁₀⁶ a

600	max sens	3 1/2
700	x8	1 1/2 - 2
750	"	1 1/2
800	"	1 3/4
850	"	2 - tending to over load.

(A) 10⁷ a R₁₀ pattern saturates 10³ v > 1000 drops off to lower level pulses.

(A) 10⁵ a R₁₀ operates off up to v² 1100 v. pulses much shorter

6-20 - Alarm Circuit Check - applied bias to input (5 mag) res.

Battery connected across resistors. read panel meter, & voltages on grid of 6577 & on grid of 2050.

Bias:	0	.1	.2	.3	.4	.5	.6	.7	.8	.9	1.0	trip
6577 Plate	45	53	61	69	77	85	93	100	111	122	130	132.5
2050 Grid	-92	-83	-75	-67	-58	-50	-42	-32	-25	-13	-6.5	-2.5
Motor	.82	.80	.774	.76	.735	.715	.694	.665	.64	.625	.60	.58


readings taken bias varied by changing sens. control

6-20-52. Second Run. Bias externally varied. Sensitivity at maximum.

Bias	0	.1	.2	.3	.4	.5	.6	.7	.8	.9	10	4 trips
Plate 1	46	50	41	70	80	89	$\frac{99}{102}$	112	117	125		
Grid 2	91	93	75	64	56	48	$\frac{37}{33}$	23	17	10		
Meter	82	80	78	75	72.8	70.5	$\frac{67}{68}$	64	62.5	60.6		

* scale change 92

Circuit:



Photomultiplier: 5819 + anthracene. Bleeder 5 meg. units in divider.

Voltage range for $R_a = 10^6 \Omega$ 900-950 v.

5819 with bleeder of 10 meg. Ω

Range for $R_a = 10^6 \Omega$ 700 - 950 volts. -

Tubes are of two types & cannot be interchanged without changing socket connections.

6-22-52. Invented a HV fil. cf. f. 7, V_{a1} , & PM-1 detector alarm.

meter reading on alarm not rise above 20 μ 1800.

When all power in or system, alarm on without changing setting.

Operating range 750 - 1000 volts.

Notes light leak thru socket.

light shielded partially from sun range 750-1200.

Changed divider on (A) (Was 410 meg Ω 8.2 meg.)
very sensitive to B.

Range 750-950 v.


6-24-52 Mounted collimator follower on PM head.
 With operating range with 100' HV line,
 in original line (82kV) & exp. to battery supply.
 Range from 800-1200 V/Hs. Could operate
 up to 1300 & above but safe range is 800-1200
 Left HV at 1200 to see if false scan seen.

6-25-52 Response time check.
 source in block in test rig.

Drop time (sec)	High Voltage	Total Time
.41	—	—
.468	900v.	.485
.412		.478
change l.s. .405		.437
.404		.468
.406	1050	.448
.413		.444
.413		.440
at same distance		.444
940v firing threshold	1150	.428
same carrier		.424
3" from PM base		.428
dist N 2 1/2" to		

7-2-52 Power failure - Replaced 45V battery (4) in DC-3
 other batteries check OK.

7-22-52 log N chamber - battery box

log N works on +
polarity

7-23 Made up battery box for need - req. 600V.

7-28-52 RCL rate meter not reading on meter - to shop
for servicing & relocation of recorder of scale joints.

RCL

Rate
meterto shop - worked after removing & checking dials - all
good. After being on for while stopped.7-29-52 Instrument on h.v. & calibrated - checked with 600
mg pulses 1 μ sec - 4V peak. OK - traced signal
for reference.

OK. 0917 - OK 1205 130 P. ok turned off

9-30-52 Relocated recorder & scale outputs - on top of
chassis behind meter -
scale - coax receptacle
recorder - phone plug

1222 power on -

1223 600 μ s, 4V neg. pulses in,
switched 30-10: meter off head of negative
back to 431 negative meter -

power off then on other N 15 sec. OK

Took voltage reading when removal - Recorded
in booklet RCL

Voltage changes noted when reading bad;

		Good	Bad
V ₆	1	21.4	36.5
	2	10.4	8.1
	3	23.8	36.5
	4	31.0	56.5
	5	10.8	83.5
	6	23.8	36.5
V ₇	1	.01	22.8
	2	5.58	36.6
	3	21.9	22.8
V ₅		OK	OK

With no signal - scale changes don't affect meter -
change from 100 → 30 O.K.

30 → 10 O.K.

10 → 3 Bad

30 → 100 bad


Selector switch open - driving grid of V7B up.

8-11-52 Off-scale response noted when turned on.

R/L

Retrometer

Brown Check Signal input for test



$$I R_C + E R_V = 1.5$$

$$I (R_C + R_V) = 1.5$$

$$I R_C = 0.70$$

$$\frac{R_V}{R_C} = \frac{19}{1500}$$

$$R_o = 150k$$

$$R_C = 1000 \Omega, R_o = 150,000 \Omega$$

Applied signal (O.K.) to Brown on DC-2. Reader checked
OK. Fed output of DC-3 into reader. OK. Trouble seems
to be in DC-2.

8-7-52 Hand Counter Pu 201 in test. Shop 9787 for servicing.
Hand 14k voltage transformer had ~~to~~ burned out
Counter D.F.C. installed heater - timer unit 8-5/16

8-12-52 Cured off-scale response by replacing V_g.
RCL-RM. Could not zero. Proper choice of V_g fixed this.

8-14-52 Installed RM in Rack in 201 of test. O.K.

9-2-52 Reactivation of B-Y scintillation counter project
BY semi ch: (Reference II p. 228-249)

Voltage divider consideration:

Ball uses: $\frac{50k}{50k, 50k, 100k, 100k, 100k}$ in net, with
1.5V from test diode to grid.


negative H.V. supply

Wired tube socket as above:

Anode R_o = 150k output coupling of 100k

output to scales thru 47pF - oscill scope on
output. RSM Reading HV

Position HV NCC 1090.



Scaler for 100 sec. gain 2 divs +25
 HV = 700 60 x 256 + 135
 61 x 256 + 179
 59 x 256 + 113
 0 x 256 + 116
 0 x 256 + 146

9-4-52 Readings from instruments in 215.

End window counter: 1	Std	2571 %m	902/1
2	blk	28.5 %m	
2	Std	311.2 %m	139/1
For Reference	Std	2698 %m	107/1
	blk	25.1 %m	
	Std	2570 %m	95/1
	blk	27.1 %m	
Propagational counts	Std	17780 %m	250/1
	blk	71.1 %m	

Stats on PM, etc -

NO.	Source	Time	Scaler	Gain (M)	Count	Rate %m	sig/blk	net. d
600	blk	20	10 ¹⁶ + 13	2 254	653	21.8	5.71	702.7
	std	4	4 ¹⁶ + 2	"	498	124.5		
640	std	2	17 ¹⁶ + 52	"	11508	5754	102.35	5700.4
65	blk	5	16 ¹⁶ + 12	"	268	52.6		
680	blk	3	12 ¹⁶ + 13	"	205	68.3	125.67	9197.1
	std	1	144 ¹⁶ + 50	"	9266	9266		
720	std	1	183 ¹⁶ + 8	"	11780	11780	162.3	11647.9
	blk	5	22 ¹⁶ + 9	"	361	72.6		
760	blk	3	17 ¹⁶ + 14	"	296	95.3	152.19	14477.7
	std	1	227 ¹⁶ + 45	"		14573		
800	std	1	295 ¹⁶ + 50	"		15730	145.6	15622
	blk	2	13 ¹⁶ + 8	"	216	70.8		
850	blk	2	18 ¹⁶ + 11	"	299	145.5	112.9	16283.5
	std	1	256 ¹⁶ + 45	"		16429		
900	std	1	265 ¹⁶ + 31	"		16991		16789.5
	blk	2	25 ¹⁶ + 3	"	203	206.2	84.22	100

Scaler on best -
 Standard accuracy
 1000
 1000
 1000

9-4-52

9-4-52 Run using 1060 scaler after adjusting pres set - to accept smaller pulses coupling ~~direct from~~ mode PM than 50µm do inputs.

(WIC)
(RSM)
H₂ &
H₁ Fe

Source	Time of Count	Scaler Reading	Gain Att. %	Total Count	% Rate	57/6kg
900	std	1 m 263 56	2 25	16888	16888	96.7
"	"	" 247 45	4 25	15853	15853	123.8
"	"	" 222 34	8 25	14242	14242	176.3
"	"	" 150 24	16 25	9624	9624	212.9
"	"	" 87 61	32	4929	4929	134.9
"	bkg	5 m 5 10	10	90	18	
"	"	" 14 3	16	226	452	
"	"	" 25 4	8	404	808	
"	"	" 28 12	4	620	124	
"	"	" 54 9	2	973	1746	
720	bkg	5 m 10 1	"	570	1140	
"	"	" 17 15	2	289	578	
"	"	" 11 1	4	177	354	
"	"	" 4 13	8	77	154	
"	"	" 2 1	16	33	66	
"	"	" - -	32	-	-	
"	std	5 m 0 2	23	2	29	
"	"	5 m 2 12	14	44	88	6.1
"	"	1 m 5 46	8	366	366	28.8
"	"	" 97 48	4	6254	6254	176.7
"	"	" 191 57	2	12281	12281	313.9
"	"	" 238 54	1	15286	15286	184.1

Source	Time of Count	Scaler Reading	Gain Att. %	Total Count	% Rate	57/6kg	net
900	std	5 m 13		338	66.6	3.5	48.0
"	bkg	" 1 29		98	18.6		
"	bkg	" 3 68		181	36.2	4.3	1540.8
"	std	1 m 24 43		1579	1579		
"	std	" 101 6		6470	6470	192.7	6423
"	bkg	5 m 3 48		285	47.0		
"	bkg	5 m 5 60		380	760	187.5	1471
"	std	1 m 258 89		14247	14247		
"	std	1 m 356 95		22789	22789	202.2	22679
"	bkg	2 m 3 28		220	110		
"	bkg	2 m 4 60		306	153	187.9	26589
"	std	1 m 112 70		28742	28742		
"	std	1 m 146 15		32271	32271	159.6	32069
"	bkg	1 m 3 10		202	202		
"	bkg	1 m 3 10		202	202	166.1	23746
"	std	1 m 138 152		33948	33948		
"	std	2 m 268 119		43723	43723	152.4	33502
"	bkg	1 m 3 28		221	221		

These readings on the 1060 with lin. amp.

9-5-52

Made cover for P-M

PM BT ch

brass sleeve to fit over
 lower cylinder



receptacle

82-505 for H.V.
88-112 for output

P.C. grounded anode R=150k 50µm coupling & out.

H.V. Source Time Scale Count Rate net ct sig/bkg

650 670 std 1m. 5.18 338 338
4.25 251 251
4.13 269 269

" " bkg 2.11 43 43
(spring cutting
key out) 1.8 24 24
2.8 40 40

700 720 bkg 3.3 57 57
5.2 82 82

std. 6.79 17

750 770 std 172 6

bkg 1 42

800 820 std 319 23

bkg 2 4

850 870 std 10.3 39

bkg 7 9

900 940 bkg 6 24

std 11.3 234

950 980 std 10.8 184

bkg 4 20

900 bkg 4 1

900 4 28

870 4 19

800 2 12

750 1 39

700 std 1 7

700 69.8 0 52 4424/52 85.1

710 177 28 1 36 11366/100 113.8

800 311 29 2 2 19933/170 117.3

850 104 15 2 6 26639/134 198.8

900 112 103 2 40 28775/148 194.4

950 108 71 3 35 27719/227 122.1

1000 98 207 4 48 25290/352 72.1

ESOM document

Scale not
responsive
as should
be. Count
pressure is.

Scale on 4400^F seems to check stopping

9-5-52 adj. zero to stop stray counts

850 bkg	1864	+ 58	2 1/2 m.	122	118
	1864	+ 50		114	47.2 c/m
850 std	165	34	10594		
	168	45	10477		224/
	165	1	10561		

9-8-52 Checked operation of PM counter as left 9/5/52

H.V.	Source	Counting Time	Scatter reading	Count	Rate	sig/bkg
9:00 A.M.	850 bkg	5m	16 23 13	221	44.2	
	std	1m	64 152 44	9772	9772	221.1
900	std	1m	199 48	12784	12784	241.6
	bkg	5m	4 56	204	60.8	
950	bkg	5m	6 52	426	85.2	170.8
	std	1m	232 49	14897	14897	
800	bkg	5m	16 8 4	132	26.4	
		5m	8 7	135	27.0	
		5m	7 15	127	25.4	
			Average		26.2	
	std	1m	64 73 38	4710	4710	
			75 80	4830	4830	185.2
			77 44	4972	4972	
		5m	380 49	24497	4899.4	
850	std	1m	155 30	9453	9453	
			154 26	9382		
			151 40	9704	9819.2	
			152 19	9747		
			153 18	9810		222.9
	bkg	5m	16 15 4	244	48.8	
			12 15	207	41.4	
			13 3	211	42.2	
				(44.1)		

109 + 852.6

Discriminator level (zero set) no low count
not working - too fast

2:20P	850	b	5m	15 0	205	41.6
-------	-----	---	----	------	-----	------

9-8-52

Adjusted zero set for proper 600 operation of sider -
 Higher voltage than required for PM operation

AV	1075	1100	1125	1150
	5 1/2 13	2 1/2 8	3 3/4 9	4 5/8 15
blg	45 c/m	40 c/m	57 c/m	69
	10250-35	13600-21	16500-31	18000-13
stg	6583 c/m	8765	10591	11533
S/O	< 200	219	< 200	< 200

one minute
 counts - all

Two minute check at 1100

blg	10 ¹⁶ - 10	170c	84 c/m	256/1
stg	680 ⁶⁴ - 23		8268.6 c/m	

9-9-52

0819	5 min blg	13 ⁶⁶ - 1	209c	41.8 c/m	
0825	1 min stg	138 ⁶⁴ 57	888.9	888.9 c/m	.212/1
0857	10 min blg	26 ¹⁵ - 1	417c	41.7 c/m	

9-9-52

Constructed preamp to go with PM - 3⁰⁰ PM

Int.

100V	5	140 ⁶⁴	47	b 3 ¹⁵	8	56
750		201 ⁶⁴	44	4	12	76
800		190 ⁶⁴	4	0 ⁶⁴	63	62
950		212	39	6 ⁶⁴	34	
900		241	42	1	38	
750		75	42	0	34	
800		170	19	0	58	

Input lead to preamp unshielded - picking up -
 stray. Try properly shielded.

9-11-52

Blank counts @ 8000 with preamp (unshielded)

400P	5	211 ⁶⁴	+33
	6	1	+8
	1		+20

9-11-52

0810	5	221	50
	1		15

110

Wmax Singapore

Voltage Range	Plate V _{g1} Ampl.	Output Ampl.	P ₁			R ₁ (pms)			G ₁		G ₂		Cathode		*reset	Output			C			R ₂			R ₃				
			pk	off	on	pk	off	on	on	off	on	off	on	off		pk	off	on	R ₁	R ₂	R ₃	R ₁	R ₂	R ₃	R ₁	R ₂	R ₃	R ₁	R ₂
5.1 — 5.2 — 5.1	19.5	200	207	272	136	207	137	277	63.5	38	31.5	72	64	62	69	273	257	290	1.7k	500k	40k	100k	25k	25k	15k	15k	500k	500k	500k
8.6 — 8.6 — 8.6	23.0	232	205	283	136	209	146	282	84	59	47.5	84	82	72	94	282	258	289	1.2k	500k	40k	100k	20k	23k	75k	200k	100k	100k	
6.9 — 6.8 — 7.0	22.0	230	209	276	141	216	142	276	74	45	45	90	73	77	90	275	261	291	310k	110k	10k	22k	4.5k	15k	230k	110k	110k	110k	
not pulling properly - light on most of time						measured from source												158k	120k	8k	22k	5k	15k	170k	170k	170k	170k	170k	
6H6 1/2 cad			5kll 200			res 6.5V 119																							
keeps falling but not reset																													

9-11-52 Remained over amp for further record -

HV cut at 1100v.

Standard run for 5 min - why 10 min This cycle

611x64 + 36 39140 7828.0

24x16 + 1 385 38.5

595x64 + 38 38118 7623.6

19x16 + 8 212 21.2

581x64 + 33 37217 7443.4

23x16 + 14 382 38.2

572x64 + 7 36615 7323.0

22x16 + 14 366 36.6

560x64 + 24 36082 7206.4

21x16 + 8 344 34.4

560x64 + 27 35867 7173.4

1m. std.

112 x 64 + 25 7193

111 + 14 7118

113 + 54 7246

117 + 36 7160

114 + 19 7315

112 + 39 7207

112 + 9 7177

113 + 3 7235

112 + 19 7257

111 41 7145

GM 553x64 + 15 35407 7081.4

12 113 43 7275

557 27 35675 7135.0

Left HV & LV on - ~~etc~~

9-12-52 0804 why - 20¹¹⁶ + 13 10w. 333 33.3

62x std 642⁶⁴ + 17 41105 6850.8

HV off check against check etc - none

9-12 3⁵P- 22x16 + 4 (10m) 352 35.6

528x64 + 24 36376 7275.0


523x64 + 10 36042 7208.4

526x64 + 24 36248 7249.6

561x64 41 35945 7189.0

9-12-52	std (5m)	562x64 + 57	36022	7204.8	3.57 P
		538x64 + 13	35735	7145.0	
		578x64 + 28	35740	7148.0	
		561x64 + 25	35929	7195.6	Aug 7200.62 320/1
	blk	18x16 + 10	298	29.8	Aug 32.7

NV off 4:26 P L.P. on over weekend

9-15-52	0817 blk 30m	61x16 + 5	981	32.7	
	std 5m	565 ⁶⁴ + 47	36207	7241.2	0837 NV depth
	lead shield	4 bricks -		22/1	

3.59 P blk	15m	23 ⁶⁴ + 6	374	24.93	
std	5m	536 ⁶⁴ + 37	34341	6868.1	275/1

NV off 4:22 P

9-16-52	0825 std	582 ⁶⁴ + 55	5m	37421	7486.0	Aug 7271.0
	blk	14 ¹⁶ + 12	10m	236	23.6	3.08/1
	std	537 ⁶⁴ + 17	5m	35281	7056.0	NV off


09:15 stacked more Pb bricks around case



0918 blk	15m	15x16 + 0	160		
0925 std		577x64 + 9	33088	6617.6	413/1

Shielded pre-amp and inserted it in output of PM!!

12119 (200)	blk	2m	1x16 + 12	28	14 c/m	
	s	1m	43x64 + 48	2800	2800	2002.1
725	s	1m	95x64 + 31		5983	
	b	2m	22x16 + 13	15	22.5	2662.1
750	b	2m	3x16 + 13	51	35.5	
	s	1m	139x64 + 21		8923	2492.1
775	s	1m	101x64 + 21		10965	
	b	2m	2x16 + 15	41	23.5	4652.1
800	b	2m	5x16 + 6	54	27.0	4742.1
	s	1m	200x64 + 21		12602	
800	b	15m	15x16 + 9	295	29.5	



PM CURVES
9-16-52

1.00	S-S	10.00X04 + 57	1.0170
338P	6-10	22X16 + 3	35.5
353P	S-S	10.20X04 + 49	1.327016

HV off

9/16/52	1:20 P	800	S-1	199x64+21		check on previous
		825	S-1	221x64+53	14197	57521
			b-2	3x16+7	55	27.5
		850	b-2	4x16+1	65	32.5
			S-1	230x64+16	14736	
		875	S-1	231x64+0	15296	340.1
			b-2	5x16+10	90	45.0

some lead removed Reproducibility check @ 800v.

45.54 ± 1
1329.9 ± 0.6
30.25

13161
13929

14.0 on L.V. on

1:27	b-10	18x16+13	301	30.1 ± 1.7	{ 28.4 } { 31.8 }
1:48	S-5	1033x64+16	66128	13225.6 ± 51.4	{ 13277.0 }
	S-5	1037x64+25	66463	13280.6 ± 57.5	
	S-5	1042x64+17	66705	13341.0 ± 51.7	{ 13289.3 }
	S-5	1039x64+32	66528	13305.6 ± 51.6	
	S-5	1040x64+28	66588	13317.6 ± 51.6	{ > 400.1 }
	b-10	50x16+14	324	32.4 ± 1.8	{ 30.6 } { 34.2 }

3:43	b-10	18x16+5	293	29.3 ± 1.7	{ 27.6 } { 31.0 }
	S-5	1041x64+24	66648	13329.6 ± 51.6	

HV off. overnight

9-17-52

0806	HV on				
0807	S-5	1050x64+9		13441.8 ± 51.8 X	warmup
	b-10	18x16+12		30.0 ± 1.7	
0824	S-5	1038x64+3		13287.0 ± 51.5	
1022	S-5	1036x64+22	66198	13239.6 ± 51.4	OK
1022	S-1	206x64+2		13196	
		210x64+50		13490 X	
		208x64+29		13353	13276.2
		207x64+35		13283	
		204x64+15		13271 X	13297.1
		202x64+15		13327	
		206x64+9		13193	
		207x64+47		13295	13318.0
		210x64+23		13463	
		208x64+33		13312	

1:07	S-5	1038x64+3	59	13170.2	
2:28	b-10	22x16+3		35.5	
2:53	S-5	1026x64+44		13270.6	HV off

9-18-52 0805 HV on:

0816 5-5 1034x64 + 23 13239.8

0822 b-10 19x16 + 7 27.9 HV on

1052 5-5 1024x64 + 1 13107.4

222P 5-5 1019x64 + 35 13050.2

b-10 4x69 + 9 26.5

9-19-52 0806 HV on:

0807 b-20 36x16 + 13 589 29.4

- 0829 5-5 1027x64 + 7

0835 5-5 1015x64 + 29

- 0841 5-5 1017x64 + 32

1035 5-5 1021x64 + 7

HV off to insert ESUM in supply line at tube -

24v 822 volts

24v 822 volts

1720 5-5 1010x64 + 50

24v adj. HV - to 822 1024x64 + 25

} removed
written check

Range over-voltage range using ESUM for HV readings

5x64
b-x70

ESUM
700 5-5 140-58 9018 1803.6 } 123.5

(675) b-5 4-9 73 14.6

720 5-5 108-17 6829 2464.5 } 298.7

(695) b-5 3-10 58 11.6

740 5-5 206-13 13197 6598.5 } 437.9

(710) b-5 4-11 75 15.0

760 5-5 237-44 15212 7606.0 } 408.9

(730) b-5 5-13 90 18.6

780 5-5 290-34 18594 9297.0 } 442.7

(750) b-5 5-6-9 105 21.0

800 5-5 344-13 22029 11014.5 } 529.5

(770) b-5 16-8 104 20.8

820 5-5 420-8 25888 13044.0 } 470.1

(800) b-5 15-2 242 48.4

840 5-5 486-25 27289 13644.5 } 530.4


(810) b-5 8-15 145 28.6

860 5-5 449-16 28752 14376.0 } 453.5

(830) b-5 5-15 95 31.7

ESUM indicated
rise over 50

PM Curves
9/17/50
A. H. ...



100

HIGH VOLTAGE
(ESVM)

^[234]
 880 S-2 467 - 51 29939 14969.5 } 424.1
 (870) b-3 6 - 10 104 35.3 }
 900 S-1 239 - 0 15246 } work on 9/19 stopped here
 (870) b-3 - - - - - } went to assist for in 202.

* re-run 800 bly 2 1/2 m. 8-12
 checked ends W & E E @ 2W.

820 b-5 after E shutdown.

920 S-1

b-2

⁰⁸⁸⁵
 9-22-52 900 b-2 9-9 121 40.3 374.6 } set power off over weekend

920 S-2 496 - 43 31781 15893.5 } 362.9
 (885) b-5 13 - 11 219 43.8 }

940 S-2 499 - 62 31999 15999.5 } 406.1
 (900) b-5 12 - 5 199 39.4 }

960 S-2 501 - 53 32117 16058.5 } 424.1
 (910) b-5 11 - 11 189 37.4 }

980 S-2 504 - 11 32267 16133.5 } 384.13
 (940) b-5 13 - 2 210 42.0 }

1000 S-2 510 - 58 32698 16349.0 } 388.6
 (950) b-5 13 - 4 212 42.4 }

set HV @ 900 0921

4:20 PM - set at 920v.

4:17 PM HV off.

9-23-52 0821 HV on reading W946v.

0826 bly for 2000 sec 78716 ± 10 32.7 q/m

0944 S-1 248 - 51 15923 ± 126.2 series of 1-m cts

248 - 21 15893 ± 126.1

8:44 p.m. 252 - 30 16158 ± 127.1

248 - 6 15878 ± 126.0

249 - 37 15973 ± 126.4

248 - 25 15897 ± 126.1

249 - 62 15998 ± 126.5

245 - 54 15728 ± 125.5

251 - 17 16081 ± 126.8

246 - 05 15749 ± 125.5

Avg: 15,928.8

9-23-52 Made Microflex Tuning unit for single reader -
checked ± 4 with detector 600.

1228 5-1 248-56
250-0
248-55
247-18
248-46
247-55
250-60
246-52
246-20
242-16

1289 b-10 22-4

1350 " 25-9

1307 " 51-5

1377 " 49-1

1321 " 21-4 1321115

Test end running during
this period-

Rate Meter & E.H. Recorder added to circuit

1352 b-2 4-0 66 32.0 <40 N.18

" 4-9 73 36.5 <40 N.18

1357 40 " 4-3 67 32.5 <40 N.18

1357 45 " 4-9 73 36.5 <40 N.18

1401 50 " 4-15 79 39.5 45 .20

1406 " 4-10 74 37.0

1306 00

1289 HV 946.

checked Tuning with 600 ⁵⁶⁻¹⁶ { ⁵⁸⁻⁴⁵ 54-38 } checked with watch
₅₈₋₅₇ at source

disconnect Rm - 582 still high

source not under PM. 58-23

904 5-1 241-60 15484 15484

243-29 15581

248-10 15882

244-13 15629

245-57 1542 15737

311 0-5 { 1226-47 15702.2

1231-48 15766.4

1234-6 15796.4

b-5 11-14 38.2% : HV off on minute

9-24-52 0807 HV on
 0833 b-5 10-6 33.2
 1233 S-5 1243-29 15916.8

9-25-52 0802 HV on
 0814 b-5 12-14 41.2
 0820 S-5 1238-40 15854.4

Check on speed of response of registers on AIC scales -
 used audio gen, pulse gen, feeding scale 1 minute count

Diad	1/2 Stage	Scale	Register
300	"	4	184 ²³
Reading 300	"	"	277 ¹¹
400	"	"	369 ⁶
500	"	"	457 ⁶⁰
600	"	"	541 ⁵⁷
700	"	"	635 ⁵⁹
750	"	"	680 ⁵⁷

1010/s OK.

PM 05-

200 P S-5 1238-39 15854.2
 206 b-5 11-11 37.4

9-26-52 0812 HV on
 0823 b-5 12-4 39.2
 0829 S-5 1236-8 15822.4
 1237 P b-5 11-12 37.6
 1⁰⁰ P S-5 1238-24 15851.2
 358 P S-5 1229-30 15737.3

9-29-52 0805 HV on
 0835 b-5 11-6 36.4
 0841 S-5 1231-47 15766.2
 1021 S-5 1226-9 15694.6
 320 P S-5 1231-58 15768.4
 330 P b-5 9-15 31.8

9-30 0801 HV on
 0815 HV reads 950
 0814 b-5 12-9 40.2
 0821 S-5 1225-38 15687.6
 105 P S-5 1233-4 15893.4

74

0827 5-5 1232-0

0827 6-5 12-0

0833 5-5 1233-34

1222 5-5 1233-02

383 P 5-5 1235-7

0806 HV a 11-7

0821 6-5 11-7

0827 5-5 1232-15

1016 5-5 1246-14

1022 5-5 1243-20

1030 6-5 11-11

removed HV-meter & CRM -

1200 6-5 11-11

1206 6-5 1253-82-P.1

0808 HV on

0833 6-5 12-2

0840 5-5 1234-34

0846 5-5 1243-21

1035 5-5 1253-9

1041 6-5 12-7

210 P 5-5 1231-31

857A HV out

1243 P 5-5 1244-46

1250 P 6-5 11-5

120 P 6-5 1238-41

230 P 5-5 1250-13

0820 HV a 12-11

1000 6-5 10-11

1006 5-5 1233-33

1205 6-5 12-12

1211 5-5 1264-11

239 P 5-5 1247-41

10-7-52 0805 HV on
0920 b-5 12-2
0942 5-5 1242-35
off in sta. during interval -
1130 5-5 1239-2
0157 b-5 11-13

10-13-52 0805 HV on
1033 5-5 1223-18
1147 5-5 1217-57
120 P b-5 12-8

10-15-52 HV on 0805
0810 b-5 10-2
0905 5-5 1219-57
223 P 5-5 1220-48

10-17-52 HV on 0807
829 A b-5 11-0
836 5-5 1217-24
(begin of time - after work during.)


11-11-52 HV on 0820
823A b-5 10-15
829 " 10-3
846 " 11-15
857 5-5 1237-24
1021 5-5 1218-17
1204 5-5 1231-20
1214 b-5 13-15
427 HV off

11-12-52 819A HV on
940 b-5 12-1
1035 5-5 1238-52
45 P HV off

11-13-52 805 HV on
808 b-5 12-8
816 5-5 1231-49

10-29-52 Check GE auto relay CR-279/C103C25

1. $R_0 = 2400 \Omega$ $I_{in} = 2.4 \text{ ma}$



3 NC
2 NO
4
10 95
lead

2500 Ω , $I_{in} 3 \text{ ma}$ pin 2-3 opens 3-4 not operate
Set to 2.2 ma 2-3 close

All three on hand checked out OK.

avg 2.9 ma in, 2.12 ma out 2-3 NC

A & D Resistor automatic first (Pup)

R_1 may need

R_2 needs to have higher point of shunt to drop out first

11-13-52 Run 215 scales 60W check C3 doubling

replaced 114 strip

	164	281 ²⁴	281 ¹⁵	281 ¹⁹	281 ¹⁵
		281 ⁴¹	281 ⁵²	281 ⁴¹	281 ⁴²
115	$\frac{62 \times 164}{5}$	8 ¹³	6 ³	8 ⁹	8 ¹
0.5	10	18 ³	13 ²	16 ⁸	16 ³
101	5-10	15	28 ²	19 ¹³	25 ⁴
10-15	20	35 ¹⁰	25 ¹³	33 ¹⁰	33 ³

60W	$\times 16$	1126 ⁷	1129 ³	1126 ⁶	1126 ⁶
	5576	70 ⁷⁸	70 ⁷⁷	70 ⁷⁸	70 ⁷¹

12-29-52 Check IC Amplifier gain

Voltage readings - output ~~switch~~ off - meter so lead is
Preamp. cable connection -

B -40.3

C -33.2

D -39.3

E +86.8


F +5.2V

H -40.1

12-21-52 Bush Recorder Tie in with Just in 202 for cherry study
 One unit of Bush recording neutron level from output of R-7.
 Second unit recording stream speed - input across series resistors in tachometer-meter circuit.

R-1 sensitivity loop for good readings -

1-22-53 Speed Controller for Agitator -
 micromax with motor ^{balance} mechanical
 but setup



mdit set	Pot
.2	0.72 - 0.92
.3	1.35 - 1.45
.4	1.75 - 1.90
.5	2.20 - 2.40
.6	2.65 - 2.85
.7	3.15 - 3.35
.8	3.60 - 3.70
.9	4.05 - 4.25
1.0	4.50 - 4.70

Pot 2.75 m .6 ± .03

~~Top of drive - to set of belt~~

Check Pot settings -	Pot	00	0
1.32	.82	1	1.3
	1.3 mu.	2	2.3
	2.75	3	3.0
	2.8 mu.	4	3.7
	1.80	5	4.15
	1.80	6	4.52
	2.05	7	5.00
	2.30	8	5.4
	2.50	9	5.5
	3.23	10	5.8
	3.65		
	3.42		
	3.80		
	4.03		

across 4K 1K = 5V ± 0.144

1-22-53

VR - Brush Pickoff


R - 1000 meter scale

Scale selector positions

Notes

zero

1 is at 1000 = 1 mv across unit resistor in selector



3-6	x 10	5-3 x 5
6-4	x 4.5	5-2 x 11
6-2	x 21	

1.58V

$247k - 1k \parallel - .525V$

Check of value of resistors: Wien's bridge

$100k\Omega = 99.4 k\Omega$

$47k\Omega = 47.9 k\Omega$

$1k\Omega = 0.999 k\Omega$

$\Sigma R = 148.299 k\Omega$

Divide current: $I = \frac{E}{R} = \frac{1.58}{1.48 \times 10^5} = 1.068 \times 10^{-5}$

80

Scale A1C
Strip 114


76 days

V_3 1: 106 V_6 1: 95 Reading from strip which "doubles"
2: 120 2: 103 when V_3 is inserted.
5: 129 5: 93 Count-off -
7: 120 7: 105

Readings on healthy strip: V_3 : 1: 100 no lite
2: 128
5: 128
7: 128

	0	1	2
V_1	102	138	102
2	125	127	124
5	133	100	133
7	125	127	124

Probed under chassis of "same" 114 strips. Separated or wires & re-installed. Operates O.K.



2-19-53 Liquid Level Indicator.


Resistive H_2O - probe - HP UTVM =

Contact closed - $\approx 2 \times 10^5 \Omega$ when touching H_2O

Contact H - good, no H_2O cond. ∞

When touch - Rdry to $2 \times 10^5 \Omega$ & rose above 5×10^5 to $9 \times 10^5 \Omega$

Salt solution setup:





soaked probe lens - still operate OK.
on fill. Hard to distinguish difference
or drain.

$R_c = \frac{12000 \Omega}{\dots}$
 $R_o = \infty$

worked fine for dilute solutions only --

for concentrated salt solution Δi 40-50 μA




$V_c \approx .4v$ $i_c = 35 \mu A$
 $V_o \approx .6v$ $i_o = 40 \mu A$

$R_c = \frac{4 \times 10^{-1}}{5 \times 10^{-4}} = .8 \times 10^4 = 80 k\Omega$

$R_o = \frac{6 \times 10^{-1}}{4 \times 10^{-4}} = 1.5 \times 10^4 = 150 k\Omega$

Measured R_c 1.6k Ω 2.4 μA
 R_o 4k Ω 48 μA $\approx 2:1$

L.H.I. circuit



$$\frac{20 \mu A \cdot 2}{2.3} = \frac{R_c - 1}{R_o - 2}$$

$$\frac{20}{2.3} \mu A = 8.3 \text{ mA}$$

$$\frac{13}{2.3} \mu A = 5.4 \text{ mA}$$

$$\frac{2}{2.3} \times 10^3 \mu A = 250 \Omega$$

$$\frac{8}{5.4} \times 10^3 \mu A = 17k$$

$$\frac{2}{2} = 1 \mu A$$

$$\frac{8}{4} = 2 \mu A$$

2-23-53 PM, divider changed so negative HV is to be measured
 This is a scintillation detector in sample slide case.


Gain of 100 thru 2 stage amplifier -

Used ESM to determine real HV on PM. HVPS reading

is 2kV true.


2/25/53

PM - detector fast response recorder - Brush
 Amplifier.




$P_{2+} = 1500$

input $0.075 \times 1000 = 75 \text{ mV}$
 output $7.5 \text{ volts} \approx \text{gain } 100$
 output P.W. $\approx 70\%$ $10 \mu\text{s}$ $\text{trials } 6 \text{ loops}$




Need to feed thru integrating circuit into Brush
 response of Brush 0-100 cps.


3-2-53 "Broad-banded" amp'd MV stages as follows:



3-2-53 Ran rough check on above - from NCC pg. HP sig gen
 input: 10 μ s, 2V pulses 2V in amplitude - output variable pulse
 500 - 10000 μ s long, 20 μ s/Hz in amplitude.
 When fed to brush off P₃ () only get
 oscillatory trace no deflection.




3-2-53 Changed supply



3-4-53

Bread Board ckt:




Return chassis wiring in parallel.

Voltage readings as in:

1 - 5.01	5 - 15	9 - 15
2 - 112	6 - 15	10 -
3 - 70	7 - 1.84	11 -
4 - 15	8 - 3.55	12 - 294

$I_f = 19 \text{ ma}$ (as read on P.S. panel meter)

PM signal - $V = .75 \text{ v}$



Coupling 2-5 .001 pf - at 5 10V input

Check in. bias dist - $\approx 20 \text{ v}$ - has off 6k Ω

Added 560k 5-8

$V_f = -18.3$; $V_s = -12$; $V = 8.34$


For $V_f = -20$; $V_s = -29$

3-5-53

PM - 1200v = \parallel Bias - 15v = $V_f = 6$ $I_{fcs} = \frac{9}{60} \times 10^{-5}$

Cap 6-7 decade box -

Coupling 6-9 - $R_1 = 6.2$.001



$$\frac{300}{40} = 260$$

$$\frac{300}{2} \times 10 = 1500$$


$$1500 \times 9 \times 10 = 9 \times 10^4$$

$2 \times 200k \text{ } 1w \text{ in } 11 -$

$$10^{-3} \times 12 = 200 \quad R_2 = \frac{200 \times 10^3}{10^{-3}}$$


$i = 1 \text{ ma}$ $R_1 = 100k\Omega = 10^5$ $P = 10^5 \times 10^{-5} = 10^{-1} \text{ w}$

$R_2 = \frac{200}{10^{-3}} = 2 \times 10^5$ $P = 10^5 \times 4 \times 10^{-5} = 4 \text{ w}$



over for complete test circuit

3-5-58




Check values of Plate resistor in V_s -

For remote operation need -

Cathode follower or PM - 70k = 30-35k


25k = 26k -

3-6-58 Cathode Follower for Photomultiplier -
From NYO 1077




On hand: 1 model 110 per-amp. (Shannon & Sande)



Trail out of overhead out.



input for vacuum tube amplifier requires capacitor coupling




30 vacuum tube




TRANSIENT RESPONSE
AMPLIFIER

3-16-53 Liquid level Warning Probe.



This probe to extend 1/2-1 in. below level indicator to warn operator of approach to set level.



3-17-53 Warning buzzer circuit wired -




3-31-53 Liquid Level Indicator Installation Rm 201-202



SP13 (E) & SP6 (L) were disconnected from terminal strip

Next reference 90

4-12-53 P.M. scales - check - Cht




using negative HV
two minute counts
standard in 55
standard out = 6

Scale may need to be recalibrated

source	reg.	voltage	count	rate	5/2	reg. #	volt	count	rate
5/5	115	31	700	4-19-53	5/2	444 ²⁷ / ₆₄	1000	28443	14221.5
6/5	5	5	"		b/2	1 ⁵⁰ / ₆₄		114	67
6/5	13	5	750		5/2	442 ²⁷ / ₆₄		28316	14158
5/5	460	14	"		b/2	2 ⁵ / ₆₄		133	66.5
5/5	1098	1	800		5/2	446 ³² / ₆₄		28576	14288
6/5	4	45	"		AV209 bare	567 ¹⁶² / ₂₅₆			
6/5	9	0	858		1/2 Pb	124 ¹¹ / ₁₆			
5/5	508	102	"		2/2 Pb	96 ⁹ / ₁₆			
5/5	282	102	800		1/2 Pb	122 ³ / ₁₆			
6/5	12	61	"		bare	573 ²⁹ / ₂₅₆			
4/5 0908	6/5	10	900		S-2	440 ³³ / ₆₄		28193	14096.5
	5/5	83	800		b-2	2 ¹ / ₆₄		129	64.5
0830	5/5	524	99						
0936	5/5	530	84						
0942	5/5	537	17						
	5/5	545	240						
1027	5/5	578	127	recalibrated 900					
	5/5	607	200						
	5/5	614	53						
	5/5	583	107						
	5/5	575	144						
During the above counts, it was noted that scales did not seem to operate normally. Some papers & more indicators observed.									
				0840 ¹⁴	5/2	440 ¹³ / ₆₄		28172	14084
				0842 ²²	b/2	1 ⁶ / ₆₄		125	62.5
				0844 ³⁰	5/2	433 ²⁷ / ₆₄		27737	13868.5
				1210	b-2	1 ²⁰ / ₆₄			
				*	S-2	421 ⁶ / ₆₄			
4-14-53	PM had directly to scale 1010.				reset	1000			
					S-2	438 ²⁷ / ₆₄		28052	14026
					b-2	1 ⁵ / ₆₄		121	60.5
					S-2	434 ¹⁵ / ₆₄		28111	14055.5
				AV209	405	214			

ROBERTS & SHAW CO., N. Y. NO. 884-714
Serial 1040101010, 5 Cycles X 10 to the inch, 4th time received
MADE IN U.S.A.

Counting Rate $\% \mu$



1000
5/4
4/4
3/4
2/4
1/4
1/5
1/5
1/5

Applied Voltage

Left scales on count overnight. No counts registered.

11-16-53		0809		14U		Reads		10154	
no.	no. source	no. source	no. source	no. source	no. source	no. source	no. source	no. source	no. source
5-2	570	5-2	570	1160	5644	18247			
5-2	482	"			1440	720			
5-2	439	1000	28120	14060	84	42			
b-2	1 3/4	"	101	53.5	2776	13826.5			
3-2	437	"	28003	14001.5	27782	13891			
b-2	1 3/4	980	95	42.5					
5-2	422	"	27064	13533					
5-2	402	960	25751	12875.5					
b-2	1 3/4	"	95	47.5					
b-2	1 3/4	940	65	32.5					
5-2	364	"	23385	11692.5					
5-2	374	980	17585	8792.5					
b-2	1 3/4	"	61	30.5					
b-2	1 3/4	860	51	25.5					
5-2	195	"	12444	6247					
5-2	67	820	4326	2163					
b-2	1 3/4	"	29	14.5					
b-2	1 3/4	780	19	9.5					
5-2	1 3/4	"	80	40					
5-2	441	1000	22272	11136					
b-2	1 3/4	"	77	38.5					
b-2	1 3/4	1020	96	48					
5-2	445	"	22508	11254					
5-2	469	1040	24741	12370.5					
b-2	1 3/4	"	118	59					
b-2	1 3/4	1060	118	59					
5-2	475	"	30441	15220.5					
5-2	483	1080	30964	15482					
b-2	1 3/4	"	125	62.5					
b-2	2 3/4	1100	134	67					
5-2	483	"	30934	15467					
5-2	498	1120	31919	15959.5					
b-2	1 3/4	"	117	58.5					
b-2	1 3/4	1140	824	412					
5-2	510	"	32665	16332.5					


200 experiments during this run

4-77 4U on 0806
adj 4U to 1000 volts
0815 5-2 439 1000
b-2 1 3/4 "

4-27-53 Level Indicator - last ref. p 87.

Installed Simplified level indicator, after need had arisen for frequent adjustment of previous setup (for each reboiler change).

Complete level system, except for warning probe, is as follows:



Procedure in modification -

201 { i - removed input & output (2 wire - probe; 2 wire - lights) cable to indicator unit, on deck above S12.

ii ran jumpers from probe terminal (red & black) to H & H - S11 & S12.


202 { iii disconnect S11, S12 leads running to lights from terminal block.

iv ran 4 wires thru to N-panel; two tied to S11-S12 block, remaining two to S11, S12 leads removed in (iii) & tied in H & N to block in N-panel.

Tested unit with dead short in 202 at.

Tested unit with probe in acid in 201 - at.

5-4-53 Level Indicator - Battery drain was too great on indicating probe circuits. Took probe circuits - warning & indicating - to 210 for reworking. Mounted on single base & installed selenium rectifier & filter as voltage supply for the three probes. Circuit:



5-5-53 { Note on above - 10k, 2w resistors to be replaced by 5k, 5w, when available. -

5-11-53 Break -

5-15 Still instable balance - non-linear - gain low

5-18-53 Checked tubes V₁ - V₄

Replaced all 4 tubes -

Ran thru calibration procedure - seemed ok. Scotty ran D.C. calibration with multivolt potentiometer

5-14-53 D.C. amplifier check out - (Begin 5-11-53)

Having difficulty balancing.

Output c.s. 12.55 voltmeter out of order.

grid = -25V; cathode + 7 volts.

cathode should be approx = grid

[E changed from 22.5 to 45 to drive grid negative]

removing capacitor & test output no help.

Revised output meter & scale selector leads on

grid = -24; cathode +9; plate +92

Revised 12.55 this reports all unit -

new 12.55 - no change

floating cathode; feedback -65 E plate +52

+ output batt -43 o.k.

V_g cathode - +52

grid +47

plate floating feedback -64

+ output batt + -42

cathode -42

grid -42

plate -42

grid floating - fb - +9

batt + - +33

cath - +33

grid - +15


plate - +92

Replaced 20kΩ with 2kΩ - output resistor -

5-18-53 Mounted 40-40 pfd caps. in place of 16 pF. - V_{GH}

Change in scale selector sw. affects meter reading -

from 100 - 1 on scale sel. increases meter reading




Gull 1955

5-14/55

Unit on test B stability, etc -
birds still marked -

6-27-55 Sealing pair test unit



Sealing Pair Components 65L7 - 6AL6: 65N7 - 6AL6

R ₁	1 meg	200k
R ₂	500k	100k
R ₃	40k	10k
R ₄	100k	20k
R ₅	25k	5k
R ₆	75k	15k
R ₇	1Mmeg	200k
R ₈	500k	100k
C ₁	40 μF	50 μF
C ₂	40 μF	50 μF
C ₃	.01	.01


Usory Grid 65L7 unit hook Voltage readings

Signal for test 100 μsec, 4 pulses @ 60 freq. @ 15 v.
 Voltage threshold 5.2 volts = operate to 20 v. same on 10 v.
 Voltage on 656 plate (scope) 15-18 volts (neg pulses)
 Grid pulse @ 30 volts { grid triode: pulse 210; off 260; on 136
 output: " 275; " 258; " 292



Same unit: 100 μsec pulse Voltage range 100 μs: 6.8 - 9.2 v.
 10 μs 6.8 - 20
 1 μs 8.3 - 20

grid triode - pulse @ 210 v.; off 275; on 142
 output " 275; 262; 290
 reset cathode 80V - from reset

Feeling Tester:



68



6-8-53 WEPB Log 11 - Had been in shop for a while to check out. Checked input stages. new 5004 in (test aged) after several hrs. would not go calibrate meter negative after several code lapse - returned to job on 6-4. put in new 5803. Put in good hi mag resistor (but 1/2 rated value). Seemed OK. Under test with Period Indicator. -

6-15. Mounted 53K10 "a" resistor (proper value) in unit -

6-17-53 Linearity Check -

(1/2 of local at 100 d. 001 %)


meter	20	10 ¹⁰	10 ¹⁰	10 ¹⁰
		Volts	10	
0001		2419	.001	} .2380
.001	.2405	2415	.2390	
.01	.2109	4584	.4449	} .2057
.1	.2079	6665	6566	
1.0	.2073	8678	8693	} .2117
10				
100				

6-23-53 D.C. Amp EP-19 Tube check:

Replaced: 6AV6, 6AV6, 12SH7, 2-6J4's - in Amp
 9002 in pre-amp.

Turned on meter reversing on fil. position

filament circuit.



Typical double indication =

$$V_2 = V_1 = 20V$$

$$V_3 = V_3 = 15V$$

$$V_4 = 10V?$$


This connected SW lead @ ①. Find 510Ω resistor direct to M+ @. Operate ok.

Resistance of $\approx 20\Omega$ or measured from mount & leads to resistor in series with meter - value $\approx 5\Omega$
 Put in 50Ω, 1W resistor in its place -

875-53 *Manfred B₂ counter*

1080: AH 4	}	<i>1/2" x 4"</i>	<i>operator. w-2000v.</i>	<i>(1700 - 3200.0v)</i>
Dec 20				
AEV				
Dec 15		<i>5/16" x 1/2"</i>	<i>- 1600v -</i>	<i>(1400 - 1800)</i>

Data of curves for runs with 2045 tube amplifier and 1010 scale as given on Manfred B₂ counter



98.

8-11-53
8-24-53

Can test using thermocouple of new millivoltmeter.
Test ok.

Developed rectifier unit to attach to output
to apply signal to recorder -


Wanted: signal to drive recorder full scale at 100°C.

For $S_{peakmax} = 7 \text{ mv}$
 $R_{max} = 10 \text{ mv}$
 $E_{thermocouple} = 1 \text{ ma}$

For Copper-Constantan 100°C @ 4.28 mv. when Ref.
 Thermocouple is @ 0°C.

With DC measuring device at .01 attenuator setting
 (10 mv full scale for indicator AC output of 1 rms volt)
 RMS output for 100°C = ~~0.428~~ 0.428 volts.

Rectifier unit:




1-15-54 - $R_A = 250,000$
 $R_B = 5,000$
 $R_C = 1,000$

$R_A = 250 \times 10^3$
 $R_C = 10^3$
 $R_B = 0$

$V_i = 5 \times 10^{-3} \text{ v}$

$V_o = 44 \text{ mV}$ $2.5 \times 10^3 \times 44 \times 10^{-3}$


$\frac{8.8}{1.0 \times 10^{-1}} = 1.10 \text{ vdc.}$



8-23 Big Sid Water Heater -


wind heater system suggested by D.F.C. for maintaining elevated temperature of reflector water

Wiring Diagram:




8-26 Tests showed that the system was not enough to heat or hold Sid water to temperature. Additional heating required.

8-27 Installed 2 immersion heaters, 9 KW each, operating from 440v, 3Ø, source. Incorporated in system safety interlock to cut off heaters when water level drops. Wiring - 440v Heaters.



8-27-53
(cont)

Control wiring completed - 110v. control system -



5
9.21

Change Over
 Job - 10/5/53
 a. New battery control on spark plug
 b. New D.C. supply for skimmer tubes

in place
 plugs ✓
 at lights
 many

valts -


amps
 mps.
 mps.

th-

18-12
finished


H
110V
60Hz

H




N

H



130
1000

H



Proposed possible change
 in sol. storage system

SKP

9-18-53

Solution Storage System - Suggested Wiring
Modifications

9.21.53

D.F.C. suggestions previous week -

✓ D.C. on Skimmer - ✓

- Additional electrical operated valves in place
of manual valves on bottom each tank

✓ Lower voltage & / or current on spark plugs ✓
a. lower voltage transformer & 2 volt lights
b. same transformer ^{with} variable in primary

9.21.53 Skimmer valve D.C. operation -

held @ 16.5 volts - 146 ma

~~20~~ 20 - 178 ma

25 - 220 ma

drops out @ 1.1 volts -

29

10v - 84 ma

For 20 skimmer valves: @ 16.5v, 146 ma - 4.1 amperes

@ 20v, 178 ma - 5.0 amperes

@ 25v, 220 ma - 6.2 amperes


9.21.53 Took DC power supply to sea for test.

13 sets (26) valves draw ~ 5 amperes @ 50 volts -

10-21-53 Poppy Disturbance -

As had been several times before suspected the IIM typewriter in 212 was cause for the Raytheon Poppy going into discharge. Poppy had been in 209 for several months for observation. Had been cutting up in 201. Possible for disturbance to get on building power distribution -

Made line fitter to visit in 212.



3-conductor cable with 3-prong plug in cord to plug in connector.

Duplex Receptacle 3-wire polarized

this cut-out interference to Poppy -

Adjustment of poppy sensitivity from max. to an intermediate level lessened likelihood of other disturbances of low amplitude -

10-22-53 Poppy check at present setting of sens.

HVc 2000 v - sets at M and max. line adj.

	1070/cm	1481/cm	1573/cm
Qty	20	50	M 20
<10	20	45	15
	30		12

2-2-54: Neutron Survey Meter (Lead ref. 2 ft.)

Serviced - new batteries 2 @ 300; 1 @ 67.5; 1 @ 3

Adjusted for proper response.

- (a) Dial ccw. adj mv to point of deflection ccw. and back off slightly.
- (b) Turn Dial cw til get bkg. $I \approx 1 \mu A$
- (c) Adj mv & dial for good response & accuracy and proper peaking on switch change.

2-3-54 Calibration with P15 222

$N_0 = 3.67 \times 10^7$ counts/sec

$N = N_0 e^{-\lambda t}$

$\lambda = \frac{.693}{T_{1/2}} = 4.95 \times 10^{-5}$

$t = 814 d.$

$= 3.67 \times 10^7 e^{-4.95 \times 10^{-5} \times 814}$

$= 3.67 \times 10^7 e^{-4.028}$

$= 3.67 \times 10^7 \times 0.0178 = 6.53 \times 10^5$ cps


$\text{flux} = \frac{N}{A} = \frac{6.53 \times 10^5}{4\pi r^2} = \frac{6.53 \times 10^5}{1.257 \times 10^3} = \frac{5.20 \times 10^2}{r^2}$

r	r ²	F	0.495 x 3	155/305 x 20
bkg	-	-	0.3	
100	10 ⁴	51.20	1.02	15/1.5
80	6.4 x 10 ³	8.12	1-2	2/1
65	4.23 x 10 ³	12.3	2-4	1.5/2
50	2.5 x 10 ³	20.8	3-5	2/3
40	1.6 x 10 ³	32.5	4-12	2.5/6.5
35	1.23 x 10 ³	42.3	8-13.5	6/10
30	9.0 x 10 ²	57.8	9-15	7/10
25	6.25 x 10 ²	83.2	11-18	13/12
20	4.0 x 10 ²	120	14-20	12/14
15	2.25 x 10 ²	231	0.5	20/05
10	10 ²	520		2.0
6	3.6 x 10 ¹	1445		3.0

8-19-54 DC Amplifier for Thermocouple -

Experimenting with sub-miniature, battery-powered, CK 572AX tubes

8-24-54



$V_{B+} = .62 \text{ volts}; V_{B-} = 42.5 \text{v}$

removed Thermocouple from $T_{1,2}$ & $T_{1,3}$

With no input element voltages on T_1 & T_2

$R_L = 0$

$R_g = 500k$

$T_{1,1} = 0$	$T_{2,1} = 0$	$T_{1,2} = 0$	$T_{2,1} = 0$
$T_{1,2} = -.82$	$T_{2,2} = -.87$	$T_{1,3} = -.69$	$T_{2,2} = -.57$
$T_{1,3} = +1.61$	$T_{2,3} = +1.61$	$T_{1,4} = +1.61$	$T_{2,3} = +1.61$
$T_{1,4} = +3.45$	$T_{2,4} = +3.5$	$T_{1,5} = +13.6$	$T_{2,4} = +13.4$
$T_{1,5} = +41.5$	$T_{2,5} = +10.5$	$T_{1,5} = +26.7$	$T_{2,5} = +14.5$
$a = +41.5$	$b = +10.5$	$a = 34$	$b = +40$

$R_L = 0$

as above - with $R_g = 100k$ ($T_{1,2}$ to E_1)

$T_{1,2} = -.21$	$T_{2,2} = -.78$
$T_{1,4} = 3.4$	$T_{2,4} = +3.4$
$T_{1,5} = 38.5$	$T_{2,5} = +10.4$
$a = 42$	$b = +10.4$

$R_g = 100k$	$R_{p1} = 0$	$R_{p2} = 500k$	$R_g = 1k$	$R_{p1} = 0$	$R_{p2} = 500k$	$R_g = 100$
$T_{1,2} = -.04$	$T_{2,2} = -.03$	$T_{1,4} = -.01$	$T_{2,4} = -.01$	$T_{1,4} = 10.9$	$T_{2,4} = 10.9$	$T_{1,5} = 10.9$
$T_{1,5} = +3.4$	$T_{2,5} = +10.8$	$T_{1,5} = +3.3$	$T_{2,5} = +10.9$	$T_{1,5} = +2.7$	$T_{2,5} = +2.7$	$T_{1,5} = +2.7$
$a = 41.5$	$b = +10.5$	$a = 22$	$b = 22$			

8-25-54 (c) GCB6 Characteristics -

Shunt-cut off Pentode

B+

R_L 100k
5
100k
0

3:17
15V

E _B	R _L	R _G	R _K	R _S	R ₂	E _F	E _S	E _G	E _C	E _W	Notes
100	max	60k	max	max	24k	79	17.5	0	0	0	Supplied to cathode - input grounded
75						75	17.2		20	78	
70.5						70.5	16.9		44	165	
65						65	16.6		65	160	
60						60	16.3		80	155	
55						55	16.0		90	150	
50						50	15.7		95	145	
45						45	15.4		100	140	
40						40	15.1		105	135	
35						35	14.8		110	130	
30						30	14.5		115	125	
25						25	14.2		120	120	
20						20	13.9		125	115	
15						15	13.6		130	110	
10						10	13.3		135	105	
5						5	13.0		140	100	
0						0	12.7		145	95	
0						0	12.4		150	90	
0						0	12.1		155	85	
0						0	11.8		160	80	
0						0	11.5		165	75	
0						0	11.2		170	70	
0						0	10.9		175	65	
0						0	10.6		180	60	
0						0	10.3		185	55	
0						0	10.0		190	50	
0						0	9.7		195	45	
0						0	9.4		200	40	
0						0	9.1		205	35	
0						0	8.8		210	30	
0						0	8.5		215	25	
0						0	8.2		220	20	
0						0	7.9		225	15	
0						0	7.6		230	10	
0						0	7.3		235	5	
0						0	7.0		240	0	
0						0	6.7		245		
0						0	6.4		250		
0						0	6.1		255		
0						0	5.8		260		
0						0	5.5		265		
0						0	5.2		270		
0						0	4.9		275		
0						0	4.6		280		
0						0	4.3		285		
0						0	4.0		290		
0						0	3.7		295		
0						0	3.4		300		
0						0	3.1		305		
0						0	2.8		310		
0						0	2.5		315		
0						0	2.2		320		
0						0	1.9		325		
0						0	1.6		330		
0						0	1.3		335		
0						0	1.0		340		
0						0	0.7		345		
0						0	0.4		350		
0						0	0.1		355		
0						0			360		

Stranded pin-3 grid & metal frame

Unshunted grid

Cathode grid

Build delay meter in diff. in E_S readings


Cathode R = 0

Re-set of point to E_S = 30V
50k grid


8-25-54 (c) 6CB6 Characteristics -

Sharp Cut off Pentode

B+



8-26-54 6CB6 - with input signal =



Adjusted R_k to point above max E_p -

$R_s \neq R_c$, R_c max, or 100K values. R_k at max 1 meg.

Readings with 0 input.

$E_p = 14.4V$ $E_g = -0.00125V$

$E_s = +0.55V$

$E_c = 12V$

Helipot rotation	E_g	E_p	E_s	E_c
1/8 turn	00043	14.4	11.8	
1/4	013	14.8	12.2	.05
1/2	024	15.6	12.4	.05
3/4	040	16.5	12.7	.05


This indicates input is greater than desired, false indication of high gain.

Why w. low resistance input


(a) 100-ohm - no change in plate voltage for 5 ~~turns~~ ~~of~~ ~~helipot~~, which is hint of signal if given proper voltage values (at 4 mv)

When tie in signal to grid reads 2.75 mv.

Measured resistance of thermocouple tube - 700



1/26/54 Measured output waveform of signal device -



Helipot turn	R_0	R_0 3achw
CW — 10k $\left\{ \begin{matrix} 5 \\ 5 \end{matrix} \right.$	5 Ω	4.4
1/8	120 Ω	
1/4 9.7k $\left\{ \begin{matrix} 275 \\ 275 \end{matrix} \right.$	270 Ω	275
1/2 9.5 $\left\{ \begin{matrix} 300 \\ 300 \end{matrix} \right.$	480	500
1 9k $\left\{ \begin{matrix} 3k \\ 3k \end{matrix} \right.$	900	1000
1 1/2	1.13k	
2 8k $\left\{ \begin{matrix} 3k \\ 3k \end{matrix} \right.$	1.68k	2.1k
2 1/2	1.94k	
3 7k $\left\{ \begin{matrix} 3k \\ 3k \end{matrix} \right.$	2.2k	3k
3 1/2	2.4	
4 6k $\left\{ \begin{matrix} 4k \\ 4k \end{matrix} \right.$	2.55	4k
5 5k $\left\{ \begin{matrix} 5k \\ 5k \end{matrix} \right.$	2.65	5k
6 $\left\{ \begin{matrix} 4k \\ 3k \end{matrix} \right.$	2.55	6k
7 $\left\{ \begin{matrix} 3k \\ 3k \end{matrix} \right.$	2.2k	7k
8 $\left\{ \begin{matrix} 3k \\ 3k \end{matrix} \right.$	1.65	8k
9 $\left\{ \begin{matrix} 3k \\ 3k \end{matrix} \right.$	910	9k
10 $\left\{ \begin{matrix} 3k \\ 10k \end{matrix} \right.$	8.6 Ω	10k

This explains change in E_p as signal helipot changed input grid resistance changed giving variable grid potential -

With 68 Ω as input resistor to test stage - no change noted in E_p as input signal varied until at end of scan ~~bottom~~ in the last 1/4 turn of helipot

E_p read - 3 mv

Readings of input signal were as follows.

E_s rose sharply to 2 mv & after 0.4 turns of pot began slowly to drop reaching null at 0.9 3/4 turns, rose sharply to 2.3 mv at extreme ccw pot position


E12

8/26/84

Measurements of network not attached to test stage gave explanation -


High Volt	1/8	1/4	1/2	1	2	4	5	6	7	8	9	9 1/2	9 3/4	10
Volts	.04	.05	.06	.06	.07	.09	.11	.14	.18	.26	.5	.95	1.8	7.0

Modified Voltage source -



Output continuous rise from 0 to 7.8 mp.

8/27/84



- E_s nearer to remaining 24k - C_p negligible.
- S tied to P: no good response
- Back to point of connection, with screen tied to divider point - 100k-24k-

E_s	E_g	E_k	$E_s R_s$	C_p	Adjusted R_s to lower static E_g & R_k
min	-1.58	.15v	13.7 mv	24.8	
max	1.64	.155	13.8	24.6	

E_s adj to lower value by shunts 24k by 30k -
 Adj E_s to 10V & R_k to get C_p 83v
 at given signal $E_g = 2.76$ mv - $C_p = 29.9$ v
 max 7.6 mv + $C_p = 38.4$ v } 1.50



tie S to K: show C_p +83. - same order of requirement given -


8/22/84

$R_{in} = 0$

$E_p = 21.8 @ E_f - 4 mV \quad E_s 9.8V$

$E_p = 20 @ E_f + 7.6 \quad E_s 9.7V$





ROK.
2-15-52

Scale of 16-15 min

#1	#2	#3	#4
1125-4	1218-15	1125-2	1124-2
1124-12	1191-2	1124-14	1124-12
1124-14	1183-6	1125-2	1124-15
1124-11	1253-8	1124-2	1124-9
1124-13	1302-13	1124-10	1124-10
1124-12	1179-13	1124-10	1124-10

Scale of 64-15 min

281-12	281-16	281-11	281-12
281-13	281-31	281-13	281-13

Std + BC.

774-14	6-11	7-15	7-9	10.54A
8-12	969-11	16-0	16-6	
17-6	6-8	872-11	23-15	
84-7	18-0	8-0	256-0	

Scale 256-15 min

70-77	70-71	70-78	70-71
70-79	70-92	70-71	70-70
70-79	70-214	70-71	70-72
70-74	70-240	70-72	70-73

Scale of 16-15 min

1124-14	1129-6	1124-13	1124-13
1124-12	1124-13	1124-12	1124-15
1125-2	1125-1	1125-0	1125-2
1124-12	1124-10	1125-0	1124-9

Std + BC

771-4	7-3	9-3	8-5	12.36A
8-6	950-7	18-0	15-9	
17-10	7-5	842-11	23-3	
25-2	15-2	8-0	738-14	

Room 215 - Record of Servicing

1-1-53 1/2

previous record: p 76

1-23-53 Counters 2 & 3 not operating - alpha not counting
 alpha - str: replaced 544 in scaler unit.
 C-2: not registering, scale 114. Replaced 114 unit.
 Pulled unit: all 6A15's 1/2 bad

6A15: 1/2, 1/2 - replaced

0-3: Put in 114 unit - still counting at extremely high rate
 original 114 unit.

Replaced input strip - count rose only, double; very
 after warming up or. on box.

Input Sensitivity check on C-4:

Signal: NISC pulse generator output fed into input of 118 strip

NISC setting: 600, 1 μ s, neg

C₄: @ .19 volts

C₃: adj to .19 volts.

C₂: adj to .15 volts.

C₁: reading .15 adj to .19 volts.

or. 11100

1-28-53

2-2-53 #2 tube on scaler 4 out - needs to be checked & replaced.

3-9-53 1614 Checked output stages of Power supplies (except HV) tubes
 in C₁-C₄ made replacements - bad tubes > 80%

This seemed to fix C₁ not lighting 2 & 8 or registering

3-26- C₂ Strip 118 - bad tubes replaced all - (had been there
 in counter HV off)

4 Replaced bad tubes in input strip (118)

Worked run, strip count check: none on C₂, 4416 on C₁, none C₃ C₄

4-8-53 Counter check - C₁ C₂ C₃ C₄

0853: 5-min 60N: 1126-12 1202-7 1200-10 1125-12

0859: " " 1125-2 1125-7 1125-2 1125-2

0906: " " 1127-0 1154-11 1126-13 1126-12

0914: " " 1124-14 1125-8 1124-12 1124-11

0920: " " 1126-7 1128-3 1126-6 1126-7

no rest shut out 807-0 8-1 7-15 8-12

816-5 380-9 16-4 15-11

823-8 387-15 883-13 24-4

(cont) 832-3 994-11 891-12 820-1

Counter Check: Std & bkg.

4-2-53

	C ₁	C ₂	C ₃	C ₄
Abulation of data p. 125	S: 807-0	972-8	862-9	795-13
	B: 9-5	8-1	7-15	8-12
	7-3	7-6	8-5	6-15
	8-11	6-12	7-15	8-9

Compare 9-23-52.

S:	{ 802-10	958-3	867-0	791-1
	{ 809-14	995-4	867-12	827-7
B:	8-14	6-11	8-12	8-3
	7-9	7-0	7-8	7-6
		6-10	8-4	
		7-6	7-15	

OK.

High

OK.

OK.

Input sense of C₁ checked. Was -11 rather changed to -19.

6000 feat. etc.

C ₁	C ₂	ΔC = 3
1126-9	1126-6	
1126-13	1127-1	ΔC = 4

9-10-53

C-2 not counting, registering on scale of 16, but OK on 6V
Noted "8" out on x16. Replaced strip 119.

C-2 not satisfactory on 6000 adj. slope sets. - OK.

4-13-53

Abundance of 119 strip of C-2 replaced tubes.

Adj to -2 - input + signal > 1 volt max away -

Put in new strip - 6000 high count - checked all strips

Replaced bad tubes. Installed strips adj. gain set. -

6-7-53

C-2: No adj on HV - High count position may same as
low-count max -

6A95 & 6A46 (V301, V303) replaced.

Still bad -

Voltage check showed only 200 + volts to supply
material tested 300

Checked 6A57 & 6A46 in 60-power supply: 6A57 dead -

OK.

6-2-53

Re-installed in 215 - HV OK.

Not scanning - scaled only thru "3". Replaced 114 strip

Put in new films after blow on reapplication of power -

6000, 1 min, count OK. for C₂ & C₁ 1314 -

6-10-52: Dead Time Scope Measurements -

Connection $\int \frac{E}{R} dt$ scope direct - no attenuator
 scaler

1	500-525 μs
2	400-425
3	460
4	425

PM 25

~~Connection~~

7-17-53 Dead Time on C₁ - Takamas above - 320 μs

9-21-53 #5 External 600 rms avg -

#1 External 600 (VCC pulse generator) quite normal -
input sensitivity is -27 volts -

10 min - 60 u - 522.464 \pm 46

1000 u - 10 min - background - 1264 \pm 62

std 697 \pm 15

std 685 \pm 15

std 686 \pm 15

1-4-53 Serviced C₃ & C₄ - tube check & replacement

140 East End Motors

6-5-53 Unit in 107. Blown fuse. Tube check - Replaced
5Y3, 6X5, 6V6, 1-6N7 that 9 tubes. Turned on VR-150 glow
657, 1-6AL5. too bright. Removed to 210.

6-8-53 Series resistors in regulation 150v. supply appeared
burned. Measured - R_1 & R_2 , each 2200 Ω , 2W. Were
each 250 Ω . R_3 & R_4 , each rated 4700 Ω , 2W, measured
<100 Ω .



6-17-55 Dumont 204H - Vertical Reflection Sensitivity on DC

horizontal deflection (cf p 152. Acc. amp rough check)

Y-Attenuator	Y-Amplitude	Amplitude (Screen Divs)	m.v. potentiometer
1	10	-	5 mV.
	50	1	
	100	1.6	
1	20 37	1	10 mV
	52	2	
	70 71	3	
	92	4	
	26	1	20 mV
	35	2	
	47	3	
	56	4	
	62	5	
	70	6	
	81	7	
	93	8	
	17	1	30 mV
	28	2	
	37	3	
	45	4	
	52	5	
	58	6	
	65	7	
	72	8	
	80	9	
	91	10	
	97	11	
	100	20	62.3 mV.
1	100	25	76.9
1	100	30	93 mV.
1	100	40	123

152
4/11/51

Calibration of Scopes

DuMont 304H:

AC volts (rms)	Amplitude p-p	Y Gain Setting	Deflection	(each small square = 1)
1mv	14.4	AC 1 - 100	1	
10 mv	14.4	AC 1 - 100	10	above center
"	"	1 - 10	1	
"	"	10 - 100	1	
100 mv	141.4	10 - 100	9.5	
1000 mv	1414	100 - 100	1.0	
		100 - 100	10	
		1000 - 100	1	

Tektronix

AC volts	P-P volts	Ampl. Deflt.	Gain	1st Att.
7mv	1mv	1cm	x100; 1	2st x1
70.7	10mv	1cm	x10; 1	2st x1
70.7	100mv	1cm	x1; 1	2st x1

Input probe Attenuator x10. Sens will prob 10mv/cm

Scope Calibration - Tektronix 2/24/53

(A) Direct - checks out as rated on front panel -

small
more

to

10 no/a

CS


CS

CS

8-90	84	88	99
16-55	87	81	
32-33	85		66

335-8D KEUFEL & ESSER CO.
10 X 10 to 6 1/2 inch, 1/4 in. lines centered
MADE IN U. S. A.

Deflection Voltage



DEFLECTION SENSITIVITY
TEKTRONIX SCOPE
2-16-52

Scope Deflection, Cm.

Check on Voltage Sens Limits of Scope, Tektronix 511A-D.

Input HP Audio Sig Gen → General Radio AF microvoltage → scope
 Known AC voltages are impressed on scope from GR instruments

Scope Input R.M.S.	Scope Input probable	Deflection
1.3 mV	3.68 μV	1cm x100 ²¹
2.85	8.06	2cm x100 ²¹
4.25	12.10	3cm x100 ²¹
6.10	17.25	4cm x100 ²¹
6.60	17.25	5cm x100²¹
14.2	40.2	1cm x10
27.7	78.4	2cm x10
44.0	124.2	3cm x10
57.5	168.1	4cm x10
57.5	168.1	9cm x1
118	339.8	1cm x1
240	701.5	2cm x1
372	1052	3cm x1
520	1470	4cm x1

With Discriminator Set at 35

Voltage	Background	Source		Voltage	Source	Background	
1350	0x16+10	0x16+10	0.0333 0.0330 0.0300	1350	0x16+7	0x16+5	0.0233 0.0167 0.0133
1400	0x16+9	0x16+11	0.0367 0.0333 0.0500	1400	0x16+4	0x16+11	0.0367 0.0667 0.0167
1450	0x16+10	0x16+15	0.0233 1.0133	1450	0x16+2	2x16+5	0.0667 0.0233 0.1467
1500	0x16+7	19x16+0	0.0267 0.0300	1500	0x16+2	2x16+12	0.0667 0.1467 0.0233
1550	0x16+8	128x16+7	0.030 19.1867	1550	0x16+7	3x16+9	0.1833 0.0967 0.0967
1600	0x16+9	89x16+160	0.0267 18.41 0.0133	1600	2x16+5	1x16+13	0.1833 0.8150 0.8150
1650	0x16+8	21x256+147	0.0233 29.543 0.0233	1650	1x16+13	3x16+9	1.400 1.220 0.080
1700	0x16+4	38x256+216	67.3000 0.0867	1700	0x256+245	15x16+5	1.916 5.167 3.320
1750	0x16+7	58x256+15	79.630 0.1133	1750	1x256+144	1x256+110	26.100 3.323 146.240
1800	0x16+7	78x256+183	97.98 0.450	1800	2x256+112	35x16+15	3.020 330.420 15.263
1850	1x16+10	33x256+81	113.710 0.970	1850	6x256+14	3x256+224	534.630 30.110 719.360
1900	2x16+2	38x256+70	114.44	1900	10x256+104	3x256+169	45.790 874.18 65.567
1950	9x16+7	44x256+107		1900	57x256+27	71x16+13	1003.05 138.04
2000	18x16+3	45x256+24		2000	129x256+18	17x256+227	
2050				2050	208x256+225	35x256+73	
2100				2100	981x256+13	53x256+9	
2150				2150	337x256+146	76x256+196	
				2200	391x256+209	161x256+24	
				2300		225x256+172	NOT accurate too erratic 963.48

Discriminator set at 40

Voltage	Background	source	counts/sec	Voltage	source	background	counts/sec.
1350	0x16+8	0x16+10	.0267 .0333	1350	0x16+0	0x16+0	
1400	0x16+10	0x16+12	.0333 .0400	1400	0	0	
1450	0x16+11	0x16+14	.0367 .0467	1450	0	0	
1500	0x16+12	3x16+4	.0400 .1703	1500	0	0x16+6	.000 .020
1550	0x16+10	64x16+9	.0333 3.4433	1550	sudden burst 1x16+11	1x16+0	.090 .053
1600	1x16+0	57x16+60	.0533 12.36	1600	0x16+3	sudden burst of fast 19x16+9 counts 20x256+77 peak re-count	.001
1650	1x16+0	13x256+69	.0533 11.3233	1650	2x256+287 in 2 bursts	0x256+24	
1700	1x16+0	29x256+81	.0533 25.0166 .0700	1700	0x256+154	1x16+2	
1750	0x256+21	44x256+195	.0700 37.546	1750	0x256+175	sudden fast bursts 205x256+138	
1800	0x256+17	59x256+183	.057 50.103	1800	worse	worse	
1850	1x16+5	78x256+225	.0700 67.3100	Stopped operation			
1900	1x16+8	97x256+162	.0800 83.3100				
1950	0x256+58	112x256+168	.1933 96.1333				
2000	13x16+6	118x256+130	.7133 100.0833				
2050	1x256+100	125x256+250	1.1867 106.67				

HANFORD STS. COUPLERS

Plateau Data				Counting Equipment			
8-5-53				AIC Model 1040			
7 min of Counts 5 min. (300 sec)				Attenuator set at 4			
9/8 x 1 1/2" Cb.:				Discriminator set at 20			
Example Scaler Reading				Example Scaler Reading			
Voltage	Background	Source	cts / c/s	Voltage	Background	Source	cts / c/s
1300		0x6+3	.01	1400		2x16+2	.34
1350		0x6+8	.02.67	1450		35x16+5	365 188.2
1400		1x16+5	.07.0	1500		52x16+4	836 278.7
1450		53x16+1	844 2.82	1550		58x16+6	924 316.3
1500		72x64+25	4633 15.44	1600		57x16+11	923 3.02
1550		52x256+71	13583 44.61	1650		51x16+3	519 2.73
1600 ✓		57x256+129	94.67	1700		32x16+8	41x16+2 1.23
1650		101x256+130	103.69	1750		133x16+0	209
1700		141x256+93	120.63	1800		488x64+18	106.3
1750		154x256+250	132.25	1850		430x256+130	367.38
1800		164x256+49	140.11	1900		754x256+104	60.76
1850 ✓		179x256+182	153.35	1950		1000x256+128	853.26
1900		195x256+29	166.50	2000		1212x256+118	1034.6
1950		276x256+112	253.89	2050		1275x256+57	1088.19
				2100		1239x256+53	1056.60
				2150		1788x64+49	1525.81

8-6-53

Discriminator setting: 20			
3/16 ✓			1/2
1350 ✓	152x16+1	123x16+0	1350 60x16+13 80x16+3
1400	150x16+0	125x16+3	1400 80x16+5 81x16+10
1450	197x16+15	168x16+9	1450 94x16+8 104x16+8
1500	101x16+2	309x16+2	1500 129x16+11 92x16+7
1550	208x16+3	1013x16+12	1550 382x16+9 439x16+2
1600	299x16+10	334x64+34	1600 62x64+57 56x64+63
1650	7x64+35	397x64+38	1650 38x64+9 41x64+47
1700	25x64+2	138x256+9	1700 10x64+41 16x64+4
1750	9x64+44	129x256+94	1750 7x64+46 95x64+20
1800			

Bad connection accounts for irregular counting.
 This recorded counts with no leads or cables

2-3/7-53

Time = 20

5/16 x 1 1/2

1/2 x 4"

Voltage	Background	Source	a/b	Voltage	Source	Background	a/b
1350	0x16+5	0x16+4	5.013 / 0.10	1350	0x16+10	4x16+5	3.053
1400	0x16+3	1x16+0	.053 / 0.03	1400	0x16+13	5x16+15	6.28
1450	0x16+3	27x16+15	1.49 / 1.78	1450	1x16+1	7x16+13	6.43
1500	0x16+9	194x16+2	.610	1500	0x16+10	6x16+14	.317
1550 (4:20 PM)	0x16+3	124x64+16	9.62 / 11.77	1550	4x16+2	8x16+15	.057
1550 (8:30 AM)	0x16+3	188x64+40 592x16+15	26.51 / 31.06	1550	3x16+2	14x16+13	used Resistor to estimate reading on tube
1600	0x16+2	269x64+6	31.66	1600	6x16+13	8x16+15	.033
1650	0x16+4	90x256+249	57.41	1650	14x16+12	10x16+15	.363
1700	0x16+10	111x256+93	77.63	1700	32x16+10	8x16+6	.477
1750	3x16+1	122x256+250	95.63	1750	9x256+80	14x16+3	1.786
1800	6x16+8 1x16+2	125x256+122	104.94	1800	104x256+106	15x16+11	.583
1850	4x16+1	131x256+120	107.15	1850	337x256+208 340x256+79	2x256+64	.174
1900 (started random Post counts)	3x16+14	148x256+115	112.14	1900	609x256+66	12x256+76	7.95
1950	75x64+35	179x256+162	126.68	1950	839x256+65 854x256+65	41x256+143	.757
			153.29	2000	1035x256+33	92x256+220	89.1
			16.12	2050	1215x256+153	213x256+203	290.4
				2100 (started random Post counts)	1286x256+55	311x256+1	57.9
				2150	1321x256+102	268x256+205	10.49
							209.0
							25.66
							88.3
							79.24
							1037.3
							182.44
							1092.6
							265.39
							1127.7
							205 229.38

With Discriminator set at 30

Voltage	Background	Source	counts/sec.	Voltage	Source	Background	counts/sec.
1350	1x16+11	0x16+10	.09 backgrnd .033 source	1350	0x16+12	1x16+14	.04 source .1 backgrnd
1400	4x16+6	1x16+9	.1267 .0833	1400	0x16+4	2x16+13	.0133 .150
1450	0x16+9	2x16+9	.030 .133	1450	0x16+7	3x16+13	.023 .230
1500	1x16+7	72x16+6	.0767 3.860	1500	6x16+4	7x16+0	.323 .3733
1550	0x16+14	247x16+13	.049 13.22	1500	6x16+14	16x16+4	.367 .867
1600	0x16+9	33x256+85	.03 28.16	1500	1x256+103	34x16+10 0x256+200	1.1967 1.256 count
1650	0x16+10	55x256+54	.033 47.113	1650	3x256+140	0x256+171	3.0266 .570
1700	0x256+14	78x256+135	.0467 67.01	1700	4x256+145	5x256+77	3.8767 4.5233
1750	1x16+9 1x16+10	97x256+9	.0867 82.333	1750	5x256+246	1x256+152	5.0867 1.360
1800	1x16+3	92x256+147	.0033 78.9967	1800	37x256+178	43x256+193	31.573 37.337
1850	1x16+3	112x256+91	.0633 96.73	1850	11x256+251	13x256+37	10.223 27.883
1900	used technique here on 0x256+49	127x256+97	.163 108.6966	1900	80x256+179	8x256+9	68.83 608.566
1950	23x16+15	133x256+11	1.2076 113.493	1950	266x256+36	19x256+200	228.213 16.88
2000	Started Supply 50x16+9	148x256+1	2.6967 136.2967	2000	554x256+33	38x256+109	509.413 32.79
2050	320x16+14	161x256+204	17.113 138.067	2050	715x256+53	53x256+158	612.870 45.953
2100	Too erratic here on			2100	2 1/2 min count 1437x256+141	35x256+88	746.53 60.986
2150				2 1/2 min 2150	502x256+25	62 1/2 sec 25x256+91	891.846 103.956

with Discriminator set at 25

5/16 x 1 1/2"

1/2" x 4"

Voltage	Background	Source	counts/second 0.33-background	Voltage	Source	Background	counts/second 0.06 - source 0.15 - background
1350	0x16+10	0x16+10	0.33-source	1350	1x16+2	2x16+13	0.067
1400	3x16+2	1x16+4	0.167	1400	1x16+1	4x16+0	0.213
1450	1x16+15	27x16+1	0.067	1450	0x16+14	11x16+3	0.1003
1500	0x16+13	152x16+15	1.443	1500	6x16+0	5x16+0	0.447
1550	0x16+15	126x16+4	0.043	1550	7x16+4	14x16+55	0.320
1600	0x16+14	52x256+114	0.05	1600	4x256+127	16x16+40	1.5067
1650	1x16+19	75x256+21	0.467	1650	6x256+33	5x256+24	2.17
1700	1x16+3	101x256+57	44.753	1700	0x256+4	2x256+11	3.8367
1750	1x16+5	96x256+21	0.2767	1750	6x256+19	4x256+34	3.5467
1800	1x16+15	110x256+161	0.07	1800	6x256+247	4x256+10	5.23
1850	3x16+8	121x256+219	0.0633	1850	39x256+130	5x256+234	4.3467
1900	14x16+6	133x256+206	0.07	1900	159x256+61	17x256+139	6.906
1950	55x16+5	142x256+234	86.2767	1900	383x256+130	27x256+146	2.33
2000	8x256+73	155x256+25	0.133	2000	596x256+227		5.1833
2050			94.403				3.5766
2100			103.983				5.943
2150			114.113				3.4467

started skipping with source
257 counts on background
14x16+6

With discriminator set at 45

5/16" tube

Voltage	background	Source	(Background) Counts/sec.	(Source) counts/sec
1350	0X16+10	0X16+10	.0333	.0333
1400	0X16+10	0X16+11	.0333	.0366
1450	^{360 sec.} 0X16+12	0X16+14	.0333	.0466
1500	0X16+14	1X16+3	.0466	.0633
1550	0X16+12	30X16+5	.0400	1.6166
1600	0X16+13	124X16+8	.0433	6.640
1650	0X16+14	8X256+16	.0466	6.880
1700	0X16+14	19X256+164	.0466	16.746
1750	0X16+14	31X256+155	.0466	26.970
1800	0X16+14	44X256+80	.0466	37.813
1850	1X16+4	61X256+65	.0666	52.270
1900	1X16+7	82X256+25	.0766	73.390
1950	2X256+36	101X256+136	.1200	86.640
2000	0X256+168	113X256+162	.5600	96.426
2050	0X256+246	119X256+180	.8200	102.746
2100	^{irregular count} 3X256+28	^{100 sec.} 42X256+263	2.953	109.150