

## BOOK93R

### *Notes:*

"1955" wrote on post-it-note on front

"INST" on spine

"NB WI-3" on spine

Blank pages: page opposite page 1, 8, 9, 161, 162, 163, 291-300, inside back cover sheets

-half sheet between front cover sheets

-page 21 has graph taped to it

-page 25 has graph taped to it

-page 28 has (8.5x11) graph taped to it

-page 32 has graph plus 1 long sheet plus 1 (8.5x11) sheet taped to it

-pages 39/40, 41/42, 59/60, 67/68, 83/84, 111/112 each have their corner turned down

-page 43 has sheet taped to it

-page 53 has sheet taped to it

-pages 65, 137, 139, 149 have long sheet taped to each

-pages 100/101 have 4 half sheet between pages

-page 123 has 2 graphs taped/glued to it

-page 142 has graph taped to it

-page 156 has sheet taped to it

-page 161 has 4 sheets stapled to it

-page 162 has 3 big graphs stapled to it

-page 163 has 3 long thin sheets taped side by side, 6 big graph sheets and 1 sheet stapled to it

-page 165 has 2 graphs attached

-page 166 and 167 have 1 graph each

-page 231 has 3 sheets attached

-page 240 has 1 graph taped

-pages 241 and 242 have 1 long thin sheet

-page 243 has 1 graph and 1 small thin sheet taped

-pages 244 has 1 long thin graph and 1 (8.5x11) graph sheet

-pages 248, 250, 254, 257, 267, 268, 281, 283 each have 1 (8.5x11) graph attached

-page 271 has 4 long thin sheets taped side by side

-page 272 has 1 long thin sheet

-page 273 has 2 sets of 3 long thin sheets taped side by side

-pages 290/291 have 2 (8.5x11) sheets between pages

*Scanned by:*

*Sheila Finch*

*RSICC /Oak Ridge National Lab.*

*September 7, 1999*

ROHREIZ

CONTENTS

NB-NI-3

Subject	Page
Instrument Summary	1
Nautilon sensitive Scintillation Detector	10
General Instrument Maintenance	12, 75, 82
Selsyn Motor Check	20
Tektronix Scope Calibration	24
G.M. pulse amplitude check	25
Preparation for APPR	26
Table Drive, Rm 107. Relay insertion	28
Scintillation Counter, C-5, Rm 215	28
Test Rack, Rm 210	24, 76, 83
Servo, 107, selsyn mounting	35
Scintillation Counters SA-1, SA 2	36-53
	55
Miniature Counters	(TUBE ASSEMBLY) 59-71 74
West Instrument Placement	54, 84, 88, 91, 112
West Instrument Placement	57, 72, 87
⇒ Bldg. Instrument Modifications Suggestions	90
Trip Circuit, A-D, 107	98
Period Timing Circuit	110
L & N Speedomax H = Check out	116
Trip Circuit, DC-2, DC-3	117, 125
Subminiature Tubes VX55, 5803, Curves	119
201 CONTROL Tie-ins	130
108 CONTROL Tie-ins	131

INSTRUMENTATION -

4/55 - 2/61



PIONEERS SINCE 1831

# Account Book

## No. S 149

NO UNITS

Journal . . . . .

Ledger, Single Entry . .

Ledger, Double Entry . .

Record Ruled (27 Lines)

Made in 150, 200 and 300 Pages

MADE IN U.S.A.

TO REORDER, SPECIFY NUMBER, RULING AND THICKNESS INDICATED ON BACKBONE OF THIS BOOK.

## Expected Deviation

90% confidence

$M = \text{AVG. CT./OBSERVATION}$

$N = \text{NO. OBSERVATIONS}$

$$\sigma_1 = \sqrt{\frac{N-1}{N}} \cdot \frac{1.645}{\sqrt{M}}$$

95%

$$\sigma_2 = \sqrt{\frac{N-1}{N}} \cdot \frac{1.96}{\sqrt{M}}$$

cf w/ observed variation -  $\frac{+A?}{-B?}$

Probable error = 1.675

50% in

Precision index - Standard stat. error

$$\sigma = \frac{\sqrt{T}}{t}$$

$T = \text{tolerance}$   
 $t = \text{fun. error}$

$$t^2 \sigma^2 = RT$$

$$t = \frac{R}{\sigma}$$

INSTRUMENTATION INVENTORY

4/55

EAST AREA

A. NUCLEAR DETECTION

- C-1: BF<sub>3</sub> proportional counter (RCL); Pre-amp (AIC); kin. Amp (AIC) Scaler (AIC)
- BF<sub>3</sub> proportional counter (RCL); Pre-amp (AIC); kin. Amp (AIC)
- C-2: BF<sub>3</sub> proportional counter (RCL); Pre-amp (AIC); kin. Amp (AIC) Scaler (AIC)
- BF<sub>3</sub> proportional counter (RCL); Pre-amp (AIC); Lin. Amp (AIC)
- C-3: fission counter (NEPA); Pre-amp (AIC); Lin. Amp (AIC) Scaler (AIC)

TIMER, MICROFLEX, for above

AUDITORY MONITOR (NEPA)

- A: BF<sub>3</sub> ion chamber (NEPA); Pre-amp (AP); Amp (AP); <sup>VR</sup>TRIP } ; RECORDER (M-H)
- B: BF<sub>3</sub> ion chamber (NEPA); Log N. Amp (ORNL); RECORDER (M-H)
- C: BF<sub>3</sub> ion chamber (NEPA); Pre-amp (AE); DC Amp (GE); RELAY AMP (GE) RECORDER (G.E.)
- D: BF<sub>3</sub> ion chamber (NEPA); Pre-amp (AP); VR Amp (AP); <sup>POWER SUPPLY (EMC)</sup>TRIP } ; RECORDER (M-H)
- E: ANTHRACENE-SB19; Amp.; Control; RECORDER (GE)
- Monitor (109)
- Monitor (107) - RECORDER (G.A)

BRUSH DC Amp; BRUSH Oscillograph - used with C above

B. CONTROL

1. TABLE

- TABLE DRIVE
- SOURCE DRIVE
- CONTROL ROD DRIVE
- SAFETY ROD POSITION
- MAGNET SUPPLY
- SCRAM PANEL
- SERVO ROD

2. ARE - PW

2. ARE (PW) PANELS

- SOURCE
- CONTROL ROD
- SAFETY ROD
- REFLECTOR LEVEL INDICATOR

b. PW REFLECTOR CONTROL PANEL.

Pump; Feed-drain-dump valves; level probe

C. GENERAL

Amplifier; CLOCK, minutes - 1/60 min.  
TEMPERATURE INDICATOR: K-2 POTENTIOMETER - GALVANOMETER

#55

INSTRUMENTATION INVENTORY:

WEST AREA -

A. NUCLEAR DETECTION -

C-1: proportional counter (Methab); Pre-amp; Lin. Amp; Scaler; Timer-Reg. <sup>Reg. timing</sup> Power Supply, H.V.

C-2: proportional counter (GE); Pre-amp; Lin. Amp; Scaler; Timer-Reg.

C-3: proportional counter (Methab); Pre-amp; Lin. Amp-Scaler; Timer-Register; Power Supply, H.V.   
 TIMER, Vender-Rot, for above.

C-4: BF<sub>3</sub> proportional counter (RL); Pre-amp; Lin. Amp-Scaler (AK)

C-5: BF<sub>3</sub> proportional counter (BL); Pre-amp; Lin. Amp-Scaler (AK)

TIMER, microflex, for above

"Popper," sodible monitor C<sub>1,2,4,5</sub> -

DC-1: BF<sub>3</sub> ion chamber; Amp; Control; RECORDER (MH)

DC-2: BF<sub>3</sub> ion chamber; Pre-amp; DC Amp; RECORDER (M-H)

DC-3: BF<sub>3</sub> ion chamber; Pre-amp; DC Amp; RECORDER (M-H)

R-1: BF<sub>3</sub> ion chamber (ORNL); Pre-amp (AP); V. Amp (AP); <sup>{Range sel}</sup> Trip; RECORDER (L.H.)

R-2: Argon ion chamber; Pre-amp (AP); V. Amp (AP); Range sel.; - - - -

PM: ANTHRACENE-931A; Amp; Power Supply, H.V.

Power Supply (Oregon), L.V. for C<sub>4</sub>, C<sub>5</sub> & PM.

Log N: BF<sub>3</sub> ion chamber (ORNL); meter; RECORDER (MH)

BF<sub>3</sub> ion chamber (ORNL); Log N Amp; meter; RECORDER (MH); TRIP

BF<sub>3</sub> ion chamber (ORNL);

BRUSH DC AMP; BRUSH OSCILLOGRAPH, used with R-1

Monitor: 1/4-m tube; pre-amp, 201 RATE METER 202, RECORDER (EA) <sup>(RCL)</sup> [for i, ii, iii]

1/4-m tube; pre-amp, 202

1/4-m tube; pre-amp 102 RATE METER 102; RECORDER (EA) [for i, ii, iii]

#55 INSTRUMENTATION INVENTORY -

WEST AREA (CONT)

B. CONTROL -

PANEL 1 (202): MOTOR DRIVE, control-indicator, S-N

MOTOR DRIVE, control, S

MOTOR DRIVE, control-indicator, S-N

MOTOR DRIVE, control-indicator, S-N

MOTOR DRIVE, control-indicator, S

SOURCE DRIVE, control, S-N

\* WATER PUMP control, S-N

FUEL PUMP VALVE, S

\* WATER DUMP, S-N

MAGNET SUPPLY, S-N

\* WATER FEED, S-N

FUEL LEVEL INDICATOR, S-N

PANEL 1a (201): For Nell only - controls marked \*

PANEL 2 (202) (102): SOLUTION STORAGE, vertical tanks; LEVEL INDICATORS, TANK SELECTORS

FEED-DRAIN CONTROL, VACUUM PUMP, DUMP WELL

PANEL 2a (202) (102): SOLUTION STORAGE, "horizontal" tanks; LEVEL INDICATORS, TANK SELECTORS

PRIMARY SCRAM PANEL

SECONDARY SCRAM PANEL

★ WATER TEMPERATURE CONTROL -101

C. GENERAL -

AMPLIFIER

TEMPERATURE INDICATOR-RECORDER, 0-2 mv.

BOGEN INTERCOM

CLOCK, minutes, seconds

4/65 INSTRUMENTATION INVENTORY -

III COUNTING ROOM

- C1 G-M end-window counter; Scaler (AIC) (AMP)
- C2 G-M end-window counter; Scaler (AIC) (AMP)
- C3 G-M end-window counter; Scaler (AIC) (AMP)
- C4 G-M end-window counter; Scaler (AIC) (AMP)
- C5 ANTHRACENE-5819; Lin-Amp (AIC); Scaler (AIC); Power Supply; Timer, V-R

TIMER, MICROPLEX

PC-1 gas-flow proportional counter (RCL); Nucleometer, power supply, pre-amp, scaler (RCL)

- \* GSC1: HV power supply (RCL); pre-amp (RCL); lin-amp (RCL); analyzer (RCL);  
2; NaI-6363 ;
- \* GSC2: HV power supply (RCL); pre-amp (RCL); lin-amp (RCL); analyzer (RCL);  
2; NaI-6363 ;

\* GSC-1 & GSC-2 under construction (4/6/55). At least 1 scaler will be required for each unit.

4/55 INSTRUMENTATION INVENTORY -

II SHOP AREA

- TEST RACK
- \* LINEAR AMPLIFIER (AIC) [East, Counting Room]
  - \* SCALER (AIC) [East, Counting Room]
  - VIBRATING REED ELECTROMETER (AP); Pre-amp;  $BF_3$  ion chamber
  - COUNTING RATE METER
  - A.V. POWER SUPPLY (NICC)
  - POWER SUPPLY (KAPCO) - - -
  - \* SCALER (AIC) [West]

SPARE NUCLEAR COMPONENTS -

THOSE ABOVE MARKED \*

- C - EAST: pre-amp; DC amp; Relay amp; control  
pre-amp; DC amp; Relay amp;
- DC 2 WEST: pre-amp, DC Amp.
- DC 3 DC Amp.
- B-EAST, or  
Log N-WEST: Log N Amplifier.
- Counter Tubes - RCL  $\frac{1}{2}$ " - 2  
1" - 2  
2" - 1  
Hard  $\frac{1}{2}$ " - 1  
 $\frac{5}{16}$ " - 1
- Ion Chambers (ORNL) - 2
- Pre. Amp (AIC) - 3
- BRUSH DC AMP - EAST
- MAGNET SUPPLY - EAST

## 4/55 INSTRUMENTATION INVENTORY —

## IV SHOP AREA (CONT)

## TEST EQUIPMENT -

SIMPSON 260 VOM - 4  
 m. m. v. Silver VTVM  
 H-P VTVM  
 DC Microammeter, RCA  
 ELECTROSTATIC VOLTMETER  
 BAILENTINE AC VOLTMETER  
 DECADE AMPLIFIER  
 DL MILLIVOLTMETER (I.C.)  
 MICROVOLTMETER  
 H-P AUDIO GENERATOR  
 NICC PULSE GENERATOR  
 MEGOHMMETER (PREED)  
 TUBE TESTER, 533  
 TUBE TESTER, portable  
 G-R IMPEDANCE BRIDGE, 650-A  
 TEKTRONIX { WIDE BAND PRE AMP.  
                   { 511AD SCOPE  
 TEKTRONIX 511A SCOPE (215)  
 DUMONT 304H SCOPE  
 REGULATED POWER SUPPLY 0-300V (OREGON)  
 DL SUPPLY, 0-100V, 0-10A.  
 PILE PERIOD SIMULATOR  
 RADIO, SX-71  
 DECADE RESISTOR BOX

## 4/55 INSTRUMENTATION INVENTORY —

## II SHOP AREA (CONT)

## TEST EQUIPMENT (CONT)

DECADE CAPACITOR BOX - 3  
 CLOCK, SECONDS - 1/100 sec. -  
 CLOCK, MINUTES - 1/100 min. -  
 TIME INDICATOR, 0.1 SEC. -  
 COUNTING RATE METER (RCL)  
 RUBICON POTENTIOMETER  
 RUBICON GALVANOMETER  
 LSA GALVANOMETER  
 RECORDER (MH) CIRCULAR 0-10 mv.  
 RECORDER (M-H) 0-2 mv.  
 RECORDER (F) 0-10 mv - 2  
 HEATER CONTROL BOX  
  
 DUMONT 241 SCOPE

## BUILDING INSTRUMENTATION, serviced by Y-12


INTERCOMS, 15 stations  
 TEMPERATURE CONTROLLERS S12, S16 103  
   S2: 107, 210, 215, 202  
   S3: 202  
   S4: 107  
 WATER HEATER, WEST (See p. 3)

WAPD-MR-45, p 96  
 WAPD-MR-46, p 149  
 ref: WAPDMR 42, p 86  
 WAPD-MR-41

SCINTILLATION DETECTOR - NEUTRON SENSITIVE

3/15/55 References suggest use of ZnS and Boric acid, anhydride, in a mixture as a phosphor for use with a photomultiplier tube.

⇒ Made up SB19 with ZnS (DuPont #1101 Phosphor, type D) phosphor: i. covered glass walls of tube with electrical tape for light seal.



ii. made phosphor screen - joined 4 strips of 3/4" cellophane tape, sprinkled on powder, shook off excess, cut to fit on face of tube, placed on tube with powder away from glass. The screen was secured to tube by cellophane tape along the edge.

iii - put mu-metal shield over tube and taped to base of tube with electrical tape.

iv. made Al cap to fit over opening over screen & taped in place with electrical tape. Aluminum was foil .0065" thick.

Ran rough check with alpha sources placed inside cap  
 Instrument - Unit II, 215 - SCDAT. Integral output,  $\bar{E} = 50$

Source	counting rate	Notes
1512 4m	2300 c/m	count not exact
14850 4m	2211 4/m	this good count
bkg	0	

counting efficiency 15%, source 3/2" from screen // E.P. Robson


3/16/55 Neutron Scintillation Detector

Added B<sub>2</sub>O<sub>3</sub> screen to type D made 3/15:

Sprinkled B<sub>2</sub>O<sub>3</sub> powder (Boric Acid, Anhydride - General Chem. Co.)

on sticky side of cellophane tape sheet -

Placed this powder side down against other screen in a loose fit by tuckin, edges around tube face.



PRELIMINARY CHECK

Time	SCDA "E"	Source	Count	Time of Count	Remarks
9:00 A	100	P <sup>232</sup> Bc (0.5x10 <sup>4</sup> / 1/2" face)	46	1 min.	
10:20		Bkg	0	1 min.	
10:30		PN-4	2612	"	
10:50		"	3207	"	
10:45		"	3062	"	
11:15		Bkg	33	1 min	
			34	2	These readings are accumulated counts for times noted.
			36	7	
			36	10	
			36	11	
			36		
9:38:20		PN-4	2751	1 min	
39:40			3168	"	
40:50			3241	"	
42:00			3236	"	
43:10			3014	"	

E.P. Robson



2/16/55 Neutron Scintillation Detector

Time	SCDA E	Source	Count	Time of Count	Remarks
9:05:10	100	Bkg	66	1	some procedure - accumulation count -
			95	2	
			99	3	
9:57:00		Bkg	62	5	Began new background -
			90	8	Calculator (Manroe) giving extra count during first 8 min. of count - not in use between 8 & 12 minutes.
			91	12	
11:11	300	PN-4 (1" foil tube)	889	1 min	
			904	"	
			866	"	
	200		1441	"	
			1400	"	
			1407	"	
	100		3000	"	
			2947	"	
			2968	"	
	50		5691	"	
			5582	"	
			5497	"	
12:03	50	10mg R <sub>2</sub>	27	5m.	
108:30	100	@ 1 inch	5	"	
114	200		2	"	
120	300		1	"	
126:30	50		2	"	
132:00	100				calculator begun during run. Spoiled count.

(cont)

Calculator (Manroe) giving extra count during first 8 min. of count - not in use between 8 & 12 minutes.

calculator begun during run. Spoiled count.

E. R. Rohrer

Neutron Scintillation Detector

Made up new detector: a - mixed type D d<sub>2</sub>O<sub>2</sub> in ratio 1:2.  
 b - Coated tube face with R-depe -  
 c - Sprinkled on mixture as evenly as possible  
 d - Covered with plastic spray can - This was too close & blew off some powder making uneven phosphor -  
 e - Covered with electrical tape -

2/22/55

Preliminary	Sample	SCDA E (EM)	Hour of Day	Count	Counting Rate /m
Check	Bkg	400	10:14	0	0
5 minute counts, unless otherwise noted		300	10:20	5	1.0
		250	10:26	3	0.3
		200	10:58	4	0.4
		175	11:04	5	1.0
		150	11:10	11	2.2
		125	11:16	21	4.2
lunch		100	11:22	15	3.0
		75	11:52	22	4.4
		50	11:58	27	5.4
6.5x10 <sup>4</sup> n/pair/sec	PoBe 222	50	12:05	3188	637.6
resting		75	12:11	2690	538.0
on clamp		100	12:17	2426	485.2
on top of tube		125	12:23	2280	456.0
unit #		150	12:29	2139	427.8
LA = 1-32		175	12:35	2001	400.2
HV not measured		200	12:41	1902	380.4
		225	12:47	1852	370.4
		250	12:53	1690	338.0
		275	12:59	1563	312.6
		300	1:05	1513	302.6
		325	1:11	1359	271.8
		350	1:17	1467	281.4
		375	1:23	1234	246.8
		400	1:29	1194	238.8

Sample	SCDA E	Hour of Day	Count	Counting Rate /m
PoBe 222	450	1:35	1135	227.0
	500	1:41	1046	209.2
	550	1:47	883	176.6
	600	1:53	835	167.0
	650	1:59	750	150.0
	700	2:05	635	127.0
PoBe 222	50	2:11	3108	621.6
	75	2:17	2583	516.6
	100	2:23	2284	456.8

Differential reading too much like integral -

Checked SCDA with pulse generator found unable to isolate single pulses at instrument settings -

Stopped run to check instruments

Checked HV: 800 volts.

E. R. Rohrer

Neutron Scintillation Detector

3/23/55 Made up out of block of Lucite a sample measuring device. Drilled a number of holes in block of varying depths and diameters. Filled with  $B_2O_3$  to determine weight of material in measured holes.

Weight of block 30.6082 g.

Sample Hole	Gross	$B_2O_3$
1	30.7975	0.1893
2	30.8801	0.2719
3	31.0286	0.4204
4	31.2835	0.6753
5	31.1093	0.5011
—	30.6098	—
6	31.3979	0.7881
7	31.6487	1.0389
8	31.9145	1.3046

reweighed block

30.6079 re-weighed block

repeated on several holes to see if any marked differences.

1	30.7918	0.1838
2	30.8838	0.2759
—	30.6072	—

weighed block

Checked wt. of "D" powder in several holes

2	31.1718 - 0.5646
5	31.6711 - 1.0639

E. R. Rohrer

Neutron Scintillation Detector

3/31/55 Tried another coating -

- a. Painted tube face with R-dope
- b. Sprinkled on "D" powder
- c. Painted on with brush dipped in thickener to thin-
- d. After drying added  $B_2O_3$  by first spraying plastic on surface & sprinkling  $B_2O_3$  over.

Balance weight 4151.

Covered with Aluminum reflector of electrical tape -

Hour of Day	Sample	SCDA E (INT)	Counts	Counting Time	Counting Rate %m	picked up counts from rest of CS -
12:23	Bkg	50	11	5 m.	2.2	
12:35	PB222	50 (I)	4430	5	886	
12:41:30	outside	50 160 (D)	750	2	375	
12:44:30		100 (I)	537	1	537	
12:46		100 160 (D)	88	1	88	
12:47:30		150 (I)	487	1	487	
12:49:30		500 (I)	337	1	331	
12:51		750 (I)	240	1	240	
12:52:30		1000 (I)	186	1	186	
12:57	PB222 @ 6"	200	49	5	9.8	source varying distance
1:03:30	3"	200	143	5	28.6	
1:09:30	2"	200	430	7	61.4	
1:17:30	1.5"	200	456	5	91.2	
1:25:10	1"	200	931	5	186.2	
1:31:10	3/8"	200	449	1	449	
1:35	-bkg-	200	3	20	0.15	
2:45:30	PB222 in H <sub>2</sub> O 1"	200	534	3	178	

E. R. Rohrer

Neutron Scintillation Detector

4/1/55 New Tube 8(a) Mixture of 1:1 B<sub>2</sub>O<sub>3</sub> - "D" powder made in paste in R-dope and poured on 6363 tube.  
(b) Let dry & covered with reflector & tape.

Can series of 1 minute counts

SCDA	I	II	III	IV	V
E	Count	Count	Count	Count	Count
50	316	335	332	314	322
100	131	130	133	152	164
150	97	112	79	107	102
200	75	68	65	68	66

Background @ 200 for E = 0 counts in 20 m.

Trouble occurred during extended background run counting "sports".

E. R. Rohrer

4/4/55 Checked system. SCDA II seemingly at fault refer to SCDA II file.

SUMMARY OF DETECTORS

DATE	TUBE	PHOSPHOR	TEST
3/15	5B19 metal shield	"D" - scratch dope	α test "E" = 50 <sup>221</sup> / 4850 ; bkg = 0
3/16	"	B <sub>2</sub> O <sub>3</sub> - scratch tape added	PB222 46 c/m @ E=100 ; bkg = 0
3/22	5B19	"D" - B <sub>2</sub> O <sub>3</sub> 1:2 (weight)	PB222 485.2 " ; bkg = 3
3/31	5B19	"D" on R-dope	
		B <sub>2</sub> O <sub>3</sub> on spray 1:1	PB222 537 " ; bkg -
4/1	6363	"D" - B <sub>2</sub> O <sub>3</sub> paste in R-dope	PB222 150 " ; " -

E. R. Rohrer

Neutron Scintillation Detector

4/5/55 6363 - B<sub>2</sub>O<sub>3</sub> - "D" paste 1:1

Time of Day	Sample	SCDA E (c/m)	Count	Time of Count	Rate c/m	H.V.	adj. HV
8:41 A	Bkg	100	X	5	0.8	800	R <sub>2</sub> = 1 meg
8:47	PB 222		806	5	161.2	800	
9:10			1333	5	266.6	850	
9:24			3021	5	604.2	900	
9:30			3029	5	605.8	900	R <sub>2</sub> = 1000 Ω
9:38			15868	5	3173.6	950	
9:45	bkg		2164	5	622.8	950	HV off
9:55	Set HV back to 800		800				check background @ E=100
10:00	bkg	100	2427	5	485.4	800	R <sub>2</sub> = 1000 Ω
			3353	5	670.6	800	← SCDA I improved II

→ TROUBLE IN WAY OF INCREASED COUNTING - TUBE?? TRIED 5B19 used 3/31/55

Time	Sample	SCDA E (c/m)	Count	Time of Count	Rate c/m	H.V.	Notes
11:05	bkg	100 (INT)	25	5	5		R <sub>2</sub> = 1000 Ω
11:11		100 (INT)	724	109	6.64		
2:00 PM		100 INT	196	42	4.67		new HV cable
3:00	PB 222 standard	100 (INT)	7413	10	741.3		

SCDA II not showing D-I difference - (as 3/22)

Operation shelved because of press of other projects -

E. R. Rohrer

5/2/55 Brought book up to date -  
 Activities during 4/55 principally routine:  
 Major effort - DC 233 - West Area } modification of trip circuit & testing  
 Others: Safety Timing Measurements - West Area  
 Instrument Maintenance: C3-W, R-1-W  
 Scintillation Counter System: material procurement.  
 APPR: preliminary design considerations.  
 Hot Critical: preliminary design considerations.

E.L. Rohrer

5/3/55 Checked out control and drive mechanism on T2W assembly before removal to Y-12 for fitting into "Hot" critical set-up.  
 East Area Maintenance - C3 ferrisim ctr.

- gas leak (w/ JFE)
- scale check (w/ VGH)
- 'A' down drift 1/2 hr after power on - 'A' on 'EP' power rather than 'CV'
- Table control and switch sticking -
- Crushed cables C3 & C1, w/ JFE

E.L. Rohrer

5/4/55 - APPR - West Experimental Console Assembly space requirements were studied. Tentative layout of arrangement of controls for West reactor Sid. Drew up request for purchase requisition -

5/5/55 - APPR - West Experimental area design considerations continued. Procured sample selsyn motors from X-10. Got hold of Veeder-Root catalogue to determine specifications of small counter to be used on console.  
 - Electrical reset mechanism for register on AIC scales 1010: two units received. Model 1290.

These are not immediately replaceable. Jones plug on new model (1290) 8 contact, whereas old (1220B) 6 contact. Switched plugs - Since mounting position of drive gear was closer to register in 1290, had to cut off two clubs of chair.

5/6/55 Drew up request for purchase requisition for Veeder-Root counters.  
 Continued APPR - West design considerations, checking out various selsyns on Rand. Detailed check to follow.  
 ARE "Hot Critical": Electrical department Y-12 measured load on emergency circuit, full 210kw, without fence lights 24kw. Generator 35kw. This reported to design group.  
 E.L. Rohrer


Hot Critical: Sketched rod drive mechanism for EMS.  
 5/9/55 APPR - West & Selwyn motors.

0930 power on - let motors stand cdc for time to check heating.

0934 power off.

⇒ unmatched pair very hot.

⇒ matched pair cool



matched pair:  
 SYNCHRO 115V TYPE 2324-1-B  
 TRANSMITTER: 60N SER. NO. 9894  
 BENDIX AVIATION CORP.

SYNCHRO 115V TYPE 2330-1-B  
 RECEIVER: 60N SER. NO. 7161  
 BENDIX AVIATION CORP.

KOLLSMAN  
 TYPE 528-0160  
 SER. NO. 523

SYNCHRO TRANSMITTER  
 115V. 60N.

1942 BENDIX AVIATION CORP.  
 C78248  
 CAL-11275  
 SER 5P130

Check is made on unmatched pair to see if such mismatch could be useable. Seems unlikely - let units cool while working on matched pair -

This pair indicates no run away as has been noted on pairs of large transmitter units.

Put plate on receiver marked in degrees to measure voltage at various positions of rotation.

Angle of Rotation	0	30	60	90	120	150	180	210	240	270	300	330
S1-S2	85.5	76.4	45.2	25	37	71	84.8	77.4	48.5	4.2	41	70
S1-S3	45.5	2.5	40	72.5	84.5	77.4	49.5	3.6	38.5	71	84.9	74
S2-S3	40	76.5	85	76	50	9.5	40	71	84.5	77	48	9.6

Vmax meter

E.L. Rohrer

5/9/55 Power on: (a) matched pair -

12:09 PM

(b) two large Bendix Synchro Transmitters -

To observe heating -

After 2 hour of operation of applied voltage - both pairs warm but not excessively hot.

⇒ Reviewed charts of monitor of regulated & unregulated voltage taken 4/20/55 - 4/27/55

CV: ran between 112-113 volts a few plateaus up to 113.5, mostly 112-112.6 no sharp changes

EP: ran between 118-121V. night 119-121 day 118-120 numerous small 1-2V spikes (voltage drops) during time

run following no particular pattern:

	M	T	W
2	7:30-8:00 PM 4/20	4/23 1142A	4/27 1:38P
1	1 AM 4/21	1:35P	3:44 9V
1	7 AM	2:56P	
1	8:25A 4/22	11:45P	
1	2 P 4/24	midnite - 5:30 A.5	2:59 8V
		1:22P	


on 4/26 2, 8 volt drops & on 4/27 once 9 volt drop these were very marked, causing dimming of lights momentarily.

Checked file on HV Power Supply, entry 2/28/55

Line	HW
110	907
112.5	905
115	904

$\frac{\Delta E}{\Delta t} = \frac{3}{5}$

E.L. Rohrer



5/10/55 Checked synchro transmitters on reject shelf -

#1. Rotor resistance 3  $\Omega$  ← low -

$S_1 S_2$  32  $\Omega$

$S_1 S_3$  32  $\Omega$

$S_2 S_3$  32  $\Omega$

Rotor stuck in commutator bearing

Taken apart - signs of excessive heating on rotor, insulation darkened & appears burnt at commutator end, shell marked

#2 Rotor open

over heating in windings

$S_1 S_2$  32  $\Omega$

$S_2 S_3$  32  $\Omega$

$S_1 S_3$  32  $\Omega$

#3 Rotor 600  $\Omega$  high

Shaft sluggish

$S_1 S_2$  32  $\Omega$

Burned out Rotor -

$S_2 S_3$  32  $\Omega$

$S_1 S_3$  32  $\Omega$

#4 Rotor 23  $\Omega$

Resistance values seem proper

$S_1 S_2$  32  $\Omega$

Some insulation worn off on stator.

$S_2 S_3$  32  $\Omega$

$S_1 S_3$  32  $\Omega$

#5 Rotor 1.8  $\Omega$  low

over heating 1/2 winding

$S_1 S_2$  34

insulating coat cracking from stator, probably from heat

$S_2 S_3$  34

$S_1 S_3$  34

The above are all Duik relays.

E. L. Lohr

5/10/55 New small Bendix synchros -

Receiver:  $R = 145 \Omega$ ;  $S_{12} = 220 \Omega$ ;  $S_{13} = 220 \Omega$ ;  $S_{23} = 220 \Omega$

Transmitter:  $R = 146 \Omega$ ;  $S_{12} = 220 \Omega$ ;  $S_{13} = 220 \Omega$ ;  $S_{23} = 220 \Omega$

Big Bendix - seemingly operable -  $R = 14$ ;  $S_{12} = 38$ ;  $S_{13} = 38$ ;  $S_{23} = 38$

$R = 14.5$ ; 38; 38; 38

Pulled from shelf one big Bendix with shaft hard to turn (had been on shirny) - Resistances: Rotor 14.5  $\Omega$

$S_{12}$  39  $\Omega$

$S_{23}$  39  $\Omega$

$S_{13}$  39  $\Omega$

Examined - front bearing stuck - commutator shaft cracked  
E. L. Lohr

5/11/55

J.F. Fox reported relay on safety drive on Soid.

All relays still on (povon on overnight) relay on safety hotter than others. Took resistance measurements in 202.

	M 1 Duik	M 2 Bendix	M 3 Duik	M 4 Bendix
Rotor	22.5	16	23	14.6
$S_{12}$	34	41.5	35	39
$S_{23}$	34	41.5	35	39
$S_{13}$	34	41.5	35	39

Since M 2 relay was excessively hot, removed for inspection & replaced w/ another Bendix.

Still not repair - inspection in 201 showed loose neutral at terminal strip on motor dolly. -  
E. L. Lohr

5/11/55 Selwyn pulled from 202 inspected, Found OK. Put back together & shelved.

There are now 4 selwyns in stock - 2 Dicks & 2 Bendigs

Forbore recorder on output DC 231 under test (had been removed from 202 because of spurious tripping):

- Take up spool clips & paper piles up.
- Put rubber cement on shaft driving take up spool.

Selwyn with broken commutator shaft (see p 23 5/10).

Patched shaft w/ rubber cement - Left to dry at room temperature for at least 24 hours (JFE)

E. R. Robson

5/12/55

Preliminary consideration control panel APPR <sup>no. of control rods</sup> sockets

G-M counters, Rm 215. Negative pulses directly to scalars AIC 1010.

Checked for pulse size.

With Nicc Pulse Generator calibrated Scope 511A, No 782 in 215.

Test Pulse	60.0 Scope Deflt	100µsec, negative, cu. Scope Setting 2 stage; X1 VAA-cw.	Scope Delay Sec.	T.P.	S.D.	S.S.	T.P.	S.D.	SS
0.1	.25		2.0	3.17	3.25	3.25, X1, X2, X4, X8	2.0	2.45	X1
0.2	.50		3.0	4.7	4.75	X1, X2, X4, X8	10.0	.75	X4
0.3 (5V)	.75		4.0	6.5	6.5	X1, X2, X4, X8	12.00	.375	X8
0.4 (5V)	.95		5.0	8.5	8.5	X1, X2, X4, X8	14.00	.45	X8
0.4 (50)	1.25	X2	6.0	1.35	1.35	X1, X2, X4, X8	16.00	.85	X4
0.5 5V.	1.2	2 stage, X1, X2, X4	7.0	1.75	1.75	X1, X2, X4, X8	18.00	1.6	X2
1.0	1.2	2 stage, X1, X4	8.0	2.25	2.25	X1, X2, X4, X8		1.85	X2
	1.25			2.05	2.05	X1, X2, X4, X8		.95	X4
	1.625			2.05	2.05	X1, X2, X4, X8		.50	X8
	1.625			1.125	1.125	X1, X2, X4, X8		.29	X8
	1.625			0.6	0.6	X1, X2, X4, X8		1.10	X4
	1.625			0.3	0.3	X1, X2, X4, X8		2.00	X2
	1.625			0.3	0.3	X1, X2, X4, X8		2.20	X2
	1.625			0.3	0.3	X1, X2, X4, X8		1.20	X4
	1.625			0.3	0.3	X1, X2, X4, X8		0.62	X8

E. R. Robson

5/12/55 Measured amplitude of pulse out of G-M counter -


#2 HV = 1400 - Deflection: 0.4 cm - 1 stage X1

0.45 cm 2 stage X8

Pulse amplitude  $\approx$  1.5 volts - negative

Cable from arrangement

this is size of pulse to scaling strips -



Plotted calibration.

Proof read paper for ADC -

E. R. Robson

5/13 Scintillation Counter - Discussed Cs counter system with

D. Scott in light of standard normalization of errors found in data -

Continued observation of scaling DC amplifier DC 231 -

E. R. Robson



CALIBRATION CURVES  
TEKTRONIX OSCILLOSCOPE  
511AD - Ser. No. 162  
Slip 58

← 2 STAGES GAIN →

DEFLECTION (CM)

PULSE AMPLITUDE (VOLTS)

25  
20  
15  
10  
0.9  
0.8  
0.7  
0.6  
0.5  
0.4  
0.3  
0.2  
0.1  
0

2.5

5

7.5

10

12.5

15

17.5

X1

X2

X4

X1

X2

X4

X3

5/16/55

Preliminary on APPR:

Checked through prints to note available power sources and wiring possibilities. APPR is to be set up in "NELL".

WIRE DESIGNATION	ORIGINAL USE	PRESENT USE			
A, B, C	25 amp. breaker 3φ, 208, distribution	same			
B, & A,	115v, 20A, distribution	same			
B/A	115v, 2A, distribution	same			
M11 H1	208, 3φ, motor, from 6 Amp. starter neutral to starter relays & limit lights relay starter, K <sub>1</sub> , to "raise" (up) switch relay starter, K <sub>2</sub> , to "lower" (down) switch raise contact to up LS, N.C. lower contact to down LS, N.C. indicator light to up LS, N.O. indice for light to down LS, N.O.	same	CONNECTOR CAP 201 RIK-M12-221 6 1/2"		
- terminal 202				H2	7
1 terminal 201				H3	12
-				H4	
-				H5	
-				H6	
+				H7	2
+				H8	4
+				H9	9
+				H10	11
M21 V1	from 3φ, 208, 6A, starter box to motor <del>neutral to starter relays &amp; limit lights</del> neutral to starter relays & limit lights starter relay, K <sub>3</sub> , to "raise" switch starter relay, K <sub>4</sub> , "lower" switch raise contact to LSE N.C. lower contact to LSL N.C. indicator light to LSR N.O. indicator light to LSL N.C.	same			
1				V2	5
1				V3	8
-				V4	7
-				V5	
-				V6	
+				V7	2
+				V8	4
+				V9	9
+				V10	11

5/16/55

WIRE NUMBER	ORIGINAL USE	PRESENT USE	WIRE NUMBER	ORIGINAL USE	PRESENT USE
+V11	indicator light to LSD (rod) N.O. switch to relay contact	same	+ P9	fuse CP 202 to w/dump sw & K12 cont.	same
-V12	relay K-11 to transformer & lamp - rect. output	same	+ P10	to relay coil K12	same
+V13		some	+ P11	to valve PB screw 202	same
1 V14		some	+ P12	w/dump sw	same
1 V15	+ rect. output	same	1 P13	PB screw & dump sw to relay K8 contact	same
M31 S1	from 3φ, 208, 6A starter box to motor	same	1 P14	SET contact (down) relay K8	same
1 S2		same	1 P15	HOLD contact (up) relay K9	same
1 S3		same	1 P16	relay K8 & K9 to K10 contact.	same
-S4	neutral to starter relays & limit lights	same	P17	fuse 101	same
-S5	starter relay, K <sub>5</sub> , to "raise" switch	same	P18	K8 contact	same
-S6	starter relay, K <sub>6</sub> , to "lower" switch	same	P19	K8 contact - K9 contact & K11	same
+S7	raise contact to LSR N.C.	same	P20	K8 contact - K9 contact	same
+S8	lower contact to LSL N.C.	same	P21	K9 contact	same
+S9	indicator light - LSR N.O.	same	P22	transformer	same
1 +S10	indicator light LSL N.O.	same	+ P23	transformer	same
+S11	indicator light LSD N.O. (140)	Level IND. PROBE	+ P24	transformer	same
1 S12	switch to transformer & lamp	tied to V13	+ PX	relay contact K12 solenoid valve	same
1 S13	- rect. output	same	+ SM1	rotor selsyns - 110v from switch	same
1 S14	+ rect. output	same	+ SM2		
Pump P1	from 15A, 208, 3φ starter box to pump	same	- SM3	selsyns M-1	
P2					
P3					
+P4	1φ of 208 to starter relay & ind. lights	same	+ SM6	selsyns M-2	
+P5	starter relay & lamps to relay K-11 contact	same	+ SM7		
-P6	relay K-11 to pump switch 202	same	+ SM8		
+P7	pump sw. 202 to pump sw. 201	same	+ SM9	selsyns M-3	
1 P8	pump sw 201 to K9 relay contact	same	+ SM10		

WIRE NUMBER	ORIGINAL USE	PRESENT USE	WIRE NUMBER	ORIGINAL USE	PRESENT USE
1715	(S) CA, 14, starter box	same	B7	156 NC	"
1716	to source motor	"	B8	lower contact to 156 NC	"
B4	fuse to st. relay light	"	B9	ind. light to 156 NO	"
B5	raise sw. starter relay K15 H	"	B10	ind. light to 156 NO	"
B6	to lower switch	"			

→ Room 107 cold - controller ok. - no voltage on relay.  
 loose wire in 110 coil circuit - fix ed -


5/17/55 East experimental Area:  
 Installed relay in table drive circuit. Relay contact normally closed, shunts table separation position of switch (points 23 & 24 on terminal board on table drive chassis). Relay powered by normal unregulated power such that contact is open when there is power.

C-5 Scintillation Counter Rn 215 -  
 Change of count with voltage line variation -  
 Power off 9:20<sup>h</sup> - Inserted source of meters -  
 Power on 9:32<sup>h</sup> HV on 0935.

Meters  
 AC volts Westco 476  
 0-150V

5 min counts

Line	H.V.	Count	Line	H.V.	Count	Line	H.V.	Count
9:41 <sup>h</sup> 105	906	1277 42	117.5	898	1274 11	112.5	908	1278 6
102.5	903	1283 35	120	896	1278 13	115	900	1275 5
9:53 110	902	1280 55	105	905	1270 10	117.5	897	1281 5
112.5	901	1279 28	107.5	903	1281 43	120	893	1278 2
115	900	1271 58	110	901	1270 47			



5/17

HV supply off 11:09 AM - Check line voltage - with meter - with  
 HV supply on - line = 112.5 ; HV = 895 - (HV just turned on)  
 capped meter 900

HV back on to stay 11:14 AM

Results:

line	HV	Counts/min
	906	16254 II
105	905	16258 II
	903	16429.4 I
107.5	903	16405.4 II
	902	16395 I
110	901	16265.4 II
	901	16376.8 I
112.5	900	16359.6 II
	900	16260.4 I
115	900	16320.4 I
	898	16309.4 I
117.5	897	16317.8 II
	896	16361 I
120	893	16364.2 II

E. R. Rohrer

5/18/55  
Lin. Amp Y100363

Standard - Imin counts @ 900v - PDI Sweep. HV. off 8:19-8:23 AM & wire set Van

LA Gain PDL Scaler Hour Sample PDL Scaler Hour Sample PDL Scaler

1-8	224	9	40	255	13	75	249	19
1-8	224	9	45	253	62	80	264	35
1-8	256	10	50	250	26	85	197	29
1-8	42	28	55	249	59	90	193	11
1-8	43	33	60	248	18	95	195	34
1-8	259	8	65	247	2	100	192	48
1-8	259	20	70	240	63	1-8	256	39
1-8	2	52	75	241	50	10	132	63
1-8	9	4	80	240	5	15	256	39
1-64	238	93	85	238	39	20	245	21
20	102	163	90	234	14	25	229	59
25	78	134	95	237	0	30	223	15
30	277	53	100	237	36	35	223	53
40	263	52	1-16	236	81	40	209	18
50	257	57	10.25	193	81	45	197	22
60	253	21	15	276	1	50	187	41
70	255	42	20	254	57	55	174	46
80	257	17	25	256	18	60	162	55
90	250	41	30	249	2	65	150	39
100	254	42	35	244	4	70	141	17
1-32	256	29	40	243	5	75	127	5
15	94	79	45	238	29	80	116	60
20	274	55	50	229	43	85	104	54
25	262	37	55	225	35	90	98	46
30	260	59	60	224	36	95	101	3
35	255	27	65	216	35	100	101	13
			70	211	43	1-410	91	236

2

0940

5/18/55

L.A. SENS.

Hour Sample LA Gain PDL Scaler Hour Sample PDL Scaler

70	28	1	70	0	8	64-15	5.1	16-60	0.245	2-10	0.069
75	18	3	50	0	7	20	10	70	0.29	15	0.255
80	11	46	40	0	10	30	0.25	80	0.842	20	0.452
85	6	63	30	0	12	40	0.368	90	0.387	30	0.83
90	6	24	20	1	4	50	0.493	100	0.439	35	1.03
10	286	10	15	1	14	60	0.60	8-10	0.183	40	1.13
15	224	18	10	2	1	70	2.77	15	0.062	45	1.48
20	178	60	40	0	11	80	0.81	20	0.107	50	1.68
25	129	12	70	0	12	90	0.92	30	0.200	55	1.11
30	86	14	50	0	7	100	0.94	40	0.247	60	1.8
32.5	65	55	35	1	5	32-11	0.1	50	0.396	70	2.18
35	51	7	20	2	6	15	0.3	60	0.45	75	
37.5	36	57	15	2	8	20	0.269	70	0.52	80	
40	26	49	10.5	3	1	30	0.51	80	0.66	90	
42.5	16	57	90	0	12	40	0.72	90	0.76	95	
45	11	20	70	0	15	50	0.95	100	0.865	100	
47.5	6	22	70	0	15	60	1.78	4-105	0.0433	105	
50	3	6	70	0	15	70	2.23	15	0.125	110	
52.5	5	8	50	1	7	80	0.42	20	0.218	115	
55	2	0	40	1	11	90	0.168	30	0.408	120	
57.5	0	11	30	2	1	100	0.169	40	0.57	125	
60	0	12	20	1	13	16	0.214	50	0.78	130	
65	0	3	10	1	13	15	0.214	60	0.985	135	
67.5	0	3	10	1	13	15	0.214	70	1.185	140	
70	0	6	10	1	13	15	0.214	80	1.41	145	
			10	1	13	15	0.214	90	1.615	150	
			10	1	13	15	0.214	100	1.86	155	

64 15  
48 (2.56) + 1.60

1123A  
std sh

Gain 20  
1-12  
2-2.4  
3.55-3.90v  
4.5v  
spread  
1.65 (8)  
(3ms)  
1.6v  
3.4ps  
8ps  
1.6v  
1.6ps

5/18/55

Comparison of "Std" with other sources

Std's Gain 2, pulse 1.7-2.1V (ls amp output)

Cs137 Gain 2, pulse 2.6  
1.3-1.8 (x2) 1.15-1.60 v/ls  
2.30

Sample	Gain	Scaler	No. 24	1-32	8-13	
100	1-4	39	314124	70	8 13	
90	3	6		80	9 3	
80	3	2		90	6 14	
85	6	0		100	7 13	
80	15	12		1-16	50	6 12
75	43	3		60	5 1	
70	117	11		70	4 5	
65	71	17		80	3 11	
60	195	46		90	3 11	
55	145	211		100	3 0	
50	397	71		1-8	50	3 10
45	10m1			60	1 12	
40	1-64			70	1 4	
35	20			80	1 11	
30	lite			90	1 4	
25	67	115		100	0 13	
20	61	59		1-4	100	0 10
15	9	16		1-2	70	0 7
10	17	8				
5	15	11				
0	16	15				
	14	5				
	14	0				
	13	3				
	12	11				
	1-32	11				
	50	11				
	60	9				

light shield  
not on  
much effect on

3:10P

back page  
de p 31 &  
bleg 3:35P

C-5

Scaler from C3 set up +  
put in cascade  
for high counts.


Curve in folder C5 33  
-215-

5/20/55

More counter	Cs137	Scaler cascade
100 0 1	20 50 5 3 2	193 76 14
90 0 1	67.5 85 6 5 5	198 135 13
80 0 2	62.5 135 10 8 7	204 24 12
70 0 4	60 339 6 21 5	209 149 11
65 0 7	57.5 140 28 8 12	214 31 15
60 0 6	55 251 28 15 11	219 86 17
55 0 9	52.5 450 19 28 2	225 64 14
50 1 3	50 196 35 12 4	229 244 21
45 5 14	48 249 49 18 11	236 195 25
40 37 13	46 455 223 28 7	242 79 30
35 268 5	44 569 184 41 10	249 183 40
30 141 5	42 - 58 1 16	256 110 80
28 327 128	40 - 80 7 15	263 184 100
26 588 244	38 105 15 14	241 202 95
24 1585	36 152 12 13	245 176 80
22 142 11	34 222 15 12	249 100 75
20 22 91	32 89 34 25	252 35 60
18 257 21	30 150 10 24	256 219 55
16 116 207	28 240 12 23	260 185 50
14 157 225	26 84 22 22	264 59 45
100 1 16 14	24 111 122 21	268 115 40
95 1 6 0 1	22 376 11 6	272 240 35
90 1 12 0 1	20 257 21	278 4 30
85 2 4 0 2	18 116 207	282 129 25
80 6 4 0 6	16 157 225	288 108 20
75 16 1 1 0	14 148 265	293 248 18
	1-8	298 108 16
	30	304 2 12
	27.5	310 22 11
	25	

found cascade scaler  
skip print  
Repeat @ 1-2:25

rough scaler  
reset



SENSITIVITY CURVE  
 LA # 17 P00023  
 (G/S - 2/15)  
 SEP 1966 JTB/SL

5/19/55 Summary of Data taken 5/18																	
DISCRIMINATOR LEVEL	COUNTS/MIN.	P.E.	DISCRIMINATOR LEVEL	COUNTS/MIN.	P.E.	DISCRIMINATOR LEVEL	COUNTS/MIN.	P.E.	DISCRIMINATOR LEVEL	COUNTS/MIN.	P.E.	DISCRIMINATOR LEVEL	COUNTS/MIN.	P.E.	DISCRIMINATOR LEVEL	COUNTS/MIN.	P.E.
"STANDARD"	(a) Bec d Cered	of $\frac{3}{4}$ " Al $\frac{1}{2}$ " Pb	.0168 <sup>5+</sup>	15,365		.0473 <sup>+</sup>	10,487		.124	1,713		.0092	211		.049 <sup>+</sup>	21	
.0062	(a) 16,344		.0176 <sup>+</sup>	15,271		.0517 <sup>+</sup>	9,639		.134 <sup>+</sup>	1,081		.0104	203		.0218	38	
	(b) 2,716		.0189	14,990		.056	9,041		.145	724		.0095	188		.0125	40	
	2,810		.020 <sup>+</sup>	15,168		.061 <sup>+</sup>	8,133		.154 <sup>+</sup>	406		.0118	152		.00433	49X	
	(c) 16,584		.0214	15,204		.066	7,484		.166 <sup>+</sup>	198		.0142	141		.076	12	
	16,596		.00103	49,489		.071 <sup>+</sup>	6,710		.175 <sup>+</sup>	88		.0165 <sup>+</sup>	147		.056	15	
	(d) 180		.00319	17,665		.076	6,218		.185 <sup>+</sup>	32		.0189	110		.0383 <sup>+</sup>	23	
	155		.0055	16,313		.081 <sup>+</sup>	6,467		.195 <sup>+</sup>	11		.0214	125		.0297	21	
STANDARD - PDL SUSSEP			.0077 <sup>+</sup>	16,402		.0865	6,477		.205 <sup>+</sup>	12		.0193	108		.0200	33	
.00078	48,221	± 148.	.0099	15,938		.00433	23,538		.225 <sup>+</sup>	3		.0245	81		.0107	29	
.00133	26,275	± 109	.0121 <sup>+</sup>	15,630		.0125	15,730		.245	6		.029	69		.0062	37	
.00184 <sup>+</sup>	20,102	± 75.6	.0148	15,557		.0218	14,634		CS 137			.0342	59		.00183	3195X	
.0025	17,781	± 89.8	.0168 <sup>+</sup>	15,261		.0304 <sup>+</sup>	13,405					.0387	59		.00269	295	
.00368	16,888	± 87.6	.0193	14,699		.0408	11,910		.186	39		.0439	48		.00368	71	
.00443	16,494	± 86.6	.0217 <sup>+</sup>	14,435		.049 <sup>+</sup>	10,349		.173 <sup>+</sup>	54		.0383 <sup>+</sup>	58		.0025	471	
.0060	16,213	± 85.8	.0245	14,370		.058 <sup>+</sup>	8,744		.1615	50		.0473	28		.00133	4052	
.0070	16,362	± 86.2	.0266 <sup>+</sup>	13,859		.067 <sup>+</sup>	7,138		.151 <sup>+</sup>	46		.056	20		.00078	12,448	
.0081	16,081	± 85.5	.029	13,547		.0773 <sup>+</sup>	5,858		.141	352		.066	27				
.0092	16,041	± 85.4	.0315 <sup>+</sup>	13,395		.0868 <sup>+</sup>	4,653		.129 <sup>+</sup>	697		.076	20				
.0107	16,298	± 86	.0342	13,091		.0975 <sup>+</sup>	3,433		.1185	1,883		.0865	13				
.00062	68,124		.0362 <sup>+</sup>	12,637		.107 <sup>+</sup>	2,538		.107	4,561		.186	10				
.00156	24,143		.0387	12,363		.1185	1,793		.0975 <sup>+</sup>	12,529		.245	7				
.00269	17,591		.041 <sup>+</sup>	12,514		.129 <sup>+</sup>	1,155		.0868 <sup>+</sup>	37,331		BACKGROUND					
.00385 <sup>+</sup>	16,805		.0439	12,336		.141	8750		.0773 <sup>+</sup>	101,703							
.0051	16,699		.00183	33,855		.151 <sup>+</sup>	497		Na 24			.245	8				
.0061 <sup>+</sup>	16,347		.0062	16,423		.1615	408					.166 <sup>+</sup>	7				
.0072	16,333		.0107	15,701		.0067	18,314		.00078	11,230		.124	10				
.0082 <sup>+</sup>	16,254		.0148 <sup>+</sup>	15,355		.0255	14,354		.00133	3,963		.083	12				
.0095	16,026		.0200	14,927		.043 <sup>+</sup>	11,452		.0025	592		.043	20				
.0105 <sup>+</sup>	15,995		.0245 <sup>+</sup>	14,325		.060 <sup>+</sup>	8,268		.00368	280		.0256	30				
.0118	15,890		.0297	13,344		.083	5,522		.00493	251		.0069	33X				
.0129 <sup>+</sup>	15,810		.0339 <sup>+</sup>	12,636		.092 <sup>+</sup>	4,215		.0060	271		.1615	11				
.0142	15,423		.0383 <sup>+</sup>	12,009		.103	3,271		.0070	224		.1185	12				
.0153 <sup>+</sup>	15,474		.0426 <sup>+</sup>	11,182		.114 <sup>+</sup>	2,361		.0081	224		.0773 <sup>+</sup>	7				



5/20/55		CS 137					
BACKGROUND							
DISCRIMINATOR LEVEL	COUNTS/MIN	DISC. LEVEL	COUNTS/MIN	DISC. LEVEL	COUNTS/MIN	DISC. LEVEL	COUNTS/MIN
0.214	28	0.245	4 <sup>(2)</sup>	0.0655	170,680	0.00355	17,282,972 <del>14,574,400</del>
0.200 <sup>+</sup>	35	0.225	7	0.0617	237,824	0.0077	15,845,888
0.184	34	0.205	6	0.058	329,472	0.0073	16,101,376
0.176 <sup>+</sup>	30	0.185	9	0.0548	433,920	0.0068	16,314,064
0.168	45	0.166	19	0.051	625,664	0.0063	16,524,032
0.153 <sup>+</sup>	34	0.144	44	0.0472	913,152	0.0058	16,833,280
0.142	31	0.123	605	0.0434	1,466,880	0.00545	17,086,720
0.129 <sup>+</sup>	32	0.103	4293	0.0408	2,460,160	0.0048	17,316,608
0.118	27	0.083	36101	0.0364	3,935,232	0.0044	17,593,088
0.106 <sup>+</sup>	40	0.073	92294	0.0325	5,561,856	0.00385	17,888,768
0.095	44	0.065	183552	0.0288	7,805,728	0.00345	18,220,032
0.082 <sup>+</sup>	51	0.057	314368	0.0249	8,590,306	0.0030	18,514,176
0.072	35	0.049	585216	0.0218	9,955,544	0.00257	18,902,016
0.061 <sup>+</sup>	43	0.042	1,541,632	0.0192	10,621,952	0.00215	19,264,768
0.051	42	0.0345	4,718,624	0.0177	11,220,736	0.00175	19,726,848
0.0385 <sup>+</sup>	51	0.0278	7,653,888	0.0160	11,790,336	0.00130	20,321,792
0.0325 <sup>+</sup>	56	0.0215	10,246,752	0.0144	12,478,208	0.00136	21,360,640
0.0269	123	0.186	29 <sup>(4)</sup>	0.0125	13,013,760	0.00113	21,862,656
0.0222 <sup>+</sup>	314	0.174	22	0.0108	13,711,360	0.0009	22,456,064
0.0178 <sup>+</sup>	1166	0.162	28	0.0200	10,779,136	0.00062	23,169,536
0.0137 <sup>+</sup>	3853	0.151	36	0.0175	11,600,896	0.00156	20,926,720
0.009	4727	0.141	100	0.0148	12,362,736	0.0020	20,264,192
0.0062	15963	0.129	257	0.0140	12,666,890	0.0024	14,751,424
		0.118	305	0.0130	13,010,688	0.0029	14,327,144
		0.113	1366	0.0122	13,375,488	0.00385	18,513,944
		0.107	2170	0.0113	13,735,168	0.0051	17,767,168
		0.103	3255	0.0107	14,032,640	0.0072	16,347,824
		0.097	5462	0.0094	14,374,400	0.0163	12,383,712
		0.092	8988	0.0086	14,763,264		
		0.0865	16092	0.0077	15,062,528		
		0.0817	28819	0.00695	15,516,416		
		0.077	50211	0.0062	15,879,936		
		0.073	76593	0.0053	16,365,312		
		0.069	116703	0.0044	16,821,504		

5/25/55

Test Rack - Check out preparations to test of BF<sub>3</sub> Stanford counters & exp. w/ small wire chamber -

AK Scale Y100355 - Model 1010 - Strips in & checks out on 60N.

Wired Veeber-Rod timer for use with scaler & connected.

Checked timer: One minute, 60N, scale 16.

	Count	Error
224 x 16	+13	-3
225 x 16	+2	+2
225 x 16	+0	0
225 x 16	+0	0
225 x 16	+0	0

4 min, 60N, scale 64

225 x 64	+0	0
225 x 64	+0	0
224 x 64	+58	-6
224 x 64	+60	-4
225 x 64	+1	+1

16 min 60N scale 256

225 x 256	+2	+2
225 x 256	+1	+1
224 x 256	+61	-3
224 x 256	+60	-4
224 x 256	+62	-2

E.R. Rohrer

5/26/55


Worked on DC amplifier, G5 type - System had been on ~~bench~~ bench for long time. for detail, refer to file folder.

Conference with Harman regarding points for air-conditioning system S-1 & S-3 in building.

5/27/55

Servo-control rod-drive.

selsyn-lead screw coupling loose, introducing error in readings.



Reassembled - trimmed al. clamps supporting Selsyn to allow better clearance.

DC-232 - refer file folder -



E.R. Rohrer

6-13-55

New Scintillation Counting System -

Two tube assemblies received 6-8-55.

Fabricated connecting cables for high voltage and signal.




Used one signal lead to tube #1

Turned on HV. measured Pot = .096 x 850V

Connected to tube .0925 x 825V -

Using Unit II.

Test arrangement:



(2 1/2" from tube)

6/13/55

Dried scan at 825V - wiring as drawn - Co137 source.

L.A. gain 32-1 to start - not satisfied -

increased HV to 900

Checked w/ Crudele on operating point ORNL test -  
x 1100 volts -

Tried 1100 volts - still not happy with pattern on scope and counting indications -

Changed Analyzers - seemed somewhat better, but not good - to further investigate further -

6/14/55

Adjust C, (AIC 1010, Y100371) to accept positive pulses -  
sensitivity 0.26 volts. [No negative acceptance]

Turn C, input to Analyzer output, wiring as drawn in red.

Pulse generator set for positive, I see pulses fed to test point on linear amplifier. This to test analyzer amplifier

(L.A. Gain 32-1, D.I.)	ANALYZER		ANALYZER		ANALYZER	
	E TEST	E SCOPE	E TEST	E SCOPE	E TEST	E SCOPE
DOES NOT MATTER	0.1	10.75	113	- (D)	0.4V	466
WHAT GAIN SETTING		(1.45cm)	100.3	(D)	460	100
IS - TEST PULSE INTRODUCED	0.2	(1st x 4)	228-241	200 (D)	449	300
AFTER INPUT STAGES IN WHICH GAIN ADJ. LOCATED	0.3	(1st x 8)	343-355	200 (D)	428	600
	0.4	(1st x 8)	458-470	200 (D)	415	800
	0.5	(1st x 8)	612-624	200 (D)	401.5	1000
	0.6	(1st x 8)	754-761	200 (D)	507.8	1000 (D)
	0.7	(1st x 8)	865-878	200 (D)	483.5	200
	0.75	(1st x 8)	960-973	200 (D)	472.5	500
					460.5	800
					454.5	900
					446.5	1000

4/14/55

Adjusted HV. Supply 102 on UNIT II

ESVM	POT. VOLTS	
800	.0896	After checking voltage-potentiometer settings - Set HV to $\approx 800$ v. (UNIT II).
400	.1106	Plugged in HV cable to TUBE I voltage dropped to 780 (.0870) -
1000	.1122	
1100	.1225	Reset to 800v (.08915) -
1150	.1282	Oscilloscope on L.A.II output, CRM & Scatter Output-SCDA II.

UNIT I

ESVM	POT. VOLTS	
800	.0893	Cs 137 placed in tray, Tube #1 against guide -
400	.1016	upward shift of sample pattern axis (gran gran) -
1000	.1120	high level source - @ 800 int. drops at $E = 130$ -
1100	.1220	not enough seen to see peak - my guess -
1150	.1281	check output pulse

jump for T on input previous to scope set

6/15/55

Made up rest of signal cables & HV. cables as well as HV junction boxes for system  
 Detectors tubes #1 & #2 removed from pig to put on leads - then replaced in lead house -

→ Note: One tube in use HV variable up to  $> 1100$  v.  
 Two tubes in use on single supply only go up to 900 v.  
 This may be cause for concern -

Tube #1 - on top; #2 on bottom -  
 Checked signal from each detector -

Signal check

6/15/55 That gain #1 seems weaker by far than than gain #2 - This may be a problem.

4/16/55 Measured Signal Resistance T-1 1 meg.  
 (Vomax) T-2 1 meg.  
 H.V. Divider R: T-1 1.5 meg-  
 T-2 1.6 meg

I HV, no load = 1100 volts  
 " #1 B. = 1100  
 T-1 only = 1060 volts  
 T-2 only = 1062 volts  
 T-1 & T-2 = 925 volts. max. setting

I HV, no load T.B. = 1100  
 T-1 only = 1065  
 T-1 + T-2 = 940 max setting

One tube load  $< 1$  ma, both tubes load  $< 2$  ma  
 2X2A should be able to push this - what is load on regulator?  
 Check 2X2A plate current -

Tube	Current	HV	No Load Ip	6B96G	Div.
No load	1.1 ma	1373			
T-1	1.25 ma	1230			
T-1 & 2	1.6 ma	940			

↑  
 this is over  
 R output Power Supply -  
 T-1 -  $> 3$  meg  
 both - 1.2 meg  
 T-2 - 1.7 meg

Power Supply

6/16 I - Noticed 5651's seemed to glow (at least there was no reflection from tube shields - when full load removed shield to check further -  
 No load < 600 - 1370 - at 60 voltage more current than 5651's.

No load - 800  
 T-1 - 765  
 T-1A2 - 770 -

Raised voltage w/ coarse adj until 5651's just glow - Still glow at full setting of resistor HV = 9400 v  
 Lowered setting until change in voltage noted -  
 No load point here = 10000 v  
 as is can get 800-900 volts

Source Check

T-1 only; HV = 900 =

Observed patterns w/ Scope 511AD (Y1003Bc) & pre-amp -

Resist T-1 = 2.4k

Scope set full gain & pre-amp full gain (att x100)  
 (gain att x2 Scope)  
 Blky pulses - varying in height most > 1cm - up to 3 cm.  
 Co<sup>60</sup> (15 mil) gain x1 (full) - 1/2 - 1.8 cm -  
 Cs137 - -

T-1 attached to pre-amp I - 2.4k Ro -

Best pattern gain 16 on L.A.

Hard to see Cs137 above blky. -

T-2: OK

T-2 to pre-amp - I and system I.

6/16/55

@ HV 900v -

able to see Cs137 & Co<sup>60</sup> well, above background.

Just scan -

Well L.A. gain at 32-1 -

Got peaks for Cs137 (sample) on door

at 50-60 # 210

Co<sup>60</sup> 380 - 420 as well as Co

lower. ~~210~~

Ran a few scans with T-2 & system 1 to get feel of circuit.

Need to check T-1 again.

Another rough check T-2

gain 32 Set HV @ 800 v. Two top peaks: 163 - 142

gain 64 HV 800 340 - 290

T-1 responding at 10000 v @ gain 64 peak 120 for Cs137  
 Tubes have different sensitivities.

E. R. Rohrer

Scintillation Counter System - 215

Co<sup>60</sup> 1.17  
1.33  
Co<sup>137</sup> 0.162

6/17/55

Further preliminary check out

Begin HV=900  
for both T1 & T2

(where not counting)  
from 5 to 1000

Source Co<sup>60</sup>  
(ismrk)

(or noted 895)

(90L)

T-2 sig  
> 4x that of T-1  
seen on scope.

100 screen  
T-1  
t=12

T-1 & T-2 instruments wired as follows:


LA	E/ΔE	2-in COUNT	LA	E/ΔE	1-in COUNT	LA	E/ΔE	COUNT (1M)	LA	E/ΔE	COUNT
100	3X16	32	25	102	94	270	198	96	NEW SIG	CABAG	T-1
200	71	32	25	102	94	270	198	96	BETTER	-	??
55	17	3	27.5	102	152	260	248	68	T-1 in Unit II		
90	52	8	30	116	111	250	344	28	T-2 in Unit I		
85	89	10	32.5	125	224	240	293	209	slight mess for T-1	to do	
80	121	3	35	140	217	230	179	182	T-2 240-260		
75	144	15	37.5	140	71	220	148	188	There at 900 v &		
70	166	2	40	121	57	210	153	78	LA=32		
65	201	10	42.5	86	75	200	182	105	T-1 @ 1000 v	at 60-20	
60	233	11	45	54	64	190	207	86	T-1 1050 gain	64 unit	
55	296	5	47.5	49	57	180	209	167	T-2 900 gain	32 unit	
50	224	5	50	36	39	170	201	144	E COUNT	E COUNT	E COUNT
45	255	10	52.5	15	44	160	155	135	AE 1m	AE 1m	AE 1m
40	231	10	55	10	21	150	194	25	300 μ	111	170
35	229	10	55	10	21	140	194	25	2	111	170
30	191	10	55	10	21	130	186	41	2	173	160
25	185	10	55	10	21	120	186	41	5	27	150
20	183	10	55	10	21	110	142	182	20	20	180
15	171	10	55	10	21	100	247	270	85	140	85
10	1284	10	55	10	21	90	183	184	69	130	170
5	104	10	55	10	21	80	184	184	260	130	160
	476	10	55	10	21	70	184	184	250	113	120
	11	10	55	10	21	60	186	285	102	114	120
	27	10	55	10	21	50	186	285	102	110	85
	1070	10	55	10	21	40	159	236	108	100	85
	167	10	55	10	21	30	146	236	108	85	30
	2295	10	55	10	21	20	146	236	108	32	237
	264	10	55	10	21	10	203	210	118	250	120
	5/100	426	125	32	7.5	133	197	210	225	240	110
	10	115	107	300	172	22.7	224	200	95	315	100
	12.5	108	60	240	292	45	255	140	73	219	220
	15	103	167	280	259	45	110	190	73	219	220
	22.5	103	161	280	259	45	331	172	75	225	240
	20	103	232	280	259	45	172	180	75	225	240
	22.5	163	119	280	259	45	172	180	75	225	240

6/19/55

Co<sup>137</sup> T-1 unit II 110-115  
T-2 unit I 100-110

6/20/55

Praxis Co<sup>60</sup> = 178-203 air conditioned  
Temp 88° - Reading 2 0820  
Air Conditioning still out  
T-2 → Unit I. Co<sup>60</sup> ΔE=200, LA=32 HV=900  
P.G. NICK 1022 E (ΔE=200)  
0.1 - 128-141  
0.2 - 256-264  
0.4 - 588-517  
T=32°C 0935 T. Controller 88.5°F



Air Cond. on 810° mu -  
10:10 - NA 0.4 E=501 - 513.5

3:15 PM Checked Analyzer calib & scanned Co<sup>60</sup> - T-2 Unit I - T=75°F

Analyzer	check	Input	0.1	133.5	Input	0.4	505
E	COUNT	LA	E	COUNT	E	COUNT	LA
400	14.4	320	136.9	240	275	8	160
390	28.0	310	141.8	230	285	10	150
380	46.8	300	150.13	220	38	13	140
370	114.8	290	162.13	210	310	59	130
360	40.13	280	175.13	200	283	135	120
350	76.6	270	187.4	190	359	120	110
340	110.6	260	201.13	180	334	120	100
330	130.3	250	216.0	170	479	120	90

140 off - some  
from on  
reading  
of 3  
Punch

2/20/54  
Summary - Scintillation Counter System - 215

(6-14-55)  
Analyzer Calibration - (AE @ 200)

ANALYZER CALIBRATION - 2/14/56

TEST	ANALYZER		ANALYZER		Test Pulse (0.4V)	WINDOW		WIDTH		
	E	T	E	T		E-I	E-T	E interval	T	
0.1	112.5-125	100.3-113	128-141	121.5-137.5	RAE	E-I	E-T			
0.2	244.5-257	228-241	256-269	-	100	502-507.8	460-466	5.8	6	
0.3	373-386	343-355.5	<del>352-365</del>	-	200	496-507.8	454-460	11.0	12	
0.4	494.5-507	458-470	505-517	485-507	300	489.6-507.8	447-460	18.2	19	
0.5	646-658	612-624	↑↑↑↑		400	483.5-507.8	441-466	24.3	25	
0.6	831-844	754-767	6-20-53	6-20-53	500	478-507.8	455-466	29.8	31	
0.7	976.5-989.5	865-878	Room HOT 90°F	Room NORMAL	600	472.5-507.8	428-466	35.3	38	
0.75	-	960-973	-	75°F	700	466-507.8	421-466	41.8	45	
POWER SUPPLY CALIBRATION					800	460-507.8	415-466	47.8	51	
SSVM	800	900	1000	1100	1150	900	454.5-507.8	408-466	53.3	58
POTENTIOMETER I	0893	1015	1120	1220	1281	1000	446.5-507.8	401.5-466	61.3	64.5
POTENTIOMETER II	0896	1006	1122	1225	1282					

6-17-55 T-1 + UNIT I ; T-2 + UNIT II

HV ~ 900V ; LA GAIN 32 (BOTH UNITS) AE = 200

E	COUNTING RATE		E	COUNTING RATE	Run one (at left) indicated drastic sensitivity difference between T-1 & T-2
	I	II			
100	245	352	3189		Fed T-1 into II & T-2 into I = T-2 peak lower (240-210) -
95	137.5	340	3406		
90	420.0	330	3835		
85	717.0	320	5023		
80	969.5	310	13601		Set up for Run two
75	1159.5	300	44259		T-1 HV(II) 1050 Gain (II) 64 unit II
70	1329.0	290	66533 74797		T-2 HV(I) 400 Gain I 32 unit I
65	1613.0	270	50784 76356		Rm. from High
60	1869.5	250	100,892 75217		
55	2370.5	230	46,006 38,076	300	623
50	7173.5	220	39,246 46,697	290	685
45	32755.5	210	53,078 53,671	280	1307
40	85490.5	200	51,600 44,945	270	5205
35	183260.5	190	49,856 47,657	260	17894
30	234709.5	180	46,600 47,032	250	29044
25	164404.0	170	47,287 47,241	240	26289
20	124933.5	160	48,422 50,250	230	27885
15	123,277.5	150	52,165 57,358	220	39286
10	137043.5	140	65,390 84,908	210	36065
5	293,942.0	130		200	24571
		120		190	18907
		110		180	19225
		100		170	22248
		90		160	23361
		80		150	22784
		70		140	21823
		60		130	21704
		50		120	21555
		40		110	21804
		30		100	21790
		20		90	
		10		80	
		0		70	

6/21/55 Calculation on Run 6/20 3:15 PM

E	Counting Rate	E	Counting Rate
400	228	190	92143
390	768	180	85716
380	1544	170	122796
370	1832	160	95327
360	1453	150	57534
350	1222	140	57630
340	1766	130	64583
330	2043	120	72762
320	2185	110	68633
310	2264	100	67132
300	2413	90	65359
290	2605	80	65688
280	2813	70	66228
270	2946	60	68758
260	3229	50	72866
250	3456	40	82511
240	3928	30	114594
230	4570	20	161844
220	5645		
210	14844		
200	72583		



6-24-56 0830 HV on

m-2 unit I, HV 900, LA 32, SE=200

Co<sup>60</sup> - cf 4/17/62 source in 0830

TEST PULSE 0958 (P.gen on overnight) 0.4 = 495-507.5

Adj HV to 900 up very slightly

Begin Count 0901

source out - 0958

Run 2 G<sup>137</sup> sample m/p p.p. begin 0953

Count	E	Count	E	Counting Rate	Counting Rate
400	453	140	237	723	60764
396	456	130	276	1686	70696
380	408	9	120	211	69427
370	82	2	110	261	66934
366	75	15	100	252	64747
350	164	0	90	249	63444
340	123	13	80	252	61685
330	133	8	70	254	65180
320	141	10	60	268	68740
310	157	15	50	281	76172
300	160	8	40	287	7568
290	175	9	140	1	2909
280	184	4	130	2	2948
270	204	15	120	8	3279
260	221	7	110	42	3545
250	250	15	100	354	4015
240	69	62	90	503	4478
230	83	20	80	142	5332
220	289	12	70	113	18188
210	235	77	60	165	60237
200	361	34	50	209	92450
190	311	74	40	242	79690
180	442	43	30	245	113195
170	438	31	20	310	112159
160	310	80	10	314	79440
150	209	25			38425

2:46 PM G<sup>137</sup> T-1 HV 2 II → I

E	count	2.46 P.M. Bkg
10	181	147
20	142	126
30	111	102
40	87	84
50	76	73
60	66	61

Left HV on I & II over night I @ 1050 II @ 900

6/22/65 Checked HV. HV I 1052 volts; .1166 mv.

0838- (G<sup>137</sup>) source in pig dropped down to bottom, by mistake

OT-1 HV 81050 Gain 64(I) SE=200

OT-1 HV 81050 Gain 64(I) SE=200

OT-2 HV 900 Gain 32(I) SE=100

OT-2 HV 900 Gain 32(I) SE=50

NOTES: (1) proceeded after... (2) Run 3 questionable cascaded

Run 4 HV 1050 Gain of 50

OT-1 1050 Gain of HV 900 Gain 32

OT-1 1050 Gain of HV 900 Gain 32

SCALER	COUNT	E	SCALER	COUNT	E	SCALER	COUNT	E	SCALER	COUNT
150	7-920	2012	85	1296	105	52	13390	130	261	66858
140	20-3	5123	80	361	102.5	94	24279	127.5	330	84709
130	135	34623	75	257	100	102	41595	125	374	97117
120	405	103905	70	231	97.5	210	61673	122.5	416	106596
110	408	104648	65	257	95	315	80724	120	420	107551
100	214	54948	60	309	92.5	324	90860	117.5	385	94689
90	162	41480	55	41	90	350	89710	115	344	88258
80	149	51174	50	75	87.5	306	78397	112.5	280	71916
70	248	63579	45	136	85	247	63318	110	228	58592
60	14	3793	40	209	82.5	218	55983	107.5	193	44463
50	89	11097	35	84	80	158	40607	105	140	35443
40	106	27221	30	86	77.5	123	31569	102.5	115	29459
30	29	48157	25	188	75	110	28175	100	103	26460
20	151	62911	20	245	72.5	104	26161	97.5	94	25390
15	232	59403	15	117	70	125	26880	95	47	24730
10	175	44399	10	175	70	125	26880	92.5	100	23631
5	21	24973	5	117	70	125	26880	90	47	24730
0	86	22205	0	86	70	125	26880	87.5	124	31744
	39	20263		39	70	125	26880	85	138	35510
	20	20263		20	70	125	26880	82.5	142	36483
	143	21448		143	70	125	26880	80	139	35402
	35			35	70	125	26880	77.5	124	31744
	86			86	70	125	26880	75	106	21372
	112			112	70	125	26880	72.5	92	23526
	224			224	70	125	26880	70	83	21503
	587			587	70	125	26880	67.5	178	45617
	1520			1520	70	125	26880			

6/23/55 Scin. Ctrs. 215

Time	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count							
0900 Tubes 394 recd																
① T-1 HV I 1166 DE 250 Gain 64 (II) Drift Check Run	135	103	24565	640	161	2588	380	31	2031	120	112	7185	520	256	48	25321
② T-2 HV II 900 Gain 32 Drift Check Run DE 250	132.5	146	37435	620	154	2479	370	32	2085	110	257	16462	315	159	40852	
③ T-2 900, 3V Po. Bl. DE 250	127.5	234	60682	610	168	2694	350	32	2086	40	121	7784	305	308	98925	
④ T-4 HV II 900 G I 32 G 5 137 DE 50	125	285	73068	590	275	4413	370	33	2122	70	194	12477	295	430	110334	
⑤ T-3 HV II 900 G I 32 G 137 DE 100	116	214	34941	560	72	4635	300	33	2174	72.5	15	4030	280	394	101114	
⑥ STD Demu T-3 HV II 900 G I 32 DE 200	100	80	20720	530	56	3641	220	35	2301	65	86	22220	265	232	59533	
700	11	15	191	440	35	7	2257	180	46	2986	42.5	205	52610	608	8	135
640	35	4	564	430	34	39	2215	170	48	3116	40	189	48506	708	4	148
680	64	11	1035	420	33	19	2131	160	58	3251	37.5	195	50158	808	4	150
670	122	3	1455	410	33	56	2168	150	57	3280	35	217	55789	908	7	114
660	150	0	2100	400	33	46	2158	140	57	3296	32.5			1008	6	102
650	177	4	2836	390	30	59	1979	130	57	3268	30			1008	4	1353

Scaler seems to have pulse bursts

6/24/55 More - 215 Scin. Ctrs

Check No: I 0.1166 41000 (1044) II 0.1010 - 900

Time	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count	Scaler Count					
① T-1 HV I 1166 DE 250 Gain 64 (II) Drift Check	135	158	39786	100	155	39863	100	209	53504	135	244	62480
② T-2 HV II 900 G 32 (II) DE 250	130	616	81054	97.5	285	73080	97.5	397	88926	130	362	92784
③ T-2 HV II 900 G 32 (II) DE 250	127.5	392	100436	95	360	92227	95	426	104146	127.5	386	101545
④ T-2 HV II 900 G 32 (II) DE 250	125	442	113333	92.5	449	114453	92.5	439	112453	125	395	101347
⑤ T-2 HV II 900 G 32 (II) DE 250	122.5	474	121407	90	473	121340	90	391	100294	122.5	371	95059
⑥ T-2 HV II 900 G 32 (II) DE 250	120	462	118423	87.5	450	115224	87.5	314	80533	120	323	82234
⑦ T-2 HV II 900 G 32 (II) DE 250	117.5	415	106368	85	383	98178	85	237	60927	117.5	208	68744
⑧ T-2 HV II 900 G 32 (II) DE 250	115	360	92389	82.5	315	80751	82.5	185	47393	115	207	53187
⑨ T-2 HV II 900 G 32 (II) DE 250	112.5	290	74343	80	250	64113	80	147	37660	112.5	165	42266
⑩ T-2 HV II 900 G 32 (II) DE 250	110	227	54255	80	250	64113	80	147	37660	110	133	34128

6/27/55 Note - Power failure over weekend 4:30<sup>P</sup> 6/25 - 1:30<sup>P</sup> 6/26

Drift check 0830 6/27, data recorded above

HV I - 0.1166 41044

HV II - 0.1010 900

Run ① T-2 HV II Gain I 32

② T-1 HV I Gain II 64

6/29/55

HV check I - .1166 ; II .1010

Gain I - 32 ; II 64

Set: T-1 → II, I HV

T-2 → I, II HV

Rough check of Cs<sup>137</sup> peaks T-1 x 120-125 (OK); T-2 x 90-92 (OK)

over for construction

6/29/58

W foil - 3662  
number down  
away from T-1

① T-1 - AE = 50  
begin 0925  
(T-2 E. 445)

Scaler	Count	Scaler	Count
170		I 48 0	768
7.5 897	101802	D 7 4	114
10 154	34570	I 28 6	774
12.5 117	30017	D 6 8	104
15 102	26625	I 47 15	767
17.5 84	24241	D 8 0	128
20 92	23773	I 6 4	100
22.5 100	25847	I 21 5	511
25 117	30044	D 0 14	14
27.5 136	35007	I 20 15	335
30 144	37046	I 24 0	384
32.5 134	34330	D 1 1	17
35 107	27550	I 28 4	468
37.5 75	14311	D 1 9	25
40 43	11096	I 31 14	510
42.5 22	5789	D 1 4	20
45 24	6143	I 38 8	616
47.5 19	1225	D 1 8	24
50 8	566	I 60 11	971
52 13	1325	D 10 0	120
55 4	297	I 257 45	16104
58 13	1325	D 10 0	13540
60 12	201		
62 1	1009		
65 8	135		

Scaler  
Repair

(Noted  
T-2  
E setting)

Integral  
Count

T-2 E set to 1000  
T-2 AE = 40  
T-2 E 0, 450

Bkg

T-2 adj up against slide  
Check all catches "1947" only range  
Noted system detects foil

6/29/58  
Gold foil  
193 (unclear)

T-2 - cloud of gold  
AE 50

8/210 → 10

8/108 245

8/108 8

Scaler	Count	Scaler	Count
168 15	43023	I 12 4	31958
112 132	28804	D 109 24	27428
53 52	13620	I 7 198	1490
7.5 5	3350	D 2 190	190
165 31	10591	I 112 5	1797
7 27	475	D 5 14	94
156 11	9445	I 44 2	1504
6 10	394	D 5 7	87
157 19	4683	I 41 14	1470
6 26	420	D 5 10	90
144 60	4276	I 83 4	1334
6 0	384	D 4 0	64
141 2	4026	I 77 3	1235
6 4	388	D 3 8	56
134 20	8596	I 46 8	1004
5 57	371	D 4 0	64
125 61	8061	I 66 2	1058
5 43	363	D 2 7	39
123 52	7424	I 60 5	965
5 26	346	D 2 10	42
121 22	7744	I 54 14	878
4 20	276	D 2 11	44
115 38	7348	I 47 14	766
3 52	244	D 3 12	60
112 33	7201	I 44 1	705
3 15	207	D 1 4	38
104 5	6981	I 40 8	648
2 26	154	D 1 7	38
105 50	6772		
2 58	166		
102 28	6556		
3 01	143		
49 4	6340		
5 29	349		
42 35	5423		
9 57	6373		
38 54	5686		
15 44	1004		
68 16	4368		
14 50	1266		
51 46	3310		
18 6	1150		
38 41	2473		
13 0	832		
24 37	1573		
7 50	448		
18 24	1176		
3 24	216		
14 15	911		
1 41	105		
13 0	832		
1 6	70		

Bkg

#3662  
Data on U-foil - exposure, taken by Crudele,  
in binder with curves.

6-30/7-17 Vacation

7-18 Check status of instrumentation -

Clean up backlog of documents

7-19-55 Check out servo chassis built by V.G.H. OK.


West end - Magnet trouble. open lead at magnet.  
Mounted 2-terminal barrier strips on plate  
supporting magnet for connection of DC leads to  
magnet leads. This should cut down on magnet  
failures.

SCIN CTRS -215. Spencer has noted tube 1  
extremely sensitive to geometry - In counter  
U-foil, each of two done separately, then stacked.  
Stacked count >  $\Sigma$  separate counts.

HOT CRIT CONTROL PANEL - Wiring in progress J.F.G.

SCINTILLATION COUNTERS -215:

Arrangement of tubes



9/19/55

T-1 - HV from PS I; SIG to PA II

T-2 - NO CABLES

T-3 - CABLES NOT CONNECTED TO ELECTRONIC INST.

T-4 - SIG CABLE CONNECTED TO PA I.

LEAVE T-1 AS IS -

MEASURED HV on PS I POT = 0.11715 volts @ 1048v, @ 1044v.

MEASURED HV on PS II NO TIE IN: 0.1141 = 1021v.

T-3 0.1105 = 988v.

CS 137 AT-2 @ 90v - scaler not right at E near 0 - check  
input at + this is how it had been set - - OK.

Check LA-SCDA calibration:

Input pulse 0.4 volts, +, from sine pulse generator -

E = 510,  $\Delta E = 100$  E = 504-510 Thus done

HV II 0.1010 4900v T-2 peaks in Cs<sup>137</sup> @ 60 (90) -

T-3 peaks Cs<sup>137</sup> 220 (290) -

shift of peak a problem -

HV = 900 - LA calibration check ok.


Possible - movement of tubes; repositioning in lead shields  
may have produced shift.

HV II ESVM = 900 (1010)

HV I ESVM 1064 (1173)

Fixed HV junction box II for parallel operation of tubes 2 & 3.

500,000  $\Omega$  potentiometers in each branch:



7/14/55

with pinch box in place adjusted HV to 900v. for both tubes & adjusted T-3 so peak on Cs<sup>137</sup> same as for T-2 (at 90)

9000

7/20/55

Set HV to get reading of .1065 @ 950v -

Check Cs<sup>137</sup> peak -

	① E	② E	③ E	④ E	⑤ E	⑥ E	⑦ E	⑧ E	⑨ E	⑩ E
① T-2 ΔE=50 beg. 0837	110	55 217	134 105	120	173 171	115	129 115	115	51	66
② T-2 ΔE=50 new source position begin 10413	107.5	116 174	275 83	117.5	288 108	112.5	144 234	110	131	92
③ T-3 ΔE=50 beg. 1200	105	232 16	331 45	115	393 195	110	274 171	105	140	132
④ T-2 ΔE=50 beg. 1221	102.5	353 11	341-93	112.5	455 10	107.5	320 116	100	174	87
⑤ T-2 ΔE=100 same as top of shield	100	494 155	308 35	110	444 188	105	320 157	95	116	80
⑥ T-2 ΔE=100 same as above	97.5	575-47	248 47	107.5	444 223	102.5	293 216	90	95	214
	95	446 81	174 72	105	460 63	100	224 120	120	34	63
	92.5	433 88	131 71	102.5	410 134	97.5	115	44	108	
	90	321 75	107 27	100	312 40	95	110	40	131	
	87.5	234 224	41 53	97.5	303 249	92.5	105	30	94	
	85	164 204	84 127	95	252 43	90	100			
	82.5	135 149	80 182	92.5	205 44	87.5	95			
	80	116 108	83 47	90	174 103	85	90			
	77.5	118 8								
	75									
	72.5									
	70									

can not tell until with same HV for  
from one of tubes

appears that  
T-3 needs to  
be adjusted  
down a bit in HV.

7/20/55

ΔE=50 T-1  
HV 1171 Cs<sup>137</sup>

U-Foils

Cs<sup>137</sup> at this

Foil #	# 11 (up)		# 3605 (DOWN)	
	INTEGRAL	DIFF	INTEGRAL	DIFF
① Foil 11 (up): T-1 (1171) ΔE=50 (NO Al. over foil)	7.5	80,701		102,589
	10	59,989		95,724
② Foil 3605 (DOWN) resting on paper slip to bottom of Al. slide.	12.5	58,148		95,278
	15	55,613		90,854
		51,244		85,977
	20	49,999		88,283
		54,095		98,437
	25	41,708		114,141
		72,107		131,752
	30	77,177		136,268
		71,044		121,550
	35	122,674	56,312	90,553
		73,982	34,969	134,190
	40	46,669	23,337	80,502
		26,857	11,446	47,322
	45	16,884	5,459	33,195
		12,405	2,918	25,568
	50	9,787	1,868	20,883
		8,275	1,508	15,338
	55	6,908	1,343	11,195
		6,364	1,193	
	60	5,270	930	11,195
				2,147
	65	3,768	652	7,784
	70	2,698	569	5,386
	75	1,954	480	1,293
	80	1,193	316	2112
	90	625	45	881
	100	482	28	698

Scaler  
43 10  
2 8

7/19/55

with junction box in place & adjusted T-3 as peak

7/20/55

Set HVII to post reading

Check Cs<sup>137</sup> peak

⊖	E	1 min ct	⊕	E	1 min ct
⊖ T-2 AE 50 beg: 10413	110	55 217	⊕	120	173
⊖ T-2 AE 50 new source position beg: 10913	107.5	116 179	⊕	117.5	288
⊖ T-3 AE 50 beg: 1200	105	232 16 331 45	⊕	115	383
⊖ T-2 AE 50 beg: 1221	102.5	353 11 341-83	⊕	112.5	455
⊖ T-2 AE 100 same on top of shield	100	474 155 308 35	⊕	110	447
⊖ T-3 AE 100 same same	97.5	535-472-48 47	⊕	107.5	447
	95	446 81 179 92	⊕	105	460
	92.5	433 88 131 71	⊕	102.5	410
	90	421 75 107 27	⊕	100	372
	87.5	234 224 41 53	⊕	97.5	303
	85	164 204 84 127	⊕	95	252
	82.5	135 144 50 182	⊕	92.5	205
	80	116 108 83 47	⊕	90	179
	77.5	113 8			
	75				
	72.5				
	70				

T-1 (now w/T4 in lead shield)

HVI: .1194

This a bit high - To check Cs<sup>137</sup> at this

⊖	E	1 min ct	⊕	E	1 min ct
140	101 107		7.5	315 69	
137.5	142 38		10	234 85	
135	209 6		12.5	227 94	
132.5	281 206		15	217 61	
130	344 52		17.5	200 48	
127.5	391 19		20	195 74	
125	407 12		22.5	211 78	
122.5	397 22		25	241 12	
120	368 15		27.5	281 17	
117.5	313 17		30	361 12	
115			32.5	277 132	
112.5			35	214 248	
110			37.5	144 105	
			40	91 41	
			42.5	44 182	
			44.5	22 83	
			47.5	48 38	
			50	24 12	
			52.5	23 36	
			55	20 63	
			57.5	18 41	
			60	15 28	
			62.5	12 11	
			65	10 11	

1313 decided to run U-fails without changing HV - Cs<sup>137</sup> peak close to early.

then at .1166

⊖	E	1 min ct	⊕	E	1 min ct
100	100 10		7.5	200 149	
97.5	122 2		10	373 236	
95	125 0		12.5	372 44	
92.5	15 0		15	354 230	
90	17.5 0		17.5	335 217	
87.5	20 0		20	344 249	
85	22.5 0		22.5	384 131	
82.5	25 0		25	445 221	
80	27.5 0		27.5	514 168	
77.5	30 0		30	530 76	
75	32.5 0		32.5	474 266	
72.5	35 0		35	563 185	
70	37.5 0		37.5	524 36	
	40 0		40	143 178	
	42.5 0		42.5	184 218	
	45 0		45	129 171	
	47.5 0		47.5	36 88	
	50 0		50	94 224	
	52.5 0		52.5	26 19	
	55 0		55	57 26	
	57.5 0		57.5	45 42	
	60 0		60	114 59	
	62.5 0		62.5	33 35	
	65 0		65	121 40	
	67.5 0		67.5	24 22	
	70 0		70	84 10	
	72.5 0		72.5	20 15	
	75 0		75	136 9	
	77.5 0		77.5	34 9	
	80 0		80	5 1	
	82.5 0		82.5	6 15	

1/21/55

Small BF<sub>3</sub> proportional counter -

Check out 5/16" counter -

Scaler 1040 check out: No. Y-89637

CR. scale 16 - internal con.

strip 115-2




tried 64-250 w/ 115-2 no hto voltage - power plug bad, needs to be replaced.

Decided to check counter w/ Pre-amp & linear amplifier using scope to check output.

Pre-Amp - AIC Ser# 335

L.A. AIC Ser# 805

HV - NICE 1090 -



	HV	LA	CRM	Scaler
6kg	1700	PA 1, 0.2, 20	X1 1100%	672/5
PB 267	1600	"	X1 - 0	14/5
Source next to counter inside paraffin	1300	"	50%	304/5
	1400	"	75	484/5
	1500	"	75	626/5
	1600	"		663/5
	1700	"	power failure	blank

BF<sub>3</sub> proportional counter.

Placed ctr in new paraffin tube #1" 55  
#1" thick w/ 3/4" center hole

	HV	Scaler 2 min.	Scaler 5 min.	Scaler #1	Scaler #2	Scaler #3	Scaler #4	Scaler #5
7/22/55								
LA: 4-1, 0.2, 1"	1300	0	548					
D 20	1350	1	599					
BACKGROUND	1400	1	677					
PB 267 31" out paraffin	1450	0	688					
	1500	2	700					
	1550	0	743					
	1600	0	803					
	1650	1	795					
	1700	2	856					
	1750	4	949					
	1800	23	1409					
	1850	453	879					

SCIN. CTR: Gold foils -

	T-1	T-2,3	T-1	T-2,3	Two min counts with T-1 & T-2,3	Two min count on T-1 (Both T-1 & T-2 displaced from tray by 5 turns)
215						
T-1 Integral at E=70 HV 81050	T-1 Au2 64	T-2,3 Au2 64	T-1 Bkg 64	T-2,3 Bkg 64	Two min counts with T-1 & T-2,3	Two min count on T-1 (Both T-1 & T-2 displaced from tray by 5 turns)
T-2,3 Integral at E=50 HV 2900	50 22	124 40	15 25	54 18	1611	4116
					4975	1737
					1113.5	2379
	N/A	CRM				ratio 5:1
T-1 Integral K=100	N/2 248-7	Bkg 48-1	N/2 3975/m	Bkg 7647/m		
			1987.5%	3845%		


7/25/55

T-1 INT BKG HV (1171) (open slide)	2-min. #16	E	4 $\mu$
92-2	20		737.0
84-12	30		678.0
74-4	40		594.0
64-5	50		514.5
52-3	70		417.5
48-7	90		387.5
counts thru $\times$	100		349.0
36-10	125		
31-8	150		
28-6	175		
25-8	200		

7/26/28

Instrument Servicing - West End  
Counter Channels & DC-1 primary  
Reference: Instrument files.

7/26 Instrument placement:




Ion Chambers DC-2, DC-3, R-1, & Log N OK. DC-1 sick  
P-M OK.

Counters: C-1 lead shot.

Got C-2 head on C-1 amp & C<sub>3</sub> & C<sub>5</sub> to work.

7/27 Reposition counters & chambers for geometry improvement



7/28 For west run C<sub>2</sub>, C<sub>3</sub>, C<sub>4</sub>, C<sub>5</sub>; DC<sub>1</sub>, DC<sub>3</sub>, R-1, Log N; PM

(7-29 sick)




8-1 Continued service work on C-1 C-2 channels in 202.  
 Refer to files.

Design consideration APPR Control

8-2, 8-3 Paper work on APPR. Well control console.  
 Set up for instrument check out in 107  
 Noted need for servicing of I & II Scm Cts in 215.

8-4-55 8-11 Checked out counters for DFC. dry stacking.  
 Instrument layout



Checked counters with "Bky" block in plate of n-source  
 from low to high series of one minute counts

	C <sub>1</sub> x64	C <sub>2</sub> x64	C <sub>3</sub> x64	C <sub>4</sub> x64	C <sub>5</sub> x64
Disc:	500	-	425	1645	
MV	4-0	760	1800	1675	
	50 33	34 47	81 0	16 58	
	49 4	33 57	82 3	18 50	
	50 4	33 28	82 11	16 62	
			80 43	15 28	
	20 24	73 60			
	20 46	77 40	5 43	57 44	
	20 32	75 62	5 26	50 53	
			5 36	49 15	
	20 47	32 9	5 36	49 28	
	20 12	33 27	5 25	48 24	
	21 7	32 1	5 10	48 54	
	21 8	33 6	5 46	48 38	
	22 39	34 15	5 40	45 7	
	22 34	34 14	5 11	46 7	
			5 38	48 62	



8-8/19 - Hot Critical Preparation - Control Panel Check Out & Instrument servicing. Refer to file.

8/14/55 - APPR console design.

8/20-8/2 - Hot Critical Run (Evening Shift) Wiring APPR panels.

8/6/55 - Scin. Anal. Chrs 215 - refer file -

9/1/55 - From results of 9-6:  
 Gain of SA2 is 2x gain SA1.  
 T2-3 had been used with SA-2  
 In earlier counting had used T-1 with SA-2 to improve operation -  
 T-1 in SA-2

Co<sup>60</sup> 1.171; 1.332

SAE=100 HV#1050 @ 2914 (3786) T1 & T4

E	Scaler	cpm	E	Scaler	cpm	E	Scaler	cpm
100	109 51	7027	80	137 49	8445	160	80 19	5134
5	113 3	7235	85	167 13	10701	65	72 41	4649
10	114 53	7349	90	171 21	10965	70	61 47	4335
15	116 15	7439	95	161 11	10315	75	73 5	4677
20	118 0	7552	200	139 50	8946	80	88 61	5093
25	119 5	7621	5	126 49	8113	85	110 23	7063
30	121 56	7806	10	134 50	8626	90	147 4	9414
35	114 43	7339	15	138 17	8849	95	173 39	11111
40	114 34	7370	20	126 51	8115	200	170 44	10924
45	110 12	7052	25	102 33	6561	5	144 14	9230
50	101 16	6480	30	77 4	4432	10	114 33	7329
55	96 50	6144	35	57 61	3704	15	111 5	7109
60	94 39	6055	40	49 49	3884	20	125 17	8019
65	92 4	5892	45	44 46	3862	25	132 20	8468
70	98 62	6334	250	45 31	7411	30	115 4	7364
75	115 38	7398				235	83 46	5258

9/7/55 SA 2 T1-4

@ 1050V - Gain 64 - pulse T-1: 21cm, 1.7cm (scope 2-x2)

Adjusted R44 to bring pulse on T-4 down to that of T-1

Ren Scen on T-4 alone.

E	Scaler	cpm	E	Scaler	cpm	E	Scaler	cpm	Scaler	cpm
170	77 57	4979	170	73 55	4727	180	72 43	4651	96 32	6176
75	73 10	4682	75	76 17	4881	85	82 36	5284	121 5	7749
80	70 16	4496	80	87 31	5544	90	104 59	6713	154 53	9909
85	73 56	4928	85	103 27	6619	95	141 14	9038	166 62	10686
90	83 7	5319	90	125 17	8017	200	168 10	10762	165 2	10562
95	96 61	6205	95	154 63	9419	05	172 12	11020	138 13	8845
200	118 25	7577	200	169 22	10838	10	147 46	9454	113 22	7254
5	144 33	9244	05	157 54	10102	15	114 21	9317	105 56	6776
10	162 18	10386	10	126 54	8119	20	103 24	6616	118 10	7562
15	154 53	9909	15	107 15	6863	180	73 31	4703	79 7	5063
20	130 14	8334	180	97 42	6258	85	88 13	5645	91 25	5849
25	103 11	6603	85	121 12	7756	90	115 34	7394	113 62	7294
30	47 12	6220	90	152 20	9748	95	115 17	9297	136 49	8753
35	108 34	6446	95	170 42	10922	200	164 30	10846	161 46	10350
40	112 59	7227	200	165 46	10606	05	170 31	10911	164 46	10542
45	107 53	6401	205	135 24	8664	10	138 63	8895	144 61	9277
50	86 49	5681	210	109 22	6498	15	112 12	7180	113 43	7275
			215	108 25	6937	20	106 5	6661	101 27	6491

12:20P - Noted that HV to tubes had increased from 1047 to 1054, due to increased R44 load. (8200) check T-1 & T-4

MP. 46  
P. 53

72-73 SAL  
(5802)  
set 1050V  
Gain 32  
GE=100

(no adj on R33  
from HV 8960)

E	Scaler	4m	E	Scaler	INT	GRAL	INT	
190	91 42	5866	50	77 10	2			
95	105 11	6731	45	114 3	1			
200	114 8	7304	40	163 4	13			
5	160 13	10253	35	293 11	7			
10	225 67	14463	30	193 14	33			
15	251 10	16074	25	107 9	117			
20	249 43	15979	72-73 SAL GE=50					
25	211 44	13548	#49	15 57	9 12			
30	167 27	10715		17.5 65	7 40			
35	146 8	9352		20 90	11 63			
40	152 33	9761		22.5 122	4 31			
45	186 10	11914		25 149	6 5			
50	169 33	12643		27.5 131	15 27			
55	168 22	10224	45	82 12	47			
60	119 4	7620	50	11 5	60			

U bits  
SA-2  
#26 (up) T1-4  
GE=50  
HV .5817

E	Scaler	INT	E	Scaler	INT	GRAL	INT	
2.5	392 9	46	50	19 13	56			
5	183 7	30	45	27 14	59			
7.5	141 14	31	40	36 6	63			
10	136 4	49	35	53 1	9			
12.5	130 11	11	30	107 7	34			
15	130 0	19	25	16 25				
17.5	145 4	24	25	115 9	205			
20	167 5	41	OKS INTEGRAL					
22.5	183 4	57	25	9 3	14			
25	176 9	31	25	7 0	0			
27.5	142 12	18						
30	94 11	18						
32.5	55 15	56						
35	110 3	12						
37.5	57 4	6						
40	36 15	1						
42.5	29 4	2						
45	25 5	15						
47.5	21 8	8						
50	18 0	11						

4/8/55 U-Foil #24

SA-2 Gain 64 (A) INTEGRAL 0833	E	Scaler 1	Scaler 2
77"	50	+167	184+15
"	"	+123	180 5
"	"	+63	196 6
100	+33	93 0	
381,542 c/m	"	+181	93 5
"	"	+93	93 1
100	X64	X64	
"	15	93 6	
#49 up (INT)	100	X16	
970	"	62 1	
"	"	61 4	
"	"	58 10	
BKG (INT)	100	X16	
(743)	"	46 15	
"	"	47 5	
"	"	45 2	



9/12/55	$\Delta E$	$E$	$\odot C_{60}^{60-4}$ SA-2	$\odot C_{60}^{60-1}$ SA-2	$\odot C_{60}^{60-1}$ SA-1	$\odot C_{60}^{60-4}$ SA-1	$\odot C_{60}^{60-4}$ SA-2	$\odot C_{60}^{60-1}$ SA-2	$\Delta E$	$E$	5% off ctr	$\Delta = 1cm$	$\Delta = 1cm$	$\Delta = 2cm$	
	100	170	56899	6990					50	60				543,124	
		75	69656	7394 4370										587,423	
		80	86418	8593			89355	9148		65				623,032	
		85	102236	10815			107868	12014						647,406 SA-2	
		90	(111,195)	14151			117395	15199		70			376,060	649,217	
		95	108506	16293			110081	16938					426,279	635,967	
	200		97474	16783			95997	16700		75			478,437	590,067	
	5		86461	85135		84492	84614	14452				416,149	529,401	549,581	
	10		84546	13019		21687		12221		80	333,583	465,105	552,473	238,225	
	15		86180	11634	11083	77861			SA-2	85	399,329	503,725	553,327	247,110	
	20		84642	12189	13793	82730				85	439,013	515,341	532,617	237,767	
	25		74980	12888	15681	83181					85	465,450	506,012	491,971	<del>248,196</del> 235,822
	30		58481	12172	14880	82000					90	-467,518	471,454	437,097	236,186
	35		(40,029)	9350	12458	78284					95	442,941			230,487
	40				10293	73221								222,173	
	45				9320	69922							169,520	201,387	
	50				9582	65808				100	191,563	221,500	252,394	210,706	
	55				11139				SA-1	105	212,284	237,477	256,511		
	60				11750					105	223,456	248,875	258,818		
	65				10839						110	241,137	253,902	256,849	
	270				7807						115	245,757	260,086	245,758	
											120	-248,025	245,269		
										125	243,350	229,109			
										120	230,517				

9/12/55

5 7/16 05.

SA1 C<sub>8</sub><sup>137</sup>  
DE-50  
(70m 32.1)

E	16 64	16 64	16 64	16 64	16 64	16 64	16 64
70							
75							
80							
85							
90							
95							
100	165 8 40			225 15 29			
105	187 1 11	216 4 60		246 7 42			
110	207 2 60	231 14 37		257 7 63			
115	218 3 32	243 0 43		252 12 1			
120	235 7 44	249 15 14		250 13 17			
125	239 15 52	253 15 54		258 15 62			
130	242 3 25	239 8 14					
135	237 10 22	224 0 33					
140	225 1 53						

refer p. 65, 61

9-15-55 1025 SA-2 160-1 peak -  
Amplifier 64-1  
Analyzer DE-100

H.W. .58208

E	scaler	Ym
INT 170	X256 607 22	155,620
170	28 X64 91	7259
175	126 40	8104
180	152 35	9763
185	192 28	12316
190	250 48	16048
195	281 57	18041
200	264 16	16912
205	223 9	14281
210	187 2	11970
215	186 6	11910
220	207 48	13296
225	207 7	13255
230	176 24	11288
235	121 52	7796
240	73 19	4691
INT 240	X256	
245	149 81	38225
250		

TIMER CHECK  
600 - X16 - 1 Min  
1. 225 + 1  
2. 225 + 1  
3. 225 + 2  
4. 225 + 4  
5. 225 + 4

20-170-235  
172-237  
7 170-234  
177-375  
172-10972

9-16-55

Scintillation Counter Assembly

Base change - shield change - visual check of leakage path & crystal.


T-2 removed from mounting frame -

removed tube & crystal from Al shell -

NOTE (3) one strand of lead from photo cathode to crystal folded back - possible source of corona & leakage to mu-metal shield.

(4) only 1/8" insulation along wire beyond mu-metal shield. VERY SHORT LEAKAGE PATH

removed mu-metal shield - to mount with black tape -



after mounting shield - ran tape around tube - covered crystal flange with one layer of tape - To reduce leakage problem

removed anode lead wires.

T-3:

replaced assembly in shell.

9/18/55

T-3: tube didn't appear to be so snug as was T-2 in shell. change as T-2

Power off 9-17-58 - Equipment on 9-19 HV high & gradually dropped.

9-21-55

SA-2: Gain 68.1; HV = .58205

SA-2 on Solo regulator -

60-1

ΔE	E	X64 Scaler	
100	160	129	13
	65	151	26
	70	184	4
	75	235	15
	80	276	44
	85	281	44
	90	280	30
	95	200	62
	200	195	59
	25	212	42
	10	214	53
	15	129	42
	20	119	2
	25	70	55
	<del>25</del>		



9-21-55

High Voltage Supply SA-1 - (CVBS) not thru Solo. (~~to be used~~)

ESUM Pot Gain left right

1100	.6080		
1110	.6136		
1090	.6020		
1100	.6098		
	.60645	5	
	.6067	4	
	.6069	3	
	.60725	2	
	.6075	1	
	.6078	0	⊕
	.6080	3	1
	.6083	2	2
	.60875	3	3
	.6090	4	4
	.60925	5	5

T-3 load

9-21-55

T-2 & T-3 checks individually with SA-2 channel of


SA-1 HV power supply.

AE	E	H.V.	Scaler	AE	E	H.V.	Scaler	AE	E	H.V.	Scaler
100	290	1100	286 39	100	415	1100	179 184	40	340	1100	165 58
	5		282 58		20		142 214		25		197 17
	300		281 39		25		129 49		45		184 40
	5		278 12		430		105 90		75		195 15
	10		284 7		35		76 195		50		206 42
	15		281 21	100	320	1100	86 45		25		219 36
	20		81 223		5		96 41		55		225 58
	25		91 18		20		111 206		75		238 46
	30		106 156		35		128 170		60		249 18
	35		125 96		40		145 244		25		249 7
	40		148 175		45		161 68		65		254 15
	45		173 206		50		172 85		70		250 49
	50		191 238		55		172 119		75		249 37
	55		201 14		60		160 200		25		242 58
	60		194 17		65		143 87		75		232 36
	65		172 133		70		121 189		80		219 26
	70		144 171		75		103 141		40	350	142 63
	75		115 37		80		94 221		25		164 16
	80		92 48		85		95 54		55		182 58
	85		80 92		90		101 247		75		198 49
	90		82 243		95		110 169		60		220 61
	95		93 119		400		119 268		25		246 18
400			104 7		5		124 63		65		263 25
05			124 87		10		124 10		75		278 40
10			144 29		15		113 239		80		284 62

Penhadj. ST3

9/22, 23, 24/55 working in APPE-NEEL CONSOLE.

9/27/55 Instrument placement.



10 min. counts counters - background.

C <sub>1</sub>	(6-20)X64	3-0	300
C <sub>2</sub>	0	9-0	34
C <sub>3</sub>	(1-1)X16	750v	
C <sub>4</sub>	(1-0)X16	2000v.	16-50
C <sub>5</sub>	(1-14)X16	2000v.	8-50

Experimental Run - observations.

C<sub>1</sub> - ok.

C<sub>2</sub> - out of order

C<sub>3</sub> - check HVPS & plates

C<sub>4</sub> - check

C<sub>5</sub> - check

9-27-55

DC-1 ok

DC-2 ok

DC 3 seems to give shift at odd intervals -

R-1 ok.

PM ok

Log N noisy

9/28/55

Log N: Substituting chamber - still noise

Signal from simulator - still noise - trouble in amplifier.

Replaced Log N amp for NEPA model. E-181 with

~~Log N~~ Log N amp, Y-99087. no noise ok.

checked periods of 3 & 7 seconds VV.

10/11/55 Ref p 68

T1 - T4 Tube Check & changes - { Remove anode lead resistor }  
{ isolate mu-metal shield }

Noted on T-1 4 threads exposed on cap.

T-4 2 1/2 threads

T-4 threads not snug - nor T-1

Removed Cap (base) T-1

Crystal centered ok

1kV lead to crystal exposed - 1/4" anode lead ok.

Tube out of socket seat by 1/8"

Top of shield to tube - covered flange of crystal & 3/16" of crystal

T4: Crystal not taped to tube as was others (cross)

Photo cathode lead held to crystal out by 8-wire

Tube set high in socket as T-1

in paper crystal shield

10-12-55

12-12-55 Set up in 210 for checking miniature counters -

Check period 202 log w -

{ Stock room

12-13-55

Semiconductor Counter 215 - Supply - Scott

{ house keeping

12-14-55

Semiconductor Counter 215; Console panel 202

12-15-55

"

; Contact sheet (202)

12-16-55

Semiconductor Counter (215); DC-2 (202) -

12-19-55

Semiconductor Counter (215); Mini chamber set up (210); Console (202)

12-20-55

600 check C1-4 (215); 600 check SA1 & 2 (215); Mini ch set up (210)

Scrap controller (215); 607, file contacts 4k. relay; adj point

12/21/55 Counter placement 201; C4 & 5 in SID. This leaves only  
C1 in Hall for entrance tube W side of Tanker 2hr.

Prop. Counter 215 2hr

Scrap Ch 215 4hr

12/22/55

Cascade of counters 6

Scrap Chs - calib - 2

12-23-55

Scrap Ch ; 202 3, 5

12-27

Scrap Chs - 3; Gal 5;

12-28

215- 4; gal 4.

12-29


Gal 1

; Scrap Chs

Checked out push button stations in Rm 216 - All functions normal - fans run & dampers operate when buttons pushed in either 217 or 216 - In Jim's past, had had difficulty in getting emergency fan to operate properly -

Check-out test stand for small counters.

12-29-55 Test Stand:



Check out 1/2" RCC tubes to see if system is operable

Tube 1084 - attached thru 6' cable -

WV turned on - high freq signal developed - ~ 500 Kc.

debe off - cable still on - H.F. still there

Applied HV from NICC supply HV gone - disturbance fed from scaler HV supply -

USING: H.V. SUPPLY - NICC MODEL 1030

PRE-AMP - AIC MOD. 205B; SER. 335

LIN. AMP - AIC MOD. 204B; SER. 366

SCALER - AIC MOD 1010; "Y-89637"

Timer minutes of 1/100<sup>th</sup> min  
Runs when counts.

Counter & source placed in center of paraffin slab pile.

Thick news around is 3" or greater - open at ends -

Noted w/ source in certain pulse height & counting rate.

→ this kept up when turned off H.V. & when disconnected

LA. sw - 1" (ohm RC)

HV cable. Measured voltage on 10V input connector to pre-amp meter up range & return to zero -

run from 10 to hi V.

Check pre-amp. - seems OK

LA  
6AV6VX1  
  
Tube  
RCC  
1"

H.V.	Time	Scaler	Disc	H.V.	Time	Scaler	Disc
1150	10m	45K+0	15	1400	2m	99+3	15
1250	5m	44+2	15	1450	2m	32+11	17.5
1300	5m	68+13	15	1450	2m	11+9	20
1350	5m	5+5	20	1450	2m	3+2	21.5
	5m	125+11	15	1450	2m	0+8	25
	5m	9+5	20	1450	2m	121+6	15
	2m	73+12	15	1450	2m	51+7	12.5
	2m	41+3	17.5	1450	2m	18+0	20
	5m	15+3	20	1450	2m	4+15	22.5
	5m	15+3	20	1450	2m	(?) 2+11	25
	5m	15+3	20	1500	1m	90+4	15
	5m	15+3	20	1500	2m	42+1	17.5
	5m	15+3	20	1500	2m	20+15	20
	5m	15+3	20	1500	2m	12+5	22.5
	5m	15+3	20	1500	3m	3+4	25

12-3  
Disc  
ch  
176  
H.V.  
Disc  
SCAL  
TIME  
GND-

12-24-55

HV	Scaler	DISC.	HV	Scaler	DISC.	HV	Scaler	DISC.	HV	Scaler	DISC.
1600	149+3	15	1350	28+6	14	1600	2+6	22			
1650	29+14	20		16+9	15		1+8	23			
	190+3	15		11+7	16		0+7	24			
1650	(124+11)	20 (1.5m)		4+7	17	1650	187+13	14	with ESUM		
	235+15	15		3+3	18		141+2	15	Measure delay		
1700	121+7	20		1+10	19		97+14	16	at input for preamp		
	276+12	15		0+10	20 (3m)		56+3	17			
1750	160+9	20					42+10	18			
	301+14	15	1400	45+9	14		28+1	19			
1800	175+0	20		21+0	15		17+8	20	705V-		
	348+2	15		13+8	16		10+7	21			
1850	217+0	21		7+1	17		8+13	22			
	386+2	15		3+2	18		6+7	23			
1900	268+13	20		1+5	19		2+11	24			
	405+12	15		1+2	20		0+15	25			
1950	349+10	20	1450	61+0	14	1700	248+15	14	cont - next pg		
				30+13	15		187+15	15			
Bkg	1450	14+11	00	19+11	16		138+3	16			
				9+13	17		98+13	17			
				1+13	18		68+10	18			
				3+3	19		48+0	19			
				1+8	20			21			
12-20-55				88+11	14			22			
set scaler for + input pulser				53+4	15			23			
discriminator (-25)				35	16			24			
check with no source & 700 HV.			1500	18+13	17			25			
LAG 64X1				11+7	18			26			
this sets min. DISC setting of 14				6+6	19			27			
				4+0	20			28			
SOURCE 10. (9.21)				1+13	21			29			
(TIME 2m under auto)				0+7	22			30			
								31			
								32			
								33			
								34			
								35			
								36			
								37			
								38			
								39			
								40			
								41			
								42			
								43			
								44			
								45			
								46			
								47			
								48			
								49			
								50			
								51			
								52			
								53			
								54			
								55			
								56			
								57			
								58			
								59			
								60			
								61			
								62			
								63			
								64			
								65			
								66			
								67			
								68			
								69			
								70			
								71			
								72			
								73			
								74			
								75			
								76			
								77			
								78			
								79			
								80			
								81			
								82			
								83			
								84			
								85			
								86			
								87			
								88			
								89			
								90			
								91			
								92			
								93			
								94			
								95			
								96			
								97			
								98			
								99			
								100			

with ESUM  
Measure delay  
at input for preamp  
705V-

cont - next pg

10 min  
w/ HV off  
discriminator

1-4-56  
1700  
266+11 14  
217+14 15  
165+8 16  
114+2 17  
60+10 18  
65+9 19  
48+12 20  
33+15 21  
24+15 22  
19+12 23  
16+10 24  
8+13 25  
45? 26 -  
379+10 14  
HV off

2m 374 14  
30m 472+8  
2m 262

12-20-55  
set scaler for + input pulser  
discriminator (-25)  
check with no source & 700 HV.

LAG 64X1  
this sets min. DISC setting of 14  
SOURCE 10. (9.21)  
(TIME 2m under auto)

GND to Bus - Schv line

31  
20  
14.75

1-4-52

Having found HV lamp in pre-amp; modified set-up to measure HV with ES5M.

Raul meter reading of 1500 requires to get ES5M to read slightly over 600V.

Put in new 2V3G.

Now can get HV reading on ES5M to correspond to Raul meter

Checked cable to see if a section had leakage path. <sup>one section</sup> reflects

Now get full HV to pre-amp & fall off after HV deenergized.

① Apparently lost ground - may be throwing in counts -

② Plugged HV into CV line -

Power Off. 4:12 PM

1-5-52

Source in over night

③ run continued - two checkpoints at 650V.

stop 4:12 P leave source in

Linear Amp.  
64x1  
0.2µs  
reg. input

HIGH VOLTAGE	DISC	SCALER	HIGH VOLTAGE	DISC	SCALER	HIGH VOLTAGE	DISC	SCALER	HIGH VOLTAGE	DISC	SCALER
4500	14	15+10	650	16	84+5	900	36	3+9	1100	70	2+2
	16	28+0		20	14+2	(822)	38	1+3		74	0+15
	18	13+8	(680)	14	211+15	950	14	577+5	(1130)	14	807+2
	20	4+4	700	16	127+0	(920)	16	444+15	1150	22	638+13
	22	2+1		18	60+2		18	700+0		30	412+12
	24	2+2	(10500)	20	28+6		20	322+9		38	341+6
	26	1+5		22	12+10		22	266+4		46	210+15
	28	1+11	(735)	24	4+13		24	207+3		54	159+9
	30		750	26	0+12		26	120+6		62	90+4
				14	278+8		28	91+7		70	40+12
				16	180+14		30	65+1		78	16
				18	108+7		32	41+4		86	3+2
				20	51+4		34	24+4		94	0+13
				22	25+5		36	15+3	(1185)	14	803+14
				24	12+11		38	7+14	1200	22	708+1
				26	4+11		40	3+7		30	584+13
				28	1+3		42	1+1		38	483+2
				30			44			46	377+6
				14	360+2		14	634+4		54	285+12
				16	250+6		18	401+13		62	209+7
				18	168+7		22	331+3		70	145+13
				20	108+13		26	224+4		78	91+12
				22	61+3		30	145+11		86	57+10
				24	36+6		34	88+15		94	28
				26	16+14		38	40+5			
				28	9+8		42	19+7			
				30	2+15		46	7+5	(1178)		
				32	0+13		50	1+7			
				14	476+13		14	708+15			
				16	304+1		18	583+7			
				18	224+10		22	454+2			
				20	154+14		26	348+4			
				22	106+0		30	262+9			
				24	66+9		34	190+5			
				26	36+11		38	140+3			
				28	18+4		42	86+9			
				30	8+12		46	48+14			
				32	4+4		50	25+3			
				34	0+15		54	13+12			
				14	476+13		58	4+6			
				16	304+1		62	0+11			
				18	224+10		14	708+15			
				20	154+14		18	583+7			
				22	106+0		22	454+2			
				24	66+9		26	348+4			
				26	36+11		30	262+9			
				28	18+4		34	190+5			
				30	8+12		38	140+3			
				32	4+4		42	86+9			
				34	0+15		46	48+14			
				14	477+9		50	25+3			
				16	393+2		54	13+12			
				18	272+13		58	4+6			
				20	224+2		62	0+11			
				22	168+6		14	708+15			
				24	125+3		18	583+7			
				26	82+4		22	454+2			
				28	51+3		26	348+4			
				30	33+14		30	262+9			
				32	18+7		34	190+5			
				34	10+8		38	140+3			

2 pins  
jacks

AC supply  
to C.V. line

(ESVM)

(630)  
630

900  
(872)

1100  
(1080)

continuation of 1/5/52

1/6/52

drew up curves on data 1/5/52 -  
check point @ 1200v -

DIPS Scale 89637 -

check 6A95 & 1X24 - OK

high frequency present in this output still -



LA: AIC 204B - SER 364  
 SCALER: AIC 1040 "Y89637"

HVPS: NICL 1090

HV	DISC	SCALER	HV	DISC	SCALER	HV	DISC	SCALER	HV	DISC	SCALER
		(584)	1430	90	666+8	1750	14	1140+9			
1200	30	583		100	670+15	(1734)	20	875+4			
(1240)	14	799+0	1500	14	834+15		30	832+15			
9545	1250	22		20	825+9		40	810+8			
		30		30	793+3		90	809+11			
		38	(1418)	40	793+13		100	805+9			
		46		50	796+14	(1790)	20	1308+8			
		54		60	794+9	1800	30	807+6			
		62		70	763+4		40	852+12			
		70		80	750+5		50	820+10			
		78		90	736+10		60	816+6			
		86		100	703+5		70	817+9			
		94	(1505)	14	905+13		80	817+10			
			1550	20	844+9		90	824+1			
		14		30	805+12		100	794+11			
1300		20		40	802+1						
		30		50	784+7		14	13+7			
(1245)		40		60	803+6	1800	16	9+14			
		50		70	795+5	Bkg	18	6+8			
		60		80	786+9	(1790)	20	6+1			
		70		90	772+11		25	2+15			
		80		100	758+6		30	3+15			
		90	(1580)	14	940+14		40	2+8			
		100	1600	20	884+5						
(x4)		14		30	808+4						
1350		20		40	-		60	1+2			
(1340)		30		50	-						
		40		60	785+1						
		50		70	788+4						
		60		80	-						
		70		90	741+1		80	1+7			
		80		100	775+15						
		90		14	943+0						
		100	1650	20	907+9						
		14	(1635)	30	817+14		100	1+5			
1400		20		40	815+9						
(1340)		30		50	-						
		40		60	-						
		50		70	-						
		60		80	788+5						
		70		90	-						
		80		100	791+13						
		90		14	927+6						
		100	1700	20	846+7						
		14	(1690)	30	815+8						
		20		40	-						
		30		50	-						
		40		60	-						
		50		70	-						
		60		80	-						
		70		90	901+9						
		80		100	795+11						
		90									
		100									
11:40		14									
12:45		20									
		30									
		40									
		50									
		60									
		70									
		80									
		90									
		100									
		14									
		20									
		30									
		40									
		50									
		60									
		70									
		80									
		90									
		100									

Stop 4:01 PM  
1/6/86

1-9-56 <sup>High</sup> Draw up bias & voltage curves for Rk, 1" & counter under test. #1087

Suggest further tests of this size be at gain of 32 on L-10

Investigation of ripple on HVPS scales

- 150-200 KC - slight on output of supply (2-stage att'n x1)  
on gain 64 - signal drops out very pulse signal.

1-10-56 Contract System - solution storage - fixed-diam switch -  
Contract Rod Drive (R) 107 - dead - gone -  
Scintillation Counter 215; Counter set-up 202.

1-11-56 202: Reported low N difficulties & switch trouble  
Put in replacement switch - one capable of handling up to 10 comp  
Same wiring as with telephone switch.

1-12-56 Sealing strip 114-7 from 107 up to 210 for use in rack.

215: Check out Scintillation Counter -

210: Serviced Tektronix Scope Ser # 1722 -

107 - Channel "C" getting some - start up range instability

~~1-12-56~~ Serviced  
work on pen holder "A"

1-13-56 107: "C" reported improved

215: Sim. etc. check-out.

202: R2 - set-up -

210: Sim-delay motor checked - can get down to

1 second delay -


Back to proportional counter set up

108  
116x



2-16-56

BF<sub>3</sub> (Operating 500-1200v)



- 1: AIC Preamp #327
  - 2: AIC Lin Amp #366
  - 3: Decade scaler
  - 4: Power Supply
- BF<sub>3</sub> NICE 1090  
 fiss. Kepco

without shield, pickup transients badly.  
 with shield and ground lead fixed to prevent shorts, & B+ lead, got signal pulses. These pulses occurred with or without

source -  
 Test:  
 HV = 1000 v Bkg = 4844 c/m; Source = 6642.4 c/m; Bkg > 10,000 c/m??  
 LAG = 84 (RT 0.2 ps)  
 PDL = 50

Sum of HV-counted spurious pulses at same settings of amplifier < 1 c/m over 60 min period. (This counter #3.)

Put this aside to check fission counter - (100-200v)

Source 9.4 c/m Bkg 0.27 c/m Source 10.0 c/m  
 (LAG 16  
 0.8 ps RT  
 PDL = 40)

To get better source - present me PB 307 [1.6 x 10<sup>6</sup>, 11-4-54]  
 @ 1-2 x 10<sup>5</sup> now.

2-17-56 New Source PN-57 (used in "HOT" unit & Sid)

Strength (DPC) 3.7 x 10<sup>7</sup> on 6-2V-5V

now 2.183 x 10<sup>6</sup>

Take series of readings with no source in shop.

Roughly, seems that gain 64 PDL 50 (RT 0.2 ps) cut out most

Bkg @ 200v.

Instrument Settings: LA - Gain 64  
 R.T. 0.2 ps  
 PDL Variable

Kepco Powersupply -  
 HV, variable.

10100 HV=0	Source	14V 20
PDL 7m	12	7954
* 12 18800	15	4118
15 3560	20	1754
20 11700	25	1008
10141 HV 20	30	671
12 7913	40	320
15 3915	50	147
20 1603	55	92.7
25 919	60	80.1
30 601	65	91
35 376	70	83.2
40 380	75	81
45 152	80	78.7
50 73	85	68.2
* 55 11.4	90	71.3
60 0.3	100	64.8
HV 50	HV 100	
12 7466	30 691	
15 3820	40 360	
20 1555	45 271	
30 634	50 180	
40 261	55 106.6	
50 12.0	60 101.9	
60 .87		
HV 100	HV 93	
15 3652		
20 1634	80 55.2	
40 272	85	
55 15.6	90 91.8	
60 0.85	HV 200	
HV 200	20 360	
15 3561	45 289	
20 1576	50 162	
40 269	55 109	
55 146	60	
60 0.6	70 95.7	
HV X1	90 98.5	

- \* appear to be occasional bursts @ PDL 12
- \* pulses > 2.1 cm trigger scope 1X4 (33 u output)
- \* noted HV lead off (How Low G?)

2-20-56 Continuation of 2-16-56 MINIATURE COUNTERS  
 Check a few points of data run 2/17. Same counter, same equipment.


Source in 9:07 HV on allows warm-up	FDL	c/mg
9:08	30	742.6
	50	164.2 <sup>(6)</sup>
	70	84.0 <sup>(6)</sup>
	90	53.3 <sup>(6)</sup>
GAIN from 44 to 52	40	94.5 <sup>(6)</sup>
	50	86.8 <sup>(6)</sup>
	60	85.0 <sup>(6)</sup>
	70	83.2 <sup>(6)</sup>
	80	82.2 <sup>(6)</sup>
GAIN from 22 to 16	90	77.1 <sup>(6)</sup>
	90	38.5 <sup>(6)</sup>
	80	47.8 <sup>(6)</sup>
	70	53.6 <sup>(6)</sup>
	60	63.6 <sup>(6)</sup>
4:39	50	78.6 <sup>(6)</sup>
	40	84.0 <sup>(6)</sup>
	30	90.3 <sup>(6)</sup>
	25	90.7 <sup>(6)</sup>
	20	86.9 <sup>(6)</sup>
	18	84.9 <sup>(6)</sup>
	16	84.0 <sup>(6)</sup>
	14	162.7 <sup>(6)</sup>
	12	392.5 <sup>(6)</sup>
	10	932.5 <sup>(6)</sup>
	9	1778 <sup>(6)</sup>
BKG. (Source in 209)	4V 100	
	9	1306 <sup>(6)</sup>
	10	632 <sup>(6)</sup>
	12	265.9 <sup>(6)</sup>
	14	68.1 <sup>(6)</sup>
	16	10.2 <sup>(6)</sup>
	16	1.4 <sup>(3)</sup>
	17	0.07 <sup>(3)</sup>

2-21-56 Set up for time tests on dump and drain valves in 201-202  
 Object: a. Determine time required for dump valve (fuel) to open.  
 b. Determine rate of fall of dump drain.  
 c. Determine rate of fall of normal drain.

Panel modification of this is:  
 Fuel dump valve switch - one section used, closing when dump release. Disconnected leads at this point (w. sections paralleled) run leads to timing equipment.

Probe panel - modify 115v. leads to pin 4 of relays N & S so that could separate relay functions. Isolated NO contact (pins 3 & 4) of N- to drain timer. N-coil paralleled w/S.

Drain switch. Tie-in leads across drain valve to activate relay --



Ready -  
 [Need to give thought to instrument response tests]  
 Data recorded in file of "Safety Test Measurements" -

~~Recorder Drive Time - (Period key N in particular)~~

2-22-56 MINIATURE FISSION COUNTER - same arrangement #1 as 2/20

Obtain data for "plateau" at reasonable PDL.  
h.a. Gain = 16 ; PDL = 20 - This above bky. cut off.

HV.	Count Rate (c)
0	0
10	50.4 (6)
20	70.4 (5)
30	84.7 (6)
40	93.8 (5)
50	90.0 (5)
60	81.6 (6) 90.8 (6)
80	89.0 (6)
100	95.0 (6)
120	93.6 (6)
140	95.9 (6)
160	92.4 (6)
180	91.9 (6)
200	89.6 (6)
60	90.8 (6) ↑

stopped to check scale - replacing 2-12AT7's

Fission Counter #2 -

Background w/ Gain 16, HV = 100

PDL	COUNTS/MIN
8	2989
9	1384
10	680
11	390
12	252
13	144.8
14	63.8
15	11.8
16	1.07
17	

2/22/56 Counter #2 -  
1:18 P.A.G. 16; HV = 100

PDL	COUNTS/MIN
8	2989
9	1297
10	682
12	315
14	152
16	88.4
18	93.9
20	88.1
30	85.9
40	78.1
50	71.8
60	68.3
70	57.5

Source had been left in place over night.

2/23/56 Min. Fsn. Cr #2

Data for Count rate vs voltage

HV	counts/min
10	51.3
20	70.2
30	82.2
40	81.7
50	88.1
60	86.9
80	88.4
100	87.9
120	93.0
140	90.9
160	81.3
180	92.7
200	83.1

Suggested Inst. Modifications of Adair Series

2-28-56 Notes on Staff Meeting 2/27/56 - Recheck of 2/1/56

ADC: "The following items need attention as you can get to them:

- 1- Insensitive building alarm (see Vic, Jim, Scotty)
- 2- West end. Trip point to follow seal changes.
- 3- East end. Charalls A+D, make fault signal.

UGH-

- a. Arrange to have vent fans in both ends of building tied together.
- b. An instrument in each end of building for:
  - i. Low sensitivity \* see v.
  - ii. Capable of being checked in operation (drift, etc.) once daily.
  - iii. Each end capable of setting off building alarm.
  - iv. Exclusive of all instrument screams.
  - v. Sufficiently low in sensitivity so that normal operation will not scream the system.
  - vi. Sensitivity low enough so that burst on other end will not scream this instrument.
  - vii. Visual, or other, indication of which instrument caused trip.

3-13-56

Min. Fission Counter in probe - #1 -

Rough check HV=100; LAG 16, PDL=20  
 Source: \* 70 c/m (2m)  
 bkg. 0 c/m (27m)

Backgrounds: HV 100, G2M 16, PDL variable.

PDL	7	8	9	10	11	12	15	20	25	30	35	40
c/m	1390	125	0.8	0.0	1309	142	0.6	0.1				

Source: PN-57

PDL	8	9	10	11	12	15	20	25	30	35	40			
c/m	321	90	172	94	87.5	86	85.2	78	73.2	57.6	52.2	44.6	28.8	16.6

PDL	45	50	55
c/m	108	2.7	0.75

3/15/56

Miniature Fission Counter #1

Volts	PDL	8	9	10	11	12	13	14	15	17.5	20	22.5	25	27.5	30	35	40	45	50	55	60	65	
10		179.4	87.4	44.4	21.6	12.8	7.3	4.5	0.0														
20		299	121	120.9	108.8	97.4	80.8	68.8	44.8	37.6	10.8	5.0	0.4										
40		230	139.2	131.0	131.8	121.2	117.3	112.8	98.4	99.2	70	45.2	32.2	19.4	12.4	2.8							
60		219.6	148.8	137.8	145.0	118.8	127.4	128.4	118.8	104.2	106.7	80.2	65.4	52.2	41.4	21.3	7.4						
2m. → 80		278	150	143.5	157	142.5	127	120.5	130.5	114	109	92	70	75	58.5	38	21.5	8.5					
120		314	149	142	143.5	142	144	135.5	127	118	107.5	120.8	104	82	74	24	36.5	21	9				
(196) 200		169	153	145	146.5	146.5	153.5	131	144	102.5	83.5	66	52.5	45	13.5								
150																							
175																							
200																							


3/16/56

Plotted above data -

Took pre-amp (AIC #327) to 201.  
 Pulled bc-cable (50') w/ Jones plugs - for pre-amp thru of Bt.  
 Put new connector on R462/U line in 201 - this for signal from p-2 to linear amp (AIC 505) in 201

3/16/52

With this set up:



Low counts w/ ctr #1 in paraffin w/ PN 57 source

LA gain = 16 0.2 ps rise time (sw 1)

PDL = 10 41x16 + 12 (5 min) ~ 130 cpm  
20x16 + 0

PDL = 10 0x16 + 1 Background

PDL = 5 0x16 + 2 Background

2/27/52

Tested miniature fission counter #1 in Jarrow Shielding Reactor  
Critical Experiment. - With DWM, run up to power  $\log N = 1.1$   
DC 3 = 92 (2016)

Linear Amplifier: Gain 16; 0.2 ps rise time (sw 1), neg, PDL = 10

Scaler: C4, Disc -25

Original data in notebook for above series of experiments (p 39)  
Calculated counts of curve in file on miniature counters - OENL

3-29-56 Miniature Fission Counter #2.

Check Gamma Response

Counter #2 + 327 P-A + LA-366 + Decade Scaler:

Gain of system seems lower. LA now @ 64 PDL 16

Count w/ PN 57 (Po Be n source): (5 minutes)  $\frac{2887}{16} = 180.4$

Background (PN 57 & B n 8'): (10 minutes) 0

PN 57 (5 min) 390

784

5.0 mg. Ra (30 min) 1

Noted gain loss probably in amplifier system.

4-2-56 Check amplifier & pre-amp. refer to file

4-3-56 Ctr #1 in place pulses @ gain 16 w/ PN 57 2-3 cm 1x2  
Now #2 for comparison. 64 1.5-2 cm -

ctr #1 50 p.p.f.  
ctr #2 50 p.p.f.

4-4-56 Miniature Fission Counter #1 - Check counter & LA.

Equipment: P-A 327

Gain 16-1, 0.2 ps R.T., neg. (sw 1)

LA 366

HV: 2-67 1/2 batteries in series 135V -

Decade Scaler

PDL	5	6	7	8	10	12	15	17.5	20	22.5	25	30	35	40
sw 1 0.2 ps nt. Background	(15)	(20)	(45)	(60)	(85)									
	1631	46	8	2	0									
	(25)	(6)	(10)	(85)	(11)	(25)	(25)	(18)	(13)	(21)	(14)	(12)	(22)	(7)
PN 57	6400	1611	1371	4559	1428	3106	2909	1938	1258	2755	1058	744	955	210
4/5/56 PDL	5	6	7	8	10	12	15	17.5	20	22.5	25	30	35	40
sw 1 0.8 ps nt. Bkg. PDL	5,585	14681	3725	5048	1298	2919	1015	48	15	17.5	20	22.5	25	
	(5)	(6)	(7)	(8)	(10)	(12)	(15)	(17.5)	(20)	(22.5)	(25)	(30)	(35)	(40)
PN 57 PDL	7694	3592	1836	1220	622	1039	1335	1407	1364	4245	3159	1288	2081	1456
	(5)	(6)	(7)	(8)	(10)	(12)	(15)	(17.5)	(20)	(22.5)	(25)	(30)	(35)	(40)
" PDL	2646	1204	1193	1205	1115	1163	2181	1802	984					
	(5)	(6)	(7)	(8)	(10)	(12)	(15)	(17.5)	(20)					
5.0 ps PN 57 x	223	229	234	253	258	245	264	270	270	289	284	310	310	722








Trip circuits (refer p 90)

4-23/24/56 Check out vibrating feed Electrometer for use as test instrument in developing modified trip circuits for channels A&D, Cu 107.

4-25-56: Bread board trip circuit as in use on A&D.




P <sub>1</sub>	P <sub>2</sub>	VR Range	(Vmax) SIGNAL	TP 1	TP 2	TP 3	TP 4	TP 5
0" CCW	64	100	.02	1.76	2.5	6.8	184	192
0" CCW	0" CCW	"	"	1.84	3.23	7.15	186	193
0 (CW)	102	"	"	1.80	1.91	6.58	183	181
105 (CW)	102	"	"	.02	1.88	6.60	183	181
0 705	0 705	"	"	.02	1.87	6.54	183	190.9
105	18	"	"	.02	1.80	6.54	182.5	190.9
105	50	"	"	.02	1.13	6.24	181	190
0	70	"	.82-.80	1.86	2.44	7.55		
0	70	"	.89-.82					
0	70	"	.85	1.93	2.54	7.25	182	191
			.89-.84					
			1.05-1.01					

R<sub>2</sub> trips only  
R<sub>1</sub> trips  
R<sub>1</sub> & R<sub>2</sub> both fired  
R<sub>1</sub>  
R<sub>2</sub>  
150K across R<sub>2</sub> coil

P <sub>1</sub>	P <sub>2</sub>	VR	SIG	1	2	3	4	5
0	70	100	.95	1.84	2.55	7.32	183	191

474K both R<sub>1</sub> & R<sub>2</sub> fire about same.  
220K R<sub>1</sub>@.90 R<sub>2</sub>@.100  
Set to trip R<sub>1</sub>@.90 & R<sub>2</sub>@.100 v signal. P<sub>1</sub>=0  
P<sub>2</sub>=70

Try circuit with signal polarity reversed.  
Note channel is floating as is power supply -




could not control @ B<sub>2</sub> & 200; reduced B<sub>2</sub> to 100 -  
Now: E<sub>B</sub> = 100 & E<sub>SIG</sub> = 0 (100)  
w/ P<sub>1</sub>@0 (500Ω) & P<sub>2</sub>@60 same trips R<sub>2</sub>@.55; R<sub>1</sub>@.65 E<sub>B</sub>  
no sig. applied - relays reset (energized)  
P<sub>1</sub> 0-500Ω no change - can trip relays by rotating P<sub>2</sub> CCW  
increasing cathode resistance. @ 420 P<sub>2</sub> relay "trip"  
cannot reset w/ P<sub>2</sub> but w/ P<sub>2</sub>@100 & P<sub>1</sub> rotating 20 relays "in"  
P<sub>1</sub> 0; P<sub>2</sub> 100: { P<sub>1</sub> 0; P<sub>2</sub> 0. } P<sub>1</sub> 100; P<sub>2</sub> 100 P<sub>1</sub> 100; P<sub>2</sub> 100 P<sub>1</sub> 0 P<sub>2</sub> 20 P<sub>1</sub> 20  
TP<sub>1</sub> = 1.78 { 1.30 } 1.78 .025 1.34 1.61  
TP<sub>2</sub> = 1.80 { 2.62 } 1.80 .05 2.54 1.63  
TP<sub>3</sub> = 85 { 89 } 85 .74 88 84  
TP<sub>4</sub> = 92.5 { 95 (tripped) } 93 87 94 92

4-25-56 replaced 2-500Ω pots with one 2.5K. & moved B+ to 100 - adjustment not so good - to return to 100 volt operation.

4-26-56

Circuit



Signal measured w/ Simpson 10073C  
 Just readings taken w/ Vomas  
 Different from 4/25, had noted Simpson trigger relay

Adj B+ to 100v. - Set signal to zero. Relays are energized with pot, P @ 60.  
 Turn P<sub>ccw</sub> trip R<sub>2</sub> & R<sub>1</sub> @ 22-19 }  
 Turn P<sub>cw</sub> reset R<sub>1</sub> & R<sub>2</sub> @ 65 }  
 Set P to 18 - Both relays de-energized  
 Set P to 66 - Both relays re-set

E<sub>K</sub> = 2.54 reset voltage  
 E<sub>P</sub> = 88 5.5v  
 E<sub>R</sub> = 92.5 6.5v firing voltage  
 E<sub>K</sub> = 1.58 R<sub>1</sub> 8 volts  
 E<sub>P</sub> = 83 R<sub>2</sub> = 9 volts  
 E<sub>R</sub> = 91

Extremes of P - 0 - E<sub>K</sub> = 2.67  
 100 - E<sub>K</sub> = 0

Adj B+ to 120. Extremes of P/100 E<sub>K</sub> = 3.10 ; E<sub>P</sub> = 107.6 ; E<sub>R</sub> = 113.5  
 w/ P=0, Trip point, R<sub>2</sub> = .75v signal E<sub>R</sub> = 194.8  
 R<sub>1</sub> 80v. E<sub>R</sub> = 194.8  
 P=20 R<sub>2</sub> 1.07v 1.13v  
 P=40 R<sub>2</sub> 1.68v R<sub>1</sub> 1.76v @ P=10 resets at same signal.


P=60; Trip @ 2.3 volts - reset @ .80 volts signal  
 ⇒ Increase R<sub>K</sub> - w/ setting 470Ω on ground side of pot, P.  
 Now can adjust & cutoff relay w/ P  
 E<sub>sig</sub> = 0 P<sub>ccw</sub> R<sub>2</sub> 24; R<sub>1</sub> 26 trips (cutoff)  
 P<sub>cw</sub> R<sub>1</sub> 84; R<sub>2</sub> 86 reset (energize)  
 P=20 = .30, .40 ← R<sub>K</sub> added = 270Ω  
 40 = .85, .95V  
 60 = 1.50, 1.55 - no reset -

⇒ B+ adj to 200 R<sub>K</sub> 1000 not enough want to just almost cutoff (reset) relays at P=0

Now Vomas across R<sub>2</sub> - E<sub>K</sub> 6.7-9.6v (R<sub>2</sub> energized) +  
 R<sub>K</sub> = 4700, too much E<sub>K</sub> = 3.2-3.6v (R<sub>2</sub> cutoff)  
 82k across 4.7k across E<sub>K</sub> to 5  
 5.6k " " 5.6  
 3.9k " " 6.3  
 2.0k " " 6.3  
 (P<sub>0</sub> 5.4 not cutoff -  
 P<sub>00</sub> - 8.0 give both relays) E cutoff 5.6-5.7 v.

R<sub>K</sub> = 2000 - "reset" @ P=40 -  
 6.2k across 2.0k in series w/ 1K give wide range of control  
 w/ Pot P - P=0 trip 2.25 P=100 trip > 2.5 no "reset". Good  
 E<sub>R</sub> range 5.8-7.75 volts.  
 With P @ 30 trip @ 20-100 - recheck this.


Q/2) To check Synchroni & other relay time of action.



$2000 \Omega - 2 \text{ mH}$

$2 \times 10^3 \times 2 \times 10^{-3} = 4 \text{ V} - 6 \text{ V}$

need to drop  $\approx 18 \text{ V}$



C-21

SIGMA

~~2000 G - A Fire 1.87~~

2000 G - SIL - I

	FIRE	RELEASE
1	1.95	1.12
2	1.90	1.17
3	1.90	1.15
4	1.8	1.1
5	1.85	1.15
6		
7		

5000 G




1.1

0.16



5000 G

0 1 2 3

C/21



6-21





4-27-56 Check circuit on bread-board for stability power on 840

Signal zero:  $P=0$ ;  $E_2 = 5.85V$

$P=100$ ;  $E_2 = 7.60V$

Trip points:	$P=0$	10	20	30	40	50	60	70	80	90	100
	$R_1=30$	32	43	46	123	1.53	1.79	2.12	2.60	2.95	3.20
	Sig = $R_2=20$	24	31	36	108	1.40	1.72	2.00	2.44	2.85	3.10

11:24  $P=30$   $R_2=85$ ;  $R_1=95$   
3:15 .85 .94

Power off WSP

5-2-56 Power on 8:25 - 8:37A 85, 95 @  $P=30$  -  
3:50P 85-95 @  $P=30$  - first - power on over night

5-8-56 8:00 A/ 78, 88  $E_2 = 200V$ ,  $I_2 = 6mA$   
3:10P 70-87 -

5-9-56 8:15 A - Trip @ 33-36 - Trip point has drifted -

Installed on bread-board trip indicator lights 6.3V, 150ma. tied to AC contacts on relays -

After power shut-down - Trip @ 75-88 @ 1:20 PM -

Power left on

5/8/56 8:22 A/ 78-90  
11:2 P/ 77-90

5/9/56 8:15 A/ 82-92


5-11-56

Modify trip circuit - Channel "A" -

Isolated input & output from chassis.

Vacuum tube supply common (B-) isolated from chassis

Wiring diagrams:



$B+ = 200V$

Can trip & reset relays by adjustment of "trip level" pot. Check this -

Relays -  $R_1$  - contact opens upon "TRIP" Trip when power off

$R_2$  - contact closes upon "TRIP"

i. - turning "Trip level" counter-clockwise -  $R_1$  trip, then  $R_2$

ii. turning " " clockwise -  $R_2$  resets, then  $R_1$

5-14-56 (c)

Voltage Readings -

"T.L." ccw.;  $E_k = 7.3v$

Tripp - 4.7.3

cw;  $E_k = 6.0v$

Reset - 6.5v

This with relays on unit "A" in 107 - put in units in bread board.  
no such trip ~~unit~~ can reset once tripped.

Voltage Readings "T.L." ccw;  $E_k = 7.25$

"T.L." cw;  $E_k = 6.0$

Source in for tripping -

Trip level	<del>R1</del> R1	R2
ccw.	B1 0.61v	B2 0.39v
	B2 0.73v	B1 0.50v
	B1 0.57	B2 0.52
	B2 0.65v	B1 0.45v
	B1 0.49	B2 0.57
	B1 0.50	A1 tripped
	B1 0.53	A2 tripped

O.C.  
8696

Left B1-B2 in - Power on - overnight

5-15-56

Check - old circuit for operation characteristics after morning check of modified A.

Modified A

ccw.	B1	B2
0	.49	.50
1	.50	.52
2	.78	.50
3	1.02	1.06
4	1.35	1.35
5	1.62	1.62
6	1.92	1.92
7	2.27	2.27
8	2.58	2.58
9	//	//
10	//	//

Reset

0.4

0.4

0.4

< 0.4

"

"

Not reset 'A' if hard to tell trip.

3-207

5-15-56

D - Trip circuit - found that ground has connection thru in large signal - this because output from this model VR is not grounded on either side - left off ground.  
AC on - DC on - adj to B+ 200v.

T.L. R1 R2

"50" .20 .30

"75" .55 .65

R1 trips if units self Rk's not sharp - "m-iff" indication should be

"100" 1.00 1.10

No click on R1 - resets 95?? R2 normal - D1

"150" 1.90 2.05

R1 resets 105 R2 will reset on P.B. @ 200

putting A1 in place of D1 gives click for both relays (check D1)

Back to Modified A -

Removal - B+ lead to R2 of inserted DC matrix - to measure

Relay currents

T.L.	Relay used	Current	Voltage
	R1 R2	R1 R2	R1 R2
0	B1 B2	2.22 2.22	.78 .58
1		" "	" "
2		" "	.66 .60
3			.95 .91
4			
5			
6			
7			2.27 2.28
8			

Comments -

Start I = 2.25

not much change near "0"

$I_2 = 2.36$   $I_{rs} = 2.9$  (Insert required 3.04 ma)

2.39

2.67 ma

2.16

Current range  
T.L. 0-10 2.20 - 3.13 ma

5-15-52 Reset button across fixed cathode resistor -  
 $R_1(B1) E_1 = 2.36, I_{T2} = 2.2$   
 $TL = 7, R_2(B2) E_2 = 2.42, I_{T1} = 2.2$   $I_T @ E_3 = 2.0$  44ma reset's OK. NOW  
 start signal. 23  
 start current 2.68 ma


TL 2	$R_1$	$E_1$					
		.93	.96	.86	.90		
	$R_2$	1.05	.98	.92	.96		

TL	0	2	3	4	5	6	7
$E_1$ B1	50	.82	1.11	1.38	1.70	2.08	2.70
$E_2$ B2	50	.90	1.20	1.45	1.76	2.10	2.35

10/52 Leave set @ 2" - check stability  
 3/10P  $R_1 = .90; R_2 = .92$   $\parallel .85, .91$

5-16-52 wire modified D -  
 9:54A Check A - .94 - .94

re-mounted D on panel.  
 Before further check, run test on each relay -  
 Used breadboard as test panel, with following changes.



Relay	TRIP (6557)		RESET		Notes
	V	I	V	I	
D2	5.3	2.17	7.56	3.04	Start w/ trip $R = 2.440/2.480$
D1	4.2	1.7	6.45	2.75	Start w/ reset - reduce set to trip.
	6.0	2.54	6.53	2.60	
D3	5.38	2.30	7.73	3.25	Trip out before reset throws in <u>relay</u> <u>source</u>
	5.4	2.3	7.76	3.26	

$R = 2.350/2.380$

Relay	Trip		Reset		R
	V	I	V	I	
A1	5.73	2.75	7.50	3.03	$R = 2.440/2.470$
A2	5.62	2.33	7.62	3.10	$R = 2.410/2.460$
B1	5.43	2.20	2.90	3.10	$R = 2.460/2.540$
B2	5.28	2.20	2.67	3.13	$R = 2.360/2.450$

If  $R_2$  is lower level trips - need relay w/ high  $I_T$  such as D3, A1, A2  
 then for  $R_1$  use B1 or B2

Try on A' circuit A2 as B2 & B1 as B1  
 D' D3 as R2 & B2 as R1

UNIT	A	Spd. B1(B2)	Cond. RE(B2)	V R1	V R2
ow(11)	S	S		7.9	7.7
1.5	S	T			5.7
2.2	S	S			7.5
0	S	T		5.8	5.6

but D' R trip @ P. set 73 set next by pot -  
 with B1 & A2 in A' as R1, R2 & B2 & D3 in D' difference in trip levels was too great,  $R_2$  tripping @ say .3 while  $R_1$  at 1.0 - This gap too far.

Re-arranged Relays A unit, A2-D3  
 D unit B2-B1 This better -  
 Set up down checks

~~5-17-56~~

5-28-56 Power on 8.42 Km -

"A" - A2 - D3 relays

9:12	TL	R <sub>1</sub>	R <sub>2</sub>
	3	.32	.2
	4	.62	.5
	5	.92	.78
	6	1.28	1.17
	7	1.61	1.52
	8	1.88	1.74
	9	2.35	2.30

"D" - B2 - B1 relays

9:03	TL = 3	R <sub>1</sub>	R <sub>2</sub>
	4	.49	.49
	5	.78	.76
	6	.99	.99
	7	1.32	1.32
	8	1.55	1.55
	9	1.79	1.82
	10	2.20	2.20

9:16

set @ 5

	.96	.86
10:44	1.04	.99
	1.02	.89
	1.04	.89
	1.02	.94
	1.04	.88
	1.04	.90

OK 9:10 -

set @ 6

	1.17	1.17
10:47	1.15	1.15
	1.17	1.17

5/21

	1.08	.95
	1.08	.98
	1.08	.98
	1.08	.95

2:54

	1.25	1.25
--	------	------

5/24 8:37

	1.06	.95
--	------	-----

7:24

	1.24	1.24
--	------	------

5/31

	1.04	.95
--	------	-----

1:28

	1.28	1.28
--	------	------

5/31 "D" 470K avon R2 - serapant - 50 R. free @ 12  
4 R. @ 13

[Faint, mostly illegible handwritten notes on lined paper]

Period Measuring Device

7-19-56 D.F.C. suggested need for device to measure periods to a greater accuracy than the meter (too much "grass" for reliability) and the manual clock timing methods. The suggestion included the idea of some type of electronic circuitry which would start & stop the clock automatically when a pre-set range of power is covered by the log N amplifier channel.

7-20-56 log N amplifiers have terminal (B.N.C. type) on rear of chassis marked "OUTPUT". This is on the "top" of the leg including the power meter, the resistance of this leg being  $\approx 8$  meg. The voltage between this point & ground (chassis) varies between  $\approx 35$  to  $\approx 180$  v over the entire power range (.0001 to 300) of  $6\frac{1}{2}$  decades. Voltage measurements with the HP UTVM on the unit mounted in Pm. 202 gave readings of

	$\approx 35$ v.	at	.0001	
Lo Cal.	56 v	at	.001	1 decade $\approx 26$ v
Hi Cal.	156 v	at	100	5 decades $\approx 100$ v (20 v/decade)

A. First thoughts are to build circuit fed by this signal of high enough impedance to have no effect on the amplifier and to include sensitive relay to activate the timer.

7/20/56 A Resistor network with total value of 10 meg may be first considered-



Current range

$$I_{35} = \frac{35}{10^7} = 35 \times 10^{-7} = 3.5 \mu\text{a}$$

$$I_{180} = \frac{180}{10^7} = 180 \times 10^{-7} = 18.0 \mu\text{a}$$


AT "GND SET" V=35 I = 3.5  $\mu\text{a}$  .0001

"Lo CAL." V=55 I = 5.5  $\mu\text{a}$  .001

"Hi CAL." V=155 I = 15.5  $\mu\text{a}$  100.

Sensitive relay (Assembly Products Inc) Bulletin 102 12/55

8/9/56 Checked current thru external circuit



@ .001 7  $\mu\text{a}$

@ 0.1 12  $\mu\text{a}$

9-6-56 Set up to check out miniature Fission Counter #1.  
 Had been in use with cube table experiments. Seemed  
 → to exhibit instability to extent that counts were  
 apparently not reproducible at fixed foil exposure levels.

a. Check out linear amplifier - pre-amplifier.  
 Pre-amplifier from field - # 327  
 Linear amplifier, shop, # 366. refer: file on  
 Lin Amp. 366

9-7/9-17 Counts with B-Bc source to check drift. No marked  
 drift. refer: file  
 min. counters serv.

11-14-56 Set-up for testing new miniature fission counters from  
 Instrument Shop - X-10.

- Depleted counters from X-10 to be checked as well as extra  
 enriched samples.
- Procedure. a. Mount counter in test mount. Attach to pre-amp.  
 Insert in paraffin housing. ~~with~~
- b. Pre-amp is attached to linear amplifier and  
 scaler and timer.
- c. Data on tubes are obtained with and without  
 source.

Equipment: Pre-amp # 349 - adapted for counter use (see file  
 pre-amps)

Linear Amp # 366

Scaler # 4-89641

Voltage Source - Kepco H.V. Power Supply -

(Note - When meter on supply reads 100v, volts <sup>read</sup> at connection  
 of counter to pre-amp reads 92v.

Preliminary -

Check count with counter mount in place, but no counter -

TEN MINUTE COUNTS PDL

HV	60IN RISETIME	6	7	8	9	10	
0	10 / 5 ps.	29,667.2	29,113	28.4	0.1	0.5	Scaler in unreg. line of disc 16 Plug in reg. line DISC-10 check - ONE MIN.
100	"	-	29,240 28,93.7	28.5	0.5	0.0	
200	"	-	4,062.6	36.0	1.1	0.3	
0	"	-	3,978.3	48.6	0.6	0.1	Same = 81.9
100	"	-	3,667.1	45.2	6.7	3.5	
100	-	-	-	-	-	2.8	
100	-	-	-	64.5 72.1	-	-	

ctr. 238 #2

ARE same  
 PB 124

Min. Fission Counter -

11-15-52 Did not seem to get any response with depleted counter.  
 Checked test setup with enriched counter - Monitoring amplifier output with Tektronix scope.

With amplifier gain of 16 (setting for operation of counter # 1 -), signal at 100 volts seemed to overdrive. Set gain at 8 & rise terminated. First to eliminate overdrive, second to shorten pulse width.

Output pulses with source about 100 up to 100 volts amplitude.

Removed PA 124 & put in ARE source (used in previous tests) (Set Amp XB 6kg (etc in bu) HV=0) at PDL 5  
 PN-57  
 begin counts at PDL 6

Counter #2 (PDL #)	Amplifier HV	9; 0.8µs	8; 0.8µs	3min	11/14	ROM(A)	(D)	
100	100							
6	485.1			PN 67 Source	PA 124 Source	7	74.58	
7	329.9					8	116.3	
8	119.7					9	6.2	
9	69.3					10	3.3	
10	26.0					11	2.7	
11	9.5					12	2.8	
12	1.2	72	34.4	276	321	15	0.6	3.5
13	0	75	32.1		365	20	0.0	3.3
6	509	80	32.3		317	25		2.0
		85	30.9		280	30		1.8
		90	28.1		304	40		0.6
		95	24.2		254	50		1.5
		100	23.0		254	60		1.8
		105	20.2		270			
		110	21.7		241	20		2.25
		115	19.4		210			
		120	19.7		194	174		
		125	19.5		167	70		0.05
		130	17.8		159			



11-16-52 Fast fission counter - depleted -

Run A - Background. Amp 16, 0.8 ps. HV = 100. (Found in

preliminary check with hot source that signal too high for gain 32. Got reasonable pulses at gain 16)

B. SOURCE - PB -124 - same amplifier & HV -

11-19-56 Mounted fast fission counter in probe, removing counter 5-3,

(etched counter) Counter now in probe 8-2

70L	A c/m	B c/m
7	849.4	551.11
9	347	2.4
9	2048 32.8	0.3
10	1471 30.5	0.1
11	2146 342	
12	1711 28.3	
13	1612 26.8	
14		
15	1712 38.4	
20	1479 23.3	
25	1672 25.8	
30	1811 22.2	
35	1213 20.5	
40	1372 21.0	

Run A - { source # PN-213 4/3/56  
Amp 16, 0.8 ps, HV = 100. 3.2x10<sup>7</sup>

Run B - BACKGROUND

LAN SPEEDOMAX H. = Test =

11-16-56

Received, on loan, one instrument. From R.G. Affel 9201-3.

Items: I. SPEEDOMAX H STRIP CHART RECORDER.

Catalog No. 3-30-41-044-6-02-0

Serial No. 36-39671-2-1

Chart # 600006 Chart Speed 2 in/hr.

Response Time 5.0 sec full travel.

Primary Element GMP source

Range 0-10 mv. - Ref. Junction ---

(1200, 600, 40 v.a (recording features only))


- 1- 124104 General Accessories kit envelope
- 1- roll chart paper
- 1- take-up spool.
- 1- battery.
- 1- INKING PARTS (pen, ink) for strip chart record (+ labels)
- 1- Instruction Manual

12-11-56

Imp Circuits for DC-2, DC-3 Super amplifiers -

Wind resistor board for use with 0-100 microampere APE meter.

Using battery as voltage source took measurements - w/ HP UTOM.



$V_{AB} = 1.27$ : Switch position 1,  $V_{A2} = .638$ ,  $I = 0.8 \mu a$ ,  $V_{A3} = 1.02$ ,  $V_{A4} = 1.14$

$V_{A5} = 1.215$ ,  $V_{A6} = 1.25$

$V_{AB}$	SW	I	A-1	A-2	A-3	A-4	A-5	A-6	A-7	A-8	A-9	A-10
1.07	1	0.8		.638	1.02	1.14	1.215	1.25				
	4	11.2	low .01				.646	1.03	1.15	1.22		
	6	65.0							.68	1.00	1.20	1.27
20.8	1	20.0		10.3	16.4	18.4	19.35	20.0	20.7	20.3		



12/13/56 Set up instrument DC-231 on bench. This is DC-2-3 type.

Wired resistance bank on previous page & meter to output bus, in parallel with output meter circuit on panel of instrument. (Recorder output to 10mv  $\times 10^5$ )

Took series of readings, adjusting zero to simulate signal.

Switch position	API Meter					
R= 1k	10	100	-	-	-	-
2k	9	48	100	-	-	-
5k	8	19	41	100	-	-
10k	7	9	20	51	100	-
20k	6	-	10	26	52	100
50k	5	-	4	10	20	40
100k	4	-	2	5	9.5	19.5
200k	3	-	-	2.6	5	10
500k	2	-	-	-	2	4
1Meg	1	-	-	-	-	2

12/17/56 Comparison VX-55 - VX-32 (5803) log N application.



Reading grid voltag with Simpson, F- w/ Simpson, Plate w/ Simpson; B+ and F+ w/ HP VTVM (not fixed in ckt)

Adjusted  $R_f$  to  $\approx 120\Omega$ ,  $R_p$  to 470k $\Omega$ .  
Readings taken of  $E_g$ ,  $E_p$ .

A. $E_g$	0	.1	.2	.3	.4	.5	.6	.7	.8	.9	1.0	$R_g \approx 120 \Omega$
$E_{s-}$	1.29	1.29	1.29	1.30	1.30	1.30	1.30	1.30	1.30	1.30	1.30	$R_p \approx 470k \Omega$
$E_{s+}$	2.53											5803
$E_p$	6.7	6.85	7.06	7.20	7.35	7.50	7.65	7.85	7.97	8.15	8.30	
$E_B$	40				40						40	

B. $E_g$	0	.1	.2	.3	.4	.5	.6	.7	.8	.9	1.0	$R_g \approx 120$
$E_s$	1.30	1.30	1.3	1.30	1.30	1.30	1.30	1.30	1.30	1.30	1.30	VX55
$R_p$	5.04	5.17	5.33	5.50	5.67	5.80	6.00	6.16	6.30	6.45	6.60	$R_p \approx 470k \Omega$
$E_B$	40	40	40		40						40	

Adj  $R_p$   $E_g = 0$ ;  $E_p = 6.7$  @  $E_g = 1.0 - E_p = 8.40$   
 C. Measure  $R_p \approx 200k \Omega$   $\bar{E}_g$  1.39  
 $\bar{E}_p$  1.27

D. Variation of  $E_p$  with changes in  $E_g$  VX55  $R_p \approx 200k \Omega$

$E_g$	$E_p$	$\Delta E_p / \Delta E_g$
1.175	3.3 - 5.15	(1.85)
1.150	5.8 - 7.60	(1.80)
1.125	7.70 - 9.40	(1.70)
1.10	9.7 - 11.7	(2.0)

E. Variation of  $E_p$  with changes of  $E_g$   $R_p \approx 200k \Omega$

$E_g$	$E_p$ range ( $\bar{E}_{g=0}$ )	
1.33	5.05 - 6.85	(1.9)
1.30	9.3 - 11.5	(2.2)
1.275	12.2 - 14.0	(1.8)
1.250	14.0 - 16.0	(2.0)
1.225	17.2 - 19.0	(1.8)

12/19/56 modify test ckt for 5803 - VX-55. Jic. bias reference to f-point for further readings. First, a recap of data opposite page.


A.  $E_g$  values reference to B-. So  $E_g = 0$  means  $|E_g - E_{s-}| = 1.29V$ .  
 $E_g = 0$   $i_p = \frac{40 - 6.7}{4.7 \times 10^5} = \frac{33.3}{4.7} \times 10^{-5} = 7.09 \times 10^{-5} = 70.9 \mu A$   
 $E_g = 1$   $i_p = \frac{40 - 8.3}{4.7 \times 10^5} = \frac{31.7}{4.7} \times 10^{-5} = 6.75 \times 10^{-5} = 67.5 \mu A$

B.  $E_g = 0$   $i_p = \frac{40 - 5.04}{4.7 \times 10^5} = \frac{34.96}{4.7} \times 10^{-5} = 7.44 \times 10^{-5} = 74.4 \mu A$   
 $E_g = 1$   $i_p = \frac{40 - 6.60}{4.7 \times 10^5} = \frac{33.4}{4.7} \times 10^{-5} = 7.11 \times 10^{-5} = 71.1 \mu A$

C.  $E_g = 0$   $i_p = \frac{33.3}{2 \times 10^5} = 16.65 \times 10^{-5} = 166 \mu A$   
 $E_g = 1$   $i_p = \frac{31.6}{2 \times 10^5} = 15.8 \times 10^{-5} = 158 \mu A$

New test circuit:

- 0-200 Nechry house NX-33 ("JN 6130")
- 0-50 Weston 0-50  $\mu A$  (from high range meter)
- 0-50 mA - Marion Elec Inst. Co.



12/14/56

Data for 5893 characteristics - Tube used removed from Log N.


$E_b = 40$  ( $E_{g1} = 1.53$ ;  $E_{g2} = 2.72$ ;  $I_f = 9.3$  ma.)

Bias voltages taken with reference to (f-) point.

volts $E_{g1}$	volts $E_p$	ma $I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$
0	2.46	20	2.8	30	3.10	40	3.35	50	3.6	60	3.95	80	5.72	200
2	6.15	18	6.64	26.6	6.93	35.8	7.2	44.4	7.53	53	7.9	71	10	177
4	9.92	16	10.6	23.5	11.0	31.3	11.4	38.4	11.6	46.5	12.2	62	14	152.5
6	14.0	14	14.4	20.2	14.8	27	15.2	33.2	15.5	40.5	15.8	53	17.7	130
8	17.5	12	18.1	17.5	18.3	22.9	18.6	28	19	34	19.3	45	21.2	108
10	21.1	10	21.5	14.4	21.8	19	22.2	23	22.5	28	22.8	36	24.5	89
12	25	8	26.3	11.4	25.7	14.75	25.8	18	26.3	20	26.6	27	28.3	66
14	28.5	6.1	28.8	8.5	29.3	10.8	29.6	13	29.6	15	30	19	32.6	45.5
16	33.3	4.2	33.0	5.9	33.5	7.4	33.5	8.6	34	10.5	34.5	12	35.5	29
18	36.1	2.3	36.5	3.4	36.6	4.1	36.7	4.6	37	6	37.2	6.5	38	13
20	38.6	1.0	38.8	1.2	39.0	1.3	39.1	1.4	39.2	3	39.4	2.5	39.7	3

Raw data for test on VX-55 Tube used was not tested shortly in Log N.

$E_{g1}$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$
0	4.2	10	4.8	20	5.52	40	6.1	60	6.9	100	7.36	125	7.7	150
2	7.58	9.3	8.26	18.1	9.1	35.7	9.68	52	10.6	90	11.2	113.4	11.4	132
4	11.2	8.2	12.0	16.1	12.8	30.4	13.6	47	14.4	78	14.9	98	15.3	116
6	14.7	7.3	15.6	14.1	16.4	27.1	17	41	18	66	18.5	84	18.7	100
8	18.1	6.3	18.8	12.2	19.7	23	20.4	34	21.3	55	21.7	70	22.2	83
10	21.4	5.4	22.3	10.3	23.1	19.2	23.6	29.5	24.6	46	25	58	25.2	68.5
12	25.2	4.5	25.8	8.1	26.8	14.9	27.5	31	28.2	34	28.5	42	28.8	51
14	28.6	3.5	29.3	6.2	30	11.1	30.9	16	31.3	25	32	30	32.5	35
16	32	2.2	32.6	4.5	33.4	7.5	33.6	11	34.4	16	34.6	18	35.2	23
18	35	1.4	35.5	2.5	36.5	4.5	36.6	7	37	9	37.4	10	37.4	12
20	39.4	0.8	38.2	1.0	38.5	1.5	38.7	3	39	4	39.2	4	39.4	4




12/14

VX-

3252

04

12/19/55  
VX-55  
CURVES



12/20/56 5803 same tube

grid voltage reference  $E_f$  (only change)  $E_b = 40V$   
 $E_f = 1.50$   
 $I_f = 9.6$

$E_g$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$	$E_p$	$I_p$
0	2.85	10	4.39	20	5.05	40	5.76	70	6.45	110	6.95	150	7.46	200
-2	7.5	9.1	8.12	18	8.9	35.8	9.66	63	10.6	99	11.2	131	11.8	175
4	11.4	8.1	12.0	15.9	12.9	31.3	13.8	54.5	14.5	85	15	113	15.7	152
6	15.2	7.2	15.8	13.8	16.7	26.8	17.4	46	18.2	75	18.7	98	19.5	128
8	18.7	6.2	19.3	11.8	20.2	22.7	20.8	39	21.6	60	22.2	80	22.7	106
10	22.2	5.3	22.7	9.8	23.5	18.5	24.4	31	25.1	49.5	25.5	64	26.2	82
12	25.9	4.2	26.8	7.5	27.5	14.1	28.2	23	28.8	35	29.3	47	29.9	64
14	29.8	3.0	31.7	5.5	31.3	10.1	32	15	33	23	33.2	31	33.7	40
16	33.6	2.0	33.9	3.7	34.7	6.4	35.3	10	35.5	14.5	36	18	36.6	22
18	36.6	1.3	36.7	1.8	37.5	3.2	37.8	5	38.2	6	38.5	8	38.8	8


Bias to (Grid) (B-)

$E_g$	$E_p$	$I_p$	$E_p$	$I_p$
0	4.75	10		
2	10.5	9.1		
4	14.3	8		
6	17.8	6.9		
8	21.9	5.8		
10	24.7	4.8		

12/20 - 1/5/57 Check out of new RCA instruments for gas-flow proportional counter. "Bread boarded" timer circuit. Refer to file 4096 scale. Dr. Ben Hermann, B. G. Motors -

1/9/57 Received 0-10 mv. meter from Assembly Products. This to be compared with 0-100 microampere movement for use in trip circuits of DC amplifiers in 202.

Meter internal resistance = 40  $\Omega$ , checked value as follows:



WITH SERIES RESISTANCE at 5770  $\Omega$ , meter reading 10mv, placed resistor box across meter and adjusted so meter read 5mv. Value of resistor box is equal that of meter.  $R = 40 \Omega$ .

Panel meter of DC-2 type amplifier has resistance somewhat less than 100  $\Omega$ . Can insert "test" meter in place of panel meter and provide necessary series resistance for total "meter" resistance of 100  $\Omega$ . Meter should then work with divider-range selector on instrument panel.

Disconnected panel meter. Its series resistor was 33  $\Omega$ . Added in series 27  $\Omega$  for total series resistance of 60  $\Omega$ .

{ Sets on X1 range. Background recorder X1 range. Ext meter (p117-118) range (10) " meter 1, range X1 meter 2, range X1 (10) recorder, range X1 } small source plastic black


2-3 mv.
4.5-7.5 mv.
4-6 $\mu$ a.
2.5-3.5 mv.
4.5-6.5 $\mu$ a.
5.5-9.0 mv.

1/9/57 10mg. Ra w 6' from shell of iron-chamber  
 m-1 (x1) 6.2 - 6.8 mv.  
 m-2 (x1) (x10) 14-15 pa  
 R (x2) 7-8.5 ma  
 m-1 (x1) (x1) 9-10 x1 m-2 (x1) (x10) 20-22 (x1) (x10) 4-4.8 (x5)  
 Ra source @ 10' 12:25  
 With zero adj set Recorder runs full scale on x100  
 max. recorder signal 86 x50 91 x10 9.3 (100)  
 (82 x50) (45) (20)

Have been able to operate up another factor of 10 by attenuating input to recorder x10. ~~This is~~ Assuming this the case M-1 would read 8.6 on x500 and have another factor of 2 to go before the max.  
 M-2 would read 91 (x40) - would have factor of 10 to go before the max.

1/10/57 Recorder chart showed steady decrease in level beginning at 4:00 AM 1/10 dropping to 20-28 (from 39-47) at 8:00 AM. Had been steady from 6 PM (1/9) to 4:00 AM (1/10).  
 Set recorder @ mark @ 10:00 AM. reading (20-27) scale (x5)  
 m-1 reading 45-5.0 mv (x1)  
 m-2 reading 10-12 pa scale (x1)

1/10/57 Circuit as under test 3



Voltage signal for full scale reading at various ranges

	M-1	M-2	R
x1	1.025	1.00	.010
x2	.050	1.200	.020
x5	.125	1.500	.050
x10	.250	1.100	.100
x20	.500	2.000	.200
x50	1.25	5.00	.500
x100	2.50	10.00	1.00
x200	5.00	20.00	2.00
x500	12.50	50.00	5.00
x1000	25.00	100.00	10.00

Cathode current, output tube for various values E.

E <sub>0</sub>	0	4	6	8	10	12	14	16	18	20	22
I <sub>c</sub>	1200pa	1000	900	800	700	600	500	400	300	200	100

For ranges as now limit M-1 = x500 M-2 = x200 R = x1000 (only go to 100)



128

## Trip Circuits DC-2 type

1/10/57 To check saturation of amplifier:

10mg Rv against chamber-

M2 &amp; R "off" m-1 reads 35.5 (x1000) &amp; 70 (x500)

a Switch m2 through scales until m-1 drops-

Scale	m2 $\mu$ a	Cb/c C.
1000	6.5	6.5
500	14.0	7.0
200	37.3	7.46
100	74.5 (cons)	7.45
50	149	
20	380	
10	760	$\leftrightarrow$ here m-1 drops to 68
5		
2		
1		

If m-1 reads 7mv on x500  $I_{m-1} = \frac{7}{10} \times 250 = 175 \mu$ a

$$E_0 = 0.7 \times 12.50 = 8.75v$$

$$I_{c2} = 762.5 \mu$$

$$I_s = 700 + 175 = 875$$

b Switch recorder selector

Scale	m2 $\mu$ a	amm 8.00v	amm 8.75v	M-2 reads $\mu$ a
1000	32			
500	64			
200	160			
100	320			
50	640			
20	1600			
10		8.75 volts		(970 $\mu$ a) 4 $\mu$ a
5		5.4 volts		(1110 $\mu$ a) 2 $\mu$ a
2		3.2 volts		
1				


Measured  $E_0 = 47 = 89.3$  volts

1/10/57

continuation -

Set ion chamber on paraffin pile


Set sources as shown



each layer  $\approx 1\frac{1}{2}$ "

		M-1	M-2	R
		x5	x1	x10
PB 307	$\frac{2.67 \times 10^4}{(2.8 \times 10^4)}$	6.4-7.0	70-76	7.0-7.6
		x50	x10	x100
PN 57	$\frac{3.4 \times 10^5}{(3.4 \times 10^5)}$	7.2-7.4	75-78	7.7-8.0
		x500	x100	
PN 124	$(3.2 \times 10^6)$	5.4-5.5	57	

Relocate - chamber



	M-1	R
PB 307	8 (x10)	9 (x20)
PN 57	8.2 (x100)	
PN 124	5.8 (x1000)	

④ PN 124  $\times 1000$  2.6  
add PN 57  $\times 100$  3.0

PN 57 increases by 4 or 4 x 100

PN 57 alone .94 (x1000)

bkg .05

.89 x 1000 or 8.9 x 100

Value of input circuit  $4 \times 10^{10}$

Nell Superstructure -

1/11/57 D.F.C. drawing to re-mount old superstructure over Nell tank.  
APPR rig is removed.

Reference to print C-10551 "Elementary Diagram Control Apparatus -

		SID # NELL "		Motor 2 (BLUE)		Motor 3 (RED)	
Motor 1:	H1	6	V1	5	S1	5	
Motor	H2	7	V2	8	S2	8	
	H3	12	V3	7	S3	7	
h.s. common	B1A	3	B1A	3	B1A	3	to M-239 tied in frame
h.s. Raised NC	H7	2	V7	2	S7	2	
h.s. Raised NO	H9	9	V9	9	S9	9	
h.s. lowered NC	H8	4	V8	4	S8	4	
h.s. lowered DO	H10	11	V10	11	S10	11	
h.s. Red Down to		10	V11	10		10	
DC Magnet	AAAA(N 5)		V14	12	(S13)	12	
DC Magnet		8	V15	6	(S14)	6	
GND		1	<del>GND</del>	1	<del>GND</del>	1	
Sm 1a	A		Sm 1a	A	Sm 1a	8	
N	B		N	B	N	B	
Selsyn	Sm 2a	C	Sm 5a	C	Sm 8a	C	
	Sm 3a	D	Sm 6a	D	Sm 9a	D	
	Sm 4a	E	Sm 7a	E	Sm 10a	E	

Wiring OK.

Source Motor:	T15	F	1	KSR NO	B9	C
	T16	G	1	LSL NC	B8	D
	N	B	1	LSH NO	B10	A
	h.s. Common(N)	H	1			
	h.s. R NC	B7	F			

1/11/57 WITH VGH. ran cables from terminal box at base of APR5 test stand to P.W. control and source and drive motor assemblies. Mounted source assembly, not drive. The cable tray had been checked earlier & OK'd. Assumed OK. Now after VGH repaired break on center limit switch source.

1/11/57 BACK TO TRIP CIRCUITS. RUN SOME DATA ON LINEARITY

By zero adj; Set to read on scale	w/o sources	set on 10% (1005 F; 1005 C)	set 10% (1005 F; 762 C)	set 10% (1005 F; 627 C)	set 10% (1005 F; 589 C)	set 10% (1005 F; 571 C)	set 10% (1005 F; 520 C)
scale	1000	10	5	2	1.4	-	-
	500	20.3	10	4	2.1	-	-
	200	50.2	25	10	5.2	2.5	-
	100	98.8	49	18.8	10	5.1	2
	50		96.5	39	19.8	10	4
	20			95	48.5	25	10
	10				95	49	20
	5					96	39
	2						
	1						

contact bad - - - 93

SOURCE STRENGTH	#	Room	Original	Date	1/10/58
<del>PN 213</del>	PN 213		3.63	6-13-56	1.26 x 10 <sup>7</sup>
<del>PN 214</del>	PN 214		1.36	6-14-56	4.75 x 10 <sup>6</sup>
	PN 124	201	3.62	9-14-55	3.2 x 10 <sup>6</sup>
	PN 123	209	3.56	5-12-55	1.686 x 10 <sup>6</sup>
	PN 58	108	3.74	6-8-56	3.25 x 10 <sup>5</sup>
	PN 57 (A)	209	3.70	6-24-54	3.49 x 10 <sup>5</sup>
	PN 307	209	3.35	2-16-53	2.67 x 10 <sup>4</sup>
	PN 267	201	2.87	6-9-52	6.49 x 10 <sup>3</sup>

## EXPERIMENTAL PLANNING-

1-11-57 Comments from ADC on forth coming project.  
Some of needed instrumentation considerations - pulse oscillator measurement of DR to  $10^{-6}$   
Rod movement presently in "ball park" what about instrument sensitivity.

1-14-57 Consideration of instrumentation for new test cell. - delayed  
Check out superstructure over Nell.

- "Blue" motor OK. in direction of limit switches
- "Red" motor limit switches OK, direction reversed.
- a. At control panel, 201: reversed S1 & S2.
- b. Selwyn indicators reversed in direction.

In panel 1, on console 202:

swapped Sm 6 & Sm 7 black & orange

swapped Sm 9 & Sm 10 red & yellow

True, of electrical engineering, over to talk w/dg. modifications.

more on Nell:

- "Elephant Gun" requested. Began making up cable for tying in to motor #1 (Yellow) connector on panel 201.

Discussion w/ ADC on office space for instrument section.

1/15/57

0800-1100  
1200-1400

Completed wiring for "Elephant Gun" wired in & checked out.  
Cable got caught in gears on top Motor 2 - repaired cable.  
Magnet supply operation faulty, Observed two seemingly bad rectifier sections. VGH replaced. Now OK.  
For detail refer to file on 201 Control Circuit Modification.

Consideration to new spaces - office for Jim & Vic.

1/16/57

Fader - problem of noise bursts from RMS on top assembly.  
Checked socket in 210, w/ 1000 volts of no 6292 OK  
When tube in no change in divider voltages.  
With scope no apparent noise pulses - source pulse normal.

Returned to test area - seems OK.

202 - Modified magnet circuit so "ROD DROP" switch opens magnet supply line. Refer to file.

ADC - reference to request for circuit information from UK.  
Item of interest - scintillation trip circuit.

Drew sketch of circuit, including notes on operation.  
East end instrument used as guide.

DWRM - 202. Paper not loud enough when used with C4.  
Mounted speaker used with C3 on console near intercom speaker.

1/17/57 AM - Hanover. Preliminary discussion new project.  
 PM - Summary of Present Instrumentation East End  
 1/18/57 AM - Summary of Present Instrumentation West End.

10:15 DC-2 Trip & Ion Chamber Considerations -

Test setup: Meter #1 zeroed on scale 2  $R = 200 \Omega$   
 Meter #2 zeroed on scale 1  $R = 1000 \Omega$   
 Recorder zeroed on scale 4  $R = 12.5 K \Omega$

Source #307 @ 1 foot = 30.5 cm.

M1 = 65 mv.	$E_0 = 27.5 \text{ mv.}$
M2 = 27 $\mu$ a	$E_0 = 27. \text{ mv.}$
R = 3.6 mv	$E_0 = 28. \text{ mv.}$

$R_i = 4 \times 10^{10} \Omega$   $I_i = \frac{28 \times 10^{-3}}{4 \times 10^{10}} = 7 \times 10^{-13}$  amps.

$N = 2.6 \times 10^4$  n/cu.

$$\phi = \frac{N}{A} = \frac{2.6 \times 10^4}{4\pi r^2} = \frac{2.6 \times 10^4}{1.257 \times 10 \times 960 \times 10^3} = \frac{2.6}{1.257 \times 960} = 2.08 \text{ n/cm}^2/\mu$$

Source PMS7 @ 50" approx. same readings  $\frac{2.5 \times 50}{12700} \text{ cm}$

$N = 3.4 \times 10^5$

$$\phi = \frac{N}{A} = \frac{3.4 \times 10^5 \text{ n/cu}}{1.257 \times 10 \times (1.27 \times 10^2)^2} = \frac{3.4 \times 10^5}{1.257 \times 11614 \times 10^5} = 1.7 \text{ n/cm}^2/\mu$$

$$\frac{i}{\phi} = \frac{7 \times 10^{-13}}{2} = 3.5 \times 10^{-13} \text{ amp/unit flux}$$

1/18/57 PMS7 @ 1'  $N = 3.4 \times 10^5$

$$\phi = \frac{3.4 \times 10^5}{1.257 \times 10 \times 960 \times 10^3} = \frac{3.4 \times 10^5}{1.206 \times 10^4} = 2.82 \times 10^{-2} = 28 \text{ n/cm}^2/\mu$$

M1 = (50) 29 mv	$R_T$ (5000 $\Omega$ )
M2 = (5) 64 $\mu$ a	(5000 $\Omega$ )
R = (50) 68 mv	(125 K $\Omega$ )

M1:  $I_m = \frac{2.9 \times 10^{-3}}{4 \times 10^2} = 7.25 \times 10^{-5} = 72.5 \mu\text{a}$   $E_0 = 72.5 \times 10^{-6} \times 5 \times 10^3 = 362.5 \text{ mv.}$

M2:  $I = 64 \mu\text{a}$   $E_0 = 64 \times 10^{-6} \times 5 \times 10^3 = 320 \text{ mv.}$

R:  $I_R = \frac{6.8 \times 10^{-3}}{2.5 \times 10^3} = 2.72 \times 10^{-6}$   $E_0 = 2.72 \times 10^{-6} \times 125 \times 10^3 = 340 \text{ mv.}$

$E_0 = 340 \text{ mv}$   $I_c = \frac{340 \times 10^{-2}}{4 \times 10^{10}} = 8.5 \times 10^{-12}$

$\frac{i}{\phi} = \frac{8.5 \times 10^{-12}}{2.8 \times 10^{-2}} = 3.04 \times 10^{-13} \text{ amp/unit flux}$

PMS7 @ 6"

M1 = (100) 35	R	10 K	Note drop in read.
M2 = (10) 79		10 K	
M3 = (100) 8.0		250 K	

M1:  $3.5 \times 10^3 / 4 \times 10^2 = 8.75 \times 10^{-5} = 87.5 \mu\text{a}$   $E_0 = 87.5 \times 10^{-6} \times 10^4 = 875 \text{ mv.}$

M2:  $I = 79 \mu\text{a}$   $E_0 = 79 \times 10^{-6} \times 10^4 = 790 \text{ mv.}$

R:  $I = 8 \times 10^3 / 2.5 \times 10^3 = 3.2 \times 10^{-6}$   $E_0 = 3.2 \times 10^{-6} \times 250 \times 10^3 = 800 \text{ mv.}$   
 $I_c = 8.30 \times 10^{-12} / 4 \times 10^{10} = 2.07 \times 10^{-11}$



$$\phi = \frac{3.4 \times 10^5}{1.257 \times 10 \times 232 \times 10^2} = \frac{3.4 \times 10^5}{2.92 \times 10^3} = 1.16 \times 10^2 = 116 \text{ n/cm}^2/\mu$$

Am. Hawman -  
DC-2 Type Trng - Output -

1/21/57

Expanded scale operation -

Circuit:



Made rough check with expander in M-1 circuit -  
did not, but - needle of meter is very quippy - Try in M-2  
A: With expander in M-2 circuit, between Y & Y' (short removed)  
made set of readings to see amount of expansion - Varying  
zero adj on DC amp. Data on attached sheet =>

1/22/57 Continuation of 1-21-57

Brought in 2ms. Brown Recorder from 202. Attached to recorder

output. Helipot still in M-2 circuit. ~~Scale~~

amp setting	Scale	M-1 (calc.) meter	$E_0$	Scale <del>Scale</del>	M-2 Helipot Meter	(calc) $E_0$	Scale	Recorder (calc) Reading	$E_0$
5.9.58	5	10mv	125mv	2	0.00	58µa	100	101mv	105ms.
5.77									


Calculate network for 2-ms. Brown -

10ms Brown unit  $R = 2500 \Omega$  @ 10mv  $I_B = \frac{10^{-2}}{.25 \times 10^3} = 4 \mu a$ .

For 2-ms. - using units  $R = 2000 \Omega$

$I_{B1} = \frac{10^{-2}}{.2 \times 10^3} = 5 \mu a$  -

Try:



$I_B = \frac{10^{-2}}{.1 \times 10^3} = 10 \mu a$

Removed M-1 from circuit (0-10ms API)

Put panel meter back in service.

M-2 used with Brown 2ms. recorder.

Grid expanded on Brown circuit -

1/21/57	DC-AMP	M-1		M-2		(1000 $\mu$ a) Meter	$E_{0m1}$ ↓	$E_{0m2}$ ↓	$R_{eff}$	$E_R$		
	Zero-Adj.	Scale	meter	Scale	Helipot							
	5.585 830	x1	1 mv	x1	000	2 $\mu$ a	2.5mv	2mv				
	5.585	x1	4.4 mv	x1	000	10 $\mu$ a	11mv	10mv				
	5.610	x1	10 mv	x1	000	25 $\mu$ a	25mv	25mv				
	"	x2	5.2 mv	"	"	"	26mv	25mv				
	5.650	x2	10 mv	x1	000	48 $\mu$ a	50mv	48mv				
	"	x5	4.2 mv	"	"	"	52.5	48mv				
	5.745	x5	9.0 mv	x1	"	100 $\mu$ a	112.5mv	100mv				
	"	x10	4.5 mv	"	"	"	112.5mv	100mv				
	"	x10	4.5 mv	x1	0.74	10 $\mu$ a	112.5mv		73 $\Omega$	.098 v	100 mv	
	5.915	x10	8.9 mv	x1	"	100	222.5mv					
	"	x20	4.45 mv	x1	"	"	222.5mv					
	"	x20	4.45 mv	x1	1.51	10 $\mu$ a	222.5mv		148 $\Omega$	.198	200 mv	
	6.10	x20	6.8 mv	x1	1.51	100 $\mu$ a	340mv					
	6.10	x20	6.8 mv	x1	2.32	10 $\mu$ a	340mv		231 $\Omega$	.296	300 mv	
	6.30	x20	9.15 mv	x1	2.32	100 $\mu$ a	457.5mv					
	6.30	x50	3.65 mv	x1	2.32	100 $\mu$ a	456.1mv					
	6.30	x50	3.65 mv	x1	3.19	10 $\mu$ a	456 mv		318	.405	400 mv	
	6.52	x50	4.65 mv	x1	3.19	100 $\mu$ a	582mv					
	"	"	"	x1	4.06	10 $\mu$ a	582mv					
	6.76	x50	<del>5.7</del> 5.7	x1	4.06	100 $\mu$ a	713mv		403	.575		
	6.76	x50	5.7	x1	4.94	10 $\mu$ a	713mv					
	6.895	x50	6.7	x1	4.94	100 $\mu$ a	838mv		491	.627		
	6.995	x50	6.7	x1	5.855	10 $\mu$ a	838mv					
	7.245	x50	7.7	x1	5.855	100 $\mu$ a	967mv		582	.745		
	7.245	x50	7.7	x1	6.755	10 $\mu$ a	967mv					
	7.50	x50	8.7	x1	6.755	100 $\mu$ a	1.088v		675	.855		
	7.50	x100	4.35	x1	6.755	100 $\mu$ a	1.088v					
	7.50	x100	4.35	x1	7.625	10 $\mu$ a	1.088v					
	7.745	x100	4.85	x1	7.625	100 $\mu$ a	1.212v		762	.962		
	7.745	x100	4.85	x1	8.495	10 $\mu$ a	1.212v					
	7.975	x100	5.30	x1	"	100 $\mu$ a	1.325v		849	1.09		
	7.975	x100	5.30	x1	9.32	10 $\mu$ a	1.325v					
	8.19	x100	5.75	x1	9.32	100 $\mu$ a	1.438v		932	1.20		



1/21/57

Calculation  $E_0$  -

$$M-1: E_{fs} = 10 \text{ mv}, R_m = 40 \Omega, I_{fs} = 250 \mu\text{a}$$

$$E_{fs}(\text{unit}) = 25 \text{ mv}, R_{\text{unit}} = 100 \Omega$$

$$\text{Scale } 5 \quad R = 500 \Omega, I = 250 \mu\text{a}, E = 125 \text{ mv}$$

$$M-2: I_{fs} = 100 \mu\text{a}, R_{(1)} = 1000 \Omega$$

$$R_2 = 2000 \Omega, I = 58 \mu\text{a}$$

$$E_0 = 10^{-4} \times 2 \times 10^3 = 200 \text{ v}$$

$$E_0 = 58 \times 10^{-6} \times 2 \times 10^3 = 116 \times 10^{-3}$$

R :

$$E_{fs} = 2 \text{ mv}$$

I :

$$E = 1.05 \text{ mv}$$

$$R_{(1)} = 2.5 \text{ k} \text{ (recorder/res)}$$

$$R = 2.5 \text{ k}\Omega$$


$$R_{(200)} = 250 \text{ k}$$

$$\frac{1.05 \times 10^{-3}}{2.5 \times 10^3} = .42 \times 10^{-6}$$

$$E_0 = .42 \times 10^{-6} \times 250 \times 10^3 = 105$$


1/23/57 Am. Standard - looked @ small recorders - Varian - Bristol - Westronics - L&N.

pm Re-created test rig for handie manipulation  
Brown Recorder @ 100  $\mu$ a meter on output



DC amplifier panel meter - 200  $\mu$ a fs. with  $R_{in} = 100 \Omega$  20 ms fs.  
 Microammeter (100  $\mu$ a) - 100  $\mu$ a fs. with  $R_{in} = 1000 \Omega$  100 ms fs.  
 Recorder (2 mV) - 2  $\mu$ a fs. with  $R_{in} = 1000 \Omega$  2 ms fs.  
 Multirange meter (10 mV) 250  $\mu$ a fs. with  $R_{in} = 100 \Omega$  25 ms fs.  
 with  $R_{in} = 50 \Omega$  12.5 ms fs.

Expander in Brown circuit



DC AMP	Paul meter			Microammeter			Recorder			
Zero Adj	Scale	meter	E <sub>0</sub>	Scale	meter	E <sub>0</sub>	Scale	Wdg pot	Reading	E <sub>R</sub>
5.675	5 (500)	35 $\mu$ a	175 mV	1 (1000)	20 $\mu$ a	20 mV	50 (510%)	0.000	19.5%	19.5 mV
5.795	5 (100 mV)	20 $\mu$ a	20 mV	1 (100 mV)	91 $\mu$ a	91 mV		61	95%	95 mV
"	10 (200 mV)	88 $\mu$ a	88 mV	2 (200 mV)	45 $\mu$ a	90 mV		0.610	5%	(90) 5 mV
not recorded		177	177 mV	40%	85 $\mu$ a	170 mV		73	25%	(100) 95 mV
	20 (200 mV)	95	170 mV	5 (300 mV)	37 $\mu$ a	185 mV		1.34	48	(100) 4 mV
6.090	20 (200 mV)	130	260 mV	10%	53 $\mu$ a	265 mV		67	95.5	(200) 95.5 mV
6.240	100 (2 ufs)	25 $\mu$ a	250 mV	20 (2 ufs)	15 $\mu$ a	300 mV		2.01	5%	(27.5) 5 mV
		33 $\mu$ a	330 mV	30	18.5 $\mu$ a	370 mV		67	94.5	(366) 94.5
6.400		41	410 mV		23 $\mu$ a	460 mV		2.68	4.5	(361) 4.5
6.560		50	500 mV		27 $\mu$ a	540 mV		65.5	94.5	(405) 94.5
		58	580		31	620 mV		333.5	6.0	(449) 6
6.730		66	660		35.4	708 mV		67	96	(545) 96
6.900		74	740		40	800 mV		4.01	5.5	(570) 5.5
7.08		83	830		44	880 mV		67	94.5	(625) 94.5
7.27								4.67	6	(629) 6
								67	96	(725) 96
								5.24	5.5	(720) 5.5
								660	95	(815) 95
								600.5	5.5	(80) 5.5
								6.67	95.5	(905.5) 95.5
									5	

Keep records in scale 50 as 100 mV = f.s.

stop - VSA changing in source.

~~1/24/57~~ 1/24/57 Work in 107-108 - wiring additions to control cabinets -

1/28/57 AM East End

PM West End - Cabinets 1 & 2 - repair.

1/28/57 Instrumentation Considerations - New Jersey -


AM - w/ Hanauer discussion of control console layout

PM - instrumentation needs listed & approx. prices set.

1/29/57

Drop Circuit Considerations -

Supply lead 0-10 mv - has internal resistance of 40 ohms.  
 Try something like the following:



(no expansion considerations here)

2/2/57 (12-8) Check over record of Jan & Feb. No outstanding instrument problem.

Chemists in 211 reported poor gas supply - Exterior tank ok. Drop in lab "flooded."

Check at 0200 - all apparently normal.

→ Checked over Hammer's console order suggestion & put in office for Dr. Callahan.

→ Received comment prints from June - electrical tie-in of new wing to present structure.

Reviewed prints. One question "Spac" conduit up from 103 thru 205.


Wiring divider, shown p 140

2/5/57 (12-8) Conf. w/ADC re console order -

Very rubicon potentiometer w/ thermocouple to check flow from enricher to swing. - check on how running.


Brought RCA Handbook up to date

TRIP ⇒ <sup>0140</sup> Put resistance divider across 0-10mv output to drop Recorder signal to 0-2 mv. range.



2/5/57 (12-8)

Pen, ink, clogging. Put back in ball point circuit.



With 0-10 millivolt meter in divider shown above of recorder attached at points noted - much oscillation on large changes. Meter undamped. This seems to affect recorder.

Disconnect meter, substituting 40Ω resistor.

Source by Chamber

Recorder does not "hunt".

M2	26µA x50	26µA x50	26µA x50	same	63µA x20
M2	OUT	1/60Ω shunt	X200	2.3mV	5.5mV x200
R	65% x200	32% x200	70% x200	28% x500	69% x200
IM	120µA x100	120µA x100	120µA x100	same	120µA x100

Removed Source IM out

M2	16µA @ x1	
M1	1.5mV @ x10	19%
	4.2mV @ x3	54%
	5.1mV x1	25%

Recorder off 0270

M1 on x1 upsets M2 & Recorder

2/6/57

0010 Check Amplifier


M2 = 8 µA (x1)

M1 = .9 mv (x10)

(Recorder off)

Calculations from readings 2/5/57:

METER	SCALE	TOTAL RESISTANCE	CURRENT	METER VOLTS	METER RESISTANCE	E <sub>o</sub> volts	I <sub>R</sub> µA	E <sub>50Ω</sub> R- 5.89 x10 <sup>3</sup>
IM	x100	24mS (200) 10,000	1.2 x 10 <sup>-4</sup>	-	-	1.2	-	-
R	x200	10 <sup>4</sup>	127	.0013	1.2 x 10 <sup>3</sup>	1.27	1.065	6.37mV
M1	-	-	-	-	-	-	-	-
M2	x50	5 x 10 <sup>4</sup>	26	-	-	1.30	-	-
IM	x100	10 <sup>4</sup>	1.2 x 10 <sup>-4</sup>	-	-	1.2	-	-
R	x200	10 <sup>4</sup>	132	.00069	1.22 x 10 <sup>3</sup>	1.32	.557	3.23 (200) 4.97
M1	x200	10 <sup>4</sup>	136	.00275	40Ω	1.36	-	3.44
M2	x50	5 x 10 <sup>4</sup>	26	-	-	1.30	-	-
IM	x100	10 <sup>4</sup>	1.2 x 10 <sup>-4</sup>	-	-	1.2	-	-
R	x200	10 <sup>4</sup>	-	1.4mV	1.22 x 10 <sup>3</sup>	-	1.15 x 10 <sup>-4</sup>	6.85mV
M1	x200	10 <sup>4</sup>	136	5.45mV	40Ω	1.36	-	6.81mV
M2	x50	5 x 10 <sup>4</sup>	26	-	-	1.30	-	-
IM	x100	10 <sup>4</sup>	1.2 x 10 <sup>-4</sup>	-	-	1.20	-	-
R	x500	2.5 x 10 <sup>4</sup>	-	.54 x 10 <sup>-3</sup>	1.22 x 10 <sup>3</sup>	-	.443 x 10 <sup>-4</sup>	2.64 x 10 <sup>-3</sup>
M1	x500	2.5 x 10 <sup>4</sup>	57.5µA	2.3 x 10 <sup>-3</sup>	40Ω	1.43	-	2.88 x 10 <sup>-3</sup>
M2	x50	5 x 10 <sup>4</sup>	26	-	-	1.30	-	-
IM	x100	10 <sup>4</sup>	1.20 x 10 <sup>-4</sup>	-	-	1.20	-	-
R	x200	10 <sup>4</sup>	-	1.30mV	-	-	1.17 x 10 <sup>-4</sup>	6.76
M1	x200	10 <sup>4</sup>	137	5.5 x 10 <sup>-3</sup>	40	1.37	-	6.88
M2	x20	2 x 10 <sup>6</sup>	63	-	-	1.26	-	-



2/5/57  
 Range change  
 oscillation  
 Am-1 & Recorder

2/2/57 Rearranged components of test - putting amp. chassis on holder so chassis vertical. Allow time for stabilizing (00%)

While waiting for instrument to warm-up. Made up 7 foot power cable. Always need extensions.

Used stepping switch to operate 0-10 mv. meter & indicator lights. Made separate selenium-rectifier power supply. Noted interaction on panel meter on range changes by means of stepping switch.

(Summary of types of dividers)

2/8/57

More re-arrangement -

added 0-100  $\mu$ a meter - interaction sometimes present - not extremely serious.

added recorder for same leg as millivoltmeter

large amplitude hunting, due to mv. meter. Need to separate recorder & mv. meter dividers

Source in place - Panel meter 105 (100)

0-100  $\mu$ a 59  $\mu$ a (x20)

0-10 mv. 4.9 mv (7) (x200)

Recorder 59-63%

Put 51  $\Omega$  resistance across 50  $\Omega$  meter branch.

Meter - from 4.9 mv to 2.65 mv

Recorder from 59-63% to 29%

to range 8 -

back to range 6 (x100) meter 4.9 mv  
 recorder 60-61 greatly reduces hunting


$E_0^-$

1.05 volt

1.14 volt


1.22 volt

2/8/57 Disturbance meter - recorder off-scale range 6  
 To range 7 - Recorder reads 58% no spread.  
 Without voltmeter recorder reads as one would expect.  
 Try recorder on separate divider -




if use 1 k as unit  
 $\frac{10 \times 10^{-3} \text{ V}}{10^3} = 10 \mu\text{a}$

Re-arrange mv. divider so:






2/12/57 DC-2 Type  
Summary of arrangements tested:




(A) PANEL METER, AS BUILT:




Scale:	1	2	3	5	10	20	50	100	200	500	1000	2000
$R_1(\Omega)$	$10^2$	$2 \times 10^2$	$5 \times 10^2$	$10^3$	$2 \times 10^3$	$5 \times 10^3$	$10^4$	$2 \times 10^4$	$5 \times 10^4$	$10^5$		
$E_0(V)$	.02	.04	.10	.20	.40	1.0	2.0	4.0	10	20		

(B) RECORDER, AS USED




Scale:	1	2	3	5	10	20	50	100	200	500	1000
$R_1(\Omega)$	$2.5 \times 10^3$	$5 \times 10^3$	$1 \times 10^4$	$2 \times 10^4$	$5 \times 10^4$	$1 \times 10^5$	$2 \times 10^5$	$5 \times 10^5$	$1 \times 10^6$		
$E_0(V)$	.01	.02	.05	.10	.20	.50	1.0	2.0	5.0	10	

(C) MICROAMMETER, similar to (A)




Scale:	1	2	3	5	10	20	50	100	200	500	1000
$R_1(\Omega)$	$10^3$	$2 \times 10^3$	$5 \times 10^3$	$10^4$	$2 \times 10^4$	$5 \times 10^4$	$10^5$	$2 \times 10^5$	$5 \times 10^5$	$10^6$	
$E_0(V)$	.100	.200	.500	1.0	2.0	5.0	10	20	50	100	

(D) MILLIVOLTMETER, similar to (A)



Scale:	1	2	3	5	10	20	50	100	200	500	1000
$R_1(\Omega)$	$10^2$	$2 \times 10^2$	$5 \times 10^2$	$10^3$	$2 \times 10^3$	$5 \times 10^3$	$10^4$	$2 \times 10^4$	$5 \times 10^4$	$10^5$	
$E_0$	.025	.05	.125	.25	.5	1.25	2.5	5.0	12.5	25.0	


(E) RECORDER



Scale:	1	2	3	5	10	20	50	100	200	500	1000
$R_1(\Omega)$	$10^3$	$2 \times 10^3$	$5 \times 10^3$	$10^4$	$2 \times 10^4$	$5 \times 10^4$	$10^5$	$2 \times 10^5$	$5 \times 10^5$	$10^6$	
$E_0$	.002	.004	.010	.02	.04	.10	.20	.40	1.0	2.0	


→ added decade possible here

2/9/57 (F) MILLIVOLTMETER & RECORDER:




Scale:	1	2	3	5	10	20	50	100	200	500	1000	2000
$R_1(\Omega)$	50	150	500	1000	2500	5000	$5 \times 10^3$	$10^4$	$2.5 \times 10^4$	$5 \times 10^4$	$10^5$	
$E_0(V)$	.0125	.0375	.125	.25	.625	1.25	2.5	6.25	12.5	31.25	62.5	125

(G) MILLIVOLTMETER, as in (A)



Scale:	1	2	3	5	10	20	50	100	200	500	1000	2000
$R_1(\Omega)$	50	150	500	1000	2500	5000	$5 \times 10^3$	$10^4$	$2.5 \times 10^4$	$5 \times 10^4$	$10^5$	
$E_0(V)$	.0125	.0375	.125	.25	.625	1.25	2.5	6.25	12.5	31.25	62.5	125

(H) RECORDER, as (F) but NO METER



Scale:	(1)	(2)	(3)	(4)	(5)	(6)	(7)	(8)	(9)	INDICATE KAMPS
$R_1(\Omega)$	$10^3$	$2 \times 10^3$	$10^4$	$2 \times 10^4$	$5 \times 10^4$	$10^5$	$2 \times 10^5$	$5 \times 10^5$	$10^6$	
$E_0(V)$	.010	.03	.10	.20	.50	1.0	2.0	5.0	10.0	

AT PRESENT (A) can be used

(C) is WIRED

(G) is WIRED TO STEP SWITCH

(H) is WIRED TO STEP SWITCH

500  $\mu$ A  
100  $\mu$ A  
250  $\mu$ A  
10  $\mu$ A

2/9/57  
0048

Run series of readings with Panel Meter out of circuit,  
Microammeter, Millivoltmeter, and Recorder

(C) (S) (H)

Vary Zero on Amp & record readings of meters - (leave fine zero at 4.0)

- \* Pen on scale 1 chatters slightly, also on scale 2
- \* Before coming to 3 1/2 pen had gone off scale negative -
- after reading - 3 1/2 pen erratic -
- † with source added

noted that recorder did not follow meters -  
when millivoltmeter < 10% recorder on zero.


Calculation E<sub>0</sub>.

µamp - sample on scale 1  $B = 10^3$

if  $I = 2 \mu a$   $E = 2 \times 10^{-6} \times 10^3 = 2 \text{ mv}$

on range multiply by factor.


mv. :



meter reading  $\times 1.25 =$  unit reading.  
then use scale factor

if  $E_m = 1 \text{ mv}$       1.25 mv

Change recorder divider



This helps recorder readings

→ Placing 0.1 capacitor across coil cut out large kick on range changes



2/12/57 - Step switch contacts seem to be loose on several positions of millivoltmeter range. This is switch played with before use in divider.  
 Loose joints on recorder divider - Resoldered joints  
 Re-wired divider for millivoltmeter

2/12/57. Work on counters in Rm. 202.

2/14/57 Construction Inspection -  
 West End Instrumentation

2/15/57 Trip - Step Switch - loose contact on various positions

2/18-20/57 Ordering new equipment

2/20-21/57 Linear Amplifier from 107 "Y-90319" see file

2/22/57 Log Count Rate Meter (see file)

2/23/57 Linear Amps Y-90319 (see file)

Scaler 489641 from bench to tall rack

DC-2 to rack from bench

2/25-2/1 Log Count Rate Meter - (refer file)

3/4 Log Count Rate Meter file up to date

Linear Amplifier - Pre-amps - check out for gamma counter

To do - test counters 5-243; 8-304

[5-188-2  
 in 201]

Data to be recorded in file to determine sensitivity & operating characteristics.

3/5-7 Calibration and checking linear amplifiers.

3/8 Nell motor one re-installation.

3-11-57 Preparation for microvalve pressure chamber calibrators.


4/8/57 Keithley 410 check out - refer files.

4/9/57 Output meter for Keithley

$$V_0 = 5v.$$

$$I_0 = 5 \times 10^{-3} \text{ A} \leftarrow$$

$$R_0 = \frac{5}{5} \times 10^3 = 10^3 \Omega.$$



$$I = 250 \times 10^{-6} \text{ A} \leftarrow \frac{5}{.25} \times 10^3 = 20 \times 10^3 \Omega.$$

millivoltmeter - meter + 20,000 Ω - undamped - reading slightly > panel meter. Damped with shunt of 500 Ω - reading as panel meter. Workable (Prepared for DC-2 Type)


Microammeter - meter + 50,000 Ω - reads as panel meter at

This not fitted as DC-2 type of unit and. Ref to p. 196  
 $10^{-4} \text{ A} \text{ \& } 10^3 \Omega$  min. scale 100 mV fs.


4-11-57 Recorder signal divider for Keithley -

2. 0-10 mV.

Output, so far



Recorder




if run 2ma fs.

$$R_2 = \frac{5}{2} \times 10^3 = 2.5 \text{ k}\Omega$$

$$R_{10mV} = \frac{10 \times 10^{-3}}{2 \times 10^{-3}} = 5 \Omega \text{ too low.}$$


want  $R_{10mV} \approx 10^3$   $I = \frac{10 \times 10^{-2}}{10^3} = 10 \times 10^{-5} = 100 \mu\text{A}$



when meter is  $4 \times 10^{-4}$ , 40µA,  
 For 10mV read 7.5 mV.

4/14/57

Output - Keithley



halted for fusion chamber check -

C4, 202 had CS-3 &amp; Pre-amp 349

C5, 202 had CS-2 &amp; Pre-amp 349

20 channel had sharp high peak on C5.

This still present when CS-3 used w/ pre-amp 349

Is it pre-amp trouble? Refer file on pre-amps for notes.

5/1/57

Conf. DWM - Controls for new installation

Conf. OFC - 1/2 inch BF<sub>3</sub> proportional counter as normalizer in sub reg.

Check out Ion Chamber Simulator on Keithley 410-180.

Testing Beckman Model V microammeter.

5/2/57

Continuation of Beckman Model V. Trouble corrected 1020

Beckman Mod V, Keithley 128, &amp; Keithley, all w/ chambers to monitor

Robert End - D. Scott raised objection to Channel E modification

5/3/57

Talked w/ JFE regarding instrument modifications.

Beckman - drew up circuit of input. Monitor west end.

Call re. counting channel 4 - seems OK.

Run sensitivity curves on Linear Amp Y90519-

Data on RCL Counters 371 8/10/58 -

5/6/57

Continuation of data collection for RCL counters - Blought in number 369 and 1016 from 108. Refer file on counters

Re-wiring of AIC 205B pre-amplifiers for better interchangeability.

Continuation of operation checks on Beckman Model V &amp; Keithley 410

5/7/57

Further work with RCL counters

20-Channel analyzer resolution problem

5/8/57

More checks on new instruments &amp; trip design.

20-channel analyzer


Linear Amplifier, Bus 202, C5, blown fuse, fixed, replacing 5V4G which was grossy.

5/9/57

20-channel analyzer.

New equipment.

5/9/57 Keithley output circuitry -




meter  $100 \mu a f.s. \approx 10^3 \Omega \Rightarrow E_m = 10^{-4} \times 10^3 = 10^{-1} = 100 mV$   
 Removed shunt - no change in readings or meter action  
 For bench check with 2mv. Brown design shunt.

let  $R_p = 1000 \Omega$   $E_p = 2 \times 10^{-3} v$ .  $I_p = \frac{2 \times 10^{-3}}{10^3} = 2 \times 10^{-6} a$ .  
 $\rightarrow$  if shunt across meter only:  $E_f = 100 mV$ .  $R_f = 50,000 \Omega$   
 use shunt as above.


$\rightarrow$  if shunt across entire output,  
 $E_f = 5V$   $R_f = ?$   
 $\frac{R}{10^3} = \frac{5000}{2}$   $2R = 5 \times 10^6$   $R = 2.5 \times 10^6$

As wired:



5/10/57 Continuation of Trig circuit considerations -

Beckman:



Requirement (from circuit point) 205 μa f.s.,  $< 500 \Omega$ .  
 with meter directly tied as shown


Signal on Beckman meter 90-100%  $3 \times 10^{-12}$   
 Meter dancing between 80-100% - NEEDS TO BE DAMPED

$\rightarrow$  want ext. meter to read less than instrument meter (so that not trip with on-scale instrument reading)

Damping Resistor	0-10mV Beckman
$\infty$	80-100%
470 $\Omega$	75-90
390	70-90
270	75-85
200	X
150	X

Get less sensitive scale.

Damping Res.	Beckman	Ext.	
270	85-90	68-72	looks about ok.
390	"	70-75	



This works very well w/ 2mv. Brown

5/13/57 Further Readings on Beckman & Keithley Instruments -  
 weak  $\eta$ -source - Beckman - 25%  $3 \times 10^{-11}$   $.75 \times 10^{-12}$   $.52 \times 10^{-10}$   $.5 \times 10^{-11}$   
 Keithley - 24% "  $.72 \times 10^{-12}$   $.70 \times 10^{-10}$   $.7 \times 10^{-10}$   $1.0 \times 10^{-10}$

→ Scale Selector Trouble in Beckman on the most sensitive scale  
 PN-57 source Beckman 38% ( $1.0 \times 10^{-11}$ ); 12% ( $2 \times 10^{-10}$ )  
 Keithley 38% ( $1.6 \times 10^{-11}$ ); 12% ( $2 \times 10^{-10}$ )


Servicing C<sub>1</sub> & C<sub>3</sub>

5/14/57 work on several units of C<sub>3</sub> channel  
 Continued mutual observations of Beckman & Keithley

5/15/57 More on C<sub>3</sub>.  
 Received 2 new Weston Recorders -

5/16/57 Completed C<sub>3</sub> -  
 Connected New Recorders to power & signal from  
 Beckman & Keithley amplifiers.

Modified resistance networks for 10 ms. Recorders








5/12/57 Recorder - Micro-Microammeter Calibration						
(BOTH INST)	(%)	BECKMAN (%)	(%)	(%)	KEITHLEY (%)	(%)
RANGE	OUTPUT	RECORDER	TRIP	OUTPUT	RECORDER	TRIP
5x10 <sup>-12</sup>	15-17	Trace 2% 15-17	-	14	narrow trace 14	-
	28-30	Trace with R 2% 28-30	-	26-27	narrow trace 26-27	-
	49.52	Trace 2% 50-53	-	53	narrow trace 53	-
	72-76	Trace 2% 73-78	60	71-72	narrow trace 72-73	60
	adjusted recorder		adjusted recorder of meter			
10x10 <sup>-12</sup>	94-99	Trace 5% 98-99	80	98-10	Trace 1% 98-100	85
	29-30	narrow w/ split 30	35	30	29.5	29
	55-57	1% trim 55-57	45	54-55	53-54	47
	74-76	1% 75-76	62	76-77	75-76	66
	92-94	1% 93-95	77	94-96	94	81
3x10 <sup>-11</sup>	31	mic. 31	38	33	31.5	29
	71	(91.6) 71	59	70	69.2	60
	89-90	mic. 89.5	74-75	85-86	84.8	74
Brought in new fuel						
3x10 <sup>-10</sup>	72	(72.5) 72	60	73	(72.6) 72	63
	figures in ( ) scale indication, shows paper off from scale.					
(Adj. Keithley output to lower meter of same recorder)						
Found that pen holder is adjustable on Weston's recorder. Adj. so paper corresponds to scale.						
72	72	(72.1)	60	73	(71)	60 OK.
Adjust scale on this recorder to get paper of scale the same						
With weak gamma source in position between chambers, compared signals						
Beckman trace 10 div			Keithley trace 1 division			

5/16/57

Keithley output - Network across entire 5v output - cannot get recorder to track - (ref p 104 for diagram) will not track when using net across (0-100  $\mu$ a) meter. So get tracking try recorder in series w/ meter.



Tracking now all right, BUT does not zero. At zero recorder reading 2%.




5/17/57 Plugged in at 9:00 AM - Pens began to operate in 3.15 sec. Brought in gamma source (10mg Ra) to provide signal to calibrate recorders with instruments. Set source at different distances from chambers to obtain varying output readings from instruments. Data to left.


6-3/4 S. Wang wiring.

- 6-5-57 Components yet to be built for South Wing.
- # Scale Unit (V.G.H.)
  - # Scale Selector switch panel for Log C.R.M. input (V.G.H.)
  - # Strip meter panel - Log CRM, Period (2), Linear (2) PM (2)
  - # Period Amps. Backman V6
  - # Audio Amplifier (V.G.H.)
  - # Log N select switch panel for Wheeler Recorder
  - # Differential circuit for Backman V output
  - # Relay Strip Keithly 410
  - # Lo Level PM Strip.
  - # High Curve Strip - worked out - just needs be built as unit.

NOTE ON STOP SWITCH - CONTEMPLATED FOR WEST END -



6/6/57 Comparison of characteristics of sharp cut-off pentodes - consideration of strip circuit application.




Start with  $E_{B+} = 150V$

- (6AK5) 5654
- 6AH6
- 6AU6
- 6CB6

6AH6  $E_A = 150, E_S = 150, \text{Log } R_p = 7500.$

Set up changed from above



Collected data on 6AH6 & 6AU6. Began on 6CB6 drift conduct.

6CB6 better conduct.

6-7-57 Plotted data -

Took data on 6A4H6 with  $E_s$  held at 150 volts - to plot  $E_p - I_p$  curves.

6-10-57 Took data on 6A06 -  $E_B = 150$ ;  $E_s = 150, 110, 50$  v

$E_B = 110$ ;  $E_s = 110$

Plotted data -

6-11-57 Took data on 6A95  $E_B = 110$ ;  $E_s = 110, 50$

$E_B = 150$ ;  $E_s = 50, 110, 150$

Plotted data

6-12-57 Select as first choice 6A06 from among three trials. Clean plate characteristics.

6-7-57

ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES
150	3.9	4.9	6.8	8.3	10	12	15	20	27	35	47	63	83	110	150
0	74	67	60	53	46	39	32	25	18	11	4	-3	-10	-18	-27
5	76.5	69	61.5	54	47	40	33	26	19	12	5	-2	-9	-17	-26
1	78	72	65.8	59.4	53	47	41.7	36.6	31.5	26.4	21.3	16.2	11.1	6	-1
1.5	84	78.5	73	67.6	62.5	57.4	52.3	47.2	42.1	37	31.9	26.8	21.7	16.6	11.5
2.0	89	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5	28.5	23.5	18.5
2.5	94	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5	28.5	23.5
3.0	99	93.5	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5	28.5
3.5	104	98.5	93.5	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5
4.0	109	103.5	98.5	93.5	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5
5.0	122	116.5	111.5	106.5	101.5	96.5	91.5	86.5	81.5	76.5	71.5	66.5	61.5	56.5	51.5
6.0	142	136.5	131.5	126.5	121.5	116.5	111.5	106.5	101.5	96.5	91.5	86.5	81.5	76.5	71.5
7.0	167	161.5	156.5	151.5	146.5	141.5	136.5	131.5	126.5	121.5	116.5	111.5	106.5	101.5	96.5
8.0															
9.0															
10															

6-7-57  
6A116  
ES = 150

ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES	ES
150	3.9	4.9	6.8	8.3	10	12	15	20	27	35	47	63	83	110	150
0	74	67	60	53	46	39	32	25	18	11	4	-3	-10	-18	-27
5	76.5	69	61.5	54	47	40	33	26	19	12	5	-2	-9	-17	-26
1	78	72	65.8	59.4	53	47	41.7	36.6	31.5	26.4	21.3	16.2	11.1	6	-1
1.5	84	78.5	73	67.6	62.5	57.4	52.3	47.2	42.1	37	31.9	26.8	21.7	16.6	11.5
2.0	89	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5	28.5	23.5	18.5
2.5	94	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5	28.5	23.5
3.0	99	93.5	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5	28.5
3.5	104	98.5	93.5	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5	33.5
4.0	109	103.5	98.5	93.5	88.5	83.5	78.5	73.5	68.5	63.5	58.5	53.5	48.5	43.5	38.5
5.0	122	116.5	111.5	106.5	101.5	96.5	91.5	86.5	81.5	76.5	71.5	66.5	61.5	56.5	51.5
6.0	142	136.5	131.5	126.5	121.5	116.5	111.5	106.5	101.5	96.5	91.5	86.5	81.5	76.5	71.5
7.0	167	161.5	156.5	151.5	146.5	141.5	136.5	131.5	126.5	121.5	116.5	111.5	106.5	101.5	96.5
8.0															
9.0															
10															

6-7-57  
6A116  
ES = 150



$E_p = 150$   
 $R_s = 47k$

$E_p = 150$   
 $R_s = 33k$






$E_p = 150$   
 $R_s = 24k$

$E_p = 150$   
 $R_s = 15k$

$E_p = 150$   
 $R_s = 10k$

GARG Data taken 6/4/50

$R_p = 3.9k$



6A04

$R_p = 1.5K$   
 $R_s = 47K$

$E_0 = 150$   
 $R_s = 21K$





$E_0 = 150$   
 $R_s = 10K$

$E_0 = 150$   
 $R_s = 47K$

6A04G


Data taken 6/6/57

$E_0 = 150$   $E_0 = 150$  (No series R.)  
(Data 6/6/57)




6146  
PLATE CHARACTERISTICS  
 $E_0 = 150$   
 $E_s = 180$

Data taken 6/9/57









Comparison of  
 $E_p$ - $I_p$  characteristics  
 6AN6  
 6AG5

4/19/57

6AU6  
Plate Characteristics  
 $E_g = 150V$   
 $E_c = 150V$



6/10/57  
 6AU6  
 PLATE CHARACTERISTICS  
 $E_p = 150, E_s = 110$   $\circ$   
 $E_p = 150, E_s = 50$   $\times$



6/10/57

6AU6  
Plate Characteristics  
 $E_g = 110$   
 $E_g = 110$




Plate Volts

6/11/57  
 6A95  
 PLATE CHARACTERISTICS  
 $E_g = 150$   
 $E_c = 150$   
 $E_c = 50$

PLATE CURRENT

10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38  
39  
40  
41  
42  
43  
44  
45  
46  
47  
48  
49  
50  
51  
52  
53  
54  
55  
56  
57  
58  
59  
60  
61  
62  
63  
64  
65  
66  
67  
68  
69  
70  
71  
72  
73  
74  
75  
76  
77  
78  
79  
80  
81  
82  
83  
84  
85  
86  
87  
88  
89  
90  
91  
92  
93  
94  
95  
96  
97  
98  
99  
100




PLATE VOLTS

$E_g = 20$   
 $E_g = 15$   
 $E_g = 10$   
 $E_g = 5$   
 $E_g = 0$   
 $E_g = -5$   
 $E_g = -10$   
 $E_g = -15$   
 $E_g = -20$   
 $E_g = -25$   
 $E_g = -30$   
 $E_g = -35$   
 $E_g = -40$   
 $E_g = -45$   
 $E_g = -50$   
 $E_g = -55$   
 $E_g = -60$   
 $E_g = -65$   
 $E_g = -70$   
 $E_g = -75$   
 $E_g = -80$   
 $E_g = -85$   
 $E_g = -90$   
 $E_g = -95$   
 $E_g = -100$

6/10/57  
6AB5  
Plate Characteristics  
 $E_b = 110$   
 $E_c = 110$   
 $E_g = 50$

Plate  
Current




Plate Volts

6-10-59 more tube characteristics -

Tube: 6AU6

Ep - HVVTM  
Ip - Simpson 260

EB=150, E3=110

Es	3.3k	3.9	4.7	5.6	6.8	8.2	10	12	15	20	27	47	150	Es	3.3	3.9	4.7	5.6	6.8	8.2	10	12	15	20	27	47	150		
0	103	92	79.3	52.5	36.7	21.7	18.4	12	7.6	5.36	2.6	0.44		0	116	107	98.7	84.4	77.2	63.5	46.5	34.4	18.8	9.5	4.6	3.1	1.5		
0.5	117.5	115	113	112	111.5	105.6	103.6	97.8	94	87.8	81.9	74.1	66.1	0.5	123	116	108	103	92.6	82	68.5	57.3	33	16.5	8.65	3.9	1.4		
1.0	119	112	104	86	74	52.8	46.6	25.8	14.7	9.33	4.0	0.47		1.0	131	126	121	116	108	101	90	81.3	61.8	36	17.8	5.9	1.7		
1.5	125	119	113	99.2	88.3	71	66.4	43.3	22	12.6	5.26	.91		1.5	136	133	129	126	120	114	106	101	85	64.6	39.2	16.3	6.3		
2.0	132	128	123	112	104	90.2	86	68	43.3	21.6	7.9	1.24		2.0	139	137	134	132	128	125	120	116	105	91	72.6	26.7	2.6		
2.5	136	133	130	122	116	104	103	89	70.5	46.6	14.2	2.2		2.5	143	142	140	138	136	134	130	128	122	112	100	67	5.6		
3.0	141	138	136	130	126	119	117	107	93	76	31.6	3.9		3.0	144	143	142	142	141	140	137	136	132	126	119	99	14.3		
3.5	142	141	139	135	133	128	126	120	111	99.5	67	7.1		4.0	146	146	146	146	145	145	144	143	142	142	142	142	142	141	
4.0	143	142	142	138	137	134	133	127	122	113	59	12.2		4.5	147	147	147	146	146	144	143	142	141	139	137	137	137	137	
5.0	146	146	145	144	142	141	140	139	135	130	118	5.7		5.0	147	147	147	146	146	144	143	142	141	139	135	135	135	135	
6.0	146	146	145	145	144	143	142.5	141	139	137	128	8.7		6.0	147	147	147	146	146	144	143	142	141	139	135	135	135	135	
7.0	148	148	148	148	148	148	148	148	148	148	148	148		7.0	148	148	148	148	148	148	148	148	148	148	148	148	148	148	148

EB=150, E3=50V

Es	3.3	3.9	4.7	5.6	6.8	8.2	10	12	15	20	27	47	150
0	77	69	60	52	40	29	20	15.7	9.2	6.2	4.1	1.9	.32
0.5	83.5	77.5	70	64	54.6	43.5	31	23.5	13.9	8	5.2	2.3	.25
1.0	90	85.5	80.5	76	68.7	60.6	50	42	25	13.2	7.4	3.2	.31
1.5	98	92.5	88.7	85	80.4	74.5	66.7	60.3	46	27.4	14.3	5.4	.58
2.0	102	99	97.5	93.2	90	86	81	77	66.6	53	35.6	10	1.14
2.5	104	102	101	100	98.3	97	91	88	82.3	73.5	62.8	32.4	3.46
3.0	107	106	104	103	102	101	100	98.5	93	88.5	82	63	7.6
4.0	110	110	109	108	108	107	106	106	106	103	101	92	52.3





1			
2			
3			
4			
5	744	8214	8207
6	735	8217	8203
7	667	8287	8214
8	692	8207	8216
9	700	8294	8213
10	733	8204	8212
11	691	8309	8217
12	692	8264	8217
13	681	8257	8214
14	689	8256	8209
15	674	8301	8212
16	679	8166	8210
17	667	8289	8216
18	654	8172	8201
19	650	8284	8207
20	622	8091	8222
21	673	8222	8212
22	649	8312	8220
23	641	8181	8207
24	645	8386	8204
25	644	8290	8175
26	632	8427	8215
27	604	8206	8225
28	646	8388	8210
29	619	8176	8217
30	607	8176	8209
31	619	8198	8213
32	639	8238	8219
33	629	8149	8222
34	583	8184	8214
35	617	8267	8209
36	592	8361	8213
37	565	8170	8208
38	608	8365	8210
39	567	8219	8210
40	609	8181	8200
41	558	8212	8213
42	587	8307	8208
43	531	8307	8219
44	566	8223	8222
45	548	8220	8213
46	555	8194	8191
47	577	8279	8221
48	570	8208	8210
49	513	8345	8214
50	567	8288	8208
51	576	8236	8214
52	514	8218	8209
53	535	8221	8202
54	553	8325	8213
55	510	8190	8209
56	530	8351	8212
57	49		8209
58	5		8215
59	53		8216
60	52		8208
61	512	8164	8208
62	502	8329	8208
63	502	8236	8212
64	502	8109	8201
65	508	8385	8221
66	492	8210	8211
67	490	8179	8201
68	484	8235	8210
69	482	8159	8217
70	492	8242	8212
71	443	8277	8216
72	523	8209	8221
73	474	8251	8205
74	484	8242	8216
75	458	8209	8202
76	495	8268	8218
77	479	8277	8207
78	486	8330	8220
79	470	8246	8212
80	450	8281	8214
81	463	8249	8221
82	461	8299	8211
83	442	8159	8215
84	452	8298	8209
85	470	8180	8196
86	470	8275	8211
87	430	8130	8199
88	498	8272	8210
89	440	8226	8198
90	463	8202	8217
91	474	8384	8213
92	424	8320	8196
93	434	8297	8214
94	448	8199	8207
95	452	8247	8216
96	435	8125	8203
97	442	8206	8217
98	429	8187	8206
99	443		

89	440	8226	8198
90	463	8202	8217
91	474	8384	8213
92	424	8320	8196
93	434	8297	8214
94	448	8199	8207
95	452	8247	8216
96	435	8125	8203
97	442	8206	8217
98	429	8187	8206
99	443	8203	8217
100	441	8140	8216
101	435	8254	8218
102	434	8326	8225
103	413	8292	8202
104		8170	8212
105		8369	8210
106		8188	8213
107		8184	8217
108	442	8229	8215
109	423	8254	8218
110	406	8258	8196
111	404	8094	8216
112	411	8269	8203
113	408	8198	8219
114	417	8251	8210
115	403	8192	8211
116	413	8338	8220
117	413	8297	8206
118	448	8286	8206
119	386	8326	8209
120	404	8201	8194
121	403	8272	8209
122	418	8175	8226
123	391	8336	8195
124	381	8333	8209
125	405	8394	8202
126	402	8322	8210
127	407	8290	8222
128	407	8214	8197
129	402	8403	8214
130	403	8250	8216
131	391	8191	8216
132	394	8320	8216
133	394	8237	8214
134	406	8290	8210
135	383	8202	8200
136	411	8160	8219
137	381	8253	8212
138	382	8258	8214
139	421	8268	8220
140	391	8201	8199
141	391	8249	8215
142	395	8165	8217
143	395	8287	8201
144	395	8269	8211
145	399	8297	8215
146	418	8231	8212
147	382	8211	8209
148	385	8265	8218
149	371	8217	8202
150	393	8219	8210
151	362	8289	8216
152	388	8193	8211
153	378	8317	8215
154	367	8140	8206
155	381	8327	8212
156	384	8148	8207
157	360	8181	8208
158	365	8378	8220
159	354	8272	8207
160	359	8250	8217
161	346	8188	8223
162	408	8148	8213
163	340	8113	8208
164	360	8223	8208
165	379	8144	8198
166	371	8329	8220
167	343	8234	8204
168	381	8296	8223
169	375	8365	8215
170	354	8258	8210
171	362	8175	8218
172	380	8338	8204
173	361	8163	8212
174	369	8164	8219
175	317	8233	8212
176	352	8212	8216
177	353	8246	8218
178	360	8145	8212
179	317	8290	8206
180	352	8223	8203
181	339	8200	8210
182	342	8110	8210
183	333	8369	8216
184	349	8160	8209
185	328	8282	8207
186	329	8294	8207
187	338	8208	8216
188	345	8240	8209
189	356	8192	8203
190	362	8306	8222


180	352	8223	8203
181	339	8200	8210
182	342	8110	8210
183	333	8369	8216
184	349	8160	8209
185	328	8282	8207
186	329	8294	8207
187	338	8208	8216
188	345	8240	8209
189	356	8192	8203
190	362	8306	8222
191	358	8245	8195
192	362	8368	8205
193	313	8408	8209
194	333	8329	8222
195	319	8334	8200
196	332	8431	8218
197		8233	8213
198		8207	8199
199		8327	8213
200		8225	8211
201		8246	8215
202	328	8204	8197
203	323	8203	8210
204	327	8218	8206
205	338	8275	8211
206	328	8331	8217
207	296	8157	8201
208	340	8266	8217
209	310	8227	8196
210	350	8313	8224
211	327	8310	8204
212	321	8295	8214
213	310	8250	8201
214	302	8202	8210
215	324	8273	8213
216	328	8212	8208
217	313	8193	8202
218	288	8295	8220
219	322	8201	8207
220	324	8304	8207
221	270	8177	8219
222	331	8355	8225
223	336	8151	8217
224	326	8186	8209
225	280	8440	8225
226	370	8266	8209
227	313	8288	8214
228	336	8141	8194
229	321	8154	8219
230	286	8169	8226
231	317	8207	8206
232	336	8213	8212
233	295	8416	8212
234	301	8187	8196
235	291	8305	8213
236	307	8336	8209
237	315	8214	8220
238	315	8134	8212
239	279	8296	8223
240	292	8206	8210
241	299	8243	8216
242	293	8235	8209
243	301	8240	8219
244	293	8295	8221
245	282	8170	8209
246	291	8260	8213
247	288	8303	8210
248	287	8251	8209
249		8213	8212
250			8202
251			8210
252			
253			
254	319		
255	283		
256	347	8241	8211

6/13 Keithley output circuit up to file Kelly, Box 178

Week -

6/17-21/57 Trip Circuit Considerations - Keithley Output

Circuit Under Test:




Chose B+ of 110. With this B+  $I_p \approx 2.5 \text{ ma}$ ,  $I_s = 196 \text{ ma}$  when grid is grounded.

Relays A & B = SIGMA SRJ 20004 -  $\left\{ \begin{array}{l} I_{\text{fire}} \approx 19 \text{ ma.} \\ I_{\text{release}} \approx 1.2 \text{ ma.} \end{array} \right.$

Set up Brush oscillograph, dual-channel, to make some time measurements. Interest in (a) response time of circuit. (b) response time of relay.

Serviced one of Brush amplifiers (refer file on Brush)

(1) Input to grid of 6AU6 from Keithley - as shown -




Set trip level to about 60% on any range ( $\approx 2 \text{ v}$  bias to fire) with source given, reading of  $\approx 40\%$ . Changed scale to increase meter, causing trips.

Attached one channel across Keithley output. - other channel from plate 6AU6 to ground.


Results in figure 1.

6/19/57


FIGURE 1




(2) Put battery in input to 6AU6 to provide step voltage in place of Keithley. Insert push-button switch in grid to give steps.



Results in fig. 2.



Using Tektronix Scope - checked operating time of microswitch V3.  
Wiring:



③ With DC step on grid, inserted normally-opened contact of square relay B in input. The brush trace would give pip on grid, indicating time from closure of push-button to opening of relay contact - This would be equivalent to circuit response time.

Wiring:





FIGURE 3

A-B = 3-4 ms.  
B-C = 1-2 ms.



As check, put scope across grid to ground. Pattern was 2.5ms. long.

Refer to fig. 3.-

Time from contact of PB1 to point grid reaches firing bias = 3-4 ms.


Peak on plate appears to occur at firing bias point.

Grid continues negative for 1-2 ms. until striking min.

Plate trace shows action of reset push-button, PB-2.


④ Set-up to take measurement of response when starting grid bias very near firing point.

WIRING:




CONDITION	$I_p$	$I_s$	$E_g$	$E_p$	$E_s$
Rg adj. to point near trip	1.15 ma.	.46 ma	1163V.	103V.	74.5V
PB1	0	.65	(spike)	0	59
PB2 & 3	2.5	.96	0	92V.	94.


FIGURE 4




10-28-57 Notes on  $\mu$  Amplifier - feed-back differential measurement  
 Representative from Argonne on 10/25 suggested circuit as  
 indicated below to record very small changes in level.



11-11-57 Instrumentation 112-113  
 PM. Mount Aulhausen crystal no 6292-




At foil, then cover w/ black tape.




Found shield of pre-amp case on shelf in 209 which could be used  
 for this instrument. Had been used earlier in some detector  
 development.

Wired base socket as follows.




11/12/57

Tube to be used w/ PM. 6AU6



R	C	t
$5 \times 10^6$	$.010000 \times 10^{-6}$	50 msec

11/13/59 Circuit Arrangement for Test.



$V_g$  measured with HP VTVM.  $I_p$  - measured with Simpson 260  
 $V_c$   
 $V_s$   
 Simpson voltages readings in error.

$Bt = 150$ ;  $V = 6AUG$ .

H.V.	$V_c$	$I_p$	$V_p$	Plate ckt	$V_s$	Screen ckt	$V_g$
0	0	2.6	139	$R_{y1} + R_{y2}$	67	$R_{s1} = 75k$	-0.94
500	0	"	no change		68		-0.97
750							-0.98
1000	0	2.58	138		69		-0.99
1250	0	2.55	137		69.5		-1.05
1000	0.2	2.5					-.84
0	0.2	2.6	139		68		-.78
0	0.4	2.56	138		68.5		-.81
0	0.6	2.53	138		69		-.84
0	0.79	2.48	140		71		-.79


HV	$V_c$	$I_p$	$V_p$	Plate ckt	$V_s$	Screen ckt	$V_g$
0	1.0	2.41	140		73		-0.14
	1.2	2.33	140		75		-0.05
	1.4	2.20	140		79		0
	1.75	1.91	143		89.5		+0.01
	2.00	1.71	148		93		+0.01
	3.00	1.00	146		114		+0.01

Need to fix  $V_g$  &  $V_c$ .

More comments over

11-13-57

Circuit to fix value of screen of cathode




Screen tied to point A & then point B.

By fixing screen of cathode in this way major effect of signal change was present in plate circuit. -  
 Other pentodes possible - 6AH6, 6AG5, 6AK5, 6C33L.  
 Only rough comparison made between of the several.

~~Method~~


Imp. set control in cathode affects plate current thru tube. This would upset recorder zero when one such instrument is incorporated in design.

To eliminate attraction of Imp. set with zero and Ellis-type plate circuit.



11/14/57 Recorder Circuit.

Consideration given to taking off signal for recorder in cathode - This would mean changing cathode circuit.



Found to be extremely sensitive & linear at full  $R_c$ .  
 $\approx 2.5V E_c$ . Reduced sensitivity by reducing  $V_{cp}$  on  $R_c$  - & correspondingly reducing 100k values.

Had noted marked grid bias change with HV changes.

Probably poor 6292 -

Checked signal from several PM tubes -

Tube	HV	500	600	700	800	900	1000	1100	1200	1300
A	From 108				0	.01	.02	.05	.11	.30
B	From 108	.20	.25	.30	.35	.39	.44	.49	.57	.67
C	From 108				.01	.03	.17	.44	.76	1.4
D	new stock				.06		.08	.08	.11	.16

Mounted crystal on Tube D and installed in place.

Battery breaker loop 18.3 pa @ 4:20 P. Light over night.



11/15/51 Battery bucking voltage 18.3v 8.25A

As more thought is put to circuit feeling is that plate take-off point should be compared w/ cathode. If one can again "fix" cathode & get major changes in plate may get better response on legs.


Cathode circuit: (Disconnect recorder)

B+ = 150 HV = 1100 Trip set so  $I_p = 1.2$  ma.

$E_c = 3.1v$  (3.0 off set to 3.1)

$E_g = -0.55$

HT	DT	inches	wires
1100	4	4	2 7/8
		4	3 1/4
		4	2 7/8
		4	2 7/8



HT = 1200 - Reset trip so both relay resist -  
 $E_c = 3.1v$ ;  $E_g = -0.11$   $I_p = 1.24$  ma. DT 5 1/2 DT 4 1/8

$E_g = -0.2v$   $-0.78v$

w/o resistor trip HT = 1000;  $E_c = 3.1v$ ;  $E_g = -0.55$   $I_p = 1.25$  ma.

DT	$E_g$	DT	$E_g$
2 7/8	.56	1 7/8	.78
2 3/8		1 7/8	
2 3/8	.54	1 7/8	.78

11/15

Change circuit so cathode is essentially fixed.

HT = 1000;  $E_c$  set at 3.1v;  $E_g = 0.07$   $I_p = 1.25$

DT	$E_g$	DT	$E_g$
3 1/8	.36	2 7/8	

Better to have cathode fixed.

Check to see value of  $R_a$  for  $m_i$ ;  $E_g =$  for this tube -  $\approx 1000$

Suggest screw down adj for chassis  $R_a$  of 2500


HT = 1200 DT  $\approx 7$  in.

HT 1000 adj - 5" - 3 1/2"

HT 1300 no adj 16" - 9"

Assume 2.5 ma  $I_p$   $R_p$  for 10mw.  $R = \frac{10^{-2}}{.25 \times 10^{-2}} = 4 \times 10^0 = 4 \Omega$

Try 10  $\Omega$ :



$\frac{1.4v}{40 \mu a} = \frac{1.4}{.4 \times 10^{-4}} = 3.5 \times 10^4 = 35k$

$40 \mu a - E = 100 mw$   
 $R = \frac{10^{-1}}{.40 \times 10^{-4}} = 2.5 \times 10^3 = 2.5k$


11:38

$E_g = -1.3v$ ; HT = 1300;  $I_p = 1.18$  ma.

T = 6A46

$E_c = 3.15v$  Set mode to 10%  $I_B = 16.2$

11:42 set HV 1000 recorder 50  
 11:59 re-adj to 50 set m up



- Components - top chassis  
 Isolate transformer  
 Filament transformer  
 Rectifier  
 Capacitor  
 Battery  
 Relay
- Rear - bra-lyp connect  
 Recorder output  
 secondary adj cathode
- panel -  
 I<sub>B</sub> meter  
 Helipot (Recorder Zero)  
 Trip set.  
 power switch  
 Recorder on  
 reset P.B.  
 fuses

Downward drift noted during afternoon - Needs further cleaning →

11/10/57 9:20 AM Power on Equipment. B+ = 148  
 HV = 1000 I<sub>r</sub> = 1.13 ma I<sub>B</sub> = 16.3  
 17 ma  
 E<sub>g</sub> = -0.07  
 Balance recorder @ 9:47 to 50%  
 @ 11:30 still -50 had gone up to 70% & back down -

Recorder trace for 11/14/57 fairly steady, drift-min.  
 11/14 Power on 8:15 HV = 1000; B+ = 150; I = 17 ma; I<sub>r</sub> = 1.2 I<sub>B</sub> = 16.2  
 E<sub>g</sub> = -0.055  
 8:25 E<sub>g</sub> probe removed.

Power on overnight  
 11/20/57 8:05 Rezero - Had gone off-scale neg. B+ 150 HV 1000  
 E<sub>g</sub> = -0.050  
 2y 4 200 & 100 as well as 150 100 10k.

4/18/57

DWN - need for instrument to read sinusoidal ~~1%~~ to show power level changes of order of .001.  
 Purpose: To measure reactivity of order  $10^{-6} \frac{\Delta k}{k}$   
 Differential circuit of Argonne suggested.

4/21/57

Fission Counter Check

{ Pre Amp 327  
 LA 490319  
 Scaler 4100371

Source Pu-Be  $1.28 \times 10^6$

LA Gain 16x1  
 2  $\mu$ s Res. Time  
 Disc. = 7

Disc	Source	Background	Counts	Time	Notes
		Background	0 counts	in 3 minutes	
	Source	8x16 + 4		in 2 min	
10		6x16 + 11		2 min	Counts 5-2
20		7x16 + 1		3 min	
30		0x16 + 15		2 min	
6	Blkg	0x16 + 2		2 min	Counts 5-1
6	Source	10x16 + 2		2 min	
7		8x16 + 11		2	
10		8x16 + 6		2	
15		5x16 + 13		2	
20		5x16 + 3		2	
30		0x16 + 7		2	

Counts 5-3

Disc Time 15  
 Changed to 20  
 0.8  $\mu$ s 30  
 Counts 5-1:

Disc	Scaler	Time	Source
7	15x16 + 3	2	Blkg
	20x16 + 11	2	Pu Be
8	0x16 + 1	3	Blkg
	10x16 + 3	2	Pu Be
10	52x16 + 12	11	"
15	9x16 + 2	2	"
20	11x16 + 10	3	"
30	5x16 + 0	2	"
40	4x16 + 0	3	"
50			

1/21/57  
C 5-3

Gain 16-1 - RT 0.1g

Disc	Scaler	Units	Per Bar only
7	29x16+11	2	
8	30x16+5	4	
10	23x16+10	6	
15	47x16+9	11	
22	77x16+6	19	
30	27	1	10
40	2	13	2

C 5-2  
Gain 16-1  
RT 0.8g


Disc	Scaler (x16)	Units	Relay	Source	Scale	Units
6	277	8	← ✓	✓ →	19+5	1
7	13	2	← ✓	✓ →	26+10	3
8	2	1	✓	✓	16+11	3
10	21	5		✓		
15	13	1				
20	10	12				
30	18	6				
40	8	12				

Source

11/25


Set up for tuning test on photomultiplier. (Refer to time test file)

Built selenium rectifier power supply for triax circuit:



11/26

Circuit under test:



Head: Anthracene Crystal  
6292 Photomultiplier  
6446 Amplifier

HV - Hamman N401

Relay - P.S. as shown in diagrams

Initial Adjustments:

1. Set Control "B" a.c.w. - max. trip relay current
2. Adjust "C" a.c.w. until  $I_p$  drops to 2 mas, at which point relay can trip & reset by adjusting "B".
3. Set trip level by turning "B" clock wise to proper point. Relay trip at  $\approx 1.15$  ma  $I_p$ .

During time test trip point will be set at 1.2, 1.3, 1.4.

Signal, $E_g$ , for triax	Set 1.2	Set 1.3	Set 1.4	same for all HV.
	.22 ; .34	.40 ; .50	.55 ; .65	


11/20/51	Background signal -	HV
	.075	1000
	.09	1100
	.10	1200
	.12	1300
	<u>.155</u>	<u>1400</u>
	.255	1500

4-7-58 PM-circuit - Has been in use in South Wing.

Certain modifications are necessary. -


- a. wider separation of low & high trip points.
- b. Reproducible trip points -

Set up "bread board" for further tests.



\* 7k resistor same as resistance of 2 relays presently in PM.

Plate meters - 0-5 ma - also use API 0-100  $\mu$ a - rated 950  $\Omega$  or 31000



$$10^{-4} \times 5 \times 10^3 = 0.5 \times 10^{-1} = 0.05$$

$$E = .1V$$

$$R = \frac{10^{-1}}{5 \times 10^{-3}} = 10^{-1} \times \frac{1}{5} \times 10^3 = \frac{10^2}{5} = 20$$

Shunt meters with 22  $\Omega$  / 100  $\Omega$  in parallel. This gives approximate 5 ma full-scale =

$I_m$  &  $I_r$  = each 0-100  $\mu$ a API w/ shunt noted above.

$I_p$  = 0-50 ma.

$E_c, E_p, E_B$  = Simpson 260 V-100302

$E_g$  = DC millivoltmeter I.C. Type 200-A, Ser # 152. Y93796 -

184

PM-

3:20P

Tube = 6AH6

4-7-58

$E_B = 201v \quad I_p = 8.5$

$E_p = 24v \quad I_m = 85 \quad (4.25)$

$E_c = 0 (L.05) \quad I_r = 87 \quad (4.35)$

$E_g = 0$

With no signal, adjusted  $R_c$  for max  $I_p$ . -  $R_f$  set at mid-range

$E_g$	$R_c$	$E_p$	$I_m$	$I_r$	$I_p$	$E_B$
0	0.54v	21.5	<del>87.5</del> 87.5 440	<del>(4.375)</del> 90 4.5	8.9	101
-0.1	0.51	21.3	88	90	9.0	
-0.2	0.49	21.8	87.5	89.5	8.8	
-0.3	0.44	23.0	<del>87.5</del> 86 4.3	<del>89.5</del> 88 4.40	8.6	
-0.4	0.45	26.5	<del>86</del> 82 4.1	<del>88</del> 84 4.2	8.2	
-0.5	0.42	32	<del>82</del> 76 3.8	<del>84</del> 78 3.9	7.6	
-0.6	0.40	37.5	<del>76</del> 81 3.55	<del>78</del> 73 3.65	7.1	
-0.7	0.38	42.8	65.3	67.7	6.5	
-0.8	0.35	47.8	59.5	61.5	6.0	
-0.9	0.33	52	<del>59.5</del> 54 2.7	<del>61.5</del> 56 2.8	5.3	
-1.0	0.31	57	44.5	51.5	5.0	
-1.1	0.29	62	42.5	44.5	4.2	
-1.2	0.27	67	<del>42.5</del> 38.5 (1.92)	<del>44.5</del> 40 (2.0)	3.9	
-1.3	0.25	70	34	35.5	3.3	
-1.4	0.23	74	<del>34</del> 30.5 (1.52)	<del>35.5</del> 31.5 (1.58)	3.0	
-1.5	0.22	78	<del>30.5</del> 26 (1.3)	<del>31.5</del> 27.3 (1.36)	2.6	
-1.6	0.21	81	23	24	2.2	
-1.7	0.20	83	<del>23</del> 20 (1.10)	<del>24</del> 21 (1.05)	2.0	
-1.8	0.19	86	17.3	18.5	1.7	
-1.9	0.18	88	<del>17.3</del> 15 (0.75)	<del>18.5</del> 16 (0.8)	1.5	
-2.0	0.17	90	<del>15</del> 12 (0.6)	<del>16</del> 13 (0.65)	1.2	

4-8-58


Noted change in cathode so grid signal introduced.

Bleeder current only the order of 5 ma.

Increase bleeder current to 100 ma - of re-run.

(Put ammeter in B+ line)

Change:



$E_p$   $E_B$  measurement on Vomax VTVM

$E_g$	$E_c$	$E_p$	$I_m$	$I_p$	$I_p$	$I_t$	$E_B$
0	.04	24	40.25	4.25	8.28	130	101
0	.54	21.8	80.5	8.5	8.5	128	102
-0.2	.58	24.6	41.75	4.75	8.55	128	* change $E_g$ Bat = 3V Rt = 10k kilohm
-0.4	.52	35	83.5	8.5	8.5	128	
-0.6	.51 <sup>+</sup>	49	40.25	4.25	8.28	126	
-0.8	.51 <sup>-</sup>	62	89.5	8.5	8.5	126	
-1.0	.50 <sup>+</sup>	72	34.75	3.70	7.175	124	
-1.2	.50	81	69.5	7.0	7.0	123	
-1.4	.50	87.5	27.5	2.95	5.8	123	
-1.6	.50	92	55	5.9	5.7	123	
-1.8	.49	95	21.5	2.30	4.45	122	
-2.0	.49	98	43	4.6	4.3	122	
			1.640	1.765	3.40	121	103
			32.8	3.53	3.3	121	
			1.180	1.275	2.46	120	
			23.6	2.55	2.4	120	
			82.5	8.0	1.72	119	
			16.5	1.8	1.7	119	
			.55	.625	1.18	118	
			11	12.5	2.2	118	
			37.5	4.25	0.8	118	
			7.5	8.5	1.0	118	
			26.0	3.00	.56	118	
			5.2	6.0	0.7	118	103

4/8/58

Relay Operation Checks:

Relay: 201		202		203		<del>501</del>		<del>502</del>		<del>503</del>		<del>504</del>	
$I_o$	$I_r$	$I_o$	$I_r$	$I_o$	$I_r$	$I_o$	$I_r$	$I_o$	$I_r$	$I_o$	$I_r$	$I_o$	$I_r$
(4.1)	(2.2)		(2.35)	(4.3)	(2.15)	(6.5)	(3.1)	(6.1)	(2.8)	(6)	(3)		
1.9	1.03	2.0	1.13	2.0	1.1	1.25	.6	1.2	.55	1.15	.6		
(4.1)	(2.2)												
1.9	1.04	1.9	1.13	2.0	1.1	1.25	.6	1.2	.55	1.15	.6		
(4.1)	(2.2)												
1.9	1.04	1.9	1.13	2.0	1.1	1.25	.6	1.2	.52	1.15	.6		
								(6.0)					
1.9	1.04	1.9	1.13	2.0	1.1	1.25	.6	1.18	.58	1.15	.6		
1.9	1.05	1.9	1.13	2.0	1.1	1.25	.6	1.18	.55	1.15	.6		
1.9	1.05	1.9	1.13	2.0	1.1	1.25	.6	1.18	.55	1.15	.6		
1.9	1.05	1.9	1.13	2.0	1.1	1.25	.6	1.18	.55	1.15	.6		
		(4.0)											
1.9	1.05	1.9	1.13	2.0	1.1	1.25	.6	1.18	.55	1.15	.6		

$I_o$  = operating current  
 $I_r$  = releasing current  
 ( ) = Voltage across solenoid

4-8-58

Inserted relays 201 & 202 in plate lead of test circuit, replacing 7000 ohm resistor. Relay meter =  $I_m$ .


- Rc 201 "fine" @  $E_g = -1.2v, I_m = 23.5 (1.175)$
- Rc 501 "fine" @  $E_g = -1.55v, I_m = 13 (0.65)$
- Rc 501 "parts" @  $E_g = -1.28v, I_m = 21 (1.05)$
- Rc 201 "rest" @  $E_g = -1.0v, I_m = 32.5 (1.625)$

4-9-58

Set-up for operation with photo-multiplier head. - HV power supply not up-to-par - 2V3G weak. Unable to get 2V3 from V-12 & X-10 stock - both stores have cancelled the item - Changed power supply to use 2X2A -

4-10-58


Circuit for test -  
 - new: signal from PM head (Anthracene + 5819)  
 - new meter for  $I_p, I_o, I_{high}$  (0-10ma Weston 301)  
 - new to relay 502 for 203





4-11-58

Chart:



New  $\Rightarrow$  dropping resistor in low level legs - to reduce current. This an effort to widen interval between low & high level legs.  $R_{g-L} \approx 2000 \Omega$ ;  $R_{g-H} \approx 5000 \Omega$   $R_b \approx 3,300 \Omega$

Voltage reading now all taken w/ HP VTVM. (Vomax developed trouble).

$\Rightarrow$  Behavior of circuit as  $R_b$  is varied.

( $E_{g0} = 0$ ,  $E_{g1} = 0.09V$ )

Dial	$E_{g1}$	$I_P$	$E_L$	$I_L$	$E_H$	$I_H$	$E_R$	$E_D$	$I_T$	$E_{B1}$
$R_b$	volts	ma	volts	ma	volts	ma	volts	volts	ma	volts
ccw -										
100	50	19.5	58.5	3.48	110	7.0	58.7	110	17.5	6.5
90	20.8	11.2	75.5	1.95	105	7.95	76.5	110	22.0	
80	60	10	64	1.58	91	6.15	68.5	110	20.8	
70	58	8.6	61.8	1.1	82	4.68	67.5	111	20.8	6.55
60	56.4	8.2	56.5	.95	71.7	3.4	67	111	20.9	
50	57	8.5	58.5	.95	70	2.85	73.5	111	21.3	
40	65	9.4	66.5	.98	74	1.8	87.3	111	21.4	
30	85.6	12.5	89.5	.2	91.5	.96	111	111	20.4	
ccw -										

Note - sudden rise in total current as  $R_b$  turned slightly off full ccw position

4-11-58

Rise in total current not shown in  $I_P, I_L, \text{ or } I_H$  - Probably screen current -

$\Rightarrow$  try another type tube to see if rise occurs -  
Circuit similar; Added - 0-50 ma - meter for  $I_S$   
to that opposite - 0-10 ma - meter for  $I_R$

Components:  $R_{g-L} = \#201 (2150 \Omega)$  Control plate resistor =  
 $R_{g-H} = \#502 (5100 \Omega)$  57a (55  $\Omega$ )  
 $R_b = 3300 (3200 \Omega)$   
 $R_b [pot] \text{ set @ } (18,000 \Omega)$

Change	measuring	plate voltage	at junction	of meters	ABT	plate					
Dial	$E_{g1}$	$I_P$	$E_L$	$I_L$	$E_H$	$I_H$	$E_R$	$I_R$	$I_S$	$I_T$	$E_{B1}$
100	75.2	13.0	80.4	2.28	110	9.00	80.5	1.67	5.5	151	$E_{B1} = 6.5$
80	46.6	12.4	51.5	1.95	84.3	7.56	56	2.96	5.7	150	$E_B = 110$
60	36.7	11.4	33.6	1.4	52.6	5.34	48.4	4.80	6.1	150	$I_{B1} = 135 \text{ ma}$
40	31.3	11.0	33.4	1.04	51.5	3.97	50.6	6.28	5.7	149	
20	40	11.4	41.4	.73	54	3.1	64.5	8.07	5.4	149	
0	65.5	12.0	66.2	.48	73	1.5	95.6	2.10	5.2	149	

Set up for use with PM head - Results fair  
 $NF = 1200$   $R_b = 100$   $I_P = 7.8$   
 $I_L = 1.34$   $I_H = 5.55$   $I_R = 1.10$   $I_S = 3.3$   $I_T = 143$   
 Low legs 22" }  
 High legs 4" }

$E_{g0} = -.83V$  - This is unusually high for background.

Set  $E_{g0}$  at  $\approx 1.0V$  reduced  $E_{g1}$  to .15 volts background.

4-11-58 Replaced Re-1 #201 with #503 Trips:  $b=9"$ ,  $H=4.5"$   
 Added input capacitor of 0.1  $\mu$ f  $b=6"$ ,  $H=3.5"$


→ Trip adjust range is limited.

4/14/58 Tried increasing  $E_B$  to get wider low-high trip.  
 Seemingly - the higher the supply voltage the narrower the interval.

To make this test - added 1000  $\Omega$  to bleeder -

4/15/58 Single pentode with parallel plates relay is not satisfactory method of getting wide range trip - try new approach.

- A - Dual triode - one relay for each plate - for trip
- B - Dual triode - for recorder output -



Found one needed  $R_c \approx 500$  for low-half. To get good control of  $I_L$ .

Replaced 500  $\Omega$  pot w/ 25000  $\Omega$ .

Increased  $E_B$  to 250 volts -

Get good reproducible trips @ 12" & 3" -  
 Trip signal 15v. & 17 volts


limited 570  $\Omega$  in  $R_c$  - high to give max  $I_H \approx 10$  ma - 1.91

4/15/58 During testing - found could get only max signal of 10  $\mu$  volts  
 Some component failure -

Tube (6M) or Power Supply ? or tube under test

Brought in Hamner Power Supply from Pm 215 (50 channel)

→ Replaced 12BH7 - now ok.



with HV @ 1050 has trip range 22" - 5"  
 High trips 3"

$E_B = 255$  v

Left power on over night to check drift.

4:25P  $I_H = 9.8$  ma  $I_R$  adj to zero - by  $R_p$  -

$I_L = 1.2$  ma

$I_T = 20$  ma

$E_B = 257$  v.

4-16-58

8:10A

$I_H = 9.7$  ma

$I_L = 1.2$  ma

$I_R$  - off scale neg.

$E_B = 259$

$I_T = 19$  ma


adj  $I_R$  to 0

4-16-58 Trip point (Snyka) HV=1050  $I_R$  (blk) 0.18 units  
 Low trip = 15.75"  
 High trip = 3"  
 Variation of  $I_R$  with HV - Source @ 5 feet -

HV	$I_R$	
900	2.5	
950	5	
1000	7.8	
1050	10-13	
1000	0.5	Blk
0	-0.4	

8:30<sup>A</sup> adj -  $I_R$  to +1.0 no HV-

Provision for 0-10 mv. recorder -



If recorder resistor is max - 1000  $\Omega$   
 10 mv on recorder if  $I_R = 10 \mu\text{a}$ .  
 This makes for extremely sensitive recorder - good for measuring drift

set up circular Brown for run - seemed to be responding after battery change & new 7F7 -  
 11:30<sup>A</sup> HV=1000 - NO source -  $I = 6 \mu\text{a}$   
 $V = 6 \text{mv}$  ( $R_{rec} = 1000 \Omega$ )  
 (Thought saw rise in gamma level on West end)

3:15 Noticed recorder trace steady -  
 checked w/ source - NO RESPONSE on Recorder  
 Recorder extremely sluggish -  
 Quickly activated small Foxboro to run trace over night -  
 Trace from 4:15<sup>P</sup> to 7:48<sup>A</sup> Reading 50 to 51%  
 very little fluctuation. Good -

4/17/58

$I_H = 9.8$

$I_L = 1.2$

$I_R = 8.5 \mu\text{a}$  -  $R = 51$

$I_T = 20$

$E_B = 257$

Comparison of Recorder with  $I_R$  used Co-source

61	20
70	30
79	40
96	60

Why?  $\rightarrow$  44 0 manual  
 Reduced  $R_{rec}$  to minimum - Recorder still reading 20%  
 Noticed recorder frame was grounded - disconnect ground per to near zero.

4/17/58

9:10 A / In re-locating PM housing - noticed noise in ~~input~~  
Removed base of slide to check crystal -  
Found loose member was mu-metal shield.  
Took out shield.

Put black tape over opening & powered - large  
indication on  $I_p$  - ????

Light leak just suspected - Is tube damaged?  
Kept getting high background current indications -

New tube the same - -

No response to source ? -

Found that signal & HV leads were reversed  $\frac{00}{00}$

Went for new tube -

4:15 Rec = 5 mV

4/18/58

7:45 A  $I_H = 9.8 \text{ ma}$ ;  $I_L = 1.2 \text{ ma}$ ;  $I_T = 20$ ;  $E_B = 257$

$I_R = 3.5 \mu\text{a}$

HV = 1000

Rec = 3.3 mV

8:11 A checked response to Co. source Source @ 12" R = 8 mV.

8:50 Brown Recorder check. Circular unit failed after several hours  
operation 4/16.

Zeroed ok.

Response & signal ok.

Standardize ok.

Lied Brown input across 1000  $\Omega$  pot. in test circuit.

(For loss - one end to center wire of same)

Effect on For loss - - initial upward step of trace. Settled  
down after 15 minutes to normal signal  
- width of trace narrowed -

Pkg F-38

B-38

10:10 -> Co @ 12" F-82; B-88

(Noticed during 1st hour of operation timing mechanism  
lagged. Set-up chart 4 times between 8:50 & 10:05)

Running source in place to see source trace -

10:20 -> Source away -

10:55 - Source @ 12" away ~ 11:15, up 11:45 - away 12:00 (WGV)

4:15 P Pkg - HV = 1000

Leave on over week-end

4-21-58 7:56 off scale neg - adj zero only slight adj  
needed in Rp to balance recorder to near mid scale  
 $E_B = 257$ ,  $I_T = 20$   $I_L = 1.2 \text{ ma}$ ,  $I_H = 9.9 \text{ ma}$ .

4/22/58 8:15 change charts - very little drift for 24 hour period.  
Change may be as much as .4 mV in 6mV -

n meter .4  $\mu\text{a}$  / 100 .4  $\mu\text{a}$  x 50  $4 \times 10^{-7} \times 5 \times 10^4 = 20 \times 10^{-3} \text{ V}$

4/23/58 8:12 change charts.


4/24/58 8:22 change charts

Last adjustment was Monday (4-21) @ 7:56 AM.

Seems reasonably stable at this sensitivity.

(over)

4-24-58 Sketch of Circuit that has been under test -



from p 140 Trip points 1.5V & 17V on signal

Adjust Brown to read on scale at low level trip -

Bring  $R_a$  (5mg) into shop -

Adjustment 100µa @ 2.4mv  $R_r = \frac{2.4 \times 10^{-3}}{10^{-4}} = 2.4 \times 10 = 24 \Omega$

@ 10mv -  $I = \frac{10^{-2}}{2.4 \times 10} = \frac{1}{2.4} \times 10^{-3} \approx .4 \times 10^{-3} = 400 \mu a$

1115P Removed 0-100 RPI & put in 0-100µa curjin.

No effect on Recorder.

4-24-58

Instrument Response to 5mg  $R_a$  source -

Source Distance	Recorder	Meter	[R]
24"	25%	100µa	25.2
22	29.5	115	25.6
20	34.8	140	24.8
18	42	165	25.4
16	57.3	210	24.4
14	65	290	24.1
13	75.3	315	23.9
~12 1/2"	82.5	350	23.7
12	86.3	365	23.6
11	85	430	
10	81	2500	

Adj. trip @ ~70% records (0.350-360µa)

11.5" trips

75% records

@ HV = 1000V  
F-80

2:45 Change HV to 900V

Source & source distance I = 130 - R = 28

Trip - R = 71.5 D = 6 3/4"

I = 350-360

W 1100 - Trip - R = 74


I = 360

D = 18.5"

High trip @ 4"

More sens. overnight -

4-25-58 Modified recorder circuit - to provide three ranges for Brown or other 0-10 mv recording device.




This pot. has been in test circuit & varied over entire range to give "feel" for operator points. Full 1000  $\Omega$  - high sensitivity on recorder. full scale equal to 10  $\mu$ a on meter - Good for drift check.

Referring to readings of 4-24 (p197), value of R for trij @ 2 3/4 f.s. was about 20  $\Omega$ . A range with this characteristic of on scale indication of trij is desirable.

An operator range with the trij point some known interval over full scale is needed.

To provide a trij, operator, background set of ranges, make up divider as follows, replacing 1000  $\Omega$  pot.:



Check above with Ra (5mg) source

Source Distance	Meter	Setting Range	R <sub>1</sub> Resistor	R <sub>2</sub> Resistor	Recorder	calc R <sub>3</sub>
24"	82	1	MIN	MAX	187.8	21.7/20
"	"	2	"	"	39	47.5/45
22"	96	2	"	"	45.2	47.1/45
"	"	1	"	"	20.3	21.1/20
20"	115	1	"	"	24.3	21.1/20
"	"	2	"	"	53.5	46.5/45
18"	138	1	"	"	29.2	21.2/20
"	"	2	"	"	64.5	46.7/45
16"	172	1	"	"	36.1	21/20
"	"	2	"	"	80	46.5/45
15"	200	1	"	"	41.8	20.9/20
"	"	2	"	"	92.3	46.1/45
14"	225	1	"	"	46.2	20.5/20
13"	258	1	"	"	52.4	20.5/20
12"	300	1	"	"	60.8	20.7/20
" 11 1/2"	330	1	"	"	66.8	20.7/20
" 11"	360	1	"	"	72.5	20.4/20
" 10"	428	1	"	"	85.5	19.1/20
" 9 1/4"	440	1	"	"	97.5	19.9/20

\* 11:45 set trij level for quartz dist. - Rlar. - Trij 4/60 - ma

4-25-57

NV 1000 (adj up to F-80)

Rain New Trip - paint

Source Distance	Meter	Range	R <sub>1</sub> max.	R <sub>2</sub> max.	Recorder (R <sub>2</sub> )	R <sub>1</sub>
27"	68	{ 2 2	max.	max.	32.8 19.3	48.3 28.4
25"	78	{ 1 2 2	max.	max.	17.2 37.5 21.8	22 48.8 27.4
20"	118	{ 1 2 2	max.	max.	25.2 58.8 32.0	21.4 47.3 27.4
18"	143	{ 1 2 2	max.	min.	30.5 38.6 67.2	28.3 27.0 27.1
Tripping 150 m <sub>2</sub>			max.	max.	42.3	20.6
15"	205	{ 2 2	max.	max.	94 53.8	45.3 26.2
12"	312	{ 1 2	max.	min.	63.8 81.2	20.4 26.0
11"	372	{ 1 2	max.	min.	75 92.7	20.2 26.3
10"	444	{ 1	max.	min.	89.0	19.9
AV to 900 E/166	25"	{ 1 2	max.	max.	7.8 14.4	25.8 43.1
15"	72	{ 1 2	max.	max.	16.2 34.8	22.5 28.4
10"	200	{ 1 2	max.	max.	42.2 91.7	21.0 35.9
6"	460	{ 1	max.	max.	91.8	19.9
800V	15"	{ 2			12.8	27.5
	9"	{ 2			20.9	42.5
	6"	{ 1			14.7	27.6
just about to trip	6"	{ 1			20.7	21.1
	5"	{ 2			68.6	47.4
	5"	{ 2			98.5	46.7
	3 1/2"	{ 1			44.7	21.0
	3 1/2"	{ 1			89.6	19.9

(below all max)

4/25/58

Large Rusk @ max

Source Dist	Meter	Range	Recorder	(Read)
AV-1100 very close to trip	29	{ 1 2	29	20.4
	142	{ 1 2	64	45
	15"	{ 1	95.8	19.9
Bkg - 1100 v -	38			
1000 v -	31.5 - 28.5			
900 v -	25			

4:05<sup>P</sup> set H@1020 F-80

4:10 - Chart on for weekend background results.

Monday - 4/28/58

No adjustment in meter zero has been made since last Monday morning, at which time meter was set at approximately 6 pa - Recorder reading 6 mv. across 1000 cm

Drift Summary

Date	Day	Time	Value	Day
4/21	Monday	8:15 AM	60	Tuesday
4/22	Tuesday	8:15 AM	59	
4/23	Wednesday	8:15 AM	50	
4/24	Thursday	8:15 AM	40	
4/25	Friday	8:15 AM	30	
4/26	Saturday	8:15 AM	29	
4/27	Sunday	8:15 AM	24	
4/28	Monday	8:08 AM	16	

4-9-58 With Co source near detector 60% recorder on operating  
 - added 34 1/2' RG62/U cable to N 7' already in use -  
 No change in level, indicating negligible attenuation -  
 - added 15' RG62/U - still no change -  
 Should be able to operate w/o pre-amp cathode follower -  
 unless speed of response is affected.

12:55<sup>PM</sup> HV set to 1100V. - Trips 17.5" - 3.5" OR  
 Bkg  $I_L = 1.2 \text{ ma}$ ,  $I_H = 9.75 \text{ ma}$ ,  $I_{FB} =$ ,  $I_T =$

La. source in position - over night -

4-30-58 Notes time for source had dropped from  
 50% @ 4:30 PM to  
 44.5% @ 8:00 AM )

Trips distances @ 15" & 3" →  
 Zero drift down to off scale.

Set bkg level to 45-60% by adj of Rip ccw rotation. slight  
 up to 265 shortly after adj.

Bkg avg 65 during day.

5-1-58 Over night drift negligible -

9:15 Trip Distances -

Low @ 17"  $I = 310-360 \mu\text{a}$   $I_L = 1.1 \text{ ma}$   
 $I_H = 8.3 \text{ ma}$   
 Rec = 70%

recorder fs @ Si Dist = 14 1/4"  $I = 50 \mu\text{a}$ .

HIGH @ 35'

5-2-58 No drift over night - low background most of day.  
 5-5-58 No drift over weekend.

Instrument in operation to 5-16-58, when cut-off.  
 Recorder kept of output - seems to be reasonably stable.

5-27-58 Keithly vacuum tube trip

8-2-58 PM instrument w/ sample power supply (had been used)  
 Regs regulated -

7-22-58 South wing counter - trouble w/ scales C1 & C3  
 bad detector C3 #1016  
 removed to shop 210 - Tube in use #371 from 108.



8-18-58

Notes upon return from vacation -  
West End station maintenance:

J.F.E

DC2 - DC3

DC3 when moved into "Set" found to be operating (responding to source)

DC2 - no source response - removed extension - still no response.

Scale selector check.

On sensitive scale recorder FS +

Lowered zero - (was all way up)

Operation, but wobbly

107- Monitor - slight response no trip - some resistor blown. (see file) Slightly noisy response

215 SA1 & 2 C<sub>2</sub> peak 400 very little change needed

SA2 Scaler same 16 light mil

6AL5 & 12AV7 V101-102 Strip (115)

Request for min. counters in 108 - One 238 & One 235 -

Got counters & made battery boxes.

P.Gen. .010V. 1ps, neg.

C<sub>2</sub> gain 2x1.8ps, neg PDL=54  
C<sub>3</sub> gain 4x1 18ps, neg PDL=40

Generate on input to pre-amp.

With counter no response at above settings

235 C<sub>3</sub> gain to 6V some response (RT 5ps)

238 C<sub>2</sub> excessive noise response at gain 6V.

Amplifier check

3: Preamp 4 6AK5's used

PHS V14 6AC7

Amp 1-6AC7, 2-6AQ7, 1-6SU7

C<sub>2</sub> PA: 3 6AK5.

Any 6AC7-1 ; 6AQ7-1 ; 6SU7-2 ; 6V6-1 ; 5V4-1

PH check C<sub>2</sub> RT. 8ps, gain 2x1 PHS 64

C<sub>3</sub> 4x1 PHS 55

None of C<sub>2</sub> seems to be eliminated -

(235 counter is one pulled & fixed by D.F.C.)

8-18-58

107- Further check out of C<sub>2</sub> & C<sub>3</sub> for Jinnie ctr. operation -

Jinnie counter, 5-2 attached to pre-amp for channel C-3. <sup>PA#335</sup>

Jinnie counter, 8-2 attached to pre-amp for channel C-2. <sup>PA#317</sup>

One channel dead, other counting, with source M-227 and plane very close to counters, & gain setting of 8x1, rise time 0.8ps, input neg, & PDL of 10 & 5 (C<sub>2</sub> & C<sub>3</sub>).

Brought down counter C-5-3 for additional check out. This known to be good.

Counter 5-3 & 5-2 seem to respond on either channels, indicating that amplifiers are apparently OK.

One gets no response to source with C-8-2, as follows

C<sub>2</sub> gain 8x1 Input Neg PDL 10  
C<sub>3</sub> gain 8x1 Input Neg PDL 5

5min. Pkly 0 + 0/6  
5min. Pkly 0 + 1/6

(Gain at PDL = 7) (Alpha 0.0. @ 2)

Source M-227 0 + 0/6  
28 + 3/6

C 8-2 to shop for check - Found center wire loose - Re-soldered center wire. - Counter in 108 for response check.

107-108 lesson counters -

8-18-58

C2 (PA 313)  
GEM 8x1 PDL=10  
RT 0.8µs.

C-5-2 : 29 + 15/16  
source in plane 28 + 19/16  
5 min c2 30 + 5/16

C3 (PA 317)  
GEM 8x1 PDL 5  
RT 0.8µs

C 8-2 : 17 13/16  
0 + 19/16  
1 + 19/16

PA 327 for 313 <u>C-8-2</u>	Pkg 0 + 1 GEM 8x1 PDL=10 RT 0.8µs	C-5-2 GEM 8x1 PDL 5
Source	0 + 12	28 + 12
	0 + 14	30 + 1
Bkg	0 + 7	0 + 2
	0 + 11	0 + 3
Source	0 + 14	29 + 13
	1 + 0	29 + 6
	source + bkg = 2.84µm	source + bkg = 94.6 µm
	bkg = 1.18 µm	bkg = .5 µm
	Source = 1.0 µm	Source = 94.1 µm

Ran with 8-2 & 5-2 on C2 & C3 changed PDL on C3 to 10 so Amplifier settings were identical.

During run noted register on scaler C-3 not following as it should. Transferred signal to scaler C-1 OK.

Service Note - Check output of scaler C-3.

8-18-58

DWM - reported failure of emergency exhaust dampers to open. - (this had been checked OK when serum systems were checked out)

Check w/ bldg prints showed safety switch in Rm 113. This was in "OFF" position. DWM threw "ON". Now fan on - dampers opened on third start-up. Apparently had been stuck shut.

DFC - et al. - Possible use of elephant gun safety on sphere experiment suggested. Holding power of magnet still in question -

8-19-58

West End - no tie-in for elephant gun - use 5rd magnet. DFC setting up to use Hall. Cannot change M-1 circuit.

Standardization of connectors for motors suggested.

Disconnect Box for instrumentation:

Existing channels:	Counter 5	Provide 6
	Ion Chamber 5	Provide 6
	P.M. Int 2	Provide 3

Counting Channel Connectors.

Signal	UHF-type
H.V.	BNC-
Power	AN-3102-165 (G)

Ion Chamber	-	HN-type
P.M.	Signal	UHF
	H.V.	BNC
	Power	AN 3102 (145) (G)

9-19

(107 - C3 scaler. Tube check output  
Replace 6A95, 12AU7, 6AV6 (?)

West End - safety magnet non-operable (Siel)  
Supply ok. Checked line w/o power open ckt  
near magnet.  
Put on substitute magnet.  
Repaired magnet.

8-20-58 ~~East end~~

West end - Work on sketch for junction box for 201.

South wing - With WCT & SWM observed rectifier  
113. Fan "out", just hums, with power  
applied

East wing: - Request set up for traversing fission counter  
Counter 5-2 & 8-2 to be used to travel  
 thru reactor, both to be positioned by single  
 drive mechanism.

w/SSC setup drive unit on frame of  
fixed half of table.

C-5-2 to C-3 channel

C-8-2 to C-2 channel

Brought down C-5-3 to use as  
normalizing counter. Mounted under core  
at core-reflector face on moving table.

8-20-58

C-5-3 feeding channel C-2.

During run, noted irregularity of counts on C-3.  
There was a sharp increase in scaling on scaler 3.  
Later, it appeared that the rate of count would fluctuate  
when supposedly there was a fixed neutron field.

By swapping leads from amplifier outputs, isolated  
trouble to scaler chassis.

In power supply replaced both rectifiers 6AX5  
and control tube (very bad) 6AS7

(had replaced 6AV6 earlier)

On input strip 118-#8 - replaced all tubes

V1-V4 6AH6

V5 & V6 12AT7

Strips 114 & 115 seemed to be normal, so no check  
of tubes made.

Overnight 60v test ok.

8-21-58 East end - Dealer test -

x 3000 c/s signal for 1 min counts

$$695 \times 256 + 31 = 177,951 \%$$

$$696 \times 256 + 161 = 178,337 \%$$

x 4000 c/s

$$443 \times 256 + 235 = 241,638 \%$$

8-21-58 With analog generator & pulse generator, checked counter channels.

A. Input sensitivity of scales

- C-1 +2 volts OK
- C-2 neg & pos ?
- C-3 +5 volts. OK

Brought down input strip #3 (11B) from 216 strip.  
(Could not get + only input with strip already in C2 #7)  
Set input strip for + pulses > 3 volts.

B. Signal from generators to scales only.

	C1 x256	C2 x16	C3 x16
2500N →	585 +148 (149,760%)	225+7 ← 60N →	225+1
60N	225+2 (2500N)	585+149	60 224+9
		586+13	224+12
60N	224+14	224+13	2500N 586+102

C. Signal input to L.A. input.

	L.A. setting 8x1 0.8µs neg in	Threshold
C1	PDL = 20	±0.02 v
C2	PDL = 10	very small
C3	PDL = 10	±0.02 v

	C1	C2	C3
2500 C3	224x16+13	224x16+12	586x256+91
2500 C2	224x16+3	584x256+193	224x16+2
2500 C1	586x256+25	225x16+2	224x16+14

8-21-58 During 1st 10 m. count at power .017 noted that scale 2 seemed to be "lurching" counts, throwing in more than expected, interpolator lights skipping.

Changed PDL for 10 to 15. These threshold was ±0.01 volts

West end - C2 register reported spurious response.

Tube check 5965 & 6A95 in output circuit.

6A95 OK.

5965 bad, Replaced.

Sketch for J-box 201 to WCT for fabrication

calculations of 8 sur of Jox chamber

8-22-58 West End  
Strated sur fusion probe =

Made up 3' cable to connect counter C-1 to pre-amp. Compared count with & without extra cable. There was no significant difference = limited in operation

Considerable to West End mod probe.

8-25-58 - Bring J.F.G. up to date on instrumentation activities in East End previous week.

- Conversation w/ "Doc" Smith re Savannah River visit.

- West End: (a) Consideration of instrument modification.  
(b) Request small diameter counter.

All enriched fission counters are tied up & TRF would like an additional small bore counter to replace C5-

$5.29 \times 10^5 \text{ n/sec}$

→ Mounted new Pu-Be source M-229 ( $5.29 \times 10^5 \text{ n/sec}$ )

for use as int. check source.

Prepare to check out RCL BF<sub>3</sub> & counter #1099 d1084

Auxiliary Equipment.

Pre-amp: BA #45B

Lui. Amp: BA Y123800

Scaler: BA Y123799


NVPS: BA Y123811

Approximate sensitivities:	Amps		Puls. Inj.		O.8 μs.	
	Gain	PDL	Gain	PDL	Gain	PDL
1 mV	4	15	8-30		16-60	
2 mV	4	30	8-60			
3 mV	4	45				
4 mV	4	60				
5 mV	4	75				

(From file BA Lui. Amp 204B Y123800)

8-25-58 Along with checking out counter, take readings on air chamber instruments.

Detector ~~(4x10<sup>-13</sup> a/d)~~  
 Small dia. BF<sub>3</sub> in 3" paraffin  
 Chamber E-100 (4x10<sup>-13</sup> a/d)  
 E-10 (3x10<sup>-13</sup> a/d)



"Bkg" Source 209	K-L	K-124029	K-124028
	$4 \times 10^{-13}$	$.25 \times 10^{-13}$	$1 \times 10^{-13}$
M 229 @ S	$2.1 \times 10^{-11}$	$7 \times 10^{-12}$	$3.6 \times 10^{-11}$
@ 1	$5 \times 10^{-12}$	$6 \times 10^{-12}$	$1.6 \times 10^{-11}$
@ 2	$8 \times 10^{-12}$	$6 \times 10^{-12}$	$.65 \times 10^{-11}$
@ 3	$2 \times 10^{-12}$	$1.5 \times 10^{-11}$	$3 \times 10^{-12}$

(File - RCL BF<sub>3</sub> & Ctr) Counter check  
 Sensitivity 1 mV from 4-PDL 15 (RT=0.8 μs) counter  
 Data in file

8-26-58 Center readin. of small BF<sub>3</sub> counter check (1 1/2 x 10" RCL).  
Mounts counter #1081 on connector adaptor.

Good curve data on #1081 - see file - { Op. voltage 1800V.  
Threshold - 10mV.  
w 9000 C/m M229 @ 4.5"

Connections for #1079 -  
Begin data on 1079

8-27-58 West End: Set-up C5-3 gamma counter on channels 5.  
East End: Reconnect gamma counter for 108  
C5-2 C8-2  
C5-3

Reconnect C-1 & C2 2" BF<sub>3</sub> & thro.

Continuation of check of BF<sub>3</sub> counter.

8-28-58 BF<sub>3</sub> Proportional Counter Characteristics  
New P-M circuit construction finished. Tests begun.  
(See file)

8-29-58 Jan Chamber Reading 8:05 M-229 & 5mg Ra in 209

KL -  $1.6 \times 10^{-13}$

K29 -  $.1 \times 10^{-13}$

K29 -  $< .1 \times 10^{-13}$


P-M circuit tests continuing - refer to P-M file.

8-2-58 Check drift of P-M circuit & spot check trip.  
Conference on self-multiplication - criticality  
experiment -  
Consideration given to new trip-output circuitry 202.

Output circuits for Keithley instruments -

Keithley  
log


Recorder output is 50 mV, same as for Beckman VL  
Raw output to network shown -



Balance recorder against panel meter, but may need  
to increase Rc.

Raw output to network as used in S. Wang & to purple pen  
of 2-sec Brown.

Keithley  
410



9-2-58 Noted that trace on K-28 was indicated large fluctuations compared to K-29. Trace on K-28 showed similar fluctuation, however not so extreme as on K-29. Chart record in file =

9-3-58 Check units - 8:45 AM.

KL output - removed 20k shunt of 5.6k resistor. This gave better balance, calibration point near center of pot. adjustment.

Output circuitry -

(Noted - fluctuation on K-28 channel 9-2 due to input cable & detector switched heads at inst. (see swapped channel.)


Disconnect E-100.

9-3-58 Set up for Self-Multiplication Measurements -

Apparatus: Source: Pa-Bz M-229  $5.24 \times 10^5 \text{ n/sec}$   
 Detector: Rch BF<sub>3</sub> filled proportional counter  
 1/2" d. (1800V op., 10mV sens.)  
 Cable: RG 62/U Length -  
 Pre-amp: BA 219B 45B  
 Counting Amp: BA 204B Y123800  
 Channel: Scaler: BA 134 Y123799  
 Tuner: B-A 960  
 Rate Meter: Y123878  
 HVPS: BA-318 Y-123871  
 Recorder: 10mV wheels

Ion Chamber: Detector: BF<sub>3</sub> filled chamber E-9  
 Channel: HV: 300V battery  
 Amplifier: Keithley 410 - Y124029  
 Recorder: 10 mV Brown.

Fixed Source - Detector




4-3-57

# Amplifier out to 10 mvr. Gain 2, RT 0.8 ps, PDL 73. | Long Cable

Long cable	HV	Scaler	CRM R
	1750	655 624 646 700 693	600
	1800	2234 2230 2239 2240	2100-2200
	1850	2544 2596 2565	2500
	1900	2673 2624 2750	2600
	1950	2743 2727 2722	2600

Operation point -

1900V



Remarks

HV=1900	3129	10m	Scaler	CRM R	B-710	2999.3	S: Air B: Aluminum
	3129	10m	23593	-	37.1 x 10 <sup>-13</sup>	2999.3	
		1m	32755	-		32755	

	1	3591*	
	2	3368*	
	3	3243	
	4	3224	
	5	3195	(2.75%)
	6	3074*	3214
2147	7	3137	±88.4
	8	3065*	
	9	3183	
	10	3050*	
	11	2998	
	12	2963	
	13	2953	
	14	2956	2939.8 ± 288%
	15	2820	
	16	2892	±84.7
	17	2946	
	18	3006	
	19	2990*	
	20	2944	68%

Short cable	#08	1m	Scaler	CRM R	ml - M
		1	2706	52	-21.77%
		2	2774	52.7	+47.17%
		3	2641*	51.4	-26.31%
		4	2734	52.3	+7.26%
		5	2730	52.3	+3.11%
		6	2733	52.3	+6.22%
		7	2792	52.8	+65.23%
		8	2727	52.2	0
		9	2734	52.3	+7.26%
		10	2699	52.0	-28.10%
411		1	2892	53.8	+55.21%
		2	2817	53.1	-19.87%
		3	2811	53.0	-25.89%
		4	2827	53.2	-9.81%

$$\sigma_r(90^\circ) = \sqrt{\frac{n-1}{n}} \times \frac{1.645}{\sqrt{m}}$$

$$= \sqrt{\frac{9}{10}} \times \frac{1.645}{\sqrt{2727}}$$

$$= 0.9487 \times \frac{1.645}{52.22} = \frac{1.5606}{52.22} = 2.99\%$$

$$95\% = \frac{1.9594}{52.22} = 3.56\%$$

S: air B: aluminum - 36%  
σ<sub>r</sub> = 2.28%

4125	5m	13833	2766.6
------	----	-------	--------

σ<sub>r</sub> air  
B: Al.



9-4-58

Amplifier 10mv rms Gain 2, RT 0.8%, PDL 73

HV = 1900

9:34 AM

Source placed @ 2" (as 9-3-58) at 4:00 PM

2nd HV to 1900

After running calculations on data of 9-3 began counts.

Series on 9:37

Series of 1-min counts - 10 to set.

Begin 9:40

	AVG	MERC. DEVIATION
2816		-22.1
* 2749		-89.1
2887		+48.9
2848	2838.1	+9.9
2862		+23.9
2832		-6.1
2832		-6.1
2878		+39.9
2825		-13.1
2852		+13.9
2799		-62.4
2876		+15.6
2811		-50.4
2831	2861.4	-36.4
2881		+19.6
* 2959		+97.6
2829		-32.4
2828		-33.4
2873		+11.6
10:42		
2827		-34.4
2971		

EXP. DEVIATION (90%)

$$\sigma_1 = \sqrt{\frac{9}{10} \cdot \frac{1.645}{\sqrt{M}}} = .9989 \times \frac{1.645}{\sqrt{M}} = \frac{1.561}{\sqrt{M}}$$

$$\pm \frac{1.561}{53.27} = \frac{1.561}{53.27} \times 10^2 = 2.930\%$$

$$p.e. = \frac{.6745}{\sqrt{n-1}} \sqrt{(n-a_1)^2 + (n-a_2)^2 + \dots + (n-a_n)^2}$$

$$= .2248 \sqrt{13578.96} = .2248 (116.27) = 26.14$$

$$\sigma_1 = \frac{1.561}{\sqrt{2861.4}} = \frac{1.561}{53.49} = 2.92\%$$

$$= \pm 83.6$$

received shipment

10:42

19%

1-4-58

Received Material: 3 Tuballoy billets for use as blanks  
2 Oralley billets for multiplication determination

Depleted material to 210 in carriers.

Move source closer to counter housing - d = 1 7/8"

A. Counts with no material - Count of one slab away

One-minute Count	dev
1 3641	+54.9
2 3541	-45.3
3 2460	-126.3 *
4 3569	-17.3
5 3686	+44.9
6 3698	+111.7 *
7 3630	+43.7
8 3500	-78.3
9 3585	-1.3
10 3585	+81.9
<hr/>	
3586.3	

$$J = .9847 \times \frac{1.645}{\sqrt{3586.3}} = \frac{1.561}{59.89} = 2.61\%$$

$$= \pm 93.6$$



B: Tuballoy slab - 8x8x1

One-minute Count	dev
1 7206	+22.1
2 7259	+75.1
3 7175	-8.4
4 7175	-8.4
5 7157	-26.9
6 7119	-64.9
7 7143	-40.9
8 7171	-12.9
9 7249	+116.1
10 7135	-48.9
<hr/>	
7183.9	

$$J = \frac{1.561}{\sqrt{7183.9}} = \frac{1.561}{84.76} = 1.84\%$$

$$= \pm 132.18$$

+212.3  
-212.3

12:26 Repeat A = 3508

12:31 Repeat B = 7145

C: Same @ edge slab




6285  
6215  
6199

9-4-58

12:37 D. slab edge 1" - source tip  
E. slab edge 2" - source tip

6641  
6309  
6630  
6927  
6988  
7149



F. Tuballoy slab 8x8x1

One-minute Count	dev
1 7012	+3.1
2 7014	+8.1
3 6949	-24.9
4 6949	-54.9
5 7039	+30.1
6 7111	+102.1
7 6953	-55.9
8 6913	-45.9
9 7086	+77.1
10 7033	+24.1
<hr/>	
7008.9	

$$J = \frac{1.561}{\sqrt{7008.9}} = \frac{1.561}{83.72} = 1.86\%$$


$$= \pm 130.68$$

Repeat A 3457  
3558  
3479

Check F 7098  
7096  
7008

1:18 10mm air count A - 35940

1:55 Oralley -




One-minute Count	dev
1 10573	+17.4
2 10488	-67.6
3 10612	+56.4
4 10709	+153.4
*5 10268	*-287.6
6 10569	+11.4
*7 10388	*-167.6
*8 10763	*+207.4
9 10611	+55.4
10 10577	+21.4
<hr/>	
10555.6	

$$J = \frac{1.561}{\sqrt{10555.6}} = \frac{1.561}{102.74} = 1.52\%$$

$$= \pm 160.4$$

9-4-58

Oralley - small



$5.5 \times 10^{-13}$

1	9874	-89.7
2	10003	+40.3
3	10106	+143.3
4	10039	+76.3
5	10036	+73.3
6	9939	-23.7
7	9954	-8.7
8	9812	-150.7
9	9927	-35.7
10	9937	-25.7
<hr/>		
9962.7		

$$\sigma = \frac{1.561}{\sqrt{9962.7}} = \frac{1.561}{99.81} = 1.56\%$$

$$= \pm 155.8$$

Summary

		kg		$\sigma$	
i. Oralley	7x8x1	17.226	9962.7	$\pm 1.56\%$	$\pm 155.8$
ii. Oralley	8x9 1/2 x 1	23.522	10555.6	$\pm 1.52\%$	$\pm 160.4$
iii. Suballey	8x8x1		7183.9	$\pm 1.84\%$	$\pm 132.2$
iv. Suballey	8x9 1/2 x 1		7008.9	$\pm 1.86\%$	$\pm 136.7$
Air		N	3586.3	$\pm 2.61$	$\pm 92.6$
2" Lead			6000		
i/ii	=		1.3868		
ii/iv	=		1.5060		

$$\text{wt } \frac{O_{ii}}{O_i} = 1.3655 \quad \frac{O_i}{C} = 1.0595$$

9/4/58

4:05 1" Lucite between source & counter

1	6391
2	6382
3	6417
4	6350
5	6398
6	6401
7	6252
8	6215
9	6435
10	6452

Noted sharp drop in CPM record?

9/5/58

Continuation of Self-Multiplication Tests -

- Procedure -
1. Determine optimum shield thickness for counter -
  2. Repeat air, suballey, oralley runs, with sufficient spacing between source & shield to insert 1" Lucite -
  3. Step 2 with 1" Lucite

Spot check - wipe test: inside knob 210 OK  
 handle cart OK  
 2 spots on floor near suballey OK

9:17 HV on - 1900v  
 Scale on

HV Power Supply - fluctuates  
 Trouble seems to be in connection adapter at power supply.

Made new HV cable to fit, BNC on both ends - RG 59/U.


10:20 PM circuit on - see file -

HV on counter - Maybe ready

This may be answer to above?

2730

Large Orally sample



6723  
6724

$$\sigma_{90} = \frac{1.521}{81.89} = 1.91\% = \pm 128.1$$

6705.9

6705.9 ± 128.1

Add 1" limits between source & slab,  
completely

$$\sigma = \frac{1.521}{87.81} = 1.78\% = \pm 137.3$$

7711.0 ± 137.3

Add 1" limits between slab & shield.


$$\sigma = \frac{1.521}{88.38} = 1.77\% = \pm 138.2$$

7811.9 ± 138.2

315

Spot blank 3924/cm

Small Orally sample



6601.7 ± 126.9

$$\sigma = \frac{1.521}{81.25} = 1.87\% = \pm 126.9$$

Spot blank 3877/cm

Small Orally sample



5842.2 ± 119.2

$$\sigma = \frac{1.521}{76.43} = 2.04\% = \pm 119.2$$

One inch limits between source & slab

dev	Counter	K	KL
✓	6589		
✓	6655		
✓	6772		
✓	6757		
✓	6741		
✓	6667		
✓	6756		
*	6843		
✓	6676		
✓	6603		
✓	7616		
✓	7671		
✓	7613		
✓	7696		
✓	7643		
✓	7738		
✓	7650		
*	7881		
✓	7754		
✓	7848		
✓	7848		
✓	7776		
✓	7831		
✓	7761		
✓	7836		
✓	7817		
✓	7839		
✓	7845		
✓	7800		
✓	7768		
✓	6528		
✓	6714		
✓	6465		
✓	6574		
✓	6592		
✓	6571		
✓	6497		
✓	6507		
✓	6601		
*	6768		
✓	5873		
✓	5876		
✓	5808		
✓	5928		
✓	5738		
*	5710		
✓	5891		
✓	5945		
✓	5850		
✓	5803		

KL 4:12 1" Lucite between source & slab.

$$\Delta_{12} = \frac{1.561}{81.19} = 1.92\% = \pm 126.6$$

6592.2 ± 126.6

Add 1" Lucite between slab & shield.

6679.2 ± 126.3

4:32

4:39 Check blank - 3869 ± 1m

Count	dev.
6589	✓
6485	✓
6448	*
6618	✓
6586	✓
6711	✓
6598	✓
6647	✓
6682	✓
6608	✓
6544	✓
6713	*
6461	✓
6669	✓
6553	✓
6550	✓
6521	✓
6599	✓
6565	✓
6619	✓

224 Ch: HV-1900 V.S. = 10mv.  
 K-410 3210-12  
 9-5-58 PM MOD N. B. 4

S-c dist 0.711  
 40% K<sub>α</sub> 1.19  
 80%

Condition  
 source M229 6" & source to  
 10:30 PM<sub>3</sub> etc w/1" shield edge of shield

$$\boxed{944.5 \text{ c/m}}$$

$$\Delta_{90} = \frac{1.561}{\sqrt{944.5}} = \frac{1.561}{30.81} = 5.07\%$$

$$944.5 \pm 48.1$$

$$\Delta_{90} = \frac{1.887}{30.81} = 6.13\% = \pm 57.25$$

Scaler	dev	K	K <sub>L</sub>	PM
807	+ 17.5			
807	- 52.5	5.7 × 10 <sup>-13</sup>	5.7 × 10 <sup>-13</sup>	45
845	- 4.5			
961	+ 11.5			
944	- 8.5			
981	+ 31.5			
934	- 15.5			
943	- 6.5			
1004	+ 54.5			
919	- 30.5			

Add 1" lucite @ source N 1/2"

$$\boxed{1680.1}$$

$$\Delta_{90} = \frac{1.561}{\sqrt{1680.1}} = \frac{1.561}{40.99} = 3.81\% = \pm 64.0$$

11:07

$$1680.1 \pm 64.0$$

1700	+ 19.9			
1650	- 30.1	5.7 × 10 <sup>-13</sup>	5.7 × 10 <sup>-13</sup>	45
1646	- 34.1			
1655	- 25.1			
1697	+ 16.9			
1717	+ 36.9			
1715	+ 34.4			
1695	+ 14.9			
1652	- 28.1			
1674	- 6.1			

Add 1" lucite @ 

$$\boxed{2002.2}$$

$$\Delta_{90} = \frac{1.561}{44.75} = 3.49\% = \pm 69.9$$

$$2002.2 \pm 69.9$$

2013		3.5 × 10 <sup>-13</sup>	4.8 × 10 <sup>-11</sup>	45
1979				
2005				
1984				
1945				
1945				
2131				
2025				
2032				
1963				

Add 1/2" lucite -


$$\boxed{2010.4}$$

$$\Delta_{90} = \frac{1.561}{48.84} = 3.2\% = \pm 64.96 = \pm 70.0$$

$$2010.4 \pm 70.0$$

2042		3.5 × 10 <sup>-13</sup>	4.8 × 10 <sup>-11</sup>	45
2013				
1957				
2047				
* 2089				
1978				
2013				
2014				
1997				
1956				

11:44  
 S-C dist 0.711  
 source

Ch in 3" paraffin shield 

$$\Delta_{90} = \frac{1.561}{63.43} = 2.46\% = \pm 99.0$$

$$4023.4 \pm 99.0$$

4029		5 × 10 <sup>-13</sup>	8.6 × 10 <sup>-11</sup>	45
3980				
4094				
* 3894				
4100				
4014				
3960				
4053				
* 4157				
3950				

3" shield not on bench source not in stamp.  
 Table built up to a bench level -

311 Row

PM

Add - 1" lucite



dev

Counter

R

KL

PM

$$\sigma = \frac{1.561}{63.98} = 2.44\% = \pm 99.9$$

4093.7 ± 99.9

- ✓ 4091
- ✓ 4044
- ✓ 4105
- ✓ 4102
- ✓ 4100
- ✓ 4089
- ✓ 4046
- ✓ 4087
- ✓ 4145
- ✓ 4128
- ✓ 4027

4.4 × 10<sup>-3</sup>

8.3 × 10<sup>-13</sup>

5

Add 1/2" lucite

$$\sigma = \frac{1.561}{61.88} = 2.52\% = \pm 96.5$$

3828.8 ± 96.5



- ✓ 3790
- ✓ 3806
- ✓ 3878
- ✓ 3894
- ✓ 3866
- ✓ 3793
- ✓ 3746
- ✓ 3810
- ✓ 3884
- ✓ 3821

4.5 × 10<sup>-3</sup>

9.8 × 10<sup>-13</sup>

1" more lower.  
3249  
KL 8.3 × 10<sup>-13</sup>  
K 3.4 × 10<sup>-13</sup>

Remove lucite =



$$\sigma = \frac{1.561}{69.03} = 2.26\% = \pm 107.7$$

4765.3 ± 107.7

- ✓ 4810
- ✓ 4826
- ✓ 4772
- ✓ 4712
- \* 4649
- ✓ 4815
- ✓ 4703
- ✓ 4745
- ✓ 4870
- ✓ 4757

4.4 × 10<sup>-3</sup>

8 × 10<sup>-13</sup>

5

Re-printer 8x8x1



$$\sigma = \frac{1.561}{65.47} = 2.38\% = \pm 102.0$$

4286.4 ± 102.0

- \* 4150
- \* 4112
- ✓ 4309
- ✓ 4277
- ✓ 4287
- ✓ 4321
- ✓ 4244
- ✓ 4294
- ✓ 4239
- ✓ 4283

Add 1" lucite between Thallium & source

$$\sigma = \frac{1.561}{63.74} = 2.45\% = \pm 99.6$$

4063.3 ± 99.6

- ✓ 4057
- ✓ 3979
- ✓ 4064
- ✓ 4002
- \* 4166
- ✓ 4103
- ✓ 4133
- ✓ 4081
- ✓ 4061
- ✓ 3989

~~3" shield~~ Remove 1" lucite  
Sub close 2 3/8" shield  
Cent away - an only 3" shield

2:21

- ✓ 4241
- ✓ 4726
- ✓ 3930

228.

9/5 Disconnect Chempo ap (<sup>w-w</sup><sub>b-b</sub> gm-frame) Connect - proton pump -

9-8-58

Am - Labelation & self-multiplication data -

Pm -

West End - C-2 head reported sputtering when in  
mode -

Log N - sick -

DC-2 - when cut off, from 9-5 did  
not get ff scale metric indication, as expected, output  
batteries are weak, I guess.

C2: Head connector - UHF - on pre-amp loose -

Upon removing cover to pre-amp noted input  
capacitor HV lead screw over to frame.

New input capacitor & HV dropping resistor -

9-9-58

Am - Conference on new table design - 9-11/55

Talk w/ Wichter re - self-multiplication -

Pm + 1/2" etc on C-1 - 12' cable jargon made up.

New resistor -

C-1 not count

A-2 throw in count -



9-10-58 Lay-out for new ponds 202

C-1 check no signal from detector  
loose center lead at connector head. Fixed.

West End Camp Dam - Month of October

{ Instrument & Equipment files to be maintained in 202.  
{ Summary log in book - Details in files -

11-6 West wing - Instru maintenance  $\frac{1}{2}$   
South wing - firm counter check  $\frac{1}{4}$

East wing - control rod syph -  $\frac{1}{2}$

11-7 South wing - new control rod counter drive - system A  $\frac{1}{4}$

West - Instrumentation

~~11-5~~

11-10-58 West - Instrumentation files up to date - files now in 202.

- DWM - possibilities of more functions - maybe W'house

- Temperature control SIA 204, etc

Controller ok. Demand at 77 reading 77

Set demand to 80° - Temperature recorder in 204

@ 1:35 P

- Decade scale 248

11-11-58 Scales 248<sup>th</sup> - Counting Room Scan for RBG in 7th

11-12-58

Self-Multiplication Measurements.

12-1-58

Record of data on insert opposite.

Run only preliminary counts to check system.

Calculation of variation.

Expected deviation  $\sigma_1$  (90%) =  $\sqrt{\frac{N-1}{N}} \times \frac{1.645}{\sqrt{M}}$   
 $= \sqrt{\frac{9}{10}} \times \frac{1.645}{\sqrt{M}} = .9487 \times \frac{1.645}{\sqrt{M}} = \frac{1.561}{\sqrt{M}}$

N = no. of observations  
M = avg.

(STANDARD STATISTICAL ERROR  $\sigma_1 = \frac{\sigma}{\sqrt{N}} = \sqrt{T}$ ; Probable error .67  $\sigma_1$ )

Case 1. Source & counter (in shield) only

COUNTS - EACH ONE MINUTE

M = 31524

$\sigma_1 = \frac{1.561}{\sqrt{31524}} = \frac{1.561}{177.55} = \pm .879\%$

$\pm 277.1$

Range 31801.1  
31246.9

COUNT	ORG. VARIATION	%
31460	- 64	
31430	- 94	
31520	- 4	
31700	+ 176	
31420	- 104	
31480	- 44	
31600	+ 76	
31440	- 84	
31620	+ 96	
31570	+ 46	

Case 2. Source & counter with cart in place. No sample.

M = 33086

$\sigma_1 = \frac{1.561}{\sqrt{33086}} = \frac{1.561}{181.9 \times 10^2} = \pm .858\%$   
 $\pm 283.9$

Range 33370  
32802

Ratio 3/1 = 1.050

33070	- 16	
33030	- 56	
33030	- 56	
32980	- 106	
32990	- 96	
32970	- 116	
33030	- 56	
33170	+ 84	
33110	+ 24	
* 33480	+ 394	


source strength reported  $1 \times 10^6$  @ 6-3-58

12-2-58 336  
184  
382 days

12-1-58

Equipment for self-multiplication measurement -

Detector: RCL Mod 10520, Ser 1771 13 $\frac{1}{2}$  counter Op Voltage 2" x 12" 1550V.  
Pre-Amp: #641  
Amplifier: Y100363  
Source shield set at 4mm: RCL Junc 0.8 $\mu$ c Gain 4, PDL 32.



210	Counts w/o cart. source & shield only	w/ cart no slabs	215	3:00
			w/o cart	Background source removed
1	3146 -64	3307 +16	3159	30
2	3143 -94	3303 +56		
3	3152 -4	3303 -56		
4	3170 +176	3298 -106		Note 12/2: Counts should be 31460 c/m, etc. NOT as recorded 3146 etc
5	3142 -104	3299 -96		
6	3148 -44	3297 -116		
7	3160 +76	3303 -56		
8	3144 -84	3317 +84		
9	3162 +96	3311 +24		
10	3157 +46	3348 +394		

Time	8:07	8:14	8:53	9:10	10:19	10:34	10:55	11:15	11:34	11:50
Source	MV am- 1550 v.	300 counts	30 cm	Cart in place	source only	Cart in place	Cart in place	Cart in place	Cart in place	Cart in place
Notes	Other equipment also energized.	200 counts background. Amp: G=4, PDL=32, RT 0.8	Same in place. Attempted to reproduce location of 12-1-58.	NOTE 1 (slabs in Rm)	NOTE 2 (slabs in Rm)	NOTE 3 (slabs in Rm)	NOTE 4 (slabs in Rm)	NOTE 5 (slabs in Rm)	NOTE 6 (slabs in Rm)	NOTE 7 (slabs in Rm)
Sample	source only	3241 x 10 <sup>4</sup>	3252 x 10 <sup>4</sup>	3252 x 10 <sup>4</sup>	3839 x 10 <sup>4</sup>	5648 x 10 <sup>4</sup>	6090 x 10 <sup>4</sup>	6118 x 10 <sup>4</sup>	6131 x 10 <sup>4</sup>	6131 x 10 <sup>4</sup>
(m) Avg.	3.0761 x 10 <sup>4</sup>	3.2317 x 10 <sup>4</sup>	3.1334 x 10 <sup>4</sup>	3.119	3.824	5.621	6.094	6.180	6.180	6.218
Time	12:06	1:07	1:22	1:41	1:58	2:20	2:36			
Sample	Cart in place	Cart (Cd shield)	Tubelloy Cd	Oralloy Cd	Oralloy Cd	Oralloy Cd	Cart Cd			
Sample	3.293 x 10 <sup>4</sup>	3.290 x 10 <sup>4</sup>	3.814 x 10 <sup>4</sup>	5.785 x 10 <sup>4</sup>	5.784 x 10 <sup>4</sup>	5.337 x 10 <sup>4</sup>	3.321 x 10 <sup>4</sup>			
(m) Avg.	3.293	3.281	3.786	5.806	5.779	5.380	3.320			
Time	12:19									
Sample	3.313	3.308	3.780	5.766	5.760	5.333	3.300			
NOTE 1*	7x8 on cart @ 4' others ≥ 8'									

\* 33480 + 394

Source Counter	3" paraffin cylinder	EF <sub>3</sub> counter, 3" para 4" cylinder + Cd sheet	
A	m = 31334 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 882%	31560 + 226 31910 - 24 31120 - 214 31280 - 54 31330 - 4 31600 + 266 31620 + 286 30910 - 424 31390 + 56 31220 - 114 32020 - 275 32420 - 325 32490 - 5 32820 + 25 32710 - 85 32580 - 215 33090 + 275 32970 + 175 33060 + 265 32940 + 145	32400 - 79 32810 - 169 33180 + 201 32860 - 119 32700 - 279 32910 - 69 33050 + 71 32980 + 1 33320 + 341 33080 + 101 38140 + 110 37860 - 170 38140 + 110 38230 + 200 38140 + 110 37680 - 350 38320 + 290 38380 + 350 37610 - 420 37800 - 230 53370 - 309 53800 + 121 53690 + 11 53550 - 129 53790 + 111 53490 - 189 53900 + 201 53820 + 141 53330 - 349 57840 + 14 57790 - 36 57600 - 226 57520 - 306 57740 - 86 58050 + 224 58180 + 354 57790 - 34 58150 + 324 57600 - 226 57850 - 20 58060 + 190 58000 + 130 57810 - 60 57850 - 80 57560 - 310 57790 - 80 57850 - 20 58270 + 400 57660 - 210 33210 + 207 33200 + 197 32680 - 323 33160 + 153 32850 - 153 33020 + 17 33130 + 127 32740 - 21 33040 + 33000 -
B	M = 32795 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 862%	32790 - 5 32820 + 25 32710 - 85 32580 - 215 33090 + 275 32970 + 175 33060 + 265 32940 + 145	M = 32979 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 860%
C	m = 38211 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 798%	38150 - 61 38190 - 21 38000 - 211 38220 + 9 38110 - 101 38160 - 51 38370 + 159 38280 + 69 38240 + 29	m = 38030 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 801%
D	M = 56824 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 655%	56480 - 344 57110 + 286 57380 + 556 56680 - 144 56680 - 144 56990 + 156 56810 - 14 56920 + 96 56710 - 114 60900 - 182 61180 + 98 61040 - 42 61010 - 72 61220 + 138 61630 + 548 60580 - 202 61150 + 68 60870 - 212 60940 - 142	M = 53679 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 674%
E	m = 61082 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 632%	61650 - 130 61310 - 470 61780 + 90 61970 + 190 61760 - 20 61400 - 300 62080 + 300 61800 + 20 62180 + 400	M = 57826 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 649%
F	m = 61780 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 628%	61650 - 130 61310 - 470 61780 + 90 61970 + 190 61760 - 20 61400 - 300 62080 + 300 61800 + 20 62180 + 400	M = 57870 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 649%
G	m = 33031 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 859%	32930 - 101 32820 - 211 33010 + 4 33240 + 209 33220 + 189 32940 - 41 33210 + 179 32740 - 291 32990 + 41 33130 + 99	m = 33003 σ = 1.561 σ <sup>2</sup> = 1.9334 x 10 <sup>4</sup> = ± 859%

12/18 Inj Honeyak Bolton - using LA V100363 ~~420-442~~  
 PA 641

Gain 4, Res - June = 0.8 μs  
 Notes 5mg Ra out-off at 108 PDL - Set PDL @ 15

3:39 Oralloy 3 (2x10%)  
 source - cti det = 1/8"

2.48 x 10 <sup>3</sup>
2.47
2.36
2.37
2.33
2.39
2.42
2.40
2.53
2.48

2423 c/m

$\frac{a-b}{b} = 1.062$

3:54 Background -  
 cart in place

2.36 x 10 <sup>3</sup>
2.32
2.25
2.27
2.21
2.30
2.21
2.26
2.26
2.37

2,281 c/m

12-4-58 Honeyak Bolton check -  
 Raw count of 10 min into counter.  
 Pb-Po source bare & shielded with 2" of lead.

LA 100363 Gain 4, RT 0.8 μs  
 PA 641 HV = 850  
 Background, with Ra, and with

	PDL	c/m		PDL	c/m	Ratio b/a
⊙ Background	5	9	⊙ 5mg Ra @ 3"	5	18984	
	10	0		7	102	
				9	3	
				10	0	
3 a. PN 215 @ 3"			b - PN 215 @ 3" + 2" Pb			
	c/m	PDL		between c/m		
	5,358	5		3,892	.726	
	2,902	7		2,228	.768	
	2,179	9		1,651	.758	
	1,186	15		906	.764	
	845	20		645	.763	
	459	30		365	.798	
	277	40		225	.81	
	169	50		134	.794	
	73	70		58	.795	
	36	90		25	-	

Should try bottom with bias of 240. This should get most gamma - counting rate rather low

12-4-58 Program to see if change in removal of plaster work.

Setup: 2x12" BF<sub>3</sub> counter. PA: 44B, LA Y123800

4mv sensitivity - Gen 4, PDL 50, RT 0.8

14V = 1550. c/m

Source @ 3" from shield - (10 minute counts)	30,881
, with cart in place	52,415
"	32,443
cart removed	30,968

12-5-58 Same as 12-4 except feed output of LA at AMP OUT - Scale

P45 @ +5.

10 minute count: Source & shield

Cart in place 9:06

9:36

9:48

c/m

30,540

30,521

32,031

31,905

32,035

32,173

32,215

32,246

32,220

32,227

Or-3/cert = 1.892 Or-3 (8x10<sup>4</sup>)

{ 60,798  
61,055  
61,035 }

60,963

Or-3 Cd sheath

{ 57,556  
57,443  
57,470 }

57,490

Or-3 Al cover Cd

{ 56,544  
56,835  
56,767 }

56,715

Source & shield - Cd

30,993

30,993

Power Supply failure @ 1:15 - Stop to fix.

One minute counts: 2:20 Turn machine - stop to fix.

Hour Condition

3:24 Source-shield - Cd

{ 3.093 x 10<sup>4</sup>  
3.060 x 10<sup>4</sup>  
3.074 x 10<sup>4</sup> } 3.0757

3:40 Or-3, Al - Cd

{ 5.658 x 10<sup>4</sup>  
5.682 x 10<sup>4</sup>  
5.640 x 10<sup>4</sup> } 5.6600

3:44 Or-3, Al - 1/16" plexiglas, top Cd

{ 5.679 x 10<sup>4</sup>  
5.736 x 10<sup>4</sup>  
5.698 x 10<sup>4</sup> } 5.7043

Or-3, Al - 1/16" plexiglas, under Cd

{ 5.632 x 10<sup>4</sup>  
5.696 x 10<sup>4</sup>  
5.682 x 10<sup>4</sup> } 5.6717

3:55 Source-shield - Cd

{ 3.033 x 10<sup>4</sup>  
3.091 x 10<sup>4</sup>  
3.083 x 10<sup>4</sup> } 3.0690

Remove Cd.

4:00 Source-shield

{ 3.056 x 10<sup>4</sup>  
3.083 x 10<sup>4</sup>  
3.086 x 10<sup>4</sup> } 3.0750

4:05 Or-3, Al 1/16 plexiglas, under

{ 6.067 x 10<sup>4</sup>  
6.026 x 10<sup>4</sup>  
6.040 x 10<sup>4</sup> } 6.0443

4:09 Or-3, Al

{ 6.082 x 10<sup>4</sup>  
6.070 x 10<sup>4</sup>  
6.092 x 10<sup>4</sup> } 6.0813

1.886

4:14 Cart in place, bkg

{ 3.219 x 10<sup>4</sup>  
3.216 x 10<sup>4</sup>  
3.237 x 10<sup>4</sup> } 3.2240

Source-shield

{ 3.083 x 10<sup>4</sup>  
3.116 x 10<sup>4</sup>  
3.061 x 10<sup>4</sup> } 3.0867

2-13-59

South ◆

D.W.M.

Why to be "down" for 10 days.

Projects: ① - Inventory chambers - if two enriched are in area, put both enriched on log channels.

② - Run counting tests on counting channels to check scatter observed during traverse runs.

③ - Make up a second panel for pump control

④ - Re-arrangement of meters -

⑤ - A 2-pen Brown would be nice for both beam instruments - (Frederick Westman for LCRM)

⑥ - Means of compensating out gamma on period measuring instruments be nice - 2 chambers, backing, perhaps -

Records: { Bot Appl -

{ Ernie Lees - 9201-3

⑦ - Make up PM operational to replace existing unit.

2-13-59

General: Inevitable water in mounts  
 → for East & West ends VGH.

→ Mount for solution storage of 105 vials.  
 → East change in Seram panel -

Call: Howard Campbell re: intercom ✓  
 Appel & E. L. re: records.  
 6-1198 7744  
 920123

2-16 Keith Martin 9:10 re intercom - A. Mills.  
 will check -

Set up suggested activities S. area for week 3/16 - VGH.  
 DFC colonimeter (b/c)  
 call for RIDL re: gamma - re. other equipment.

2/17

8:44 Healy re-purchase eq. counter equipment.  
 DFC req -

2/18 DFC req - VGH &amp; PM

Bills on counter equipment.

2/19 DFC req - West clean up -

2/20 West -

ORDER - TUBES for w/end

- NATI crystal for PM clean -

9-10-53: Scintillation Detectors:

RCA-6655-A with NE404 - fast neutron phosphor mounted

HV = 900v - Signal to PM-2 -

With source in contact - Brown reads 21 - operate range

Background @ 900v - 27 - background range

800 - 11

700 - 5.5

Neutron source in contact  $5.29 \times 10^5 n/sec$  M-229 PuBe

700v - 1.5

850 3.5

800 7

850 13

900 21


5mg Ra 900 66

@ 3.5" 850 36.5

800 22

750 12

700 5.5



9-22-53

Scintillation Detectors:

Model: Phosphor: NE404  $\eta$  Amplifier: Y123801 & 45B.

PM : 6655-A

HV = 40 ma.

5mg Ra on top of detector housing.

GAIN 8, RTO.8

PDL

3 1149080 1154930

4 273136 272910

5 29330 38660

6 7010 6960

7 1170 1210

8 260 270

9 40 50

10 20 20

3 2760/28M

4 1200/42M

5 320/14M

10 20/14M

3 621710/4M

5 15810 16390

10 4600 4620

15 2600 2440

20 1480 7560/EM

25 1820/EM

30 22970/40M

3 8810 9900

4 250/3M

5 40/6M

5mg Ra

10/20M

M

~~20~~ - NO PHOSPHOR

5mg Ra

PDL 10 - ~~10~~

10/20M.



Bg - 10 Phosphor			
Nat	PDL		
Gen 8	0.8 RT	50	199,770 201,820
	0.2 RT	6	963,190 968,590
		8	868,950 867,830
		10	677,480 678,150
		12	449,430 500,520
		14	375,480
		16	292,360
		18	218,520
		20	164,620
		22	134,340
		24	120,410
		26	105,930
		28	92,720
		30	77,020
		32	64,680
		34	58,640
		36	52,580
		38	46,800
		40	42,220
		45	33,230
		50	25,120
		55	17,990
		60	12,990

5<sub>1/2</sub> Ra @ 1 meter.

9/22

Some in 20' PDL so gen 8 RT=0.2.  $N 2.9 \times 10^3$

M229 & Bg both produce response.  $M 229$  at top  $9 \times 10^4$

Ra @ 20' on top rank 20" above floor  $2.9 \times 10^3$   
on floor  $5 \times 10^2$

Bkg - both same in 20' in shield -  $2.15 \times 10^2$  cm.  $9^{20} = 184 \%$


9-23-57 (3) Detector in C3 - NE404-66357A HV=1.4 Gen 8 - RT=0.2 PDL=25  
1m. count

(1) Min Bg in C1 can 16 RT=0.8, PDL=10 HV=500

Some of (1) in per. can - PM.

Scintillation Detector: NE404 66357A - Channel 3 -  $M 226 (N 18)$

HV=A	Gen 8 RT=0.2	PDL=25	Bkg
			3129
			3107
			3201
			3381
			3410
			3364
			3148



11868
12296
11669
11772
12533

$M 229$   
on top  
of PM

12262
11999
11220
12249
11525
11763
11844

4608
4266
4166
4514
4454

$M 226$  in can

20	26
17	21
22	28

Bkg (226 in 102  
229 @ 20')

9-23-59

M 229 @ 18"	227	336	432	287
Gain 8, 0.2 RT	255	253	268	
PDL = 25	270	280	418	

Bkg (10min) %: 18.2  
19.0  
17.9

229 @ 201

15.0

3:49

BF3


M 226 in contact 10V < 1000 Gain 18, RT 2 PDL 10

refer ORNL

9/24/59

$HV = 0.42 \mu\text{a} @ 745 \text{ volts}$

Source M-226 ( $9.30 \times 10^6 \eta/\text{sec}$ )



ARRANGEMENT

phosphor  
NE 404

Source at  $d = 4\frac{1}{2}$ " with 3" paraffin

45,157 c/m ; 43,855 c/m ; 43,437 c/m

Source at 22 ft "Bkg"

116 ; 90 ; 90

M 226 @  $d = 5$ " with 3" paraffin

91,095 ; 87,773 ; 88,194

M 226 @ 5" no paraffin

147,172 ; 147,878 ; 144,764

NE 400

gain 8 RT = 0.2 PDL = 40  
"Bkg"

286 ; 257

M 226 5" with 3" paraffin

85,550 ; 86,133 ; 85,235

M 226 @ 5" w/o paraffin

16,134 ; 16,573

Source in 102

19 ; 18 ; 18



9/24/59 Source in 202 @ 22'

305 307 318

9-25-59 2:10 Begin series with slow neutron phosphor NE400 -

HV=900 - Gain 2 RT. 0.2.

Source M-226 in 3" paraffin

can w/2"plexiglas reflector s.d. = 5"

5mg Ra at 22'

HV=900 PDL

C3 10

1 = 1" BP @ 1500v @ 4mv i.s.  
inside rock of main  
test stand

2 = NE400 @ Am  
HV=900 M226  
d = 5" @ 12

Gain 2, RT. 0.2 PDL 10

(A) 1	12560
2	18761140
1	12970
2	18955660
1	13110
2	18545880
1	1300015
2	2863750
1	12790
2	2791270
1	13080
2	2745190
1	12690
2	2617780
1	13560
2	1422000
1	13270
2	1422260
1	12990
2	1441210
1	12710
2	1159090

2	1027960
1	12510
2	950330
1	12830
2	944110
1	12670
2	895830
1	12600
2	895210
1	12150

9/24/59 Source in 202 @ 22'

305

307

318

9-25-59 2100 Began series with slow neutron phosphor NE400 =

HV=900 - Gain 2 RT. 0.2.

Source M-226 in 3" paraffin

can w/2" paraffin reflector sid. = 5"

Sample at 22'

HV=900 PDL

C3 10

1 = 1" BP @ 1500v @ 4mv is.  
inside rock near  
flat show for 30'

2 = NE400 @ 900  
HV=900 / M226 source  
d = 5" @ RT

Gain 2, RT. 0.2 PDL 10

(A) 1	12560
2	18761140
1	12970
2	18955660
1	13110
2	18545880
1	1300015
2	2863750
1	12790
2	2791270
1	13080
2	2745190
1	12690
2	2617780
1	13560 30
2	1422000
1	13270
2	1422260
1	12990
2	1441210
1	12710 3
2	1159090 3

2	1027960
1	12510 3
2	950330
1	12830
2	944110
1	12670 30
2	895830
1	12600
2	895210
1	12150 3
2	858410
1	12810
2	858600 3
1	12090 3
2	828390
1	12690
2	829030
1	12560 3
2	794290
1	12110
2	796900
1	1520
1	92770
1	12470 3
2	775000
1	12430
2	774670
1	12330 3
2	747250
1	11870
2	749450
1	12370 3
2	724660
1	12460
2	717010
1	4610
2	103650 3
1	12560
2	704960
1	12470
2	704500
1	12270 3
2	678320
1	11380
2	677070 3

1	11380
2	6770.70
1	13070
2	656710
1	12640
2	658330
1	12540
2	656370
1	<del>9470</del>
2	<del>508750</del>
1	13070
2	637820
1	12660
2	642930
1	12280
2	598590
1	12650
2	596270

PF<sub>3</sub> @ PM tube

1	352590
2	608500
1	324510
2	602280

Both sources @ 22'

1	1100
2	1780
1	1390
2	1710
1	1200
2	1410
1	1300
2	1830

5mg Ra @ 20' in w-end tunnel

(b)

1	220
1	11090
1	150
2	12830
1	240
2	950
1	190
2	870
1	150
2	260
1	150
1	170
1	120
2	200
1	130
1	110
1	180
2	<del>60</del>
1	150
2	130
1	230
2	50
1	160
2	40
1	140
2	70
1	170
2	50
1	130
2	50
1	210
2	40
1	110
2	10
1	140
2	30
1	150
2	00
1	90
2	40
1	160
2	10
1	130
2	00
1	90
2	00
1	100
2	10
1	120
2	00

5mg Ra @ 5" in paraffin (like m226) in tunnel

(c)

1	150
2	6364110
1	120

2	50
1	160
2	40
1	140
2	70 <sup>15</sup>
1	170
2	50
1	130
2	50 <sup>50</sup>
1	210
2	40
1	110
2	10
1	140
2	30 <sup>75</sup>
1	150
2	00
1	90
2	40
1	160
2	10
1	130
2	00 <sup>100</sup>
1	90
2	00
1	100
2	10
1	120
2	00

Smg Ra @ 5" in paraffin  
 mzzc in tunnel

① 1	150 <sup>90</sup>
2	6364110
1	120
2	6331640
1	130 <sup>15</sup>
2	1074610
1	80
2	1088500
1	160 <sup>70</sup>
2	276300
1	120
02	00277220
1	110
2	201200 <sup>50</sup>
1	20150
2	80010
1	120
2	79360 <sup>30</sup>
1	140
2	23210
1	110
2	23050 <sup>15</sup>
1	130
2	5330
1	120
2	6110 <sup>10</sup>
1	110
2	1830
1	110
2	1770 <sup>15</sup>
1	110
2	530
1	130
2	490
1	90
2	660
1	110 <sup>70</sup>
2	270
1	140
2	200
1	70
1	100
1	130 <sup>15</sup>
2	130
1	120
2	110
1	120 <sup>9</sup>
2	40
1	110
2	80
1	130
2	70

# 242 Scintillation Detectors

9-28-59

As noted on tape,  
 HV = 850 scale 1 records data  
 from BF<sub>3</sub> counter (RCL 1x8"),  
 at 4mv. sensitivity.

Scale 2 records data from  
 detector under test.

Amplifier: CB. Gain 2, RT = 0.2 μs.  
 HV = 850.

Run A: neutron source, M-226 (7.32x10<sup>6</sup> n/β)

@ 5" 3" paraffin can, 2"  
 plexiglas reflector in  
 3 3/4" hole in can.

5mg Ra @ 22' (SW corner 202)

Run B: neutron source, M-226

in west end of tunnel > 5' concrete.

5mg in Ph pit 201

Run C: 5mg Ra in contact with  
 Detector shield @ 3 1/2"  
 from ~~to~~ phosphor.

D: M-226 in 102 NE corner.

10 min. det. time  
 1 = Pol<sub>3</sub> on at P.M. 1500V  
 Gain 4, RT 0.2, PDL = 45 (40)

2 = NE 900 - HV = 850 (46)  
 GAIN 2, RT 0.2 PDL  
 M-226 5" PM 100  
 3" par. can 2" plex  
 5mg Ra @ 22' 4/m

1	37.5	96.5	0	100
2	45.0	33.9	0	
1	34.7	06.4	0	90
2	49.6	40.4	0	
1	34.8	14.3	0	80
2	54.3	14.8	0	
1	34.5	43.3	0	70
2	59.1	79.2	0	
1	34.4	53.8	0	60
2	65.5	90.8	0	
1	34.5	94.0	0	50
2	72.1	85.0	0	
1	34.0	93.7	0	PDL
2	79.7	52.9	0	
1	33.7	31.7	0	30
2	89.0	17.1	0	
1	33.7	87.6	0	25
2	95.8	04.8	0	
1	33.8	92.1	0	20
2	107.7	97.3	0	
1	33.5	57.9	0	15
2	136.6	34.7	0	
1	33.4	71.4	0	10
2	80.0	075.1	0	

M-226 w. end tunnel  
 5mg in 201 -  
 11.48 n/β SK4 PDL 10

1	4.0	010.4	0	
2		85.9	0	
1		9.2	0	10
2		11.5	0	
1		8.9	0	20
2		9.7	0	
1		9.4	0	25
2		3.1	0	
1		7.5	0	30
2		3.2	0	
1		8.2	0	40
2		2.1	0	
1		7.4	0	60
2		7.0	0	
1		9.4	0	80
2		1.6	0	
1		10.7	0	100
2		4.1	0	

5mg Ra shielded

1	9.8	0	100	
2	3.2	0		
1	6.9	0	80	
2	1.7	0		
1	8.1	0	60	
2	2.2	0		
1	7.6	0	50	
2	7.5	0		
1	8.3	0	40	
2	97.1	0		
1	6.7	0	35	
2	502.6	0		
1	10.2	0	30	
2	302.6.6	0		
1	6.6	0	25	
2	20.6	49.1	0	
1	8.3	0	20	
2	131.1	23.2	0	

5mg Ra shielded

1	18.1	0	60
2	13.6	0	




9-28-59

9/28/59

INTEGRAL TRANS

DETECTOR NE-400 G655A

SOURCE M-226



SEMI-LOGARITHMIC 359-81G  
 KEUFFEL & ESSER CO. MADE IN U.S.A.  
 4 CYCLES X 70 DIVISIONS

PDL

Neutron detector - 7

9-20-57

Scalogram just on  
Detector in 202.

(A) M-226 in NE corner Rm 102

(B) M-226 in Rm 201

Setup detector in Rm 201.

Channel C3

C3 gain 0.10

①

1	4m. 00
2	69.00
1	00
2	19.30
1	00
2	18.00
1	00
2	17.60
1	00
2	30.70
1	00
2	23.10

M-226 in 301

②

1	00
2	2.30
1	00
2	.10
1	00
2	.60
1	00
2	1.50
1	<del>00</del>
2	<del>70</del>

11:20 PM  
NE 405 - C3 preamp in 201  
GAIN 2 JOT 0.10 PDL 60

M-226 in "Sid"  
M-43 in "P" by Nell  
M-94 in "P" by PJ

1	00
1	200.00

Source M-226 100"

1	00
2	93.60

Source in (1 C2 of 100)

1	52.00
2	200.70

10-9-57

9:45 Am

NE 400 - N/A detector  
2, 0, 7, 2, 6  
460  
Calor. (R.D. 12)

A: -	1	0	0	0
	2	3	3	2
	1	0	0	0
	2	1	7	6
	1	0	0	0
	2	1	7	4
	1	0	0	0
	2	1	3	8
	1	0	0	0
	2	1	0	5
	1	0	0	0
	2	5	4	6
	1	0	0	0
	0	2	5	8
	1	0	0	0

244

Scintillation Counter

Automatic Scaling System on C.P. 15

10-9-57

NE 400 A: #1: no signal  
 #2: NE 400, 665-A, C3  
 Gain 2, RT 0.2ms, PDL 60  
 HV # 350V  
 Count time 1.0 minute  
 SID SOURCE DOWN  
 HIGH COUNTS - scaler 2 malfunction  
 during warmup - 1st decade transferring  
 to 2nd decade after 3 sec.

B: #1: 100KC linear signal.  
 #2: As in A  
 AS indicates seconds counting time  
 (Timer set on "Live-Timer")

C: (Timer set on "Clock-Timer")  
 During run - stopped & re-set  
 1st digit on printer.

10:18

D: Output from C3 to both SID 52  
 (During run timer hung up on  
 recycle)

E: SID SOURCE NOW UP

10:32

NE 404 F: Removed NE 400 detector -  
 NE 404 detector in place  
 NO change in gain or HV settings  
 SOURCE UP

G: SOURCE DOWN

S-1 function

Live Time

B: -	1	6	0	0	2	4	4	5	0
	2	3	7	3	0	0	0	0	0
	1	6	0	0	2	4	7	1	0
	2	3	3	3	0	0	0	0	0
	1	6	0	0	2	4	7	7	0
	2	3	6	0	0	0	0	0	0
	1	6	0	0	2	4	4	7	0
	2	3	5	6	0	0	0	0	0
	1	6	0	0	2	4	5	3	0
	2	3	5	4	0	0	0	0	0

Clock Time

C: -	1	6	0	0	2	4	4	3	0
	2	3	6	9	0	0	0	0	0
	1	6	0	0	2	5	0	1	0
	2	3	8	9	0	0	0	0	0
	1	6	0	0	2	4	8	3	0
	2	3	8	7	0	0	0	0	0
	0	0	0	0	0	0	0	0	0
	0	0	0	0	0	0	0	0	0
	0	0	0	0	0	0	0	0	0
	0	0	0	0	0	0	0	0	0
	1	6	0	0	2	4	4	4	0
	2	3	4	6	0	0	0	0	0
	1	6	0	0	2	4	4	6	0
	2	3	4	8	0	0	0	0	0
	1	6	0	0	2	4	5	6	0
	2	3	6	4	0	0	0	0	0
	1	6	0	0	2	4	6	2	0
	2	3	6	9	0	0	0	0	0

244

Semi-automation Counter

Automatic Sealing System on CP-15

10-1-52

NE 400

- A: #1: 70 signal
- #2: NE 400, 66J-9, C3

Gain 2, RT 0.125, PDL 60  
 HV # 350V  
 COUNT TIME 1.0 minute  
 SID SOURCE DOWN

HIGH COUNTS - scaler 2 malfunction during warm up - 1st decade transferring to 2nd decade after 3 bits.

- B: #1: 100KC linear signal.
  - #2: Co in A
  - #3 indicates seconds counting time
- (Timer set on "Live-Timer")

C: (Timer set on "Clock-Timer")

During run - stopped & re-set 1st digit on printer.

10118

- D: Output from C3 to both SID & S2 (during run time hung up on recycle)

E: SID SOURCE NOW UP

10152

NE 400

- F: Removed NE 400 detector - NE 400 detector in place
- No change in Gain or HV settings
- SOURCE UP

G: SOURCE DOWN

B+C - Comparison between live & clock time - No difference noted.

0 2 38  
 1 35  
 2 70

Live-Timer

B: -

1	6 0 0	2 4 4 5 0
2	<del>4 0 0</del>	3 7 3 0
1	6 0 0	2 4 7 1 0
2	<del>6 0 0</del>	3 3 3 0
1	6 0 0	2 4 7 7 0
2	<del>6 0 0</del>	3 6 0 0
1	6 0 0	2 4 4 7 0
2	<del>6 0 0</del>	3 5 6 0
1	6 0 0	2 4 5 3 0
2	<del>6 0 0</del>	3 5 4 0

Clock-Timer

C: No. of pulses recorded

1	6 0 0	2 4 4 3 0
2	<del>6 0 0</del>	3 6 9 0
1	6 0 0	2 5 0 1 0
2	<del>6 0 0</del>	3 8 9 0
1	6 0 0	2 4 8 3 0
2	<del>6 0 0</del>	3 8 7 0
0		1 0
0		1 0
0		1 0
0		1 0
1	6 0 0	2 4 4 4 0
2		3 4 6 0
1	6 0 0	2 4 4 6 0
2	<del>6 0 0</del>	3 4 8 0
1	6 0 0	2 4 5 6 0
2	<del>6 0 0</del>	3 6 4 0
1	6 0 0	2 4 6 2 0
2	<del>6 0 0</del>	3 6 9 0

DISCONNECT  
 TIME SIGNAL  
 FROM  
 PULSE  
 CONTROL

1016

1	3 6 9 0
2	3 6 9 0

E: SOURCE UP

1 9 8 0

up

DESCRIBE CT  
TIME SIGNAL  
SIGNALS  
SIGNALS  
SIGNALS

1	6 0.0 2 4 4 6 0
2	<del>6 0.0 2 4 4 6 0</del>
1	6 0.0 2 4 5 6 0
2	<del>6 0.0 2 4 5 6 0</del>
1	6 0.0 2 4 6 2 0
2	<del>6 0.0 2 4 6 2 0</del>

10.16

1	3 6 9 0
2	3 6 9 0

5: source box


1	9 8 0
2	9 8 0
1	1 0 5 0
2	1 0 5 0
1	1 1 1 0
2	1 1 1 0
1	1 1 1 0
2	1 1 1 0

6: source box  
source box

1	0 0 0
2	0 0 0
1	0 0 0
2	1 2 0 0
1	0 0 0
2	0 4 0 0
1	0 0 0 0
2	1 0 0 0
1	0 0 0 0
2	0 9 0 0
1	0 0 0 0
2	1 5 0 0

6: source box

1	0 0 0
1	9 9 0 0
1	0 0 0 0
2	9 3 0 0



10-9-55 Took data for determination of gamma cut-off with NE404 crystal - 5mg Ra. on top of shield - SID source down.

10-12-55 8:50 Automatic System On - -

9:40 Set for 10 minute one-shot count.

(WCT SUGGESTS BUZZER ON WASTE TANKS)

NE404 detector + 6655A HV 850V. Gain 2, RT=0.2  
PDC=60


SID SOURCE UP: 9.1 c/m

DOWN: 102.1 c/m

DOWN: 102.0 c/m (M-226 ~ 15' in Run 201)

M-226 @ 23' : 5,730.8 c/m -

Ion Chambers:



	K-1 mk	K-2 mk
5mg Ra - in contact with paraffin shields	$4.2 \times 10^{-11}$ (+)	$2.25 \times 10^{-11}$ (-)
Background	$7 \times 10^{-13}$ (+)	$1.8 \times 10^{-13}$ (+)
M-226 ( $2.5 \times 10^6$ n/s)	$7.3 \times 10^{-10}$ (+)	$2.3 \times 10^{-12}$ (-)
10mg Ra in contact	$9.1 \times 10^{-11}$ (+)	$5.3 \times 10^{-11}$ (-)

These readings taken preliminary to consideration of parallel operation of 2 ion chambers for compensating gammas.

10-12-57 Further Scintillation Detector -

Attach 6655A PM tube to PM-2 channel.

Set PM-2 zero. Skelport = 345.

→ Purpose: to measure dark-current signal & compare with several 6292 tubes.

( $E_0 = 1250$ ) ( $E_M = 1800$ )

HV	6655-A	6292	2	3	4	5	6
0	.05 mV	.07					
600	neg	"					
700	neg	"					
800	neg	"					
900	.075	"					
950	.140	"					
1000	.400	"					
1050	.600	"					
1100	.850 mV	.80 mV					
1200	"	.10 mV					
1300	"	.50 mV					

(1 mV = 1  $\mu$ A)

\* When first set HV at 1300 there was sharp rise in reading. Then signal tapered down to recorded reading.

→ Check response to gamma source with no crystal


HV = 1100v

6655-A -  $\approx 15 \mu$ A on panel meter

6292-1  $\approx 1 \mu$ A (recorder read 1 mV)





10-12-57 Place NaI crystal on 6292-1



with 5mg Ra - got readings on meter

10-13-57 more of above


NaI + 6292-1



PM2 w/o M @ a given HV = 42% on meter  
 w/ M = 21% on meter

10-13-59 Jon Chambers:

5mg Ra lined w/ chambers as follow



12:50 P.

$$\left. \begin{aligned} K1 &= +1165 \times 10^{-12} a \\ K2 &= -1.7 \times 10^{-12} a. \end{aligned} \right\}$$

R-105

A K1 Both  $6.7 \times 10^{-10} a$  M226 @ 3

add  $\left\{ \begin{aligned} E 100 \text{ only } & 6.8 \times 10^{-10} a \\ 5mg Ra @ 2 & 6.9 \times 10^{-10} a \end{aligned} \right.$

Both K1  $6.65 \times 10^{-10} a$   
Remove Ra  $6.8 \times 10^{-10} a.$

⊙ Ra bot IC  $7.1 \times 10^{-10} a.$

Ra @ 2  $6.75 \times 10^{-10} a$

@ 3  $6.65 \times 10^{-10} a.$

Remove Ra  $6.85 \times 10^{-10} a.$

Ra @ 4  $6.9 \times 10^{-10} a$

E 100 only  $7.05 \times 10^{-10} a.$

Remove M226

Ra @ 4  $1.7 \times 10^{-11} = .17 \times 10^{-10}$


Both IC  $6.4 \times 10^{-12} = .06 \times 10^{-10}$

$- 1 \times 10^{-12}$

EUGENE DIETZGEN CO.  
MADE IN U. S. A.

NO. 340 10 DIETZGEN GRAPH PAPER  
10 X 10 PER INCH

$I_{\text{PM}} (10^{-8})$



10-14-59 P.M. Dark Current -

	6292-1 (348.5)	6292-2
500	0	0
500	.01 $\mu A$	.02
600	.01	.02
700	.01	.02
800	.01	.02
900	.02	difficult to read reading $\rightarrow$
1000	.05	
1100	.13	
1150	.13 (+.54 back down)	
1200	.23	
1250	.17 (+.1 $\mu A$ back down)	
1300	.20	

10:35 put 20  $\mu A$  meter in series with scope panel meter - To do this

cut power - Read of lit current  $\mu A$  - up


20  $\mu A$  meter  
set zero  
40  $\mu A$  check  
20  $\mu A$   
Load 1050

HV	6292-2	6292-1	6292-3	6292-4	6292-5	6292-6	6292-7	
500	<.1		1.1	.6	.3	.2	<.1	
600	<.1		1.5	.8	.4	.2	<.1	
700	.15		1.7	1.0	.5	.2	<.1	
800	.1		1.9	1.2	.6	.4	<.1	
900	.25	<.1	2.1	1.5	1.0	.4	<.1	
950							2.1	
1000	.30	.1	2.5	1.9	1.4	.5	.1	$5 \times 10^{-9}$
1050	.40	.1					.2	$15 \times 10^{-9}$
1100	.30		3.0	2.7	1.70	.70	.8	$46 \times 10^{-9}$
1150	.34		3.3	3.3				
1200	.30	.2	3.4	4.4	.90	1.40		
1250	.30		3.6	5.7				
1300	.25	.2	3.8	7.7	.95	3.90		

IM


6655-A mly - 5mg Ra -	7.7 $\mu$ a	1100V
	3 $\mu$ a -	1000V
	1.1 $\mu$ a	900V
6252-2 mly - 5mg Ra -	0.3 $\mu$ a	900V
	0.5 $\mu$ a	1000V
	1.0 $\mu$ a	1100V
	1.5 $\mu$ a	1200V
	2.8 $\mu$ a	1300V

1-20-60: Check pulse from first meter scrub plate with 6655-A photo multiplier tube -




Pulses much larger than anticipated - Is method correct  
Scope pre-amp is in poor condition -  
Nothing conclusive here -

Mount: 2U




This assembly mounted in socket of aluminum tube as protection shield. Top of phosphor in line w/ edge of Al. can. This end covered w 1/16" cork & black tape.



35 10 171.2


6pm 254

SEMI-LOGARITHMIC 359-81G  
 KEUFFEL & ESSER CO. MADE IN U.S.A.  
 4 CYCLES X 70 DIVISIONS



1/21/60 - Using detector assembly collect data with source  
 m-230, PuBe, -

- Pre-amp: AIC 5W641
- L.A: AIC-B, 123800
- Scaler: B-A 123748
- Timer: BA 124254
- Scope: Tektronix 160330
- HVPS: NICE 100304



source distance from  $\phi$  source  
 to top of crystal phosphor entrance

HV meter - ESVM

AV 900: LA gain 2, RT 0.2

Background = PDL

PDL	Time	Counts/m	
7	1	229	20 { 5.8
		194	
		202	
8	1	78	25 { 4.9
		81	
		70	
		45	
9	1	32	35 { 1.1
		35	
		55	
10	10	28.5	50 { 0.6
		38.6	
		33.3	
		15.2	
12.5	10	14.1	
		8.9	
15.0	10	10.11	

Source @ 1 meter

Time	Counts/m	
7	1280	50, 10m, 80.7
	1366	
10	1639.5	

11117


Time	Counts/m
8	978.2
9	824.7
10	685.3
15	459.0
20	340.3
25	267.8
30	220.5
35	171.2

done 254

1/22  
8:40 AM

Source @ 25 cm

Dist 10m	PDL	C/m
	30	2948.6
	40	1817.4
	50	1072.6
	60	
	70	



pm 62922 of down phosphor "400"

(HV 21100V 1105) = inside purephos can -

Dist 1m	PDL	C/m
	10	33063
Source	20	24913
25 cm	30	25142
Source	40	18513
Source	50	13240
	60	9855
	70	6789
	80	4751
	90	3172
	100	2197
	100	1204

	10	33,099		
Blg	10	65	60	13 C/m
	20	49	70	6
	30	21	80	8
	40	33	90	4
	50	26	100	2

1/22 - BF<sub>3</sub> α ct attachment -

Check small sensitivity at Gen 2

PDL	Input	C/m
5	1 mV	
10		2.7
20		6.9
30		11.25
50		24.7
100		63.

Set PDL @ 15 ± 4 mV - HV 1500V BF<sub>3</sub> α ct -

Blg - 96%<sup>1</sup>, 189%<sup>1</sup>, 83%<sup>1</sup> 1096-  
1125% 36% w/Source no count -  
Try new counts 10502, 1682 PDL - 12 ± 4 mV  
HV = 1500.

C/m =	14149
	13979
HV 1400 -	13866
HV 1300 -	13426
HV 1200 -	9722
1400	13806
Blg (10 m)	52.4 C/m



1/21/60 1:50: Same @ 50 cm :

PDL	Time	Counts/m
7	1	55411 24820
	10	4773.8
8	"	3463.6
9	"	2912.2
10	"	2552.5
15	"	1545.9
20	"	1200.9
25	"	960.6
30	"	780.2
35	"	597.7
40	"	486.7
45	"	3
50	"	312.5
55	"	
60	"	206.2
70	"	141.6

1/22/60

7:55<sup>A</sup> Bkg.

PDL	Time	c/m
7	10m	1.8
20	"	7.2
25	"	5.3
30	"	2.8


back to 252 please

1/21/60

1/28/60 -  $\rightarrow$   
 Bias Curves  
 6655A + NE404 (Mf)  
 $R_L = 2000 \Omega$   
 $HV = 900V$   
 $G_{m2} \cdot R_T = 0.2$   
 $M = 230$  10cm & 100cm  
 5mg Ra 10cm  
 Background

1/22/60

back to 252 please




1-28-60 LA 92m = 2 6655-A + NE-404  
 $R_T = 0.2$

A - HV = 900V. Source 5mg Ra -  
 distance = 10cm

Same units as on 1/21/60 (p 251)

PDL	4m	Time - 1 min
7	130,510	
8	430,562	
9	197,330	
10	87,920	
12.5	16,996	
15	3,023	
17.5	748	
20	227	
25	15	
30	6	
35	2	
40	1	
Background		
7	343	
8	95	
9	65	
10	37	
12.5	17	
15	17	
17.5	7	
20	(1) 14	
25	5	
30	5	
35	2	
40	3	
2.85 x 10 <sup>6</sup> n/s M. 230		
7	125,057	
8	89,246	
9	72,612	
10	58,483	
12.5	41,976	
15	33,437	
17.5	28,446	
20	24,827	
25	20,212	
30	16,384	
35	13,145	
40	10,586	
50	6,743	
60	4,702	
70	3,282	



Handwritten scribbles at the bottom of the page.

1/28/66 M-230 @ 100cm

PDL	832	890	911
10	832	890	911
20	369	395	398
30	232	244	218
40	169	157	149
50	129	102	101
60	71	72	

2 minute counts

30 455 ct = 2275 c/m

70 ct = 35 c/m body @ detector

478 ct = 239 c/m Cd cover

466 ct = 233 c/m no cover

6292 + NE400

HW = 1100 v.

101207 Background (no thermalizing medium) around detector

PDL	50	Time (m)
10	50	
15	44	
20	51 (41)	
25	39	
30	19	
40	18	
50	14	
60	13	
70	12	

5mg Ra @ 10cm

10	1163
15	124
20	44
25	37
10	4093
15	6851
20	5860
30	3997
40	2850
50	2012

All plottom - 24cm below

1" Plexiglas @ detector face 30 6,158

2 7/8" wide

2" 30 12,391


PDL 30

M230 @ 20cm

No Plexiglas

1"	1,420
1"	2,045
2"	3,251
3"	3,162
3" near source	2,967

@ detector



1/28/66  
6292  
NE400  
HW = 1100V.

Sources: M230 @ 10cm  
5mg Ra @ 10cm

sd = 20cm

DL 10

15

20

25

30

35

40

10

20

25

30

1/28/60



11:05 - 1" box 3650 c/m

3" in front 3346 c/m

Cd cover - 1254 c/m no plexiglas.

Cd + 3" face - 1095 c/m

11:21

Twin (10 mic) 1130.9 c/m

PDH=30 1188.7 c/m

1347.5 c/m

3479.4 c/m

Cd + 3" face sd=20cm

Cd only -

Cd off - no plex.

3" face -

12:10 P

(3" face) + 58.1 c/m

59.5 c/m

+ 47.9 c/m

+ 46.10

34.9

33.7

26.1

5 mg Ra - 298.8 c/m

48.6

45.2

Bkg - PDL 10

15

20

25

30

35

40

10


20

25

30

*Ben*

1/28/60



M-230 @ 100cm  
NO Therm.

5mg Re  
@ 100cm

Blg

EUGENE DIETZEN CO.  
MADE IN U. S. A.

NO. 340R-L410 DIETZEN GRAPH PAPER  
SEMI-LOGARITHMIC  
4 CYCLES X 10 DIVISIONS PER INCH

PDL

70

60

50

40

30


20

10

4m

Scintillation - Detector

5-23-60



1:32 D = anthracene S = 5mg Ra source  
 PM = 6810A  
 M = APE - 100  $\mu$ a  
 HV = NJE

(S.D. =  $\pm$  10%)

HV	M	
1500	1 $\mu$ a	Background
	3 $\mu$ a	Source, 1'
1600	5.5 $\mu$ a ; 1 $\mu$ a	Source, 1' ; Background
1700	8.5 $\mu$ a ; 1 $\mu$ a	" "
1800	15 $\mu$ a ; 1 $\mu$ a	" "
1900	27 $\mu$ a ; 1.5 $\mu$ a	" "
2000	43 $\mu$ a ; 2.5 $\mu$ a	" "
2100	62 $\mu$ a ; 4 $\mu$ a	" "
2200	90 $\mu$ a ; 5 $\mu$ a	" "

1:53 HV off  
 Nat on

1:56

NW	M
1500	28 ; 1

Scintillation - Detector

5-25-60

12:42 6810A & 16I - HV = NJE supply M = APE 0-100  $\mu$ a -

Source = 10mg Ra - ADJUST METER ZERO


Backgrounds at varying HV settings

HV	M
1500	~.5
1600	~.5
1700	~.7
1800	.8
1900	1.0
2000	2.0
2100	3.0
2200	5.0

Source @ 1'

2.5  $\mu$ a - > 100  $\mu$ a -

Source @ 1'



Source @ 1'

HV	S=1'	S=2'	S=3'	S=4'	S=8'
1100	3.5	2			
1200	7.0	2.5			
1300	17.5	5			
1400	39	11			
1500	88, 91	23	13	8.5	9.
1600		49	23	15	5
1700		99	45	28	9
1800			86	52	16
1900				88	27
2000					45
2100					72
2200				(2150	87)

5/26/60 734A Check - 508 2100 - 71  $\mu$ a

5-26-60. Need insensitiv. device for central area monitoring to catch rising pulse & burst. "incident"

7-11-60 Reports from vacator 7-1/7/60  
 ADC - 20 channel fan stalled & "hot" Sat. 7/9  
 DVM - Man from Monroe was here to adjust data-log unit. Talked with him about setting up means of identifying source.  
 Has talked with people about pre-amps. Suggests wires together of available low noise pre-amps.  
 - Has ordered 8 flat fission counters.  
 - Question about use of amplifiers with position output signal from fission counters.

JPE - Has checked TC-3 & TC-4 in '112-

7-12-60 2:30 PM - Power off for 20 sec. When on butchery alarm & cut off.

- Inspect fan - 20 channel.
- Check out R1DL unit in 210, June at.
- Check position input to 219 pre-amps - at
- Inspect check of HP amplifier.

- 7-13-60
- Scan table counter 215 - adj HV.
  - HP amplifier check
  - R1DL unit timer - service -

7-14-60 Channel C3 - 109 - high counting rate at times. Channel on "count" with scope monitoring high amp. output. Pulse amplitude increases producing counts. This followed by large transient and which upon collapse reduces signal to normal.

NICC Test Pulse Generator attached to input of amplifier. Seems to be still some instability in signal -  
 Pulse Generator in place of U233 detector. Signal: negative  
 1 pulse wide, 1 ms.

PDL	cpm	37.	44 <sup>H</sup>
28	225 <sup>14</sup>		
	225 <sup>13</sup>		
34	222 <sup>10</sup>		
	222 <sup>12</sup>		
35	160 <sup>5</sup>		
	129 <sup>13</sup>	135 <sup>13</sup>	
36	15 <sup>11</sup>		
	31 <sup>2</sup>		
	67 <sup>2</sup>		
	81 <sup>2</sup>		
	150 <sup>2</sup>		
	110 <sup>13</sup>		

C1	essentially on 60V position on bkg,	} 1722 { scope 1x1
C2	60V amp. 1/2 cm	
C3	60V amp., 8 cm	



7-14-60 Check battery for C3 - 3-40V in plastic case  
 when meter attached read 110V - dropping to 95V.  
 Such batteries - will 10,000 a load, dropped to <math>500\mu\text{s}</math>.  
 Fresh supply - box found in bench <sup>supply</sup> 107.  
 133 volts no load, 129 volts 10,000 a load. ✓  
 With new battery in place, U<sup>233</sup> detector attached took  
 extended count of background to see if noise ignored.  
 Time started 9:18:15 = Scale 16 =  $\left\{ \begin{array}{l} \text{PDL } 29 \\ \text{Gain } 8, \text{ RT } 0.8\mu\text{s} \end{array} \right\}$   
 stop 9:38:15 20 min.  
 start 9:42 -  
 @ 11:37 no counts - Good sign

check 112 distribution  
 RIDL Time report of check out.

7-15-60 AM West End - Bottle Exp - set up  
 P. 1-3 PM East End - setup 235 & 238 fission counters on  
 Channels 1 & 2 -  
 3-4 PM West End - Counter setup Bottle Exp  
 4 - write up -

7-18 - Conf on CTU -  
 7-19-62 Genl Int  
 7-25 - Inst. Rm 215.  
 7-26 Conf. on CTU - Rm 215

July 27, 60 Excerpt from WAPD-MRT-10 3/8-60 report  
 p-59: Pulsed Neutron Measurements in Multiplying  
 Media - E Carroll, N. Harkma

Detectors: 2 - BF<sub>3</sub> counters  
 2 - Boron polyether crystals  
 2 - w/ 6810 RCA photo multiplier  
 6 w/ 1002 CBS photo multiplier.

Consideration of comments on CTU spec. regarding  
 testing.


8-16-60 Consideration of detector & distributed Amplifier for TMC

TMC signal -  $\approx 1$  volt

6810A - subthreshold - in 210 operation of 6810A @ 1600V w/ 25 $\mu$ A

La - gives good pulses out through 460BR

Trial run in South wing -



Using existing cable - try to see what pulses obtained with HP units will.


6810A - Anode current of 75 $\mu$ A La in 1B connected to Top Hat

(impedance & bias normalizer circuit had been)

get 218V pulse with 460BR linear at 1700V

8-17-60 - 8<sup>th</sup> Further check of neutron detector for pulsed neutron work.

LAYOUT:



8-17-60

HV 1700V - (1mA)

HP 460BR

Blg -  $\approx 1/2$  cm (scope 2X1), with 1cm of low freq noise

Ra Gamma Source - 1 cm ~~LINEAR~~ 1X4

7 cm ~~PULSE~~ 1X2

Po Be Source - 1.3 cm LINEAR 1X4

HP 460BR to Scaler - ~~at input of scaler to scope~~ - Scale PHS

Both sources - (1 min)  $46.17 \times 10^3$  c/m

46.46

Remove Po Be - Ra  $2.18 \times 10^3$


1.95  
1.83 x 10<sup>3</sup>

HP PHS	10ky	10 $\mu$ A	Ra + Po Be M227	Po Be
NV=1700 HP 460BR + 2	8 x 10	$28.13 \times 10^5$	$53.543 \times 10^5$	$1.9479 \times 10^5$
3	5 x 10	$16.76 \times 10^5$	$32.6078 \times 10^5$	$1.3093 \times 10^5$
4	2 x 10	$5.7148 \times 10^5$	$13.9156 \times 10^5$	$.8022 \times 10^5$
5	0	$3.656 \times 10^4$	$19.368 \times 10^4$	$.844 \times 10^4$
6	-	$2.33 \times 10^3$	$24.41 \times 10^3$	-
7	-	$2.0 \times 10^2$	$38.3 \times 10^2$	-

No good -

Source were not fixed.

Clamp Po Be source to stop stud - put 3" paraffin over detector  
Scope only in HP 460BR output 75.2 on output of HP 460BR HV



Ra gamma bulb at 1cm - Remove paraffin as apparent change.

8-17-60 Condition 6810A - Aucl. <sup>(Si)</sup>  $^{60}\text{Co}$  (outg) cross - 3" par above  $^{60}\text{Co}$

HV - AMP

1400 A+B Amex - Bin.  $^{60}\text{Si}$  - Seize low (X1)

Scaler - PHS 2 - 10m - 178 c/n

Most of day spent in preliminary adjustments.  
Brightest back TMC analyzer - Checked out. (Yair)

8-18-60 Comparison of responses of amplifiers to several sources

overnight HV off TMC on	HV	IP46aA	IP46aB	TMC 212 D	(N) cycles	Project count	Blg'ct	Source	9:15A - check setup sources in place TMC 100, D=2, R=16, M 1/4 ok for 600 sys. 10g.		
To calculate count rate $\% = \frac{I \cdot C}{N \cdot \Delta t} \cdot 10^6$ $= \frac{51.2 \times 10^6}{N}$ i.	1500	max	LIN	1	2921	2 <sup>14</sup>	16385	M227 6060	9:15A - check setup sources in place TMC 100, D=2, R=16, M 1/4 ok for 600 sys. 10g.		
				2917	2906	16384	16385			% 1.755 x 10 <sup>4</sup>	
				2	3751	2 <sup>14</sup>	16384				
				3766	2 <sup>14</sup>	16385					
				3961	3771	16388				1.361 x 10 <sup>4</sup>	
				3	8284	2 <sup>14</sup>	16385				
				4776	4765	16390				1.074 x 10 <sup>4</sup>	
				4789	15380	16386					
				4	5999	2 <sup>14</sup>	16384				
				6001	5990	16384				8.53 x 10 <sup>3</sup>	
					18016	16386				Removed M-227	
				1	4231	2 <sup>14</sup>	16384			6060	
				4223	4220	16386				1.212 x 10 <sup>4</sup>	
				2	4218	16385					
				5129	17669	2 <sup>14</sup>	16388				
				3	5083	2 <sup>14</sup>	16384				
5129	5123	16384		9.98 x 10 <sup>3</sup>							
3	5781	16385									
6284	15387	2 <sup>14</sup>	16386								
4	6262	2 <sup>14</sup>	16385								
6284	6248	16385		8.15 x 10 <sup>3</sup>							
	6245	16385									
	18853	2 <sup>14</sup>	16385								
	8033	2 <sup>14</sup>	16385								
	8045	16385		6.38 x 10 <sup>3</sup>							
	7996	16384									
	8023	16384									

Reduced HV to 1350V & Res. count to  $R^{12}$


S2 M227

ROBERTSON DIEZGEN GRAPH PAPER  
DESIGNED FOR CYCLES X 10 DIVISIONS

EUGENE DIEZGEN CO.  
PRINTED IN U. S. A.

n  
g  
%

3"p



x10  
y10

Reduced HV to 1350V & Preset count to  $R_{12}$  (Presets same set preset cycles of 20,000)

APA max, APB MIN

3-18-60	212 Disc	cycles	Blq cts	source	43		
$\frac{c/s}{N} = \frac{1}{N} \cdot 12.79 \times 10^4$ (for preset of 212)	3	1	1926 1900 1933	1920	4096 4097 4096	$6.66 \times 10^3$	
		2	9080 8918 9398	9132	4096 4096 4097	$1.40 \times 10^3$	
		3	20,000 20,000 20,000	151 115 117	C 128 383	20	
$\frac{c/s}{N} = \frac{1}{N} \times C$ (for preset cycle of $2^2 \times 10^4$ )	4	1	20,000 20,000 20,000	15 18 19	19 15	2.66	
		1	4977 5873 4957	5004 1501	4096 4096 4096	M <sup>227</sup> Co measured (+3" paraffin)	$2.558 \times 10^3$
		2	8334 8166 8216	8259 2116	4096 4096 4096		$1.553 \times 10^3$
5	4	3	12991 12837 13078	12969 3890	4096 4096 4096		$9.86 \times 10^2$
		4	20,000 20,000 20,000	3693 3687 3622	3669 11000		$5.73 \times 10^2$
		1	3,762 3,762 3,709	3744 11233	4097 4096 4096	M <sup>227</sup> w paraffin	$3.44 \times 10^3$
6	4	15,868 15,438 15,818 16,403 16,345	15844 11972	4097 4096 4096 4096 4096		$8.02 \times 10^2$	
		1	20,000 20,000 20,000	41 38 34	374 1113		
		4	20,000 20,000 20,000	0 3 3	1215		

For all runs - every addition correct 64/64 on tape -

815 hrs

Background polymerized disc NO 4000 - on 2810A - P5 Ct - 2<sup>12</sup>, P504 20,000

HPA more HPB WU -

M-227 source with 3" paraffin

Counts	Counts, B	Disc	c/s
5516	4096	1	2.47 x 10 <sup>3</sup>
4947	4096		
5043	4096		
10941	4097	2	1.202 x 10 <sup>3</sup>
10518	4096		
10468	4096		
15185	4096	3	8.17 x 10 <sup>3</sup>
15643	4096		
15411	4096		
20,000	3691	4	5.94 x 10 <sup>3</sup>
20,000	3781		
14,318	4097	1	9.17 x 10 <sup>3</sup>
14,739	4096		
20,000	217	4	35.2
20,000	235		
20,000	2680	4	49.4
20,000	1653		

Remove paraffin -

14028  
129057


Background

appears to be bursts of counts on hky. Display 2<sup>7</sup> movement of  
disc is slow then speeds up & slows down  
Why irregular counts

2000v M-227 @ 1" w/paraffin 2<sup>15</sup> 794 32,770  
618 32,790

Fig 2<sup>12</sup>

0/s  
1x10<sup>3</sup>  
2x10<sup>3</sup>  
3x10<sup>3</sup>  
4x10<sup>3</sup>  
5x10<sup>3</sup>  
6x10<sup>3</sup>  
7x10<sup>3</sup>  
8x10<sup>3</sup>  
9x10<sup>3</sup>  
10x10<sup>3</sup>



8-19-60 8:04 AM - Residual background behavior could be residual light effect which is common among such phosphors.


Check use of 511 scope + gate as system trigger. In this -  
 $B=2, D=2, L=10$  -

Range	Scope Setting	Deflex per minute
10,000	1.0	5867
		3333
<del>10,000</del>	<del>10.0</del>	3805
10,000	10	328
		324

~~10,000~~ 1.0 given as 10,000 1. 2

Use 10,000, 1.0 to approximate 600 trigger 1800 8. no good.

FP 460 AR MIN Pulse from FP 460 BR ;  $N=20000$  Bkg - NE 460



Scope at 10,000 ; 1.0 - HV = 2000

Present Count =  $2^{16}$   
 Present cycles 10,000 -

Bkg	$L/D$	Cycles	Time	Channel #
BZ $L=160$	1	10,000	201.7	33463
BZ $M=44$		10,000	202.2	32177
		10,000	202.9	30855

some difference in rate of accumulation of the present and values not predictable.

Change Scope was checked was King.

	1	4425	58.7	65538	(4)
$L=40$	1	10,000	40.2	7851	(5)
		"	40.2	9810	(6)
		"	40.2	6441	(7)
		"	36.1	34585	(8)
$B=2, D=2, L=40$	2	10,000	40.1	7535	(9)
		"	40.1	1782	(10)
		"	40.0	1529	(11)
		"	40.0	1015	(12)
		"	40.0	1188	(13)



270  
present  
cycles  
10,000

6810 A + NB400 - HV=2000  
R<sub>0</sub> = 75 sec

HP 460 AR MIN  
HP 460 BR LIN R<sub>0</sub> = 75 sec

8-19-60

Run	Time	Channel I	212 D
14	40.0	235	3
15	40.0	300	
16	40.0	259	
17	40.0	935	
18	40.0	34	4
19	40.0	26	

Background in spectra -

HV=1700

HP460A MAX HP460B LIN.

Set TMC - for 500,000 cycles

RUN Time Channel I 212 D

2000 sec -

10:14

20 2000.3 8421 1

Background does not appear to  
come in spectra at this HV, at  
first -

$$\frac{0}{\%} = \frac{10^6}{N \cdot 2L} \times C$$

$$= \frac{1}{40} \cdot C$$

$$\frac{0}{\%} = 210.5$$

After longer observation, still  
irregular.

PS cycle 50,000

21 1949 11818 1

M-227 above 3" paraffin  
set auto detector -

start auto

22 1948 12113 1

11:29 23 2597.1

12250  
12400  
12201  
12004  
12217  
12176  
12019  
12352  
12309  
12114  
12455  
12123  
12053

$$\frac{0}{\%} = 3,052$$

13 ~~2597.1~~  
Access

158677  
12,206

M227 at far corner by door

10,000 405 1018

M222 in fig 696

Replace chip to run of 6810-A

Spect counts, still present

2kg 24 598.8 998

PS cycle 50,000 3 - 1488

12:44 1441

	500000	500000	500000	500000
1	2831	2791	2930	3798
2	1543	1598	1600	1934
3	1196	1213	1307	1753
4	1175	1179	1213	1817
5	1172	1139	1205	1764
6	1124	1103	1216	1758
7	1063	1070	1131	1747
8	1026	1056	1170	1768
9	1008	974	1137	1732
10	974	980	1102	1716
11	961	982	1087	1719
12	954	935	1100	1632
13	927	932	1020	1695
14	913	959	1046	1670
15	854	891	1006	1613
16	876	898	1042	1693
17	857	874	997	1630
18	857	816	992	1618
19	870	868	982	1666
20	843	833	941	1485
21	868	855	977	1565
22	813	812	914	1591
23	798	847	918	1593
24	781	755	923	1539
25	810	763	903	1557
26	738	781	923	1532
27	768	724	910	1582
28	763	732	880	1547
29	703	724	892	1529
30	730	701	838	1490
31	704	700	872	1491
32	714	632	888	1490
33	683	659	856	1520
34	678	681	870	1475
35	662	648	816	1482
36	670	608	812	1475
37	714	619	857	1467
38	636	597	804	1459
39	655	639	797	1488
40	615	627	792	1383
41	627	598	798	1447
42	600	594	784	1417
43	608	603	783	1404
44	615	547	752	1428
45	609	570	739	1455
46	586	558	767	1408
47	602	521	750	1390
48	601	539	756	1449
49	594	547	742	1378
50	583	525	745	1391
51	559	539	777	1399
52	551	591	754	1388
53	564	545	744	1405
54	567	487	779	1365
55	555	562	707	1376
56	571	550	735	1347
57	526	527	752	1364
58	506	486	693	1323
59	508	491	715	1375
60	493	483	731	1331
61	499	504	688	1369
62	523	503	723	1326
63	485	479	793	1306
64	635	617	828	1413
	0			

B

B2

ed.

1/2

1000/100

1/100

ms

ms

1/5  
1/100

1:00 Ground IP465A P212D Bkg start fast value after 70 sec.  
 preset sec. CH1  
 50000 199.6 652 1  
 199.6 1960 Super fast. slow after 160 s.

Slide out 110 to marker 4150 -1.3V or -

During one scan background rate varied 1 volt/cm scan fixed.

1:18 500,000 1996.75. 8286 I. Disc 2 - fast rate at start  
 Pkg, slow @ 18.5 fast 108.5  
 300 283  
 345 530  
 587 678  
 725 800  
 855 965  
 1020 1140  
 1210 1325  
 1380 - etc


$c/s = 207.2 -$

HP  
 Audin Sig Gen + NICE pulser gen. - 212 D = 1  $V_i = 4V$  - set @ 15V, 1/ps.  
 B02, D02  
 1:40  $5 \times 10^4$  cycles; 200.0 s; 4048 (1.012 kc) signal 1000 cycles/s  
 2:45 P/  $5 \times 10^4$  cycles; 2200 s/total 3972 (.994 kc) 4div, 1 ps 1 kc  
 3:05  $5 \times 10^4$  Preset cycles; 385 2000 s Channel 4 signal 1000 cycles  
 1.110<sup>4</sup> cycles (0.100 c/s) 2 80 274 11 100 cycles/sec, 5.0V, 1/ps.  
 1.110<sup>4</sup> (2500) 2400 s/total 199.8 sec. 1000 c/s 5.0V - 1/ps  
 5 channels { 815 3955  
 764 9910  
 784  
 784  
 208  
 3982 3986  
 3991 15914  
 3982  
 3989

$5 \times 10^4$  (2500 c/s) 2000 s/total  
 (212 disc) 1.3 20 channels - 1000 c/s 5.0V, 1/ps

8-22-66 8:32A Power in TMC -  
 Check System Dragger w/ 1/2 capacitors set separating ccw,  
 stability must range to get good sweep on peak.

System Signal	Trigger Range	mult.	stat	mag	1/2	D	L	aj/m
Direct to gate 50,000 p/s of	10,000	1	1/2	MIN	2	2	10	3389
			1/2	MIN				14991
			1/2	MIN				3412
	1,000	1	1/2	MIN				34935
			max	max				19389
	1,000	1	1/2	MIN	2	2	10	87.2
								87.3
								87.5
								87.6
								87.8
								573.4
								572.7
								571.4
								570.8
								569.5
								569.57
								568.05
								567.60
								235.91
	10,000	1	max	max	2	2	10	250.22
								250.07
								250.46
	10,000	1	1/2	MIN	2	2	10	55.65
					2	2	160 (200 sec)	55.56
	10,000	5	1/2	MIN	2	2	40	
	1000	10	1/2	MIN	2	2	40	53.18
		2	1/2	MIN				274.42



R=100K  
 signal varied  
 5V, 1000 cps  
 DISC 11/96

$R = 100 \times 10^{-12} \times 16 \times 10^6 = 60 \times 10^{-6}$

100 sec count

max max appears to be noisy scan channel  
 max max  
 seconds have varying scan of this condition, NOT RECORDED

print

sys: scope 10,000; 1  
 stab d frig max.  
 sig = border line - 5V <sup>input</sup> 1000 c/s  
 DISC 1/4

999	99999	
100	250.46	g/s
1	1291	
2	356	
3	382	
4	229	
5	269	
6	262	
7	248	
8	258	
9	289	
10	241	
11	291	
12	255	
13	292	
14	218	
15	244	
16	207	
17	212	
18	252	
19	248	
20	234	
21	216	
22	214	
23	207	
24	215	
25	188	
26	223	
27	189	
28	190	
29	190	
30	230	
31	206	
32	175	
33	235	
34	217	
35	205	
36	180	
37	183	
38	212	
39	260	
40	167	
41	171	
42	157	
43	190	
44	170	
45	188	
46	179	
47	240	
48	166	
49	162	
50	221	
51	181	
52	189	
53	183	
54	173	
55	179	
56	125	
57	177	
58	127	
59	199	
60	117	
61	169	
62	158	
63	115	
64	109	

8-22

~~Scope trigger~~

5.0V 1000 c/s - AP/NICK  
TMC DISC 1.4

Preset cycles	SCOPE GATE	TRIGGER	MIN	TMC	TIME	cyc/s
10000	1000	5.5	1/2	2, 2, 40	101.3 101.4 101.4	
5	1000	5.5	1/2	2, 2, 20	101.4	
6				40	101.45	
7				80	101.5	
				80	101.5	(in 3/4)
				80	101.5	(in 4/4)
				80	101.5	(in 1/4)
					101.55	

(160 missing 1/2 of cycle)

sw from 1/4 to 1/2, then x for 2 3/4, erasing 1/2  
record 2nd in 1/4  
2nd higher trace

8 1000 5.5 1/2 MIN 2 2 20 406.6 (full menu)

still tapes

Input sig 5.0V (1000)

4 pot (1000)

Discriminated signal in scope - top removable list  
P 212 DISC = 1.0

11:25

set 2<sup>14</sup>

1000 5.5 1/2 MIN 2 2 20

full menu

(P

signal 100 Kc

Set signal generator at 5.0V - adj 1st pot to determine range of discriminator point.

Sig 5.0V

1st pot

1000

945

943

933


930

1

normal pattern

distorted pattern

low rate, but flat.



2 5  
 3  
 4  $\frac{10}{40}$  26  
 5 2.5C  
 6 992  
 7  
 8  
 9  
 10  
 11  
 12  
 13  
 14  
 15  
 16  
 17  
 18  
 19  
 20  
 21  
 22  
 23  
 24  
 25  
 26  
 27  
 28  
 29  
 30  
 31  
 32  
 33  
 34  
 35  
 36  
 37  
 38  
 39  
 40  
 41  
 42  
 43  
 44  
 45  
 46  
 47  
 48  
 49  
 50  
 51  
 52  
 53  
 54  
 55  
 56  
 57  
 58  
 59  
 60  
 61  
 62  
 63  
 64

574 - flow, 1000 ft  
 596 - slope 1000, 575, 1/2 min  
 TMC 2, 2, 20

216  
 203  
 207  
 213  
 189  
 201  
 198  
 185  
 184  
 196  
 197  
 203  
 191  
 204  
 183  
 194  
 198  
 189  
 194  
 175  
 192  
 214  
 189  
 186  
 192  
 190  
 189  
 211  
 175  
 186  
 193  
 192  
 184  
 206  
 183  
 186  
 206  
 212  
 200  
 190  
 188  
 205  
 188  
 181  
 186  
 183  
 197  
 202  
 185  
 209  
 191  
 194  
 186  
 190  
 190  
 197  
 187  
 220  
 186  
 191  
 199  
 205

774  
 398  
 406  
 388  
 391  
 392  
 383  
 397  
 387  
 395  
 396  
 379  
 399  
 387  
 380  
 404  
 395  
 395  
 380  
 386  
 374  
 372  
 374  
 374  
 390  
 373  
 380  
 397  
 375  
 374  
 383  
 393  
 394  
 364  
 374  
 373  
 380  
 396  
 374  
 370  
 369  
 339  
 356  
 360  
 368  
 344  
 352  
 337  
 324  
 344  
 322  
 346  
 336  
 326  
 324  
 326  
 329  
 332  
 326  
 314  
 297  
 299  
 319  
 337

93  
 1.25C  
 928

494  
 360  
 310  
 254  
 218  
 167  
 126  
 118  
 123  
 98  
 93  
 88  
 77  
 62  
 54  
 55  
 47  
 43  
 46  
 50  
 29  
 35  
 28  
 30  
 25  
 28  
 20  
 21  
 31  
 23  
 21  
 21  
 24  
 22  
 19  
 20  
 17  
 20  
 13  
 20  
 19  
 16  
 7  
 16  
 24  
 20  
 13  
 20  
 16  
 11  
 13  
 16  
 14  
 7  
 15  
 14  
 10  
 15  
 12  
 13  
 12  
 12  
 14

L=80  
 1.6

100000  
 100000  
 100000

100g STAB. 1/2  
 515 TEL6 MIN  
 RAN-16  
 MULT

SIGNAL: HP + NIC  
 1/2 1/2 1000 c/s  
 1/2 1/2 1/2

7 MIC  
 DMC 1.4  
 B-2-2-D  
 L3-4/0  
 P-3-104  
 10/10/45

	100000	100000	100000
1	-784-	-823-	-789-
2	387	389	411
3	382	402	395
4	402	380	395
5	388	397	383
6	407	373	395
7	363	374	412
8	362	382	378
9	400	409	390
10	380	389	390
11	362	373	380
12	357	400	414
13	349	384	399
14	348	383	364
15	367	401	401
16	354	359	397
17	353	378	393
18	317	353	375
19	304	386	382
20	321	401	386
21	306	375	401
22	305	387	385
23	301	376	376
24	310	358	387
25	268	373	370
26	303	344	395
27	267	330	398
28	257	350	366
29	253	356	382
30	251	331	366
31	243	325	390
32	251	330	374
33	209	300	352
34	215	328	345
35	235	328	376
36	218	307	362
37	207	311	357
38	196	291	354
39	178	296	339
40	236	314	337
41	180	276	355
42	182	307	355
43	174	253	305
44	176	259	353
45	187	272	310
46	184	252	334
47	166	276	338
48	165	261	328
49	163	250	322
50	179	245	306
51	150	234	324
52	150	234	297
53	143	248	305
54	145	227	303
55	155	221	304
56	152	232	311
57	142	227	309
58	124	204	301
59	113	200	302
60	121	231	291
61	111	186	303
62	111	234	273
63	119	191	263
64	123	222	263

980

1029

980



8-30-60 Turned TMC on 9<sup>55</sup>A -

1<sup>25</sup> PM - Set up NICE pulser as Septum trigger. (capitax needs > 3v, + pulse)

Check for proper triggering -

NICE pulse rate 60 c/s.

Run for 200 sec, got 12003 counts. 60 c/s 4V

Detector: NE400 + 6.810H

Brought from 210 (had been cleaned of new mag slip put on)

Set up in 112. HV cable 23' RG 41/U

(in rack 8 by)

Same as when in 113.

(Null Detector)

516 cable 27' RG 62/U

$R_a = 75 \Omega$  (+ 270pF cap)


Tektronix 543 scope brought down from 216 -

HV = 1900

Note varying intensity of trace, indicating changing rates of background still evident -

Observed background pulses -

Direct from detector to scope.



HV = 1700


Same shape pulse - Amplitude reduced to 0.04v.

Varying intensity trace present.

When Div - count rate low

Height - high

Observed further with Neutron source & detector



8-30-60

Arrangement:  $R_a = 75 \Omega$  RG 62/U  
Signal to HP460AR (gain max)

HP460BR (linear)  $R_o = 75 \Omega$

Scopes & TMC

On TMC - got indication of count rate changes -

→ Try removal of mag slip at detector - Still changes -

HV1700 - Pkg	212 Disc	MIN	1	2
pre-set at 2 <sup>6</sup>		642	2020	
(or 10 <sup>4</sup> sweeps)		1111	1945	32 cts
		845	1867	
		610	1978	
		873	1812	
		650	2028	

8-31-60 Made up new HV cable - RG 59/U 28'

BNC connection of HV connector

Lower HP amplifiers from top of Rack 7 to writing surface.

Place Detector on writing surface -

Seem to be no bkg bursts - {need to compare counts with those of 8-18/19-}

Septem Aug 600 NICE pulser - Memory full B=2 D=2 L=40  
pre-det-ct=26 - 10<sup>4</sup> sweeps

HV = 1700	212 Disc	MIN	1	2	3
Pkg		580	1624		
		736	1906	3 counts	0 counts
		730	2023		
		733	1542		
		657	1596		
		633	1401		
		703	1644		
		642	1585		
		620	1852		
		760	1495		

M-230

1/2" write

665	2398	4156
606	2433	3444
668	2108	2876
652	1887	3137
778	2151	4299
623	2089	3625
624	2393	3840
742	1916	3562
621	1880	2932
428	2090	4016

m = 1/2

8-31-60 - try different detectors -

NE400+6292-2 @ 1500V just see neutrons on TMC with disc minimum

Li I + 665T-A @ 1200 just see neutrons on TMC

Back to 6810A + NE400

BACKGROUND HV 1700

Series on TMC -

$N = 10^4, b = 40, 2bN = 80 \times 10^4 = 8 \times 10^5$

$\% = \frac{1}{N} \cdot \frac{C}{2b} \cdot 10^6$

$\% = \frac{10^6}{8 \times 10^5} \times C = 1.25C$

HABO max  
TMC: Disc = min  
B=2, D=2, L=40  
mem = 1/4  
Resist count  
10<sup>4</sup>

C	%
1646	2055
605	756
522	652
493	554
505	631
557	684
544	680
573	641
489	611
<hr/>	
184	230
186	132
190	237
237	296
211	264
202	252
186	232
235	294
185	231
176	220
229	286
268	335
165	206

Disc = 1

dfs

TMC - B, D, L mem 1/4 -

M-230  
1/2" plastic

Counts for 10<sup>4</sup> cycles

Disc = 1		Disc = 2	
925	1156	366	460
980	1225	387	484
967	1208	389	486
981	1226	389	486
1016	1269	415	519
974	1217	389	486
1006	1257	365	456
1009	1260		
1092	1364		
1057	1320		
1020	1275		
995	1244		
1024	1280		

Bkg  
263 volts  
ampl.  
10<sup>4</sup> pulses  
→ 0.1 μs

no counts  
in 1<sup>st</sup> channel  
at Disc = 2

Detector in 113 - HV cable RG59/U, Sig RG 62/U -

Bkg at Disc = 1. Pulse height agreement about same -

Preset 10 <sup>4</sup> cycles	Disc →	1	2	3	4
		460	575		
		292	365		
	200 pm →	273	341		
		221	276		
		205	256		
		217	271		

M-2274  
3" paraffin

dfs	dfs	dfs	dfs
2425	3050	1485	1855
2376	2967	1420	1775
2445	3057	1436	1795
2483	3105	1458	1820
2505	3130	1480	1850
2434	3042	1408	1760
2471	3090		
2527	3158		

begin 3:20 am  
Bkg.

dfs	dfs	dfs	dfs
264	330	1	1.25
222	274		
276	295		
309	384		
342			

8-31-6

3:50

HUPS in 113 - Disc 1 -  $N=1700$

(Still getting count rate at 113)

220	275	218	272
312	390		
244	305		
177	221		
187	234		
183	229		

9-1-60

8<sup>42</sup> TMC power on (in 113)


8<sup>44</sup> AVDC on - 1700 v.

8<sup>46</sup> start count - (700 sec had accumulated 3600 counts)

Pre-set survey 10<sup>4</sup> or pre set count rate Disc 1 Counts in Ch1: 374, 348, 316, 214, 159, 157, 147, 138, 161, 173, 130, 122, 136, 121

% : 467, 435, 395, 267, 199, 194, 184, 172, 201, 211, 162, 152, 158, 153

Observing trace at 1 millisecon with occasional pulse with neg. band -



pulses are 0.7  $\mu$ s wide at base line - similar to reversal reaction pulses.

Big: Output of detector as measured 8-30 - .01 volts.

Output of HP4603R-AR comb.



9:31

Stop run to check HP amplifiers - Disconnect - NICE pulses

System trigger 543 + gate 1ms @ 80 p/s 1000 checks.

50 80 p/s 400.

seems to lose some count in later channel (in 1/2)

set 1/2 run 5 139 p/s -

(8 too fast)

TMC Settings Unchanged -


B=2 Mem 1/4

D=2

6=90

HP Amp -

signal NICE 1ps .4v neg.




no running over base line after pulse

OUTPUT


HP4603R: A - RG62U directly to output.

Gain: Min. amp = 4v,

MAX amp = 10v.




MAX is 15v on output - RG62U




$R_0 = 200 \Omega$  at output

C - 100' RG 119/U

amp Gain max = 4v.




HP4603R -> BR 1 AR gain MAX - BR MIN.



Gain with 2nd BR -

HP4603R B w 75a.4 RG62/U

max at 5v



F-1-60- More Polg - counts.  $\left\{ \begin{array}{l} \text{Time } 2, 2, 40, 1/4 \\ \text{HV } 1700 \text{ in } 113 - \end{array} \right.$  NICE low pulser same as earlier

Ch I  
Disc 1 counts: 102, 96, 85, 90, 92, 94.

unch-

Disc 1 more/above: 100, 68, 78, 60.

Check signals from detector on  $2.7 \mu\text{A} + R_9 62/0$  line then through  
b.  $200 \mu\text{A} + R_9 114/0$  line then hall (100)



HP460 AR at min gain. amplitude  $\approx 4$  volts  
" " "  $\approx 5$  volts.

Very little difference -

Pre-  
or

M-229 of 3/1/60

Disc	Channel 1 Counts
Disc 1	858, 879, 941, 859, 848 1073, 1099, 1176, 1074, 1060
Disc 2	511, 508, 514, 526, 536 639, 635, 642, 708, 670
Disc 3	206, 258, 235, 247, 215, 231 295, 320, 294, 302, 269, 289
Disc 4	92, 89, 87, 75, 90, 84 102, 111, 109, 94, 112, 107

HP460 AR MIN, 13R L/A

\* Sig trip  
Scope .5  
& fine sweep  
set to 100 cps

Big Discy - D. 1

HVPS back in 112 = HV @ 1700 - long cable

Background: Present cycles  $10^5$


Time  $\left\{ \begin{array}{l} \text{Menu } 1/4 \\ \text{D} = 2, \text{D} = 2, \text{L} = 40 \\ \text{Disc} = 1 \end{array} \right.$

Ch 1 = 3.  $\approx 1 \%$

M 227

Present cycles  $10^5$

Ch 1 = 8400 1050 c/s  
8516 1060 c/s



EUGENE DIETZEN CO.  
MADE IN U.S.A.

ND 340R-L410 DIETZEN GRAPH PAPER  
SEMI-LOGARITHMIC  
4 CYCLES X 10 DIVISIONS PER INCH

9-2-60 Set M227 in place on paraffin 113.

Prepared for run.

DESIRING TO RUN w/o SCOPE MONITORING SIGNAL, SET ABOUT TO DISCONNECT SCOPE. IT & HP LEAD JOINED IN "T" at SIG. INPT of TMC UNIT. GOT BAD SHOCK WHEN SEPARATED TWO CABLES, AFTER REMOVING "T." (had HP lead in one hand, scope lead in other). Broke hold by pulling away from cables on the run - could not release hand hold.

CHECK OUT - FOUND 110V AC on GND of scope & TMC (scope connected to TMC by "SYS. TRIG" cable) DISCONNECT this cable. 110V on TMC - coming from Monroe - this was plugged in same receptacle as NTE power supply. - IN another receptacle NO HOT CHASSES. - LEADS TO NTE AUPS wrong - CORRECTED FAULT QUICKLY -

10:01 Power On TMC units.

10:02 Check system trigger operation - ok.

(Scope used as trigger - "STABILITY" full c.w.)

range at .5 millisecc/sec. @ 180 c/s.

HV = 1700v.

TMC { B=2, D=2, L=40  
M=14 PRESET  
DISC = 1 cy 105  
ct 216

M227 - 564.0sec, Tmc Σ = 4514.0 Avg 564.25s.

① 19376, 18719, 18548, 18312, 17980, 18233, 18031, 17957, 18211  
% 2422, 2340, 2318, 2290, 2248, 2229, 2254, 2244, 2278

B/cg, 627, 672, 781, 552, 549, 577, 570, 507, 509.  
② 734, 84, 97.6, 64, 63.4, 64.6, 71.2, 69.6, 69.9

4-2-60 Disc = 1.5 Channel 1 = 20 2.5 1/3 10<sup>4</sup> cycles  
 Disc = 2 = 6 2 4  
 .75 .25 .50


M-227 Disc 1 17,334 2167 1/3  
 1.5 11,383 1424  
 2 8,324 1041  
 2.5 6,069 758.0  
 3 4,262 533.0  
 3.5 2,827 353.0  
 4 2,145 268.1


3128  
 Blg. 10<sup>4</sup> cycles Disc 1 See - 31  
 See 41  
 27

T Connection - 47, 39

Scope lead attached (RG 1/8) 0

Scope monitor lead  
 attached to signal




**SEMI-LOGARITHMIC 359-816**  
 KEUFFEL & ESSER CO. MADE IN U.S.A.  
 4 CYCLES X 70 DIVISIONS



9-6-60: TMC on 8:21 AM, Scope ON (SYS TRIG), HVPS on @ 1700V  
 TMC: <sup>B P W</sup> 2, 2, 40, 1/4 M. Print Cycle 10<sup>5</sup> -

1 Bkg - Disc = J. 8:33 339, 374  
 M-227 Disc J. 9:01 3219, 2272, 2218, 2258, 2260, 2260  
 17753, 18170, 17744, 18052, 18081, 18079  
 Disc 2 10:56 1153, 1176, 1157, 1180, 1172, 1161, 1143 (10.53)  
 9223, 4412, 4261, 9443, 4377, 4285, 4385

\* these points in error - Channels 2-64 indicate that  
 Channel 1 should be 9 ---

▲ This point was observed during print out.  
 Data Output Unit scale read 9385. Printer printed 4385.  
 Bkg Disc 2: 12:16 One Sample Channel 1 = 3 -

Signal: EP 4550A (see next page) Output: 7.2v - 30db, 10 - 4, 23v.  
 through 100' RG 114A/U to TP 460AR - gain min → TP 460BR linear → 75Ω RG 67/U  
 to input TMC - signal cut off before 30db or discriminator 23.7

Signal EP 4550A to TMC - after checking + output to P  
 Signal 7.2v = exceeds all discriminator settings -  
 Determine EP settings for Disc 1, 2, 3, 4  
 18db, 6; 13db, 6; 10db, 6; 8db, 5  
 (.7v, 1.3v, 1.8v, 2.3v)

Signal thru 4P's signal for threshold @ Disc = 3 30db 4 = 18v  
 -12db, 8 - Disc 1 1.6 volts [ RG 114 A/U terminated ]  
 3db 8 4 5v. [ @ 200Ω at TMC ]

~~EP~~ E-P signal generator

9-6-60

Look EP 4J30A, 492 to 112 to check out with TMC &

other counting instruments

TD111 fixed at (1+0)x.1 ps	TD411 P.S. μs	PW111 μs	0A601 Attn	Ampl
neg - (9+1)x1	(5+0)x0.1	0	10	
		100	10	
		200	10	
		300	10	
		300	0	
positive		0	10	

7.2V  
100 R6114A/5 term @ 50μs 200


2.1V

0.72V

0.23V

0.15V

7.2V



EP  
{  
f  
30

9-6-63 TMC - with signal from E-P 4550A-492 -

---  
sweeps  
Pres-set  
Z<sup>16</sup>

EP: TD111  
{ fixed @ (.1±0)x1 ps  
0A601  
30dB, 10, neg.

Pulse Spacing, Pulse Width  
(.9±.1)x1 ps (5±0)x1 ps

TMC					sweeps	Pres-set
B	D	L	M	N	Z <sup>16</sup>	
2	2	40	1/4	804	65536	
				804		

$$f_{mc} = \frac{1}{N} \frac{C}{ZL} = \frac{1}{N} \frac{655.36 \times 10^2}{80} = \frac{819.2}{N} = 1.019 \text{ Mc}$$

(.5±0)x1 ps	(.5±0)x1 ps	2	2	40	1/4	407	2.012
(.2±0)x1 ps	(.5±0)x1 ps	2	2	40	1/4	164	4.995
(.1±.05)x1	(.5±0)x1 ps	2	2	40	1/4	112	7.31 Mc
(.1±04)x1	(.5±0)x1 ps	2	2	40	1/4	102	8.03
(.1±03)x1	(.5±0)x1 ps	2	2	10	1/4	398	

volts

9-16-60

Setup in 108

Own Reports varying counter rate.

Evidence: different reading for given N-source  
varies unlesch of signal trace on scope

Made up new detector NE400 + 6810A 3-60-199

(Older was NE400 + 6810A 3-60-226)

10:10 <sup>PM</sup> Del to DWM

9-19-60

8:08<sup>PM</sup> Scope HVPS, HP's on -  
scope settings for sweep

Trig Slope + line +  
Mode - are slow  
stabilize

Trig level slightly +  
Sens/cm 0.1 sec.


10:05

Detector in place in 108

N-source at paraffin face  
near base of detector tube

To compare pulses for varying HV - gain combination.

HV	A-Gain	B-Gain
1750	MIN -	LOW
	MAX -	



1650	MAX	LOW
	MIN	
1550	MAX -	LOW

peak still 4.9v  
peak slightly under 6v  
peak 6.2v

N= 100 swings  
D= 679 ch rate  
CT +106

9-11

4-14-60

1038 - HV = 1550, HP max, LIN

Source at base of PM  
DETECTOR IN PARAFFIN ON  
assembly.

TMC DISC 6M

B=256, D=2, L=2560, M=1/4, Scavenging rate 1/2 sec

Note counts in channel 7 (Bkg) for 10 sweeps

No. 10 sweeps  
 B = Bkg ch rate  
 for analysis channel length  
 for N=10  
 L=256  
 R=256

DISC		"BKG"	SOURCE	
1	7.74%	51 55 51 52 47 48 61 41 20 16	10150 10237 10058 10199 40674	9983 10283 10012 32278 40674 73952
				10132
2	1.46%	6 9 90 10 58 10 13	4701 4731 4825 4699	4617 4656 4680
				4701.3, 71.7%
3	.99%	10 6 7 5 45 6 5	2098 2064 2171 2135 2148	2055 2106
				2111 322
4	.35%	4 4 1 2 2 2 1	860 871 881 879 923 889	891
				884.9 135

CT +10%  
 NBL  
 CT  
 2055  
 6.556  
 c/s  
 c/s

plotted, see graph p 282-283

9-19-60  
11:30 A

HV 1550  
 HP A MAX  
 HP B LIN  
 DISC TMC 2

550V  
max



4-27-60

(For pulse refer to Semitellatrons Detector File)  
 OUTPUT FROM DETECTOR TO SCOPE

$I_p = 6810A(3-60-226) + NE400$  HV=1500v,  $I_p = 0.9ma$  pk pulse 6mv, N pulse .10v.

Low Current 1700v, 0.9ma, .03v, 0.4v.

Divider 1900v, 1.0ma, .15v, 2.0v.  
 (40ms)

and R=100k

acc. anode  
 tied to Dy 14.

Noted that N pulse never overshoot reference line.

- N pulse returns to reference in  $\approx 10\mu s$ .

@ 1700v & P460AR: N-pulse amplitude  $\approx 4$  volts. (Gain 10)

Pulse tail different. Overshoots in 7 $\mu s$ , reaches peak overshoot in 15 $\mu s$ , & settles down at 40 $\mu s$ .

@ 1700v & P460BR: slight overshoot evident.

N-pulse amplitude  $\approx 2.0$  volts (Gain 5)

@ 1500v & both P points: N-pulse amplitude =  $\frac{64}{8}$  volts.

Returns to 25% of peak in .15 $\mu s$ .

Tail decays to reference at 10 $\mu s$ .

Rey ~~peak~~ at 20 $\mu s$ .

Reaches reference at  $\approx 30\mu s$ .

Put same detector on divider for high current.

HV=1500v. Direct to scope.

$I_p = 4.55ma$ . Pulse shape same, amplitude about same.

Adjusted accelerating quad for max. gain of pk pulse.

N-pulse  $\approx .06$  volts Gain just about the same.

Tube 3-60-179  
 +NE 400

Sig. Trig.  
 0.1 sec - limit

HV=1500v	Ch 1	Ch 2	Ch 3	Ch 4	Ch 5	Ch 6	Ch 7
TMC							
A=252 n=10	39644	6049	26,823			19,843	7329
D=2	39648	6050	27,426	27,222		19,905	7286
b=2560 DBC=1	39651	6050	27,417			19,841	7280
			381,666			19,915	7875

Reset acc. anode at max mark.

	Bkg
37412	20
37755	19
37754	16
112921	18

3-14-61 Use SA-2 as scintillation detector checker.


NTE NIPS; Signal to 10-amp input (pair in SA-2 discrimin) A- with JFE rate meter & recorder run scan of Cs<sup>137</sup> 6363 in 2-channel eq.

SCAN	Source	HV	GAIN	OE	RATE METER		PEAK	RESOLUTION	
					DAMPING	CPM		6363 + 353	
UP	<sup>137</sup> Cs @ 9.3a	910	32	100	FAST	30,000	64% @ (NOT RECORDED)	NOT RECORDED	
500 → 125					MED	30,000	62% @ 397	38/397	4.57
500 → 25					SLOW	30,000	58% @ 392	39/392	9.95
25 → 475					"	"	58% @ 402	40/402	9.95
415 → 325					"	"	58% @ 392	39/392	9.95
300 → 150					MED	"	64% @ 393	37/393	9.41
450 → 300					"	"	64% @ 395	37/395	9.32
800 → 20	<sup>137</sup> Cs @ 9.3	990	32	100	MED	30,000	38% @ 660	61/660	9.24
800 → 300	<sup>137</sup> Cs @ 0.0	990	32	150	MED	10,000	92% @ 667	54/667	8.10
45	6292 + NaI								
400 → 800	<sup>137</sup> Cs (small)	1350	320	150	MED	1,000	>100% @ 685		
800 → 350	"	"	"	"	"	3,000	34% @ 670	70/670	10.45
800 → 350	<sup>137</sup> STD @ 2am	"	"	"	"	10,000	67% @ 684	69/684	10.08
400 → 50	"	1175	"	"	"	30,600	42% @ 340	37/340	10.9
600 → 50	Pube	"	8	50	"	10,000			



6363 + 3x3 NaI  
(20 cm Detector)  
HV = 990V  
Cs<sup>137</sup> (card) contact

FULL SCALE = 10,000 C/M



Pu-Be Source

GAW 8

6292 + NaI

HV = 1175

(Cs = 240, gain 32)

